



19528 Ventura Blvd. #379, Tarzana, CA 91356 | P: 800-799-5873 | F: 818-345-9011

**Date: 06/10/2013**

**Recall Notice for FBS for IVD applications.**

RE: Use of Fetal Bovine Serum for in vitro diagnostic applications

Dear Valued Customer,

We have received a notification from one of our previous suppliers of fetal bovine serum that any use of FBS for in vitro diagnostic purposes should be discontinued and any unused product returned.

This notice affects a small percentage of FBS that was manufactured by one of our previous suppliers between March 2008 to 2013. This notice does **not** apply to all lots from Omega Scientific, Inc.

The reason given for this action is that the FBS from this supplier may contain adult bovine serum albumin of US origin, some added water and cell growth additives. We have tried to obtain more specific information about what was added and the amounts and the supplier has refused to provide that information. We regret having to take this action but the supplier has left us no choice. Please review this notice and contact Omega Scientific if you have any questions and/or concerns.

Sincerely,

Robert Danziger  
President  
Omega Scientific, Inc.

*Revision 1 from Letter Dated 5/17/2013.*