

Uni Hospital Limerick

Report Name:POCT Monthly Feedback

Created by:John Doe

Date/Time:09/23/2025 09:19

Note:Information shown in this report depends on the user's permissions to access the related information.

Report Filters:

Site:

HA

Month:

September 2025 - September 2025

Location:

Instrument Type:

Instrument Name:

Serial Number:

Test:

Number of Users:

0

Number of Patients:

0

User 1:

Patient 1:

User 2:

Patient 2:

User 3:

Patient 3:

User 4:

Patient 4:

User 5:

Patient 5:

Include empty results:

No

Reviewed by:

Review Date:

Signature:

Observations:

Site:

HA

Location:

HA - NURSE STATION

Month:

September 2025

Instrument

ACI II

Test:

Glu2

Instrument:


UU14022001-1

Patient Results

Total	Rejected	STAT	With Comment	Without Patient
0	0	0	0	0

Quality Control Results

Level	N	N (Rejected)	Mean	SD	CV [%]	Bias [%]
1	22	1	43.1	2.9	6.73	-4.22
2	22	0	313.5	5.0	1.59	2.79



Observations:

Reviewed by:

Date:

Signature:

Date:

Site:

HA

Location:

HA - ICU

Month:

September 2025

Instrument

cobas b 123 system

Test:

COHb

Instrument:


HA-IC 123007

Patient Results

Total	Rejected	STAT	With Comment	Without Patient
0	0	0	0	0

Quality Control Results

Level	N	N (Rejected)	Mean	SD	CV [%]	Bias [%]
1	1	0	22.10	0.00	0.00	-1.34
2						
3						



Remark: Data values outside the upper/lower boundaries are drawn on the graph borders.

Observations:

Reviewed by:

Date:

Signature:

Date:

Site:

HA

Location:

HA - NURSE STATION

Month:

September 2025

Instrument

ACI II

Test:

Glu2

Instrument:


UU14022002

Patient Results

Total	Rejected	STAT	With Comment	Without Patient
1	0	0	0	0

Quality Control Results

Level	N	N (Rejected)	Mean	SD	CV [%]	Bias [%]
1	23	0	43.2	2.9	6.71	-4.00
2	23	0	313.5	4.8	1.53	2.79



Observations:

Reviewed by:

Date:

Signature:

Date: