Filename: 2021-03-31 - 16.41.45.D5000



Default image (Contrast 100%)

Sample Info

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
A1	92.1	Ladder	⚠	Caution! Expired ScreenTape device; Issue with ladder peak detection (too many peaks detected); Ladder
B1		1	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
C1		2	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
D1		3	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
E1		4	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
F1		5	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
G1		6	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
H1		7	⚠	Marker(s) not detected; Caution! Expired ScreenTape device
A2		8	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

© Agilent Technologies, Inc. 2019

A1: Ladder



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
A1	92.1	Ladder	⚠	Caution! Expired ScreenTape device; Issue with ladder peak detection (too many peaks detected); Ladder

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
15	10.2	-	1050	-		Lower Marker
30	7.28	-	371	7.91		
55	7.42	-	208	8.06		
100	7.74	-	119	8.41		
250	7.80	-	48.0	8.47		
400	7.72	-	29.7	8.39		
600	8.28	-	21.2	8.99		
1000	7.83	-	12.0	8.50		
1500	7.64	-	7.84	8.30		
2500	8.01	-	4.93	8.70		
3500	7.41	-	3.26	8.04		
5000	7.27	-	2.24	7.90		
7025	7.67	-	1.68	8.33		
10000	3.25	3.25	0.500	-		Upper Marker

© Agilent Technologies, Inc. 2019

B1: 1



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
B1		1	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

C1: 2



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
C1		2	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

D1: 3



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
D1		3	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

E1: 4



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
E1		4	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

F1: 5



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
F1		5	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

G1: 6



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
G1		6	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

H1: 7



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
H1		7	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019

A2: 8



Sample Table

Well	Conc. [ng/µl]	Sample Description	Alert	Observations
A2		8	⚠	Marker(s) not detected; Caution! Expired ScreenTape device

Peak Table

Size [bp]	Calibrated Conc. [ng/µl]	Assigned Conc. [ng/µl]	Peak Molarity [nmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	-		Lower Marker
-	-	-	-	100.00		

© Agilent Technologies, Inc. 2019