

3199 De La Cruz Boulevard • Santa Clara, CA 95054-2483
TEL: (408) 988-0770 FAX: (408) 988-0762
E-MAIL: test@quantalabs.com

Certificate of Conformance

This is to certify that the results from the test(s) requested by					
BACL are on file under					
Quanta Laboratories Job No. QL-19-1066 and conform					
to the specification(s) stated in P.O. No. AE19092367					
These results apply to the following equipment and are					
available for review upon request.					
Model No: M20 Bioamplifier System					
S/N: N/A					
*** Random Vibration and Shock Test ***					
Quanta Laboratories Quanta Laboratories Date					

SHOCK TEST DATA



3199 De La Cruz Blvd, Santa Clara, CA 95054 Phone: (408) 988-0770 FAX: (408) 988-0762 test@quantalabs.com www.quantalabs.com

CLIENT:	BACL		P.O. NO: AE19092367			
SPECIMEN:	M20 Bioam		JOB NO: QL-19-1066			
SPECIFICATION: Client Specification				PAGE 1 OF 1		
DATE	S/N	AXIS	SHOCK SPEC.	REMARKS		
09/24/2019		+X	Half-Sine Shock 150 m/s², 11 ms 3 shocks/direction 18 shocks total			
		-X		Non-Operational Test Test completed to specification requirements. Functional test was performed after each direction.		
	N/A	+Y				
		-Y				
		+Z		DEFINITION OF AXES		
		-Z		See photo page		
TEST ENGINEER	: Anthony Y	′un	b.c.	DATE: 09/26/2019		

QL-TTS-008 Rev: 08/14/17

RANDOM VIBRATION TEST DATA



3199 De La Cruz Blvd, Santa Clara, CA 95054 Phone: (408) 988-0770 FAX: (408) 988-0762 test@quantalabs.com www.quantalabs.com

			tootte-quaritatabo.	Com www.quantalabo.com		
CLIENT:	BACL		P.O. NO: AE19092367			
SPECIMEN:	M20 Bioamplif	ier System	JOB NO: QL-19-1066			
SPECIFICAT	TION: Client Specific	ation		PAGE 1 OF 1		
DATE	S/N	AXIS	FREQUENCIES & LEVELS	REMARKS		
09/24/2019		x		Non-Operational Test Test completed to specification requirements.		
	N/A	Y	Random Vibration Test 10-2000 Hz 10 Hz – 100 Hz @ 1.0 (m/s²)²/Hz 100 Hz – 200 Hz @ -3 dB/octave 200 Hz – 2000 Hz @ 0.5 (m/s²)²/Hz Overall: 32.6 m/s² _{rms} At full level: 30 minutes	Functional test was performed after each axis.		
				DEFINITION OF AXES		
		z		See photo page		
TEST ENGIN	NEER: Anthony Yun		PC	DATE: 09/26/2019		

QL-TTS-007 Rev: 08/14/17



BACLRandom Vibration and Shock Test







X-Axis Y-Axis Z-Axis

QL-118-152

JOB NO. : QL-19-1066

QUANTA LABORATORIES EQUIPMENT LIST



3199 De La Cruz Blvd, Santa Clara, CA 95054 Phone: (408) 988-0770 FAX: (408) 988-0762 test@quantalabs.com www.quantalabs.com

Client:		BACL		P.O. NO:	AE190	92367
Client:	JOB NO.:	QL-19-1066				
•		DIGITAL S	YSTEM LIST			
Device Type	Description	Make & Model	Range	Asset #	Serial #	Due Date
Shaker Control System	VC-07	JAC VICO-8 8 inputs	0.1Hz - 3 KHz RES. 0.1dB	QL-0786	177247011	06/26/2020
Current Source	VC-07	Dytran 4123 8 Inputs		QL-0218	119	04/24/2020
		MECHANICAL	SYSTEM LIST			
Device Type	Description	Make & Model	Range	Asset #	Serial #	Due Date
Shaker Amplifier	Red/Green	Ling DMA-48		QL-0504	46	Calibration Not Required
Electrodynamic Shaker	Red/Green	Ling B-335 (Red)	5Hz - 3 KHz	QL-0256	3	Calibration Not Required
Electrodynamic Shaker	Red/Green	Ling B-335 (Green)	5Hz - 3 KHz	QL-0253	91	Calibration Not Required
		SENS	OR LIST			-
Device Type	Description	Make & Model	Range	Asset#	Serial #	Due Date
Accelerometers	Single-Axial	DYTRAN 3256A2	5~2000 Hz 50 G	QL-0878	10868	01/09/2020
and the second s			neous List			
Device Type	Description	Make & Model	Range	Asset #	Serial #	Due Date
		Customer Sun	plied Equipment			
Device Type	Description	Make & Model		The second secon	Serial #	Due Pete
Device Type	Description	Wake & Wodel	Range	Asset #	Serial #	Due Date

Notes

- 1. This report may not be reproduced, except in full, without written approval by Quanta Laboratories.
- 2. The information in this report applies only the items tested or calibrated.
- 3. Measurements in this report are traceable to SI units via national standards maintained by NIST or derived from acceptable values of natural physical constants that comply with ISO 17025:2017 and A2LA requirements.

- 4.In Tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.
- 5. The estimated measurement uncertainty (EMU), if reported on this certificate, is being reported at a confidence level of 95% or K=2 unless otherwise noted in the comments section.

QL-TTS-091 10/23/2017



3199 De La Cruz Boulevard • Santa Clara, CA 95054-2483 TEL: (408) 988-0770 FAX: (408) 988-0762 E-MAIL: test@quantalabs.com

Quanta Laboratories Test Report

Quanta Laboratories submits this report with our Certificate of Conformance to the requirements of the applicable specifications and with appropriate supporting data, but with no other expressed or implied warranty. Customer assumes full responsibility when using or interpreting the data herein for evaluation and/or reporting purposes. The contents of this report apply only to the sample(s) as received and were provided to Quanta Laboratories by the Customer. Sampling methods are unknown unless data is provided by the Customer.

Quanta Laboratories is only responsible for the processes and data resulting from testing at Quanta Laboratories. Quanta Laboratories is not responsible for verifying data supplied by the Customer. Customer supplied data, equipment, items, and personnel are identified in the report by the symbol "*" and accompanying footnote.

End of Report QL-19-1066

QL-TTS-150 Rev: 07/22/2019