THERMAL CYCLING TEST UNIT for PIPE SYSTEM

DESCRIPTION

- High constancy of temperature by large reservoirs
- Static pressures up to 10 bar
- Accuracy pressure regulation
- Automatic failure detection
- Simultaneous testing of two seperate pipe systems independendly.
- Automatic test procedure with programable numbers of cycles and cycle periods, temperatures, etc.
- Test chamber (optionally with fixing lates)
- Cold water tanks with chiller or heat exchanger, hot water tanks with heaters
- Flow measurement and regulation automated.
- Automatic calculation and application of prestress needed by the standardss via Stepper Motors

Accessories

THERMAL CYCLING TEST UNIT for PIPE SYSTEM

Standards	
ISO 19893	DVGW W 542
ISO 10508	DVGW W 543
DVGW W 534	

Technical details

Pressure range	3 up to 10 bar, 15 to 95 °C
Test circuits capacity	up to 63 mm
Flow rate	Max. 1.0 l/sec
Max. test sample volume	12 l
Cycle period	Selectable between 1 and 9,999 min
Number of cycles	Max. 99,999 per test
Temperature range (hot)	Selectable between 50 and 95 °C
Temperature range (cold) with cooling unit	Selectable between 15 and 30 °C
Regulating accuracy	Temperature controller: approx. 0.5 °C In the tanks: approx. 1.0 °C
Permissible ambient temperature	5 - 25 °C
Max. relative humidity	70 %
Connection	
Power supply	230/400 V, 50 Hz; approx. 25 kW; constructed according to VDE 0100; protection class of switch box: IP 55
Water	G 3/4 inch external, min. 2 bar, max. 5 bar
Drain	HT pipe DN 2 x 40
Test sample	G 3/4 inch internal

POLİNAR PLASTİK MAKİNE ve KALIP SANAYİ
TİCARET LİMİTED ŞİRKETİ
İkitelli O.S.B Eskoop San. Sit. D Blok
No: 34 Başakşehir / İSTANBUL - TURKEY

Phone : +90 212 549 88 20 – 21
Fax : +90 212 549 88 19
E-Mail : info@polinar.com.tr

www.polinar.com.tr

PTE (POLİNAR TEST EQUIPMENT)
İkitelli O.S.B. Eskoop San. Sit.
D Blok No.34 Başakşehir
İstanbul / TURKEY

Phone : +90 212 549 88 20 – 21
Fax : +90 212 549 88 19
E-Mail : info@polinar.com.tr

www.pte.com.tr

POLINAR GmbH
THE ICON VIENNA, Tower 24
Wiedner Gürtel 13 1100
Vienna AUSTRIA

Phone : +43 (0) 1 253535 4044
Fax : +43 (0) 1 253535 4044
Mobile : +43 676 5983869
E-Mail : office@polinar.at

www.polinar.at