END CLOSURES for PIPE TESTING

DESCRIPTION

The internal pressure creep test is a test procedure for determining the strength of thermoplastic pipes to constant hydrostatic internal pressure at a constant ambient temperature. The samples are tested either for a specified period or until they fail. The test duration is subject to the tension generated by the internal pressure and the temperature.

Accessories

Simple and safe operation

- Quick assembly thanks to proven end closure design
  Simple and reliable venting directly at the sample

Lasting efficiency

- High quality unit components guarantee high reliability, a long service life and low maintenance costs

END CLOSURES for PIPE TESTING

Standards
ISO 1167
ASTM D 1598
ASTM D 1599

Versions

For pipe diameter DN	mm	20-40	50-90	110-315	350-630
Maximum test pressure	bar	60	60	60	40

End closure material	Stainless steel/AISI 304/S30300
Suitable for PE pipes	●
Suitable for PP pipes	●
Suitable for PVC pipes	●
Vent screw	●
Ring nut for suspension	●
Pressure connection	Quick-release plug

● included ○ available / optional □ eligible - not available

POLİNAR PLASTİK MAKİNE ve KALIP SANAYİ
TİCARET LİMİTED ŞİRKETİ
İkitelli O.S.B Eskoop San. Sit. D Blok
No: 34 Başakşehir / İSTANBUL - TURKEY

Phone : +90 212 549 88 20 – 21
Fax : +90 212 549 88 19
E-Mail : info@polinar.com.tr

www.polinar.com.tr

PTE (POLİNAR TEST EQUIPMENT)
İkitelli O.S.B. Eskoop San. Sit.
D Blok No.34 Başakşehir
İstanbul / TURKEY

Phone : +90 212 549 88 20 – 21
Fax : +90 212 549 88 19
E-Mail : info@polinar.com.tr

www.pte.com.tr

POLINAR GmbH
THE ICON VIENNA, Tower 24
Wiedner Gürtel 13 1100
Vienna AUSTRIA

Phone : +43 (0) 1 253535 4044
Fax : +43 (0) 1 253535 4044
Mobile : +43 676 5983869
E-Mail : office@polinar.at

www.polinar.at