Baxter

SIGMA SPECTRUM Infusion System and SPECTRUM IQ Infusion System Compatible Cleaners

- 10% Solution of **Clorox** Bleach and 90% Water
- **3M** HB Quat Disinfectant Cleaner Concentrate
- A-456 II Disinfectant Cleaner
- **Clorox Commercial Solutions Clorox** Germicidal Wipes
    - Available in Canada as **Clorox Healthcare Professional** Disinfecting Bleach Wipe
- **Clorox Healthcare** Hydrogen Peroxide Wipe
- **Dispatch** Hospital Cleaner Disinfectant Towels with Bleach
- Oxycide
- **PDI Sani-Cloth** Bleach Wipe
- **Super Sani-Cloth** Germicidal Disposable Wipe
- Virex II 256
- **MICRO-KILL** Germicidal Bleach Wipes
- **MICRO-KILL** Two Germicidal Wipes
- PCS 5000 Oxidizing Disinfectant/Cleaner

Note:

- The listing is strictly to identify materials compatibility with the **SIGMA Spectrum** Infusion Pump and **Spectrum IQ** Infusion Pump.
- Consult the respective Operator's Manual for instruction and cautions for cleaning.

This document is subject to change, please visit https://service.baxter.com/tsportal/#!/ or contact Technical Support for information:

- **Phone: (800) 356-3454 and follow prompts**
- **Email: MedinaTechSupport@baxter.com**

Rx Only. For the safe and proper use of this device, refer to appropriate operator's manuals

Any other trademarks, product names or brand images appearing herein are the property of their respective owners.

Baxter, Sigma Spectrum and Spectrum IQ are trademarks of Baxter International Inc. or its subsidiaries
Baxter Healthcare Corporation, One Baxter Parkway, Deerfield, IL 60015 www.spectrumiq.com
Document Number DOC 11318 Revision Q