

# Analytical and Monte-Carlo modeling of Multi-Parallel Slit and Knife-Edge Slit Prompt Gamma Cameras

Brent F. B. Huisman<sup>1,2</sup>, É. Testa<sup>2</sup>, D. Dauvergne<sup>3</sup>, J. M. Létang<sup>1</sup>, and D. Sarrut<sup>1</sup>

<sup>1</sup> CREATIS, Université de Lyon; CNRS UMR5220; INSERM U1206; INSA-Lyon; Université Lyon 1; Centre Léon Bérard, Lyon, France

<sup>2</sup> Univ. Lyon, Univ. Claude Bernard Lyon 1, CNRS/IN2P3, IP2I Lyon, F-69622, Villeurbanne, France.

<sup>3</sup> LPSC, Laboratoire de Physique Subatomique et de Cosmologie, CNRS/IN2P3, Université Grenoble-Alpes, Grenoble, France  
E-mail: e.testa@ipnl.in2p3.fr

November 13, 2019

## Abstract

*Purpose:*

*Materials and Methods:*

*Results:*

*Conclusion:*

## Contents

<b>1</b>	<b>Introduction</b>	<b>3</b>
<b>2</b>	<b>Materials and Methods</b>	<b>4</b>
2.1	Analytical model for spatial resolution and detection efficiency . . . . .	4
2.1.1	Spatial resolution . . . . .	4
2.1.2	Detection efficiency . . . . .	6
2.2	Monte Carlo simulations . . . . .	9
2.2.1	Simulation tool . . . . .	9
2.2.2	Prototypes comparison . . . . .	10
2.2.3	Analytical Model Verification (AMV) . . . . .	11
2.2.4	Background modeling . . . . .	11
2.2.5	Beam and target . . . . .	12

2.3	Figures of merit . . . . .	12
2.3.1	Analytical Model Verification . . . . .	12
2.3.2	Prototypes Comparison . . . . .	13
<b>3</b>	<b>Results</b>	<b>13</b>
3.1	Analytical Model Verification . . . . .	13
3.2	Prototype Comparison . . . . .	14
3.2.1	PG profiles . . . . .	14
3.2.2	FRP . . . . .	16
<b>4</b>	<b>Discussion</b>	<b>18</b>
4.1	Clinical implications . . . . .	18
<b>5</b>	<b>Conclusion</b>	<b>19</b>
<b>6</b>	<b>Acknowledgments</b>	<b>19</b>
<b>A</b>	<b>Fall-off position and width estimation procedure</b>	<b>21</b>
A.1	Fall-off position estimation procedure . . . . .	21
A.2	Fall-off width estimation procedure . . . . .	22
<b>B</b>	<b>Verification of the cameras</b>	<b>23</b>

# 1 Introduction

The well-defined range of particles in matter is the main reason they are used in cancer treatment today. Unfortunately we are not able to take full advantage of this property, because of treatment uncertainties, e.g. uncertainties in patient positioning, changes of patient anatomy between treatment fractions and uncertainties in the Hounsfield unit to particle stopping power conversion (Paganetti, 2012). Often, medical practice leads to the use of several irradiation fields to ensure robust treatment planning at the expense of larger healthy tissue volume irradiation in respect to ideal treatment with single irradiation field. Moreover margins around the tumor are added, reducing the potential benefits of particle treatment (Knopf and Lomax, 2013). Ion-range verification could permit more precise planning which could take maximum advantage of the steep Bragg peak (BP) fall-off and reduce damage to tissues surrounding the tumor. A promising modality to perform this verification consists in detecting prompt gammas (PGs), a natural byproduct in particle treatments (Krimmer et al., 2017b; Parodi and Polf, 2018). Various modalities are currently under investigation and can be classified in two categories: imaging devices using either mechanical or electronical collimation (respectively collimated (Min et al., 2012; Perali et al., 2014; Pinto et al., 2014) and Compton cameras (Krimmer, 2015; Kurosawa et al., 2012; Llosá et al., 2016; Polf et al., 2015; Thirolf et al., 2016)) and non-imaging devices such as Prompt Gamma Spectroscopy (PGS) (Hueso-González et al., 2016), Prompt Gamma Timing (PGT) (Pausch et al., 2016) and Prompt Gamma Peak Integral (PGPI) (Krimmer et al., 2017a). Regarding cameras using mechanical collimators, two types of designs have been considered: Multi-Parallel Slit (MPS) and Knife-Edge Slit (KES). The latter has been chosen by the IBA group to be prototyped and has been tested clinically with pencil beam scanning (Xie et al., 2017). Both types of collimators provide part or the whole of the one-dimensional PG profile along the beam direction.

To date, three publications (Lin et al., 2017; Park et al., 2017; Smeets et al., 2016) have attempted to compare these two types of collimators with various geometries. Indeed, although they refer to the same KES design from (Perali et al., 2014) some geometrical parameters are slightly different in (Lin et al., 2017; Park et al., 2017) (i.e. the source-collimator distance and the collimator-absorber distance). The same applies for the MPS design: Smeets et al. (2016) uses a modified MPS design from Pinto et al. (2014), Lin et al. (2017) used the design from Gueth et al. (2013) and Park et al. (2017) their own specific design. It is worth to note that the authors of the latter publication justified the need for a complementary study by the fact that the MPS geometries used in Lin et al. (2017); Smeets et al. (2016) were not “optimized”. Unfortunately the fact that the optimizations of KES and MPS designs were not performed according the same figures of merit can not ensure a fair comparison of the two designs. Actually from a general point of view the design of any collimated device is a compromise between detection efficiency and spatial resolution. While collimator features have been extensively investigated in the context of nuclear imaging (Gunter, 2004), no theoretical considerations have been proposed for the specific 1D collimation systems developed in the context of ion-range verification during hadrontherapy. Therefore, the first objective of this article is to present an analytical model based on geometrical considerations that allows for a theoretical estimation of the detection efficiency and spatial resolution, for both MPS and KES

designs. The intrinsic features of MPS and KES collimators are then derived and the model is validated by means of Monte Carlo (MC) simulations. In light of the predictions of this analytical model, the PG prototypes developed by the CLaRyS collaboration (Krimmer, 2015; Pinto et al., 2014) and the IBA group (Perali et al., 2014) are compared and their performances are assessed in terms of detection efficiency, spatial resolution and fall-off retrieval precision thanks to MC simulations.

## 2 Materials and Methods

### 2.1 Analytical model for spatial resolution and detection efficiency

In order to compare performance of different camera types and configurations, we derive an analytical model that predicts camera performances. Such a model allows to estimate variation in performance according to variation in geometrical parameters assuming perfect gamma detection in the absorber (full gamma absorption on the absorber entrance surface). Figure 1 presents the schemes of the MPS and KES cameras gathering the geometrical parameters defined in Table 1.




Figure 1: Schemes of the Multi-Parallel Slit (MPS) and the Knife-Edge Slit (KES) cameras. The definition of the various parameters is given in Table 1. The red line corresponds to a linear source perpendicular to the slit camera planes.  $\Lambda$  is the distance between the source point and the absorber entrance surface in the KES via the middle of the collimator slit.

#### 2.1.1 Spatial resolution

From a geometrical point of view, the spatial resolution, denoted Res, is characterized by the *detector unit field of view*. This is the portion of the source that can be seen through

	MPS	KES
Source-collimator distance	$d_1$	$d'_1$
Collimator-absorber distance	$d_2$	$d'_2$
Source-absorber distance	$L = d_1 + D + d_2$	$L = d'_1 + d'_2$
Collimator thickness	$D$	$T$
Camera pitch	$p$	$p'$
Knife-edge slit angle	–	$\alpha$
Slit width	$s$	$s$
Camera height	$H$	$H$

Table 1: Geometrical parameters of the MPS and KES cameras. One can also define the fill factor  $f$  of the MPS collimator:  $f = (p - s)/p$  (ratio between the septa width and the pitch). It is worth to note that the source-collimator distance and the collimator-absorber distance are not defined exactly in the same way in the two collimator designs.

a single camera unit: a single slit for the MPS and a single detector unit for the KES. The probability of a photon emitted at a given point along a linear source perpendicular to the slit plane to reach this detector unit can be described as an isosceles trapezoid whose width of the top segment corresponds to the slit width while the width of the base segment is equal to the sum of the slit width and the penumbra. We defined Res as the FWHM of this trapezoid, see eq 1 and 2.

$$\text{Res}^{\text{MPS}} = s \left( 1 + \frac{d_1}{D} \right) = p(1 - f) \frac{D + d_1}{D}, \quad (1)$$

$$\text{Res}^{\text{KES}} = s \left( 1 + \frac{d'_1}{d'_2} \right) = \frac{sL}{d'_2}. \quad (2)$$

If the collimator transparency is neglected, Res is fully defined by geometrical parameters, in particular the slit width  $s$ . However, prompt gamma have high penetration capability that can not be neglected in the case of KES where the collimator depth is very small in the region of the knife edge around the slit. Indeed, we define the *Effective Slit Opening*  $s_e$  that can be used in the evaluation of the field of view and the efficiency in place of the geometrical slit width. Metzler and Accorsi (2005) proposed a method to estimate the effective slit width, specifically to calculate the spatial resolution accurately. The proposed expressions were based on one-dimensional cuts through the pinhole geometry and can be applied directly to a knife-edge geometry without modification. Their approach is for a point source and dependent on the location of the source within the Res. For a source in the center, it simplifies to eq. 3 with  $\mu$  the linear attenuation length of the collimator material.

$$s_e = s + \frac{\ln 2}{\mu \tan \alpha}, \quad (3)$$

$$\text{Res}^{\text{KES}} = s_e \left( 1 + \frac{d_1'}{d_2'} \right) = \frac{s_e L}{d_2'}. \quad (4)$$

Hence, eq. 1 and eq. 4 represent the spatial resolution of the MPS and KES camera systems according to simple geometrical and gamma attenuation parameters (collimator distances, angle and collimator linear attenuation length). The spatial resolution is expressed in millimeters.

### 2.1.2 Detection efficiency

The probability that a gamma reaches a given detector unit is described by the solid angle  $\Omega_D$  of the detector unit in respect to the gamma emission point.

For MPS, the solid angle is composed of the azimuthal angle and the polar angle. With small angle approximation, it is described by eq. 5. Note that this is true only when  $\Omega_D$  is limited by the slit width, i.e. the absorber size and the distances (e.g.  $L$  and  $D$ ) are such that all photons that cross the collimator impinge on detector material.

For KES, the solid angle under which a point of the source sees a detector unit depends on the location of the point source,  $x$ , since the distance between source and detector changes significantly over the field of view of the camera. We consider the solid angle for a point of the source that is within the central part of the field of view of the crystal at location  $x$  on the detector plane (the origin of the detector plane facing the center of the slit). Under small angle approximation, the solid angle is given by eq. 6.

$$\Omega_D^{\text{MPS}} = \frac{H}{L} \frac{s}{D + d_1}, \quad (5)$$

$$\Omega_D^{\text{KES}} = \frac{H L p'}{\Lambda^3(x)} = \frac{H p'}{L^2 \left( 1 + \frac{x^2}{d_2'^2} \right)^{3/2}}. \quad (6)$$

with  $\Lambda$  the distance between the source point and the crystal in the KES.

Since we are interested in extended gamma emission source, let us consider now a linear source in front the cameras (red lines in Figure 1). In this case, the camera detection efficiency is determined not only by the point source detection efficiency (PSDE) characterized by  $\Omega_D$ , but also by the fraction of the source seen by the camera (the FOV factor)  $f_{\text{FOV}} = \frac{\text{Res}}{p}$  where  $p$  is the pitch of the camera. It is, in principle, larger than 1 since cameras are usually designed to see all the source without any hidden regions.

Hence, the detection efficiency of a detector unit (Eff) and therefore the detection efficiency of the whole camera located in front a linear source perpendicular to the camera plane can be expressed as:

$$\text{Eff} = \text{PSDE} \times f_{\text{FOV}}, \quad (7)$$

with PSDE the detection efficiency for a point-like source that sees a detector through the slit, and  $f_{\text{FOV}} = \frac{\text{Res}}{p}$  the FOV factor. The detection efficiency of MPS and KES are given by equations 8 and 9, respectively.

$$\begin{aligned}
\text{Eff}^{\text{MPS}} &= \frac{\Omega_D^{\text{MPS}} \text{Res}^{\text{MPS}}}{4\pi \frac{p}{p}}, \\
&= \frac{1}{4\pi} \frac{H}{L} \frac{s}{D+d_1} p(1-f), \frac{D+d_1}{D} \frac{1}{p}, \\
&= \frac{Hs(1-f)^2}{4\pi LD};
\end{aligned} \tag{8}$$

$$\begin{aligned}
\text{Eff}^{\text{KES}} &= \frac{\Omega_D^{\text{KES}} \text{Res}^{\text{KES}}}{4\pi \frac{p'}{p'}}, \\
&= \frac{s_e H}{4\pi d_2' L} \frac{1}{\left(1 + \frac{x^2}{d_2'^2}\right)^{3/2}}.
\end{aligned} \tag{9}$$

The formulas of detection efficiency and spatial resolution of MPS and KES cameras are gathered in Table 2. The most striking feature of these formulas is their similarities.

	MPS	KES
Effective slit width ( $s_e$ )	$s$	$s + \frac{\ln(2)}{\mu \tan(\alpha)}$
Spatial resolution (Res)	$s \left(1 + \frac{d_1}{D}\right)$	$s_e \left(1 + \frac{d_1'}{d_2'}\right)$
Detection efficiency (Eff)	$\frac{Hs}{4\pi LD}(1-f)$	$\frac{Hs_e}{4\pi Ld_2'} \left(1 + \frac{x^2}{d_2'^2}\right)^{-3/2}$
Collimator effective thickness ( $T_e$ )	$D \times f$	$T$

Table 2: MPS and KES detection efficiencies and spatial resolutions from the analytical model. The parameters of the cameras are defined in figure 1.  $\mu$  is the linear attenuation length of the collimator material.

To be more specific, let us define the following conditions that we will refer to as “MPS and KES Similarities Conditions” (MKSC):

- same distance between the source and the MPS collimator entrance and between the source and the middle of the KES collimator:  $d_1 = d_1'$  ;
- MPS collimator thickness equal to the KES collimator-absorber distance:  $D = d_2'$  ;
- same source-absorber distance  $L$  (which entails  $d_2 = 0$  with the previous conditions), same height  $H$  and same slit width  $s$ .

We can also consider perfect collimator condition (PCC) for which the collimators present infinite absorption capacity. In this condition,  $s_e = s$  and the filling factor  $f$  of the MPS can be decreased down to zero without degrading the collimation properties.

Combining these conditions (MKSC and PCC) then MPS and KES have strictly the same spatial resolution (Res) and the same detection efficiency (Eff) in the central part of the cameras ( $x = 0$ ). The main deviations between the actual MPS and KES performances are the following:

- the partial transparency of the collimators leads to:
  - an effective slit width of KES larger than the one of the MPS. In the case of the KES prototype of the IBA group using a tungsten alloy collimator (density of 18.5 g/cm<sup>3</sup>), the term accounting for collimator transparency  $\left(\frac{\ln(2)}{\mu \tan(\alpha)}\right)$  is about 5 mm (with  $\mu \sim 8 \times 10^{-1} \text{ cm}^{-1}$  in the 3-6 MeV range and  $\alpha = 63^\circ$ ) meaning that the effective slit  $s_e$  is about twice the actual slit width ( $s = 6 \text{ cm}$ ).
  - an actual fill factor that cannot be neglected: in the case of the MPS prototype of the CLaRyS collaboration it can be adjusted but a standard value is 0.4, leading to an actual detection efficiency about half the one of the perfect MPS collimator with  $f = 0$ .
- the KES design intrinsically leads to a detection efficiency that decreases when the source moves away from the center of the camera while the MPS detection efficiency remains constant over the whole camera field of view.

It is worth to note that the partial transparency of the collimators leads to an increase of the KES spatial resolution (Res) and a decrease of the MPS detection efficiency (Eff) with similar amplitudes (factor of 2). Since both Res and Eff are proportional to the slit width  $s$  one can expect that KES and MPS designs fulfilling MKSC should have very close actual performances if the KES slit width corresponds to the half of the MPS slit width.

For illustration purposes, Figure 2 shows the analytical model predictions for the MPS and KES designs compared in (Lin et al., 2017; Park et al., 2017; Smeets et al., 2016). Parameters have been extracted from the publications (Table 3). We can observe the compromise between detection efficiency and spatial resolution: the designs chosen in these comparisons lead to  $\text{Eff}^{\text{KES}}$  and  $\text{Res}^{\text{KES}}$  larger than MPS ones by factors ranging from  $\sim 3$  to  $\sim 7$ . The detection efficiency obtained by Park et al. (2017) is much larger than the other ones due to camera height ( $H$ ) increased by a factor of 2.

Parameters		(Smeets et al., 2016)		(Lin et al., 2017)		(Park et al., 2017)	
MPS	KES	MPS	KES	MPS	KES	MPS	KES
$d_1$	$d'_1$	250	220	200	300	180	200
$d_2$	$d'_2$	0	176	200	300	0	200
$L$	$L$	350	396	600	600	280	400
$D$	$T$	100	40	200	40	100	40
$s$	$s$	2.4	6	2	6	2	6
$H$	$H$	100	100	100	100	200	200
$f$		0.4		0.5		0.5	

Table 3: Parameters values of the MPS and KES designs compared in (Lin et al., 2017; Park et al., 2017; Smeets et al., 2016). All values are in millimeters except  $f$  without unit.

In the following, we compare the analytical model with MC simulations. We estimate, at first order, the intrinsic efficiency of the MPS and KES absorbers with the Beer-






Figure 2: Analytical model predictions for the MPS and KES designs compared in (Lin et al., 2017; Park et al., 2017; Smeets et al., 2016).

Lambert law assuming that a photon is detected as soon as it reaches the detector. At 4 MeV, the MPS absorber of the CLaRyS collaboration and the KES absorber of the IBA group have intrinsic efficiencies of 56% and 57% (respectively 3 cm thick BGO scintillators and 3 cm LYSO scintillators). *To be updated if we keep in MC simulations a BGO thickness of 5 cm*

## 2.2 Monte Carlo simulations

The objective of MC simulations is twofold: the analytical model verification (AMV) and the comparison of the MPS and KES prototypes developed by the CLaRyS collaboration and the IBA group, respectively. These two studies use slightly different setups that will be detailed in the following subsections.

### 2.2.1 Simulation tool

Imaging paradigms such as PG detection are evaluated against experiments, and often also with Monte Carlo (MC) simulations. For rarely occurring processes such as PG simulation, convergence to the truth to within acceptable statistical error can be slow. Therefore, we used in this study the vpgTLE variance reduction method described in Huisman et al. (2016). vpgTLE is a two stage process, where firstly a PG yield distribution image is estimated, which in the second stage is used as a PG source. Gate 7.2 (Sarrut et al., 2014) with Geant 4.10.02 and the QGSP\_BIC\_HP\_EMY physics list, commonly used for PG studies, are used in this work. Thanks to vpgTLE, simulations for about  $10^9$  protons (about  $6 \times 10^8$  photons) took 1-2 hours on a single core of an Intel(R) Core(TM) i7-3740QM.

### 2.2.2 Prototypes comparison

The MPS and KES prototypes are illustrated in Figure 3.




Figure 3: Schematic presentation of the two PG cameras considered in this study. All distances are in millimeters. The green arrow represents the proton beam. In red the collimation elements and in blue the detection elements. The dimensions were taken from [Pinto et al. \(2014\)](#) and [Perali et al. \(2014\)](#); [Sterpin et al. \(2015\)](#). Note that the two cameras have an identical height of 200 mm and that here they are not drawn to scale. The magnification factor of the KES prototype is 0.8. You used an absorber thickness of 5 cm for the CLaRyS prototype (value from Pinto 2014) but the actual value is 3 cm! Can it be corrected?

**The MPS prototype developed by the CLaRyS collaboration.** This camera intends to measure the whole PG profile to control ion-ranges in the patient with a field of view (FoV) of 300 mm. It consists of a bismuth germanium oxide (BGO) absorber and a tungsten alloy collimator. Each block has a sensitive volume of  $3.5 \times 3.8 \times 3 \text{ cm}^3$  and is streaked in  $8 \times 8$  pseudo-pixels of 4 mm in edge. In the optimization carried out by [Pinto et al. \(2014\)](#), parameters such as collimator pitch, axis-to-collimator and axis-to-detector were varied, and their impacts evaluated in terms of fall-off retrieval precision (FRP) and spatial resolution (sharpness of the fall-off region). Here, configuration 1 of their simulations (with relaxed constraints on spatial resolution) was chosen for its better FRP performance. One can notice that the pitch of the camera (8 mm) corresponds to an integer number of the BGO pseudo-pixel size (4 mm). The event selection from the gamma energy deposition  $E$  in the absorber is  $E > 1 \text{ MeV}$ . Finally this prototype makes use of TOF selection to reduce the neutron background. For the IBA C230 accelerator with a period of 10 ns, [Pinto et al. \(2014\)](#) chose a window of 4 ns around the PG maximum, based on experimental TOF spectra. Since the neutron background is essentially constant with the IBA C230 time structure ([Pinto et al., 2014](#)), TOF selection should allow for a  $\frac{6 \text{ ns}}{10 \text{ ns}} = 60\%$  noise reduction for this specific accelerator.

**The KES prototype developed by the IBA group** The purpose of this camera consists of verifying the BP position with a FoV of 100 mm. It consists of two rows of 20 LYSO blocs of 4 mm ( $p'$ )  $\times$  100 mm ( $H/2$ )  $\times$  30 mm (thickness) and a tungsten alloy collimator (Perali et al., 2014; Sterpin et al., 2015). The standard energy selection for this prototype is  $3 < E < 6$  MeV.

**Absorbers modeling** The absorbers of the two prototypes use different scintillators, photodetectors and data acquisition system. In this study, these differences are not implemented. Instead, for the sake of simplicity, the absorbers are modeled as monoblocks in which the event positions are stored in histograms with bins corresponding to the pitch of the cameras (8 mm for MPS and 4 mm for KES). When a gamma interacts with the absorber, an event is recorded if the integrated energy deposited in the absorber lies in the acceptable energy and TOF windows. Then the energy weighed barycenter of all interactions in the absorber is computed and the event position is assigned to the center of the bin where lies the energy weighed barycenter.

We compared the MPS and KES prototypes with their published properties:  $E > 1$  MeV and TOF for MPS and  $3 < E < 6$  MeV without TOF for KES. For the sake of completeness and to facilitate the interpretation of the camera performances, several combinations of energy and TOF selections will be studied as well.

To be updated according to Results section

### 2.2.3 Analytical Model Verification (AMV)

In order to allow for a direct comparison of MC simulations with the analytical model, simulations with perfect collimators and detectors will be performed. Table 4 gives an overview of the main cameras parameters used for AMV and prototypes comparison.

### 2.2.4 Background modeling

**Prototypes comparison** Background (BKGD) estimation in PG simulation is a difficult and still an unsolved issue (Huisman et al., 2016; Perali et al., 2014; Pinto et al., 2014; Sterpin et al., 2015). Simulations would ideally include beam nozzle and whole room modeling, but these are habitually omitted. TOF selection techniques can improve the signal-to-noise ratio (SNR) (Roellinghoff et al., 2014; Testa et al., 2008), but then it depends on the proper simulation of the beam accelerator time structure. As noted in Huisman et al. (2016), no validation for background in PG simulations has been performed at this time. In this study, we considered the IBA C230 accelerator time structure in which the neutron background is essentially constant. The estimates of background counts in the detector were taken from Perali et al. (2014); Pinto et al. (2014) and are given in Table 5

Per unit of bin length, the background yield of the MPS with TOF is therefore 4 times as low as the background seen with the KES (without TOF). For the KES camera the background with TOF can be obtained by multiplying the background with the same  $\frac{4ns}{10ns} = 0.4$  fraction as with the MPS.

		Prototypes comparison (PC)	Analytical model verif. (AMV)
Absorber	MPS	BGO	Perfect or like PC
	KES	LYSO	
Collimator		Tungsten alloy	Perfect or like PC
Energy selection	MPS	>1 MeV	>1 MeV
	KES	3–6 MeV	
TOF selection	MPS	TOF	no TOF
	KES	no TOF	
BKGD modeling		Exp. data based modeling	No modeling
Target		Cylindrical PMMA phantom	No
Beam		160 MeV proton	

Table 4: Summary of the main cameras parameters used for AMV and prototypes comparison (PC). The configuration of the column “Prototypes comparison” corresponds to the reference one with standard energy and TOF selections. In the case of AMV, the gamma source corresponds to the PGs emitted during the PMMA phantom irradiation with 160 MeV proton beam (first stage of the simulation tool, see section 2.2.1). However the target has been removed in the second stage of the simulation tool (PG detection) to avoid any attenuation and allow for a direct comparison of the detection efficiency with the AM predictions.

**Analytical Model Verification (AMV)** For the AMV, the background will be left out, as only gamma detection is modeled.

### 2.2.5 Beam and target

**Prototypes comparison** We use the test-case presented in [Perali et al. \(2014\)](#) because the KES detector properties, such as the background, were published for that scenario, again in order to remove any doubt that a difference in camera performance could be due to a difference in implementation or setup. A 160 MeV mono-energetic proton beam is shot into a cylindrical PMMA phantom (length 30 cm, radius 15 cm). Employing the batch method we performed 50 simulations for each experiment.

**Analytical Model Verification (AMV)** In the case of AMV, the same source is used (first stage of the simulation tool, see section 2.2.1). However the target has been removed in the second stage of the simulation tool (PG detection) to avoid any attenuation and allow for a direct comparison of the detection efficiency with the AM predictions.

## 2.3 Figures of merit

### 2.3.1 Analytical Model Verification

The comparison of the analytical model with Monte Carlo simulations will be performed on the two features predicted by the model, namely the detection efficiency and the

	MPS	KES
Bin width (mm) (pitch)	8	4
BKGD counts per bin	250	500
BKGD level per mm	$\sim 30$	$\sim 120$

Table 5: Background (BKGD) counts for  $10^9$  incident protons estimated from Figure 9 of [Pinto et al. \(2014\)](#) for MPS and from Figure 11 of [Perali et al. \(2014\)](#) for KES. The energy and TOF selections are given in Table 4 (column “Prototype comparison”).

spatial resolution. The detection efficiency can be computed by MC simulations as the ratio of the mean number of detected gammas over the number of emitted gammas for camera units seeing the PG profile. Regarding the spatial resolution, it is derived from the fall-off width of the simulated PG profiles. We define this width as the FWHM of the peak resulting from the computation of the PG profile first derivative.

### 2.3.2 Prototypes Comparison

In addition to the figures of merit mentioned for AMV, the fall-off retrieval precision (FRP) will be determined for its clinical relevance. The FRP is the standard deviation of the fall-off position (FOP) distribution, which is obtained by way of the batch method. Each of these can then be examined as function of the energy and TOF selections. A comprehensive description of our procedures for retrieving the FOP, FOW and FRP is provided in appendix A.

## 3 Results

### 3.1 Analytical Model Verification

Figure 4 shows PG profiles obtained with MPS and KES with perfect absorber, the same energy  $E > 1$  MeV and no TOF selection. [The figure is in principle up to date. To be double checked.](#)

Table 6 compares the detection efficiency and the spatial resolution obtained from the analytical model and Monte Carlo simulations.

[Main results:](#)

- Very good agreement between AM and MC (perfect cameras) except the discrepancy on the MPS resolution. Good agreement on the KES spatial resolution with physical camera.
- The discrepancy on the physical camera efficiency is due to the transparency of the collimators that is not taken into account in the AM.

[The emphasis in Table 6 shows that there is still a discrepancy to understand:](#)

[What we can see with the update of Figure 4 \(points in light gray of the physical camera\) is that the fall-off width of the physical camera seems slightly larger than the](#)




Figure 4: MPS and KES comparisons with perfect collimator and absorber, the same energy ( $E > 1$  MeV) and no TOF selection. The first row shows the datapoints in black crosses with the curvefit in red. The points in light gray are the data from the physical detector for otherwise the identical simulation. On the bottom row the first derivative of the red fit is shown, with the dashed line the Gaussian fit over the fall-off region (solid).

one of the perfect camera (I am maybe wrong). Brent, could you superimpose in Figure 4 the derivative of the physical camera PG profile (like you did for the PG profile graph on the top of the figure)? And double check the value of the fall-off width?

**Spatial Resolution** The results for the spatial resolution

**Efficiency**

**Fall-off Retrieval position** The MC-obtained fall-off retrieval position is also given in table 6. The model does not predict FRP, but results are given to see the effect of switching from a perfect camera model to the physical one. The closeness of the results implies that for the MPS in particular the opacity of the collimator and quality of PG detection in the absorber don't have much room for improvement.

## 3.2 Prototype Comparison

Performance under clinical conditions is the primary purpose of these PG cameras, and therefore we include the results of the clinical case study. As described in table 4, we took settings and corresponding background estimates for the camera's published settings to simulate each camera as-is during measurements of the same PMMA phantom.

### 3.2.1 PG profiles

Figure 5 shows the PG profiles with  $10^9$  protons, averaged over 50 runs. The averaging removes statistical fluctuations so that the base structure of the profiles can be best studied. The MPS camera is large enough to capture both the fall-off and fall-in positions, so the camera was displaced slightly (5cm in the distal direction) in order for the full

		MPS			KES		
		AM	MC	RD	AM	MC	
Res	Perfect camera	14.5 mm	22.5 mm	35%	13.5 mm	12.0 mm	13%
	Phys. camera		17.2 mm		23.7	19.3 mm	23%
Eff	Perfect camera	$6.66 \cdot 10^{-4}$	$6.49 \cdot 10^{-4}$	3%	$1.06 \cdot 10^{-3}$	$9.74 \cdot 10^{-4}$	9%
	Phys. camera	$3.76 \cdot 10^{-4}$	$6.61 \cdot 10^{-4}$	43%	$6.03 \cdot 10^{-4}$	$1.03 \cdot 10^{-3}$	41%
FRP	Perfect camera		0.90 mm			0.10 mm	
	Phys. camera		0.26 mm			0.33 mm	

Table 6: MPS and KES detection efficiencies ("Eff") and spatial resolution ("Res") from analytical models and MC. The predictions of the analytical model (AM column) are derived from the equations of table 2, at the center of the camera in the case of KES ( $x = 0$ ). MC data have been obtained with energy selection  $E > 1$  MeV. For the "Perfect camera" simulation, the collimator was adjusted to be 100% opaque, and detection with perfect precision and 100% efficiency. RD: relative deviation between AM and MC.

profile to show up. The camera's are positioned on the PG production fall-off, so the expected fall-off is at +50mm and 0mm for the MPS and KES respectively.




Figure 5: MPS and KES prototypes comparisons. MPS BGO absorber with 1-8 MeV energy selection and TOF selection. KES: LYSO absorber with 3-6 MeV and no TOF selection. On the top row the reconstructed profiles, averaged over 50 simulation each with  $10^9$  protons. On the bottom row in blue the first derivative of the profile, with in black the Gaussian fit in the fall-off region, with which the fall-off position and width will be estimated. Note that the horizontal scales are different and correspond to each camera's field of view.

We can see that, as expected, the MPS has higher PG counts, because of the wider energy selection (the PG production spectrum shows most production at lower energies). We see qualitatively the fall-off positions are at the expected coordinate, and that the BP is more pronounced in the MPS. At the tail end we see the backgrounds converge on

roughly the same level.

We expect different “tail levels”:

- Tail level (Figure 5) = Tail level without background (Figure 4) (tail due to collimator partial transparency – light gray points)+ background level (section 2.2.4). Table 7 shows the expected tails with and without background: the MPS expected tail with background (650) is in accordance with the PG profile tail in Figure 5. However it seems that the background has not been taken into account in the KES PG profile (tail level of  $\sim 800$  close to the tail without background of Figure 4).

	MPS	KES
Tail wo BKGD (Figure 5)	400	800
BKGD level (section 2.2.4)	250	500
Expected tail with BKGD	650	1300
“Observed” Tail in Figure Figure 5	$\sim 650$	$\sim 800$

Table 7: Expected background counts (in the tail region) for  $10^9$  incident protons. Table to be removed from the paper.

- Note that the cameras have different bin sizes (8 mm for MPS and 4 mm for KES). We need to divide the MPS counts by 2 to compare the tail levels.

Table 8 summarizes the spatial resolution and efficiency results. Even though the MPS had a more pronounced fall-off, we can see the spatial resolutions are roughly equivalent. *Some comment on efficiency?*

	MPS	KES
Res	19.4 mm	20.1 mm
Eff	$1.04 \cdot 10^{-3}$	$5.58 \cdot 10^{-4}$

Table 8: Detection efficiencies and spatial resolution of MPS and KES prototypes. Energy and TOF selection:  $E > 1$  MeV and TOF for MPS – 3-6 MeV and no TOF for KES.

### 3.2.2 FRP

Figure 6 shows the falloff retrieval precision (FRP) (standard deviation of the falloff position distributions) for the 2 prototypes as a function of the number of incident protons ( $10^7$ ,  $10^8$ ,  $10^9$ ). As expected, more primaries translate to a better FRP. Both cameras achieve submillimetric accuracy at  $10^9$  primaries, and millimetric precision with  $10^8$  primaries. With fewer than  $10^8$  primaries, the cameras, KES in particular, are outside the limits of interest of shift detection.

Figure 6 does not present all the data of Table 6. Brent, could you try to plot all the data with the following notations?






Figure 6: Fall-off retrieval precision of the prototypes as a function of the number of incident protons. Remove the title of the figure (which is btw incorrect: FRP= precision and not position).

- KES (orange), MPS (blue)
- Add the points with full symbols (TOF) and open symbols (no TOF)
- $E > 1$  MeV (full line) and  $E > 3$  MeV (dashed line)

I think that this set of FRP is important to discuss the performances of the MPS and KES prototypes as a function of energy and TOF selections.

Primaries	Time selection	TOF				none			
	Energy selection	1	3	3	1	1	3	3	1
	Camera	MPS	KES	MPS	KES	MPS	KES	MPS	KES
$10^9$	AMV perfect det.	0.90	0.10	NTNE					
$10^9$	detectors as proposed	<b>0.32</b>	0.56	0.36	0.57	0.31	0.60	0.46	<b>0.65</b>
$10^8$	detectors as proposed	<b>1.05</b>	1.69	1.30	1.96	1.04	1.93	1.34	<b>1.80</b>
$10^7$	detectors as proposed	<b>2.81</b>	7.96	3.65	12.9	2.96	14.8	19.3	<b>17.1</b>

Table 9: **THIS TABLE IS TO BE REMOVED. REPLACED WITH FRP PLOT.** Fall-off retrieval precision (defined as the standard deviation of the FOP over the number of times the simulation is ran. In bold, the cuts and TOF selections as proposed.

## 4 Discussion

Mention the limitations of the KES: detection efficiency not constant over the FOV + limited FOV

I (Etienne) plan to write this section once the Results section is completed.

### 4.1 Clinical implications

It is often stated PG cameras will be able to be used to track deviations at the spot-level. The cameras in this study required  $10^8$  primaries for an acceptable FRP. It is helpful to put this in clinical perspective, so we plot the structure of two treatment plans in figures 7 and 8. These treatment plans were produced and used for two clinical cases, the former considered of medium complexity, which is to say a fairly typical plan, and the second of high complexity. It stands to reason that from a clinical point of view, treatment monitoring becomes more interesting as the complexity of the plan increases, because plans become more complex in more difficult or sensitive cases, where a correct planned and delivered dose are desirable.




Figure 7: Treatment plan for a Meningioma case, which is considered a plan-type of medium complexity. On top: the spots are binned as function of energy and spot weight (number of protons), with high spot counts in red (hot spots) and low spot counts in blue (cool spots).

From figures 7 and 8 we can observe that spot-level monitoring is unlikely with the camera prototypes included in this study. In the complex case not a single spot has a prescription of  $10^8$  protons or more. Fortunately, nearly all energy layers exceed  $10^8$  primaries, so monitoring at the layer-level poses no issues for these cameras. Nevertheless,




Figure 8: Treatment plan for a large and complex pediatric brain and spine case. The left two fields are for the brain, the right two for the spine. The color scale chosen the same as in figure 7.

we hope the theoretical framework developed and validated in this paper will help move PG camera designs towards the realm where spot-level monitoring would become feasible.

## 5 Conclusion

We have presented and validated an analytical model that relates the collimator dimensions of Prompt Gamma cameras to a predicted spatial resolution and detection efficiency. Two collimator designs are modelled: multi-parallel slit and knife edge. The validation consisted of Monte Carlo simulations of two current Prompt Gamma camera prototypes, one of which has been employed in clinical setting. We also show the performance of these two prototypes in their proposed configuration with a mono-energetic spot in a PMMA phantom with background modelling. The prototypes reach millimetric precision with  $10^8$  primaries or more. The fall-off retrieval position performance is put into clinical context.

## 6 Acknowledgments

This work was partly supported by SIRIC LYric Grant INCa-DGOS-4664, LABEX PRIMES (ANR-11-LABX-0063 / ANR-11-IDEX-0007) and Fondation ARC. The authors would like to thank Marie-Claude Biston, Thomas Baudier and Gloria Vilches-Freixas for their

help finding the CT images and making the treatment plan. We also thank Erik Almha-gen and Uppsala University Hospital, Sweden for the treatment plan data presented in this paper.

## A Fall-off position and width estimation procedure

### A.1 Fall-off position estimation procedure

From a clinical perspective, the range estimate could be more interesting than FOP, because it can distinguish simple offset errors from patient morphological change. While the MPS camera was conceived for whole range PG profile detection, the KES camera FoV was chosen for BP region PG detection only. To make the comparison fair, only the FOP could be considered. Multiple approaches to extracting a FOP from the line profile have been proposed (Gueth et al., 2013; Janssen et al., 2014; Roellinghoff et al., 2014; Smeets et al., 2012; Sterpin et al., 2015). In preparatory work, a number of the proposed procedures were investigated. Significant sensitivity to free parameters on the final FOP estimates were seen. In summary, the FOP estimate depends greatly on the procedure, and often on having yields uncommon on the spot-level in clinical TPs, and also on an absence of unavoidable inhomogeneities.

Therefore the fitting method was not chosen as a topic for study in this paper. Instead, a simple method that works on most the data available to the authors was used: first a smoothed and interpolated spline function is fitted against the detected PG data points, after which a baseline and (distal) peak position are determined. The intersection of the spline with the half-height of the peak above the baseline is then taken as the FOP.

1. The measured PG profile is smoothed and interpolated with a smoothing spline function:

$$\sum_{i=1}^n (y_i - \hat{f}(x_i))^2 + \lambda \int_{x_1}^{x_n} \hat{f}''(x)^2 dx \quad (10)$$

where  $y_i$  is the measured PG profile and  $x_i$  the associated x-coordinates,  $\hat{f}(x_i)$  the estimate smoothed spline function and  $\lambda$  a smoothing parameter that determines the penalty for deviating from measurement in exchange for smoothness (second order derivatives are close to zero on smooth functions).  $\lambda = 0$  produces a perfect spline fit to the data, while  $\lambda \gg 1$  produces a horizontal line. We found that  $\lambda = 2$  provided an acceptable trade-off between overfitting to noise and removing too many features, which tends to happen for low statistic measurements.

2. The obtained function is plotted for 1024  $x_j$ , an number that provided a sufficiently high resolution. Any  $f(x_j) < 0$  are set to 0.
3. The global maximum is found.
4. The baseline is set equal to the lowest 25% of bins.
5. From the distal end backwards, the first maximum is taken as the distal most peak position, if it is above the threshold of 30% of the difference between baseline and global maximum. If no such point is found, the global maximum is taken as the distal most maximum.
6. The fall-off amplitude (FOA) is set to the difference between the distal maximum and baseline:  $FOA = max - baseline$ . The FOP is obtained by traversing the smoothed profile from the distal end towards the peak until  $y_j > \frac{1}{2}FOA$ .

The results of this procedure are illustrated in figure 9. Every PG profile was estimated 50 times, and so we obtained 50 estimates for the FOP. It is assumed that the FOPs follow

a Gaussian distribution, so the mean of the 50 realizations gives the best FOP estimate and the sigma gives the precision of the ability to estimate the best FOP. Comparing the 50 FOP estimates obtained from the CT with the 50 estimates obtained from the RPCT simulations, gives 2500 possible shift estimates. Again, the distribution of shifts should be centered at the true shift, while the sigma indicates how likely it is that this true shift is detected under the current conditions.




Figure 9: The top row demonstrates the fall-off determination procedure on the multi-parallel camera data; on the bottom row on knife-edge slit camera data. The left column is produced with a PG signal due to  $10^9$  primaries, while the right column was produced with  $10^7$  primary protons. In black crosses the measured PG counts are plotted. The smoothed data is shown in red. The green horizontal lines are drawn at the obtained distal maxima and baselines, while the vertical green lines shown the position of the distal maximum and the position of the fall-off. For the bottom-right plot, a history is visible where the procedure fails: the background induces an erroneous peak detection.

## A.2 Fall-off width estimation procedure

Figure 10 shows and illustration of the procedure for two selected profiles obtained with each collimator. To estimate the fall-off width (FOW), we smooth the profile in a similar manner as for the FOP estimation, as detailed in appendix A (row 1 in Figure 10). Then, a first and second order derivative is computed. On the first order derivative, a Gaussian is fitted (dashed line on second row), on the interval (solid line) between the profile maximum (Bragg Peak) and the first inflection point past the FOP on the second order derivative (here the 1D PG profile is at the baseline of the background, see third row). The Gaussian is fit with a fixed offset of zero, because the baseline of the background is




Figure 10: MPS and KES FOW estimation illustrated. idem x scale

zero. The full width half max of the fitted Gaussian is then taken as the FOW (second row).

## B Verification of the cameras

In [Priegnitz et al. \(2015\)](#) PG shifts due to beam energy shifts are studied for the KES camera: the *detectability* of the fall-off as function of the number of primaries. Here that simulation was recreated: a mono-energetic beam shoots into a waterbox at two energies. 50 realizations are generated with a 139 MeV beam energy, and 50 realizations with 144 MeV. At  $10^9$  primaries, the distributions are well separated with a shift of 8.3 mm (different from [Priegnitz et al. \(2015\)](#) because of the different material). In figure 13 in [Perali et al. \(2014\)](#) with  $10^9$  primaries a standard deviation of 1.5 mm is obtained, while here 1.21 and 1.14 mm were obtained. It is sufficient agreement to be confident of our setup and further results.

The KES prototype's sensitivity to accurate positioning with respect to the expected FOP was elaborated upon in [Sterpin et al. \(2015, Section IV.A.3\)](#): the detector response is, due to the KES collimator, not linear as with a parallel slit collimator. In this study, to make the comparison as fair as possible and avoid any bias, alignment on the FOP specific for each spot was ensured as follows: the intermediate PG source image of vpgTLE (equivalent to the PG emission) was projected on the beam axis, and then convolved with a Gaussian of  $\sigma = 8.5$  mm, which corresponds to the point spread function (PSF) with a FWHM of 20 mm used in [Priegnitz et al. \(2015\)](#) to approximate the detected profiles from the emitted profile. These profiles will be referred to as "PG + PSF" profiles. As a matter of fact, the MPS prototype has roughly the same PSF as the KES prototype so that "PG + PSF" fall-off position can be considered as the expected position for both cameras.

To verify the implementation of the MPS camera, the precision on the FOP, obtained with the procedure outlined in the previous paragraph, is compared to earlier results. In the caption of figure 9 in [Pinto et al. \(2014\)](#) it is stated that with  $10^8$  primaries a standard deviation of 1.3 mm is obtained for the detector design used here, which is about 20% different from the results obtained in this study: 1.63 and 1.54 mm.

## References

- P Gueth, D Dauvergne, N Freud, J M Létang, C Ray, E Testa, and D Sarrut. Machine learning-based patient specific prompt-gamma dose monitoring in proton therapy. *Physics in medicine and biology*, 58(13):4563–77, jul 2013. ISSN 1361-6560. doi: 10.1088/0031-9155/58/13/4563. URL <http://www.ncbi.nlm.nih.gov/pubmed/23771015>.
- Donald L. Gunter. Collimator Design for Nuclear Medicine. *Emission Tomography*, pages 153–168, jan 2004. doi: 10.1016/B978-012744482-6.50011-9. URL <https://www.sciencedirect.com/science/article/pii/B9780127444826500119>.
- Fernando Hueso-González, Fine Fiedler, Christian Golnik, Thomas Kormoll, Guntram Pausch, Johannes Petzoldt, Katja E. Römer, and Wolfgang Enghardt. Compton Camera and Prompt Gamma Ray Timing: Two Methods for In Vivo Range Assessment in Proton Therapy. *Frontiers in Oncology*, 6(April):1–13, 2016. ISSN 2234-943X. doi: 10.3389/fonc.2016.00080. URL <http://journal.frontiersin.org/article/10.3389/fonc.2016.00080>.
- Brent F B Huisman, J M Létang, É Testa, and D Sarrut. Accelerated Prompt Gamma estimation for clinical Proton Therapy simulations. *Physics in Medicine and Biology*, 61:7725–7743, 2016. ISSN 0031-9155. doi: 10.1088/0031-9155/61/21/7725.
- F M F C Janssen, G Landry, P Cambraia Lopes, G Dedes, J Smeets, D R Schaart, K Parodi, and F Verhaegen. Factors influencing the accuracy of beam range estimation in proton therapy using prompt gamma emission. *Physics in medicine and biology*, 59(15):4427–41, aug 2014. ISSN 1361-6560. doi: 10.1088/0031-9155/59/15/4427. URL <http://www.ncbi.nlm.nih.gov/pubmed/25049223>.
- Antje-Christin Knopf and Antony Lomax. In vivo proton range verification: a review. *Physics in medicine and biology*, 58(15):R131–60, aug 2013. ISSN 1361-6560. doi: 10.1088/0031-9155/58/15/R131. URL <http://www.ncbi.nlm.nih.gov/pubmed/23863203>.
- J. Krimmer, G. Angellier, L. Balleyguier, D. Dauvergne, N. Freud, J. Hérault, J. M. Létang, H. Mathez, M. Pinto, E. Testa, and Y. Zoccarato. A cost-effective monitoring technique in particle therapy via uncollimated prompt gamma peak integration. *Applied Physics Letters*, 110(15):154102, 2017a. ISSN 0003-6951. doi: 10.1063/1.4980103. URL <http://aip.scitation.org/doi/10.1063/1.4980103>.
- J. Krimmer, D. Dauvergne, J.M. Létang, and É. Testa. Prompt-gamma monitoring in hadrontherapy: A review. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, aug 2017b. ISSN 01689002. doi: 10.1016/j.nima.2017.07.063. URL <http://linkinghub.elsevier.com/retrieve/pii/S0168900217308380>.
- Jochen Krimmer. Development of a Compton camera for medical applications based on silicon strip and scintillation detectors. *Nuclear Instruments and Methods in Physics*



*Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, pages 98–101, 2015.

- Shunsuke Kurosawa, Hidetoshi Kubo, Kazuki Ueno, Shigeto Kabuki, Satoru Iwaki, Michiaki Takahashi, Kojiro Taniue, Naoki Higashi, Kentaro Miuchi, Toru Tanimori, Dogyun Kim, and Jongwon Kim. Prompt gamma detection for range verification in proton therapy. *Current Applied Physics*, 12(2):364–368, 2012. ISSN 15671739. doi: 10.1016/j.cap.2011.07.027. URL <http://dx.doi.org/10.1016/j.cap.2011.07.027>.
- Yi-Chun Lin, C.Y. Pan, K.J. Chiang, M.C. Yuan, C.H. Chu, Y.W. Tsai, P.K. Teng, C.H. Lin, T.C. Chao, C.C. Lee, C.J. Tung, and A.E. Chen. Monte carlo simulations for angular and spatial distributions in therapeutic-energy proton beams. *Radiation Physics and Chemistry*, 140:217 – 224, 2017. ISSN 0969-806X. doi: <https://doi.org/10.1016/j.radphyschem.2017.03.018>. URL <http://www.sciencedirect.com/science/article/pii/S0969806X17302980>. 2nd International Conference on Dosimetry and its Applications (ICDA-2) University of Surrey, Guildford, United Kingdom, 3-8 July 2016.
- Gabriela Llosá, Marco Trovato, John Barrio, Ane Etxebeste, Enrique Muñoz, Carlos Lacasta, Josep F. Oliver, Magdalena Rafecas, Carles Solaz, and Paola Solevi. First Images of a Three-Layer Compton Telescope Prototype for Treatment Monitoring in Hadron Therapy. *Frontiers in Oncology*, 6(February):14, 2016. ISSN 2234-943X. doi: 10.3389/fonc.2016.00014. URL <http://journal.frontiersin.org/Article/10.3389/fonc.2016.00014/abstract>.
- S D Metzler and R Accorsi. Resolution- versus sensitivity-effective diameter in pin-hole collimation: experimental verification. *Physics in Medicine and Biology*, 50(21):5005–5017, nov 2005. ISSN 0031-9155. doi: 10.1088/0031-9155/50/21/004. URL <http://www.ncbi.nlm.nih.gov/pubmed/16237237><http://stacks.iop.org/0031-9155/50/i=21/a=004?key=crossref.554d8443522872c9a60e734710db9839>.
- Chul Hee Min, Han Rim Lee, Chan Hyeong Kim, and Se Byeong Lee. Development of array-type prompt gamma measurement system for in vivo range verification in proton therapy. *Medical physics*, 39(4):2100–7, apr 2012. ISSN 0094-2405. doi: 10.1118/1.3694098. URL <http://www.ncbi.nlm.nih.gov/pubmed/22482631>.
- Harald Paganetti. Range uncertainties in proton therapy and the role of Monte Carlo simulations. *Physics in Medicine and Biology*, 57(11):R99–R117, 2012. ISSN 0031-9155. doi: 10.1088/0031-9155/57/11/R99. URL <http://stacks.iop.org/0031-9155/57/i=11/a=R99?key=crossref.4b6c83cb5125b4a360ead936994fceb4>.
- Seyoun Park, William Plishker, Harry Quon, John Wong, Raj Shekhar, and Junghoon Lee. Deformable registration of CT and cone-beam CT with local intensity matching. *Physics in Medicine and Biology*, 62(3):927–947, feb 2017. ISSN 0031-9155. doi: 10.1088/1361-6560/aa4f6d. URL <http://stacks.iop.org/0031-9155/62/i=3/a=927?key=crossref.6d51a84e851cc257d939edf081764a29>.
- Katia Parodi and Jerimy C. Polf. *In vivo* range verification in particle therapy. *Medical Physics*, 45(11):e1036–e1050, nov 2018. ISSN 0094-2405. doi: 10.1002/mp.12960. URL <https://onlinelibrary.wiley.com/doi/abs/10.1002/mp.12960>.

- G. Pausch, J. Petzoldt, M. Berthel, W. Enghardt, F. Fiedler, C. Golnik, F. Hueso-Gonzalez, R. Lentering, K. Romer, K. Ruhnau, J. Stein, A. Wolf, and T. Kormoll. Scintillator-Based High-Throughput Fast Timing Spectroscopy for Real-Time Range Verification in Particle Therapy. *IEEE Transactions on Nuclear Science*, 63(2):664–672, apr 2016. ISSN 0018-9499. doi: 10.1109/TNS.2016.2527822. URL <http://ieeexplore.ieee.org/document/7454861/>.
- I Perali, a Celani, L Bombelli, C Fiorini, F Camera, E Clementel, S Henrotin, G Janssens, D Prieels, F Roellinghoff, J Smeets, F Stichelbaut, and F Vander Stappen. Prompt gamma imaging of proton pencil beams at clinical dose rate. *Physics in Medicine and Biology*, 59(19):5849–5871, oct 2014. ISSN 0031-9155. doi: 10.1088/0031-9155/59/19/5849. URL <http://stacks.iop.org/0031-9155/59/i=19/a=5849?key=crossref.d1c598721e0b970b36f1c8a6ad1dd1a1>.
- M Pinto, D Dauvergne, N Freud, J Krimmer, J M Letang, C Ray, F Roellinghoff, and E Testa. Design optimisation of a TOF-based collimated camera prototype for online hadrontherapy monitoring. *Physics in medicine and biology*, 59(24):7653–7674, 2014. ISSN 1361-6560. doi: 10.1088/0031-9155/59/24/7653. URL <http://www.ncbi.nlm.nih.gov/pubmed/25415207>.
- Jerimy C Polf, Stephen Avery, Dennis S Mackin, and Sam Beddar. Imaging of prompt gamma rays emitted during delivery of clinical proton beams with a Compton camera: feasibility studies for range verification. *Phys Med Biol*, 60(18):7085–7099, 2015. ISSN 0031-9155. doi: 10.1088/0031-9155/60/18/7085. URL <http://dx.doi.org/10.1088/0031-9155/60/18/7085>.
- M Priegnitz, S Helmbrecht, G Janssens, and I Perali. Measurement of prompt gamma profiles in inhomogeneous targets with a slit camera. *Physics in Medicine and Biology*, 4849:4849, 2015. ISSN 0031-9155. doi: 10.1088/0031-9155/60/12/4849. URL <http://dx.doi.org/10.1088/0031-9155/60/12/4849>.
- F Roellinghoff, a Benilov, D Dauvergne, G Dedes, N Freud, G Janssens, J Krimmer, J M Létang, M Pinto, D Prieels, C Ray, J Smeets, F Stichelbaut, and E Testa. Real-time proton beam range monitoring by means of prompt-gamma detection with a collimated camera. *Physics in medicine and biology*, 59(5):1327–38, 2014. ISSN 1361-6560. doi: 10.1088/0031-9155/59/5/1327. URL <http://www.ncbi.nlm.nih.gov/pubmed/24556873>.
- David Sarrut, Manuel Bardiès, Nicolas Boussion, Nicolas Freud, Sébastien Jan, Jean-Michel Létang, George Loudos, Lydia Maigne, Sara Marcatili, Thibault Mauxion, Panagiotis Papadimitroulas, Yann Perrot, Uwe Pietrzyk, Charlotte Robert, Dennis R Schaart, Dimitris Visvikis, and Irène Buvat. A review of the use and potential of the GATE Monte Carlo simulation code for radiation therapy and dosimetry applications. *Medical Physics*, 41(6), 2014. doi: <http://dx.doi.org/10.1118/1.4871617>. URL <http://scitation.aip.org/content/aapm/journal/medphys/41/6/10.1118/1.4871617>.
- J Smeets, F Roellinghoff, D Prieels, F Stichelbaut, A Benilov, P Busca, C Fiorini, R Peloso, M Basilavecchia, T Frizzi, J C Dehaes, and A Dubus. Prompt gamma imaging

with a slit camera for real-time range control in proton therapy. *Physics in medicine and biology*, 57(11):3371–405, 2012. ISSN 1361-6560. doi: 10.1088/0031-9155/57/11/3371. URL <http://www.ncbi.nlm.nih.gov/pubmed/22572603>.

Julien Smeets, Frauke Roellinghoff, Guillaume Janssens, Irene Perali, Andrea Celani, Carlo Fiorini, Nicolas Freud, Etienne Testa, and Damien Prieels. Experimental comparison of knife-edge and multi-parallel slit collimators for prompt gamma imaging of proton pencil beams. *Frontiers in Oncology*, 6:156, 2016. ISSN 2234-943X. doi: 10.3389/fonc.2016.00156. URL <https://www.frontiersin.org/article/10.3389/fonc.2016.00156>.

E Sterpin, G Janssens, J Smeets, François Vander Stappen, D Prieels, Marlen Priegnitz, Irene Perali, and S Vynckier. Analytical computation of prompt gamma ray emission and detection for proton range verification. *Physics in Medicine and Biology*, 60(12):4915–4946, 2015. ISSN 0031-9155. doi: 10.1088/0031-9155/60/12/4915. URL <http://stacks.iop.org/0031-9155/60/i=12/a=4915?key=crossref.aabd8815e135401a22d165e343a7bac4>.

E. Testa, M. Bajard, M. Chevallier, D. Dauvergne, F. Le Foulher, N. Freud, J. M. Letang, J. C. Poizat, C. Ray, and M. Testa. Monitoring the Bragg peak location of 73 MeV carbon ions by means of prompt  $\gamma$ -ray measurements. *Applied Physics Letters*, 93(9): 1–10, 2008. ISSN 00036951. doi: 10.1063/1.2975841. URL <http://arxiv.org/abs/0809.0185http://dx.doi.org/10.1063/1.2975841>.

P.G. Thirolf, S. Aldawood, M. Böhmer, J. Bortfeldt, I. Castelhana, G. Dedes, F. Fiedler, R. Gernhäuser, C. Golnik, S. Helmbrecht, F. Hueso-González, H. v.d. Kolff, T. Kormoll, C. Lang, S. Liprandi, R. Lutter, T. Marinšek, L. Maier, G. Pausch, J. Petzoldt, K. Römer, D. Schaart, and K. Parodi. A Compton camera prototype for prompt gamma medical imaging. *EPJ Web of Conferences*, 117:05005, 2016. ISSN 2100-014X. doi: 10.1051/epjconf/201611705005. URL <http://www.epj-conferences.org/10.1051/epjconf/201611705005>.

Yunhe Xie, Hassan Bentefour, Guillaume Janssens, Julien Smeets, François Vander Stappen, Lucian Hotoiu, Lingshu Yin, Derek Dolney, Stephen Avery, Fionnbarr O’Grady, Damien Prieels, James McDonough, Timothy D. Solberg, Robert Lustig, Alexander Lin, and Boon-Keng K. Teo. Prompt gamma imaging for in vivo range verification of pencil beam scanning proton therapy. *International Journal of Radiation Oncology\*Biophysics*, 2017. ISSN 03603016. doi: 10.1016/j.ijrobp.2017.04.027. URL <http://www.sciencedirect.com/science/article/pii/S0360301617308489>.