

Pre-programmed commands

- Replace the use of remote command = no user action
- Strategy: “one-shot” (A-B-C-D…) OR “repetitive” (A-B-A-B…)

⚠ Combination of both strategies is not possible!





A
B
Event 1
A
B
C
D
Event 1
Event 2
Event 3
A
B
A
B
A
B
Change - Restore
Change - Restore
Change - Restore

One-shot pre-programmed commands

- Trigger on cycle index:
  - Define a specific parameters set for deployment
  - Switch to a standard parameters set after a while
- Trigger on date:
  - Apply winter/summer parameters set
  - Adapt acquisition methods during life of the float


_script.txt

⚠ **This file name must be respected!**




Repetitive pre-programmed commands

- Trigger on cycle periodicity:
  - Activate/deactivate a pattern
  - Activate/deactivate a sensor
  - Switch between day/night surface synchronization
  - Switch between standard/high resolution acquisition settings




Files – Script interpretation: !wait-cycle:x




(*) Pre-programmed changes are evaluated at the end of each cycle

Files – Script interpretation: !wait-date:yy-mm-dd




(*) Pre-programmed changes are evaluated at the end of each cycle

Files – Script interpretation: !wait-loop:one-out-of:x



(*) Pre-programmed changes are evaluated at the end of each cycle


!wait-cycle:4
...
Launch specific
Cycle 4
Standard
!wait-date:23-08-01
...
!wait-date:24-06-15
...
Standard
Date 1
Winter specific
Date 2
Standard
!wait-loop:one-out-of:3
...
!wait-loop:then
...
Standard
One out of 3
Then
Standard
One out of 3
Then