

FACULTY OF INFORMATION TECHNOLOGY  
BRNO UNIVERSITY OF TECHNOLOGY



**SEN**  
**Intelligent Sensors**

**Commissioning Heartbeat Sensor and Comparison  
Against Oximeter**

Jiří Záleský (xzales12)  
Adrián Tóth (xtotha01)

October 11, 2018

## Contents

<b>1</b>	<b>Project</b>	<b>2</b>
<b>2</b>	<b>Abstract</b>	<b>2</b>
<b>3</b>	<b>Introduction</b>	<b>2</b>
<b>4</b>	<b>Setup</b>	<b>3</b>
<b>5</b>	<b>Implementation</b>	<b>7</b>
<b>6</b>	<b>Conclusion</b>	<b>7</b>
<b>7</b>	<b>Appendix</b>	<b>8</b>
<b>8</b>	<b>References</b>	<b>9</b>

# 1 Project

The aim of this project is to commissioning a heart beat sensor whereby a few measurements are done, which are later compared to results of a valid device - oximeter. Prof. Ing., Dipl.-Ing. Martin Drahanský, Ph.D<sup>1</sup> is the project supervisor.

## 2 Abstract

The goal of this project is to demonstrate a method of heart beat measurement via infrared light. The next sections below, will describe further details, which were required to be done to the right functionality. The results of measurement were compared with a valid heart beat sensor, which measured heart beat in other way than the infrared method.

## 3 Introduction

The necessary facilities for the implementation of the project were:

- Device shown in Figure 1.




Figure 1: STM32 Nucleo board [4].

- Sensor shown in Figure 2.



Figure 2: Keyes KY-039 Finger Heartbeat Detection Sensor [1].

- Three female-to-female jumper wires.

---

<sup>1</sup>[www.fit.vutbr.cz/~drahan](http://www.fit.vutbr.cz/~drahan)

After, as we received the devices, we were facing to many problems. The main device representing the basic platform was not working properly so we had to ask for a new one. Later, the main device (shown in Figure 1) was replaced to a new one (see Figure 3) due to wrong functionality by the technical assistant of the project - Ing. Tomáš Goldmann<sup>2</sup>.




Figure 3: STM32 Nucleo board [2].

The device shown in Figure 3 - NUCLEO-F030R8 [6], was the main platform for the project. The device was equipped with a STM32F030R8 [7] microcontroller unit (MCU), which was designed and made to be suitable for a wide range of applications. MCU includes a set of peripherals through which it communicates with other devices such as the sensor shown in Figure 2. To create a communication pipeline, these units must be connected to each other via jumper wires. Wires provide a connection between the pins of units and so the pins have to be configured correctly.

## 4 Setup

Firstly, the components must be connected to each other in a right way. Sensor, as a slave component shown in Figure 5, is connected to the main device (master component). These two units are connected via 3 jumper wires to the corresponding pins. The Figure 4 and 5 have three common marked pins: *GND* (orange), *5V* (yellow), *PA0* (green); by which they are connected together. *PA0* indicates an analog input pin to the master component. A precise description of all the necessary pins is shown in Figure 6.

---

<sup>2</sup>[www.fit.vutbr.cz/~igoldmann](http://www.fit.vutbr.cz/~igoldmann)




Figure 4: NUCLEO-F030R8 [3] connection scheme.




Figure 5: Keyes KY-039 [1] connection scheme relating to Figure 4.

<b>B1</b>	button 1
<b>B2</b>	button 2
<b>LD2</b>	LED 2
<b>GND</b>	ground
<b>5V</b>	5 volt voltage supply
<b>PA0</b>	pin PA0
<b>mini USB</b>	mini USB port

Figure 6: Explanatory notes.

A connection scheme of the devices is shown in Figure 7. After the devices were correctly connected and attached to the computer via mini USB port (see Figure 8), they were ready to be programmed.




Figure 7: Device connection scheme.




Figure 8: Correct connection of devices.

As we were working on the implementation part of the project, there was a measurement anomaly. The results were affected by different input voltage on the pin. We asked for an advice from the technical assistant. He has recommended a different device connection which is shown in Figure 9. It required to use a protoboard with two  $10k\Omega$  resistors.




Figure 9: New device connection scheme.

The devices have been reconnected based on the given advice, which is shown in Figure 10.




Figure 10: New connection of devices.

## **5 Implementation**

...sensor code[5]...

## **6 Conclusion**

...

## 7 Appendix

Name	Value
n1	v1
n2	v2
n3	v3

Figure 11: Table of measurements.

## 8 References

- [1] *Keyes KY-039 Finger Heartbeat Detection Sensor*. [Image; Online; Accessed: 2018-10-07].  
Retrieved from:  
[https://images-na.ssl-images-amazon.com/images/I/71MfNkRMYDL.\\_SX425\\_.jpg](https://images-na.ssl-images-amazon.com/images/I/71MfNkRMYDL._SX425_.jpg)
- [2] *STM32 Nucleo board (NUCLEO-F030R8)*. [Image; Online; Accessed: 2018-10-07].  
Retrieved from: <https://statics3.seeedstudio.com/images/product/NUCLEO-F030RB.jpg>
- [3] *STM32 Nucleo board (NUCLEO-F030R8) - front*. [Image; Online; Accessed: 2018-10-07].  
Retrieved from: [https://media.digikey.com/Photos/STMicro%20Photos/MFG\\_NUCLEO.jpg](https://media.digikey.com/Photos/STMicro%20Photos/MFG_NUCLEO.jpg)
- [4] *STM32 Nucleo board (NUCLEO-F042K6)*. [Image; Online; Accessed: 2018-10-11].  
Retrieved from: <http://www.rhydolabz.com/images/ARD2648.jpg>
- [5] Luboš M.: *Arduino senzor tepu srdce*. 2016-12-13. [Online; Accessed: 2018-09-28].  
Retrieved from:  
<https://navody.arduino-shop.cz/navody-k-produktum/arduino-senzor-tepu-srdce.html>
- [6] STMicroelectronics: *NUCLEO-F030R8*. [Online; Accessed: 2018-10-07].  
Retrieved from: <https://os.mbed.com/platforms/ST-Nucleo-F030R8/>
- [7] STMicroelectronics: *STM32F030R8*. [Online; Accessed: 2018-10-07].  
Retrieved from: <https://www.st.com/en/microcontrollers/stm32f030r8.html>