
The Summary Never Reads the Mask: Localisation-Grounded Report Faithfulness in a Deployed Chest Radiograph System

Anonymous Author(s)

Affiliation

Address

email

Abstract

1 Deployed chest radiograph systems increasingly show a diagnostic prediction,
2 a localisation overlay, and a short natural-language summary together, and the
3 overlay is read as evidence that the summary is grounded in the image. We audit
4 one such deployed four-class system (Normal, pneumonia, tuberculosis, COVID-
5 19; DenseNet-121 classifier, Grad-CAM localisation compressed into a U-Net for
6 one-forward-pass serving). Its summary generator turns out to be a four-entry
7 static template keyed only on the predicted class: no property of the mask, not the
8 side, the region, the extent, nor even whether a region exists, reaches the printed
9 text. Over the system’s 245 already-scored test overlays we then show, with no
10 retraining, that gating the summary on measured mask properties changes what
11 can honestly be said; that on most films the honest output is to say nothing about
12 location, because the overlay sits outside the lung fields; and that permuting each
13 film’s mask onto a different film drops claim accuracy to near chance (15%). That
14 last result is the failure mode the current template’s silence on location happens to
15 sidestep and the one a naive descriptor-driven generator would walk straight into.
16 We report this on a deployed system and give a small evaluation protocol, crossing
17 heatmap source with generator type, for the general question of when a generated
18 radiology report is reading the image it describes.

19 1 Introduction

20 Chest radiograph systems built for low-resource triage bundle three outputs behind one click: a
21 diagnostic label, a heatmap that claims to show where the evidence lies, and a short clinical-language
22 summary. Each is usually evaluated on its own, if at all. Classification accuracy is scored against
23 a held-out set, localisation quality against some overlap reference, and the sentence is treated as
24 presentation. That split hides a specific, checkable failure: a system can be accurate, can show a
25 heatmap that looks like an explanation, and can still print a summary whose spatial claims, when it
26 makes any, land on the mediastinum or the cardiac silhouette instead of the lung.

27 We audit a system with exactly that shape. LungLens is a four-class chest X-ray classifier on a
28 DenseNet-121 backbone [Huang et al., 2017], with Grad-CAM [Selvaraju et al., 2017] localisation
29 compressed into a U-Net [Ronneberger et al., 2015] so an overlay costs one forward pass at serving
30 time, not a forward-plus-backward pass. Every prediction carries a summary. Reading the source
31 settles what writes it: a four-entry lookup keyed on the predicted class, joined to one of five fixed
32 sentences chosen by whether the prediction is confident, uncertain, or Normal. No coordinate, area,
33 side, or component count from the mask is ever an input. Nothing else feeds it, so the question
34 of whether the localisation subsystem contributes anything to the report has a direct answer, and
35 Section 4 is built to give it.

We turn that question into a measurement. Reusing the region-level statistics the system’s own failure analysis already computes over 245 scored test films (side, in-lung fraction, area fraction, plus a component count we add), we build a small descriptor-to-prose pipeline and compare it with the deployed template under four heatmap conditions (no mask, the served Grad-CAM map, the served U-Net map, and a mask lifted from a different film) crossed with two generators: the static template and one gated on the descriptors. The in-lung gate suppresses a spatial claim on 62% of films, because the served overlay’s centre of mass falls outside the lung fields on a majority of them. Permuting the mask onto a different film leaves the claim rate unchanged but drops the claim’s accuracy against the film’s own ground truth to 15%, the sanity check the saliency literature uses to tell a method that reads the input from one that does not [Adebayo et al., 2018].

Contributions. (1) An audit, from source, of what a deployed chest-radiograph summary generator conditions on, and a direct answer to whether its localisation subsystem reaches the text. (2) A retraining-free descriptor pipeline over the system’s own overlays with an in-lung gate, and a measurement of how often that gate fires on real served overlays. (3) A degenerate-mask sanity check in the tradition of Adebayo et al. [2018], applied to the generated report rather than the saliency map alone. (4) A small evaluation protocol crossing heatmap source with generator type, cells we could fill without new training or clinician time filled and the rest left open.

2 Related work

Classification and its explanations. CNN interpretation of chest radiographs matured with CheXNet [Rajpurkar et al., 2017], a 121-layer DenseNet at radiologist level for pneumonia on ChestX-ray14 [Wang et al., 2017]; DenseNet-121 [Huang et al., 2017] stays competitive with residual [He et al., 2016] and compound-scaled [Tan and Le, 2019] backbones at chest-radiograph data scales, which is why it is a fixed component here, not a contribution. Four-class Normal/pneumonia/TB/COVID work exists [Chowdhury et al., 2020, Kc et al., 2021], as do multi-task designs sharing a representation across classification and segmentation [Gündel et al., 2019]. Grad-CAM [Selvaraju et al., 2017] and its variants [Chattopadhyay et al., 2018] are the default explanation for such models [Singh et al., 2020], and a body of work asks whether the maps are faithful to the model or only plausible: Adebayo et al. [2018] show several saliency methods survive randomising the model’s weights, and Kindermans et al. [2019] extend the concern. Our degenerate-mask control carries that test from the map to the report built on it: if an unrelated map leaves the text’s claims as accurate, the map is not doing the work the text implies.

Generated reports. A separate line generates full narrative reports from radiographs [Jing et al., 2018, Chen et al., 2020], usually scored with text-overlap metrics that reward fluency without checking image support. Factual-consistency metrics over clinical entities [Miura et al., 2021], region-grounded generation that ties each sentence to an image region [Tanida et al., 2023], and biomedical vision-language pretraining [Bannur et al., 2023] all respond to that gap. LungLens predates this: its summary is closer to a caption, and we ask the same faithfulness question at the point a much simpler system first gets the chance to fail it. Earlier hand-crafted CXR features [Jaeger et al., 2014, Cortes and Vapnik, 1995, Breiman, 2001] and the COVID-era CNN surveys [Sethy et al., 2020, Wang et al., 2020, Apostolopoulos and Mpesiana, 2020, Ozturk et al., 2020, Minaee et al., 2020] are background for the backbone and dataset choices below.

3 System

Data. The classifier trains on eight public chest-radiograph sources, deduplicated by content hash to 37,553 images and sampled to 32,000, arranged so each of the four classes draws from at least two acquisition pipelines, a deliberate hedge against the source-fingerprinting failure of Zech et al. [2018]; Section 6 reports the residual. Images are border-cropped 8% at train and inference time to drop burned-in corner text, resized to 224×224 , and ImageNet-normalised. The split is patient-grouped and reproducible: because every directory listing is sorted before use, re-running split construction from the raw files reproduces the test accuracy to four significant figures on the same 4,759 held-out images. That guarantee is not free. An earlier, unsorted version of the same code moved the reported test accuracy by nearly two points purely as a function of filesystem directory-traversal order.




Figure 1: Dice between the served U-Net mask and a freshly computed live Grad-CAM map, same film and class, over all 245 served films. The distribution is heavy at zero, not merely low on average: a generator that reads a location from one of these two sources cannot assume the other agrees with it.

87 **Classifier.** A DenseNet-121 [Huang et al., 2017], ImageNet-initialised, fine-tuned under class-
 88 weighted cross-entropy (inverse-frequency weights, mean 1.0) for up to 40 epochs with early stopping
 89 on validation macro-F1, Adam at 10^{-4} , batch 32. It is a fixed upstream component; its full evaluation
 90 is in Appendix A.

91 3.1 Localisation: explanation amortisation, not knowledge distillation

92 A second network, a U-Net-style encoder-decoder (MultiTaskUNet: four DoubleConv blocks,
 93 $64 \rightarrow 128 \rightarrow 256 \rightarrow 512$ channels, skip connections, 7,705,224 parameters, counted from the
 94 served checkpoint), learns to reproduce the classifier’s Grad-CAM map in one forward pass, avoiding
 95 the backward pass Grad-CAM needs. The frozen DenseNet (no gradient enters it here) produces
 96 a Grad-CAM map at `denseblock4.denselayer16.conv2` for each image’s *ground-truth* label,
 97 ReLU’d, resized to 224×224 , and divided by its own maximum; targets are held in memory for the
 98 run and never cached to disk. The U-Net is trained against that target with an unweighted Dice+BCE
 99 loss (Adam, 3×10^{-4} , batch 8) over 3,000 images for 20 epochs; the served channel is the class
 100 being visualised.

101 We call this *explanation amortisation*, not knowledge distillation. Nothing is shared between the two
 102 networks: no layers, no logit or feature matching, no gradient path from student to teacher. The only
 103 channel is one 224×224 per-class scalar field, computed once and discarded. Very little survives a
 104 channel that narrow, and the cost is measurable: on the 245 served films, the median Dice between
 105 the U-Net’s mask and a fresh live Grad-CAM map for the same film and class is 0.22, with 27%
 106 showing no overlap and 15% above 0.5 (Figure 1). What the compression buys is a real serving-cost
 107 saving (Section 5); stating that saving together with this cost is the honest description.

108 3.2 Report generation

109 The version under review disposed of this subsystem in eleven words. Section 5’s central finding
 110 needs its exact behaviour, so we specify it here.

111 **What it is.** A static Python dictionary keyed on the predicted class (Normal, Pneumonia, Tuberculosis,
 112 Covid-19), joined to one of five fixed sentences selected by three booleans: is the prediction below
 113 the confidence threshold, does the visualised class equal a confident non-Normal prediction, and was
 114 a mask computed at all. No model, prompt, decoding parameters, or learned weights; the full listing
 115 is Appendix B.

116 **Inputs.** The argmax of the softmax output and its probability; a flag for whether the visualised class
 117 is the top prediction; a flag for whether that top prediction is Normal; and, for the trailing provenance
 118 line only, which subsystem produced the mask. No coordinate, area, side, zone, or component count
 119 from the mask is ever read.

120 **Confidence gate.** Below 60% top-class probability the report reads “Uncertain” with a fixed
 121 cautionary sentence; the only other non-standard path is a caught exception on malformed input.




Figure 2: System overview. Stage 1 trains the served DenseNet-121 classifier. Stage 2 supervises a U-Net student on Grad-CAM targets cached from the frozen classifier (stop-gradient marked): no layers, logits, or features are shared, so the only channel is one 224×224 scalar field per image, amortising the backward pass out of serving. At inference the classifier supplies label and confidence; the localiser supplies gated region descriptors; both arrive at the report generator, a class-keyed template (S0), which alone determines the printed summary. Tags mark where H0–H5 and S0–S4 enter.

Failure this produces. The “confident non-Normal” branch does not check whether a region survived the display threshold, so the text can assert “the outline marks the most influential area” on a film where nothing was outlined. This happens on 2 of the 245 scored films, one a confidently wrong tuberculosis call (95.7%) on a radiograph whose true label is Normal, with no thresholded region anywhere in the image (Appendix C).

4 Experimental design

Classification accuracy and localisation quality are already measured (Section 3, Appendix A). The new experiment asks one thing: does the summary change, and change correctly, as a function of the heatmap it is given. We cross a heatmap condition with a generator condition over the 245 test films already scored for the failure analysis (all 102 pneumonia-predicted-Normal errors, all 92 Shenzhen and 12 Montgomery test films, and 40 stratified correct films).

Heatmap conditions. H0, no mask (class and confidence only). H1, the live Grad-CAM map. H3, the served U-Net map. H5, a degenerate control: the mask of a uniformly random *different* film under a fixed derangement (seed 42), so each film keeps its own class and confidence but gets someone else’s spatial evidence.

Generator conditions. S0, the deployed generator (Section 3.2), rebuilt from source. S1, a template stating side, vertical third, extent, and focality from the region descriptors, gated to make no claim when the thresholded region’s in-lung fraction is below 0.5 or when no region survives the display threshold. S2, a text-only LLM given class, confidence, and descriptors, is out of scope here (Section 6).

Table 1: The evaluation grid. Filled cells report a measured number (Section 5); open cells are proposed and not run. S0 makes no spatial claim under any heatmap, so its row collapses by construction, which is the headline. S1’s heatmap cells are reported jointly as H_actual; a per-film H1-vs-H3 split is left for a revision.

Generator (rows) / Heatmap (cols)	H0 (none)	H1 (Grad-CAM)	H3 (U-Net)	H5 (permuted)
S0 (deployed template)	measured	measured	measured	n/a (static)
S1 (descriptor template)	measured	measured	measured	measured
S2 (text LLM)	open	open	open	open
H2 (Grad-CAM++), H4 (second backbone), S3/S4 (VLM): open, see Section 6				

Region descriptors. For each film the served overlay’s thresholded region (the app’s own thresholds and cleaning: 0.50 for a U-Net mask, 0.45 for Grad-CAM, 5×5 elliptical opening, components under 0.3% of the frame dropped) is reduced to laterality (left/right/outside both, from the centroid against a separate lung-field segmenter), vertical zone (upper/mid/lower third, from the centroid; new here), extent (focal <5%, patchy 5–25%, extensive >25%, from area fraction), focality (single vs. multifocal, from component count; new here), and in-lung fraction (the gate). Laterality, extent, and in-lung fraction were already in the system’s analysis code; the two new descriptors were checked against it by matching recomputed area fractions to a median absolute difference of 0.0001 over the 204 films with a region.

Cells run (no GPU, no training): S0 as baseline; S1 crossed with H0, H_actual (whichever of H1/H3 the system served for that film), and H5. A per-film H1-versus-H3 pairing, computing both maps fresh for every film, needs the 245 radiographs staged for a forward-plus-backward pass each and is left for a revision.

Metrics. Spatial-claim rate (how often a claim is made). Grounding precision (does a claim match the film’s descriptors). Phantom-claim rate (does the generator assert evidence that did not survive the threshold). Anatomical plausibility (does a claimed location fall in the lung, true by construction under the gate). H5 grounding accuracy (does a claim from another film’s descriptors match this film’s truth). Latency, reported in prose.

5 Results

Classifier, in one paragraph. The DenseNet-121 reaches 96.18% accuracy and 95.77% macro-F1 on the patient-grouped 4,759-image test set, with per-class recall from 92.2% (pneumonia) to 98.5% (Normal). The dominant error, 102 pneumonia films predicted Normal (56% of all 182 errors), sits in a single source folder whose label, *lung opacity*, is a broader radiological finding than pneumonia, and several of those films have clear lung fields on inspection: label noise as much as model failure. Full detail is in Appendix A.

5.1 The summary does not move with the mask, because it never reads it

Under every heatmap condition (H0, H1, H3, H5), S0 prints byte-identical text for a given film, because none of its five branches take mask geometry. The explainability apparatus and the report apparatus are disconnected. In Table 1 this is the collapsed S0 row: a property of the generator, not of our harness.

5.2 Gating on measured evidence changes what can honestly be said

S1, gated on in-lung fraction, states a spatial claim on 92 of 245 films (37.6%). On the other 153, either no region survives the display threshold (41 films, 16.7%) or the region’s centre of mass is outside the lung fields. The gate rate follows directly from an already-published measurement: 55% of the served overlays put the majority of their region outside the lung (Appendix D). Figure 3 shows a pair on one film: S0’s silence on location above, S1’s grounded “left lung, mid zone, patchy, multifocal” below, read from the film’s own descriptors. Without localisation the generator can only restate the class; with it, it can also say where, and the accuracy of that “where” is bounded by Section 3.1’s Dice.

S0 (deployed). Prediction: Covid-19 (99.1%). Findings consistent with Covid-19. Warmer areas indicate possible bilateral ground-glass opacities. Heatmap: region driving the Covid-19 prediction. The outline marks the most influential area. *Overlay source: disease-segmentation U-Net.*

S1 (descriptor, H_{actual}). Prediction: Covid-19 (99.1%). Findings consistent with Covid-19. Warmer areas indicate possible bilateral ground-glass opacities. Heatmap evidence: left lung, mid zone, patchy, multifocal. *Overlay source: disease-segmentation U-Net.*

(Reconstructed from the template in Appendix B and the film’s measured descriptors; *figures/paired_summary.pdf* overrides.)

Figure 3: The same radiograph, predicted Covid-19 at 99.1% confidence, under both generators. S0 is identical whether or not a mask is available; S1 adds a spatial claim built from the film’s own descriptors. Both are language over measured pixels, not a clinical finding.

5.3 The degenerate-mask control

Permuting each film’s mask onto a different film under a fixed derangement leaves S1’s claim rate unchanged (92/245, since a permutation preserves the marginal of which films pass the gate), but the claim’s accuracy against the film’s true side, zone, and extent falls to 14/92 (15.2%), near the chance level the skewed zone and side marginals imply. This is Adebayo et al. [2018]’s check moved from the saliency map to the report a naive descriptor-driven generator would write: if the stated location is no more accurate with the right mask than the wrong one, the localisation subsystem is not grounding the sentence. The same bound acts on any real generator drawing from H1 or H3: the two maps disagree at median Dice 0.22 (Section 3.1), so a generator reading one of them inherits that ceiling. A localisation-grounded report is exactly as trustworthy as the localisation under it, and no more.

Latency. On a CPU laptop (i7-13620H, 10 threads), the served pipeline averages 293 ms (sd 81) when the distilled U-Net supplies the mask, the common case at 91% of test films, and 560 ms (sd 22) on the live Grad-CAM fallback (forward-plus-backward, 9%), a test-set-weighted 318 ms. The roughly 2× gap is what the U-Net buys on hardware with no GPU. Report generation is a dictionary lookup and adds nothing measurable; a hosted-API generator would need its network latency reported separately.

6 Proposed protocol and limitations

What is proposed, not measured. Table 1’s open cells are not implied results. H2 (Grad-CAM++) and S3/S4 (a vision-language model given the radiograph, or the radiograph plus the overlay) are inference- or API-cost work scoped out by design. H4, Grad-CAM from a second, independently trained backbone, is the most informative open cell, since it separates “the summary tracks the pathology” from “the summary tracks this model,” and it needs a second classifier trained end to end. No clinician was available; the automated metrics are faithfulness proxies, not substitutes. An LLM-as-judge score, if used later, must report its agreement with human labels on a subsample or not be used.

Dataset-source bias. Despite multi-source classes and identical 8% cropping, per-source test accuracy runs from 83.3% (Montgomery, 12 images) to 99.5% (TBX11K). The two weakest sources fail oppositely in the localisation data: Shenzhen’s overlays are diffuse (median 41% of the frame on TB-positive films), Montgomery’s are mostly absent (a quarter of its films produce no region). Source and class stay partly confounded in an eight-source composite, so 96.18% is an optimistic upper bound, not a clean prospective number.

Label noise. 82 of the 102 pneumonia-to-Normal errors come from a folder whose label, *lung opacity*, is a mapping decision rather than a diagnosis, and several inspected films have clear lung fields. Some of this error mode is model failure and some is the model correctly reading a film that should not have been labelled positive; the two are not separable after the fact without the original read.

Reproducibility. An earlier, unsorted directory-traversal order moved the split from 22,419/4,785/4,796 to 22,420/4,759/4,821 images and a re-scored checkpoint from 96.62% to 98.11%,

219 nearly two points from enumeration order alone on an otherwise identical pipeline and seed. Sorting
220 every listing before use is now load-bearing for every number here being reproducible.

221 **No clinical validation.** LungLens is a research and educational system, not evaluated prospectively
222 against radiologist reads. Its summary, static template or descriptor-gated, is language over measured
223 pixels, not a clinical finding, and the deployed interface should say so, not only this paper.

224 7 Broader impact

225 A confident label, an apparently explanatory heatmap, and a fluent summary shown together are
226 more persuasive than their parts justify, especially to a non-specialist in the low-resource setting
227 this system targets. Our central finding, that a plausible summary can be entirely disconnected from
228 the localisation evidence beside it, is a caution against that persuasiveness, not an endorsement of
229 shipping the descriptor-gated version as a clinical tool. Any future report layer here should carry
230 the non-diagnostic disclaimer this paper argues is missing, and be checked for faithfulness before
231 fluency.

232 8 Conclusion

233 LungLens pairs an accurate classifier, a fast localisation shortcut, and a clinical-language summary.
234 Auditing the third from source shows it never reads the first two: the summary depends only on the
235 predicted class and its confidence. A small retraining-free gate on the system’s own region statistics
236 shows what changes when it does read them, and most of the time the honest output is to say nothing
237 about location, because the evidence sits outside the lung. The degenerate-mask control makes the
238 point: without a faithfulness check, a more sophisticated report generator would not be safer, only
239 more convincing.

240 References

- 241 Julius Adebayo, Justin Gilmer, Michael Muelly, Ian Goodfellow, Moritz Hardt, and Been Kim. Sanity checks
242 for saliency maps. In *Advances in Neural Information Processing Systems (NeurIPS)*, volume 31, pages
243 9505–9515, 2018.
- 244 Ioannis D. Apostolopoulos and Tzani A. Mpesiana. Covid-19: Automatic detection from X-ray images utilizing
245 transfer learning with convolutional neural networks. *Physical and Engineering Sciences in Medicine*, 43(2):
246 635–640, 2020.
- 247 Shruthi Bannur, Stephanie Hyland, Qianchu Liu, Fernando Perez-Garcia, Maximilian Ilse, Daniel C. Castro,
248 Benedikt Boecking, Harshita Sharma, Kenza Bouzid, Anja Thieme, Anton Schwaighofer, Maria Wetscherek,
249 Matthew P. Lungren, Aditya Nori, Javier Alvarez-Valle, and Ozan Oktay. Learning to exploit temporal
250 structure for biomedical vision–language processing. In *Proceedings of the IEEE/CVF Conference on*
251 *Computer Vision and Pattern Recognition (CVPR)*, pages 15016–15027, 2023.
- 252 Leo Breiman. Random forests. *Machine Learning*, 45(1):5–32, 2001.
- 253 Aditya Chattopadhyay, Anirban Sarkar, Prantik Howlader, and Vineeth N. Balasubramanian. Grad-CAM++:
254 Generalized gradient-based visual explanations for deep convolutional networks. In *IEEE Winter Conference*
255 *on Applications of Computer Vision (WACV)*, pages 839–847, 2018.
- 256 Zhihong Chen, Yan Song, Tsung-Hui Chang, and Xiang Wan. Generating radiology reports via memory-driven
257 transformer. In *Proceedings of the 2020 Conference on Empirical Methods in Natural Language Processing*
258 *(EMNLP)*, pages 1439–1449, 2020.
- 259 Muhammad E. H. Chowdhury, Tawsifur Rahman, Amith Khandakar, Rashid Mazhar, Muhammad Abdul Kadir,
260 Zaid Bin Mahbub, Khandakar Reajul Islam, Muhammad Salman Khan, Atif Iqbal, Nasser Al Emadi, Mamun
261 Bin Ibne Reaz, and Mohammad Tariqul Islam. Can AI help in screening viral and COVID-19 pneumonia?
262 *IEEE Access*, 8:132665–132676, 2020.
- 263 Corinna Cortes and Vladimir Vapnik. Support-vector networks. *Machine Learning*, 20(3):273–297, 1995.
- 264 Sebastian Gündel et al. Multi-task learning for chest x-ray abnormality classification on noisy labels. *arXiv*
265 *preprint arXiv:1905.06362*, 2019.

266 Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. Deep residual learning for image recognition. In
267 *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, pages 770–778,
268 2016.

269 Gao Huang, Zhuang Liu, Laurens van der Maaten, and Kilian Q. Weinberger. Densely connected convolutional
270 networks. In *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*,
271 pages 4700–4708, 2017.

272 Stefan Jaeger, Alexandros Karargyris, Sema Candemir, Les Folio, Jenifer Siegelman, Fiona Callaghan, Zhiyun
273 Xue, Kannappan Palaniappan, Rahul K. Singh, Sameer Antani, George Thoma, Yi-Xiang Wang, Pu-Xuan Lu,
274 and Clement J. McDonald. Automatic tuberculosis screening using chest radiographs. *IEEE Transactions on*
275 *Medical Imaging*, 33(2):233–245, 2014.

276 Baoyu Jing, Pengtao Xie, and Eric Xing. On the automatic generation of medical imaging reports. In *Proceedings*
277 *of the 56th Annual Meeting of the Association for Computational Linguistics (ACL)*, pages 2577–2586, 2018.

278 Kamal Kc, Zhendong Yin, Mingyang Wu, and Zhilu Wu. Evaluation of deep learning-based approaches for
279 COVID-19 classification based on chest X-ray images. *Signal, Image and Video Processing*, 15(5):959–966,
280 2021.

281 Pieter-Jan Kindermans, Sara Hooker, Julius Adebayo, Maximilian Alber, Kristof T. Schütt, Sven Dähne, Dumitru
282 Erhan, and Been Kim. The (un)reliability of saliency methods. In *Explainable AI: Interpreting, Explaining and*
283 *Visualizing Deep Learning*, volume 11700 of *Lecture Notes in Computer Science*, pages 267–280. Springer,
284 2019.

285 Shervin Minaee, Rahele Kafieh, Milan Sonka, Shakib Yazdani, and Ghazaleh Jamalipour Soufi. Deep-COVID:
286 Predicting COVID-19 from chest X-ray images using deep transfer learning. *Medical Image Analysis*, 65:
287 101794, 2020.

288 Yasuhide Miura, Yuhao Zhang, Emily Bao Tsai, Curtis P. Langlotz, and Dan Jurafsky. Improving factual
289 completeness and consistency of image-to-text radiology report generation. In *Proceedings of the 2021*
290 *Conference of the North American Chapter of the Association for Computational Linguistics (NAACL)*, pages
291 5288–5304, 2021.

292 Tulin Ozturk, Muhammed Talo, Eylul Azra Yildirim, Ulas Baran Baloglu, Ozal Yildirim, and U. Rajendra
293 Acharya. Automated detection of COVID-19 cases using deep neural networks with X-ray images. *Computers*
294 *in Biology and Medicine*, 121:103792, 2020.

295 Pranav Rajpurkar, Jeremy Irvin, Kaylie Zhu, Brandon Yang, Hershel Mehta, Tony Duan, Daisy Ding, Aarti Bagul,
296 Curtis Langlotz, Katie Shpanskaya, Matthew P. Lungren, and Andrew Y. Ng. CheXNet: Radiologist-level
297 pneumonia detection on chest X-Rays with deep learning. *arXiv preprint arXiv:1711.05225*, 2017.

298 Olaf Ronneberger, Philipp Fischer, and Thomas Brox. U-Net: Convolutional networks for biomedical image
299 segmentation. In *Medical Image Computing and Computer-Assisted Intervention (MICCAI)*, volume 9351 of
300 *Lecture Notes in Computer Science*, pages 234–241. Springer, 2015.

301 Ramprasaath R. Selvaraju, Michael Cogswell, Abhishek Das, Ramakrishna Vedantam, Devi Parikh, and Dhruv
302 Batra. Grad-CAM: Visual explanations from deep networks via gradient-based localization. In *Proceedings*
303 *of the IEEE International Conference on Computer Vision (ICCV)*, pages 618–626, 2017.

304 Prabira Kumar Sethy, Santi Kumari Behera, Pradyumna Kumar Ratha, and Preesat Biswas. Detection of
305 coronavirus disease (COVID-19) based on deep features and support vector machine. *International Journal*
306 *of Mathematical, Engineering and Management Sciences*, 5(4):643–651, 2020.

307 Amitojdeep Singh, Sourya Sengupta, and Vasudevan Lakshminarayanan. Explainable deep learning models in
308 medical image analysis. *Journal of Imaging*, 6(6):52, 2020.

309 Mingxing Tan and Quoc V. Le. EfficientNet: Rethinking model scaling for convolutional neural networks. In
310 *Proceedings of the 36th International Conference on Machine Learning (ICML)*, pages 6105–6114, 2019.

311 Tim Tanida, Philip Müller, Georgios Kaissis, and Daniel Rueckert. Interactive and explainable region-guided
312 radiology report generation. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern*
313 *Recognition (CVPR)*, pages 7433–7442, 2023.

314 Linda Wang, Zhong Qiu Lin, and Alexander Wong. COVID-Net: A tailored deep convolutional neural network
315 design for detection of COVID-19 cases from chest X-ray images. *Scientific Reports*, 10(1):19549, 2020.

- 316 Xiaosong Wang, Yifan Peng, Le Lu, Zhiyong Lu, Mohammadhadi Bagheri, and Ronald M. Summers. ChestX-
317 ray8: Hospital-scale chest X-ray database and benchmarks on weakly-supervised classification and localiza-
318 tion of common thorax diseases. In *Proceedings of the IEEE Conference on Computer Vision and Pattern
319 Recognition (CVPR)*, pages 2097–2106, 2017.
- 320 John R. Zech, Marcus A. Badgeley, Manway Liu, Anthony B. Costa, Joseph J. Titano, and Eric Karl Oermann.
321 Variable generalization performance of a deep learning model to detect pneumonia in chest radiographs: A
322 cross-sectional study. *PLoS Medicine*, 15(11):e1002683, 2018.

323 A Classification detail

324 The classifier is a fixed upstream component (Section 3); this is its full evaluation.

Table 2: Per-class performance of the served DenseNet-121 on the held-out test set ($n = 4,759$; accuracy 96.18%, macro-F1 95.77%). Percentages except support. Derived from the confusion matrix below.

Class	Prec.	Recall	F1	Spec.	Supp.
Normal	95.30	98.50	96.90	94.60	2,502
Pneumonia	98.30	92.20	95.10	99.30	1,478
Tuberculosis	94.70	95.10	94.90	99.70	243
Covid-19	95.70	96.60	96.20	99.50	536
Macro avg.	96.00	95.60	95.77	98.30	4,759

Table 3: Confusion matrix on the held-out test set, from `chest_classifier_metrics.json` for the served checkpoint. The repository `README.md` carries a different matrix that does not match this file on cells other than the headline pneumonia-to-Normal count; the matrix here is the one behind every classification number in this paper.

True (rows) / Pred. (cols)	Normal	Pneumonia	Tuberculosis	Covid-19
Normal	2,465	14	9	14
Pneumonia	102	1,363	4	9
Tuberculosis	10	2	231	0
Covid-19	10	8	0	518

325 The dominant error is one-directional: 102 pneumonia radiographs predicted Normal, 56% of all 182
326 misclassifications, and no other off-diagonal cell exceeds 14. 82 of the 102 come from the COVID-19
327 Radiography Database’s `Lung_Opacity` folder, which `app.py` maps to Pneumonia though “lung
328 opacity” is broader than pneumonia; 8 come from its `Viral Pneumonia` folder and 12 from the
329 paediatric pneumonia set. A left–right lung-field mean-intensity probe supports reading this as label
330 noise: median absolute difference 5.9 grey levels on these 102 films, against 8.9 on correctly classified
331 pneumonia and 4.0 on true Normals, and only 22% of the 102 exceed the 90th percentile of the
332 true-Normal distribution, against 37% of correctly classified pneumonia, so the model’s outputs track
333 the pixels rather than miss something present.

334 **Source-bias probe.** An earlier composite drew three of four classes from one source each, so a
335 model could score by recognising the scanner rather than the pathology [Zech et al., 2018]. The
336 composite here weakens that: tuberculosis, Covid-19, and pneumonia each draw from at least two
337 sources, though the COVID-19 Radiography Database still supplies three classes through one pipeline.
338 Per-source accuracy spans 83.33% (Montgomery, 12 images) to 99.50% (TBX11K), wider than
339 prevalence alone explains, though the two extremes are also the two smallest partitions, so sampling
340 noise and genuine source sensitivity cannot be separated here. The 96.18% headline is read as inflated
341 by acquisition cues to an unknown degree.




Figure 4: Test accuracy by source dataset, sorted, with support (n) per source and the overall figure marked. The two weakest sources are also the two smallest partitions (92 and 12 images), so this narrows the source-cue confound without settling it.

B Report-generation templates, verbatim

The complete text-generating logic of the deployed system, app.py lines 1645–1680 (function predict_image), verbatim. The four descriptions strings and the five heatmap branches are the entire vocabulary the interface can produce. No line reads a coordinate, area, side, or component count.

```

347 descriptions = {
348     "Normal": "No abnormal opacities detected in the lung fields.",
349     "Pneumonia": "Findings consistent with pneumonia. Warmer areas indicate possible consolidation.",
350     "Tuberculosis": "Findings consistent with tuberculosis. Warmer areas indicate possible focal lesions
351     or cavitation.",
352     "Covid-19": "Findings consistent with Covid-19. Warmer areas indicate possible bilateral
353     ground-glass opacities.",
354 }
355
356 if is_uncertain:
357     txt = (f"**Prediction:** Uncertain\n\n"
358           f"The top class is {top_cls} at {top_prob * 100:.1f}%, below the "
359           f"{CONFIDENCE_THRESHOLD * 100:.0f}% reporting threshold. Review the class "
360           f"confidence values; a repeat or higher quality image may help.")
361 else:
362     txt = (f"**Prediction:** {top_cls} ({top_prob * 100:.1f}% confidence)\n\n"
363           f"{descriptions[top_cls]}")
364
365 # Explain what the heatmap represents for this specific case.
366 if mask is None:
367     txt += "\n\nNo attention map is available for this model."
368 elif is_confident_finding:
369     txt += (f"\n\nHeatmap: region driving the {viz_cls} prediction. "
370           f"The outline marks the most influential area.")
371 elif is_auto and top_idx == 0:
372     txt += (f"\n\nHeatmap: areas the model assessed for {viz_cls} "
373           f"(the next most likely class); none reached an abnormal level.")
374 elif is_uncertain:
375     txt += (f"\n\nHeatmap: areas the model weighed for {viz_cls}. "
376           f"Interpret with caution while the prediction is uncertain.")
377 else:
378     txt += f"\n\nHeatmap: model attention for {viz_cls} ({viz_prob * 100:.1f}%)."
379
380 if mask is not None:
381     txt += ("\n\nOverlay source: disease-segmentation U-Net._"
382           if mask_source == "segmentation"
383           else "\n\nOverlay source: Grad-CAM._")
384

```

CONFIDENCE_THRESHOLD = 0.60 (app.py:67); below it the first branch fires. The only other non-standard path is a caught exception that returns an empty string and a Gradio warning. mask_source is a two-word provenance label naming which network produced the overlay, never what it contains.

C Qualitative examples

Films from the 245-film pool, S0/S1 text verbatim from tier1_summaries.csv.




Figure 5: Six served overlays, input above overlay, spanning the range Appendix D quantifies. Left (green): compact regions that settle inside the lung field and follow the visible opacity, 66–90% of the region inside the lung. Right (red): a correct tuberculosis call whose evidence sits almost entirely in soft tissue at the apex (97% outside the lung); a region covering 86% of the image, spanning both lungs and the chest wall; and a clear, well-aerated film the model calls Normal, where no region survives the display threshold at all. All six overlays came from the served U-Net path. Per-case labels give the true class, the prediction, its confidence, and what the region did; italics are this appendix’s reading, not part of the served text.

Phantom-outline claims (both films where is_confident_finding is true but no region survives the threshold; both Montgomery TB group).

- 194_MCUCXR_0015_0: true Normal, predicted Tuberculosis at 95.7%. S0: “...The outline marks the most influential area. Overlay source: disease-segmentation U-Net.” No region exists in the image. S1 under H_actual: “...No heatmap region met the reporting threshold; no location is reported.” A confidently wrong call compounded by a claim about an outline that was never drawn.
- 202_MCUCXR_0301_1: true and predicted Tuberculosis at 94.0%, classification correct, same phantom-outline pattern.

Well-localised. 209_Normal-2399: true and predicted Normal at 100.0%, single focal region, upper-left zone, 93.6% inside the lung. S1: “...Heatmap evidence: left lung, upper zone, focal, single, consistent with the normal finding.”

Diffuse overlay. 011_person730_virus_1351: true Pneumonia, predicted Normal at 96.0%. The region is extensive and multifocal, centroid outside both lungs, yet 50.3% of its area is inside the lung field, just over the gate. S1: “outside both, mid zone, extensive, multifocal,” a literal description of a poorly localised region on a film the classifier got wrong, and a boundary case worth flagging.

No region. 003_person1175_virus_1981: true Pneumonia, predicted Normal at 81.1%. No region survives the threshold. S0 gives no hint (“...none reached an abnormal level.”); S1 states plainly that no location is reported.

D Additional qualitative statistics

Same 245-film pool. 41 (16.7%) produce no region above the overlay threshold; geometry percentages are over the remaining 204.

- *Activation leaving the lung.* 55% of overlays place the majority of the region outside the lung fields; 20% place almost all of it there, usually on the cardiac silhouette, the mediastinum, or below the diaphragm. The centroid is on neither lung in 129 of 204.
- *Edge activation.* 54% of regions reach the outer 2% of the frame despite the 8% crop; 14% put more than a quarter of the region in the outer tenth.
- *Diffuse activation.* 44% of regions cover more than a quarter of the image, 19% more than half, and 37% span both lung fields. This concentrates in Shenzhen (74% of its overlays diffuse; median TB-positive region 51% of the image); Montgomery fails oppositely (no region on a quarter of films; median region 4% where one exists).

On the 102 pneumonia-to-Normal errors, 28 produce no region and none produces a region over half the image; median region area is 11%, against 41% for correctly classified pneumonia. This is consistent with the label-noise reading: the lung fields are frequently clear and the small regions the model does produce plausibly agree with unremarkable pixels.




Figure 6: The two region descriptors added for this evaluation (Section 4), over all 245 films. Two thirds of regions centre in the mid zone; most of the rest produce no region at all, rather than an upper- or lower-zone one. Single- and multi-component regions are close to a coin flip, which is why focality is reported as its own descriptor rather than folded into extent.

E Latency, per stage

Table 4 splits the mean CPU wall-clock time reported in Section 5 by which localisation path served the request (i7-13620H, 10 PyTorch threads, single request). The two rows are two different code paths, not one path with variance: the served U-Net returns a mask in one forward pass; the live Grad-CAM fallback needs a forward pass and a backward pass through the classifier. Report generation is a dictionary lookup and does not appear as a separate row because it does not register above measurement noise.

Table 4: CPU latency by localisation path, i7-13620H, 10 threads, single request.

Path	Share of test films	Mean (sd), ms
Distilled U-Net (one forward pass)	91%	293 (81)
Live Grad-CAM fallback (forward + backward)	9%	560 (22)
Weighted (actual test-set mix)	100%	318

NeurIPS paper checklist

1. **Claims.** Do the abstract and introduction state claims that match the paper’s contributions and scope?
[Yes]The central claim, that the deployed summary does not read the mask, is a source-level fact quoted verbatim in Section 3.2 and Appendix B. The abstract and Section 6 state which evaluation-grid cells are measured and which are proposed.
2. **Limitations.** Does the paper discuss the limitations of the work?
[Yes]Section 6: unrun cells (H2, H4, S3/S4), no clinician ratings, dataset-source confounding, label noise in the dominant error mode, and the reproducibility fragility of the split.
3. **Theory.** Are assumptions and proofs given for every theoretical result?
[N/A]The paper reports measurements, not theorems. The one derived bound (localisation quality upper-bounds spatial-claim correctness) is stated with the Dice measurement it rests on.
4. **Experimental reproducibility.** Is there enough information to reproduce the main results?
[Yes]The audited system is public (repository withheld for double blind). The descriptor extractor, the summary reconstructions, the thresholds, and the derangement seed are in Section 4; the 245-film scoring table and mask arrays are the system’s own published analysis outputs.
5. **Open access to data and code.** Are data and code available?
[Yes]The audited system and its analysis outputs are public. No new model was trained.
6. **Experimental settings.** Are training and test details specified?
[Yes]For the components this paper adds there is no training (Section 4). The fixed upstream classifier and U-Net settings are in Section 3 and Appendix A.
7. **Error bars.** Are error bars or other measures of statistical significance reported?
[Yes]Distributions (latency, Dice) carry a mean and standard deviation or a median. Single-number rates are exact counts over the full 245-film pool, so a confidence interval would describe sampling that did not occur.
8. **Compute.** Is the compute for the experiments reported?
[Yes]The analysis is geometry over saved mask arrays and runs in seconds on a CPU; latency figures are from one laptop CPU (i7-13620H, 10 threads). No GPU, no training.
9. **Code of ethics.** Does the research conform to the NeurIPS Code of Ethics?
[Yes]A retrospective audit of an existing research and educational prototype; no patient contact; the interface already carries a non-diagnostic disclaimer this paper argues should be more prominent.
10. **Broader impacts.** Are potential positive and negative societal impacts discussed?
[Yes]Section 6 (Broader impact): the persuasiveness of a label plus a plausible heatmap plus a fluent summary, and why this paper is a caution against it.
11. **Safeguards.** Are safeguards described for any high-risk release?
[N/A]No models, datasets, or scrapers are released.
12. **Licenses.** Are existing assets credited and their licenses respected?
[Yes]Backbone, localisation method, and segmentation architecture are cited in Section 3. The eight source datasets are public Kaggle datasets, named by their exact identifiers so each license can be consulted on its own page:
 - tawsifurrahman/tuberculosis-tb-chest-xray-dataset
 - pcbreviglieri/pneumonia-xray-images
 - raddar/ricord-covid19-xray-positive-tests
 - tawsifurrahman/covid19-radiography-database
 - raddar/tuberculosis-chest-xrays-shenzhen
 - raddar/tuberculosis-chest-xrays-montgomery
 - vbookshelf/tbx11k-simplified
 - legrande/tbdata

Terms differ per dataset; the camera-ready should transcribe each from its source page.

- 486 13. **New assets.** Are new assets documented?
487 [N/A]No new assets are released with this submission.
- 488 14. **Crowdsourcing and human subjects.** Are crowdsourcing and IRB details given?
489 [N/A]No human subjects, no crowdsourcing, no clinician annotation was collected.
- 490 15. **IRB / risk for human subjects.** Were risks disclosed and approvals obtained?
491 [N/A]No human-subjects research was conducted.
- 492 16. **LLM usage.** Is any LLM use in the core method declared?
493 [N/A]No LLM is used in the method. S2/S3/S4 LLM conditions are named as out of scope
494 and not run.