

Tune Report

Instrument Information

MS Model	G6546A	Run Date/Time	2026-02-03 14:04:53-0500
Serial Number	SG2429E201	Last Tuned by	SYSTEM
Firmware Revision	27.852	Last Modified by	SYSTEM
Source Type	Dual AJS ESI	Slicer Mode	High Resolution (Position: 5)
Mass Range	Low (1700 m/z)	Instrument Mode	Standard Resolution (10GHz)
Ion Polarity	Negative	Tune Type	TOFMassCalibration-1700mzRange
SureMass Optimization	Not Enabled		
Tune File Name	TOFMassCalibration-1700mzRange.tun		

Negative Polarity Results

TOF Results

TOF Mass Calibration Data

Theoretical	Actual	Time	Abundance	Calibration Abundance	Resolution	Primary Residuals	Corrected Residuals
112.985587	112.985586	28.493090	1,639,418	853,331	36,065	-5.21	-0.01
301.998139	301.998149	45.782764	1,590,173	896,050	52,204	-0.32	0.03
601.978977	601.978941	64.119052	4,958,984	892,482	60,026	0.43	-0.06
1033.988109	1033.988202	83.642170	5,698,625	947,314	65,923	0.46	0.09
1333.968947	1333.968811	94.832354	8,073,044	871,259	67,479	-0.01	-0.10
1633.949786	1633.949853	104.820357	7,293,105	997,153	71,336	-0.20	0.04

Setpoints

Source Settings

Gas Temperature(°C)	325	Drying Gas(l/min)	5.0
Nebulizer Pressure(psi)	20	Capillary(V)	3500
Sheath Gas Temperature(°C)	275	Sheath Gas Flow(l/min)	12.0
Nozzle Voltage(V)	2000		

Source Actuals

Gas Temperature(°C)	325	Drying Gas(l/min)	5.0
Nebulizer Pressure(psi)	20	Capillary(V)	3491
Cap Current(μA)	0.110	Chamber Current(μA)	0.55

Optics 1

Fragmentor(V)	-165	Skimmer(V)	-45.0
Oct 1 RF Vpp(V)	750	Oct 1 DC(V)	-29.6
Lens 1(V)	-28.7	Lens 2(V)	-19.7
Lens 2 RF Enable(False/True)	1	Lens 2 RF Voltage(V)	0
Lens 2 RF Phase(deg)	180		

Quad

TTI Quad AMU(amu)	89.6	Quad DC(V)	-24.7
Post Filter DC(V)	-24.6		

Cell

Collision Cell Gas Flow(psi)	18.0	Hex RF(V)	550
Hex DC(V)	-22.2	Hex Delta(V)	7.0
Cell Entrance(V)	-23.7	Hex 2 RF(V)	600
Hex 2 DC(V)	-15.2	Hex 2 DV(V)	1.0

Optics 2

Ion Focus(V)	-10.1	Extractor DC(V)	11.6
Lens 3(V)	60.4	Bottom Slit(V)	39.70
Top Slit(V)	40.45	Hex 3 DC(V)	-14.0

TOF

Pusher(V)	-1200	Pusher Offset(mV)	-26
Puller(V)	700	Puller Offset(V)	-43
Acc Focus(V)	2135	Front Mirror(V)	9000
Mid Mirror(V)	2274.5	Back Mirror(V)	-1475
Minimum Mass(m/z)	50	Maximum Mass(m/z)	1700
Acquisition Rate	1.0	Acquisition Time	1000
Acq Hold Off Delay(ns)	0		

Detector

MCP(V)	718	PreAmpOffset(DAC)	1931
Low Gain PreAmpOffset(DAC)	962	Gain Abund Ratio	12.0
Gain T0 Offset	-0.170		

Vacuum And Temperatures

Quad Temp(°C)	100	Rough Vac(Torr)	2.46E+00
Quad Vac(Torr)	5.64E-05	TOF Vac(Torr)	9.05E-08
Turbo 1 Speed(%)	100.0	Turbo 1 Power(W)	213
Turbo 2 Speed(%)	100.0	Turbo 2 Power(W)	25

TOF Mass Calibration Coefficients

a	3903.284E-04	t0	1261.047E-03	a2	-1429.032E-06
b2	2961.583E-08	c2	-1617.224E-16	d2	-1755.445E-20
e2	0000.000E+00	f2	0000.000E+00	Term Flag	0x0063
Trad	0	Poly	6		