

Instrument Information

MS Model	G6546A	Run Date/Time	2026-02-03 14:00:43-0500
Serial Number	SG2429E201	Last Tuned by	SYSTEM
Firmware Revision	27.852	Last Modified by	SYSTEM
Source Type	Dual AJS ESI	Slicer Mode	High Resolution (Position: 5)
Mass Range	Low (1700 m/z)	Instrument Mode	Standard Resolution (10GHz)
Ion Polarity	Positive	Tune Type	TOFMassCalibration-1700mzRange
SureMass Optimization	Not Enabled		
Tune File Name	TOFMassCalibration-1700mzRange.tun		

Positive Polarity Results

TOF Results

TOF Mass Calibration Data

Theoretical	Actual	Time	Abundance	Calibration Abundance	Resolution	Primary Residuals	Corrected Residuals
118.086255	118.086254	29.097864	1,757,561	957,655	35,431	-4.98	-0.01
322.048121	322.048151	47.232680	962,667	956,735	47,619	-0.23	0.09
622.028960	622.028831	65.151896	1,585,198	875,134	51,828	0.23	-0.21
922.009798	922.009887	79.047305	1,975,471	901,165	54,234	0.49	0.10
1221.990637	1221.990744	90.811960	1,903,956	919,651	55,888	0.22	0.09
1521.971475	1521.971381	101.201180	1,681,764	948,369	57,020	-0.32	-0.06

Setpoints

Source Settings

Gas Temperature(°C)	325	Drying Gas(l/min)	5.0
Nebulizer Pressure(psi)	20	Capillary(V)	4000
Sheath Gas Temperature(°C)	275	Sheath Gas Flow(l/min)	12.0
Nozzle Voltage(V)	2000		

Source Actuals

Gas Temperature(°C)	325	Drying Gas(l/min)	5.0
Nebulizer Pressure(psi)	20	Capillary(V)	3995
Cap Current(µA)	0.110	Chamber Current(µA)	0.45

Optics 1

Fragmentor(V)	165	Skimmer(V)	45.0
Oct 1 RF Vpp(V)	750	Oct 1 DC(V)	28.1
Lens 1(V)	27.5	Lens 2(V)	18.1
Lens 2 RF Enable(False/True)	1	Lens 2 RF Voltage(V)	0
Lens 2 RF Phase(deg)	54		

Quad

TTI Quad AMU(amu)	126.7	Quad DC(V)	25.5
Post Filter DC(V)	25.4		

Cell

Collision Cell Gas Flow(psi)	18.0	Hex RF(V)	550
Hex DC(V)	22.3	Hex Delta(V)	-7.0
Cell Entrance(V)	22.4	Hex 2 RF(V)	600
Hex 2 DC(V)	15.3	Hex 2 DV(V)	-1.0

Optics 2

Ion Focus(V)	11.7	Extractor DC(V)	-9.3
Lens 3(V)	-59.0	Bottom Slit(V)	-38.75
Top Slit(V)	-40.10	Hex 3 DC(V)	14.1

TOF

Pusher(V)	1200	Pusher Offset(mV)	-29
Puller(V)	-700	Puller Offset(V)	37
Acc Focus(V)	-2169	Front Mirror(V)	-9000
Mid Mirror(V)	-2272.1	Back Mirror(V)	1475
Minimum Mass(m/z)	50	Maximum Mass(m/z)	1700
Acquisition Rate	1.0	Acquisition Time	1000
Acq Hold Off Delay(ns)	0		

Detector

MCP(V)	748	PreAmpOffset(DAC)	1931
Low Gain PreAmpOffset(DAC)	962	Gain Abund Ratio	12.0
Gain T0 Offset	-0.170		

Vacuum And Temperatures

Quad Temp(°C)	100	Rough Vac(Torr)	2.47E+00
Quad Vac(Torr)	5.64E-05	TOF Vac(Torr)	9.02E-08
Turbo 1 Speed(%)	100.0	Turbo 1 Power(W)	213
Turbo 2 Speed(%)	100.0	Turbo 2 Power(W)	24

TOF Mass Calibration Coefficients

a	3903.527E-04	t0	1259.647E-03	a2	-1398.033E-06
b2	2805.123E-08	c2	-1364.107E-13	d2	-1577.871E-18
e2	0000.000E+00	f2	0000.000E+00	Term Flag	0x004B
Trad	0	Poly	6		