General Settings

Labware name: 24_tuberack_shortbase_snapcap
Type: Regular
Labware footprint: Microplate
Tips per well: 1
Sort in Labware group: Tube

Positions / Teaching

	X	Y	Z	
Dimensions	127.7	85.5	50.0	[mm]

Bounding Boxes	X-Offset	Y-Offset	Z-Offset	X-Size	Y-Size	Z-Size	
	0.0	0.0	0.0	127.7	85.5	50.0	[mm]
							[mm]

Number of wells: 6 × 4
Well alignment: regular / interleaved
Compartments per well: 1

Please select a labware on the worktable to be used for teaching ...

Current compartment: 1

Pipetting positions: All values are specified in [mm].

	X	Y		
(position #1)	18.93	76.21	<< Transfer	Move to
(position #24)	117.37	18.37	<< Transfer	Move to



	X	Y	Z		
(adjustment)				Show delta	Save
				Show delta	Save

The adjustment only affects the site hosting the currently selected labware.

Z-Coordinates

Z-Travel	58.87	<< Transfer	Move to
Z-Start	50.87	<< Transfer	Move to
Z-Dispense	48.87	<< Transfer	Move to
Z-Max	14.87	<< Transfer	Move to
Z-Bottom	14.00	<< Transfer	Move to

Y-distance of wells

Compartment Definition

Select shape: Add

All dimensions in the table below are specified in [mm].
Square measures are defined in [mm²].

	Height	Bottom diameter	Top diameter
(Truncated) Cone	34.00	9.50	9.50



Miscellaneous

Color: Gray
Mesh or Texture File: mtp_384 Browse... Import...
Vendor Name:
Part Number:
Link to: http://lifesciences.tecan.com/products/consumables/labware/microplates Go to...

Allowed Labware Types

Please select the labware types which may be placed on this labware.

Filter labware by footprint:

Labware Types: 24Microplate_CoverSite_5

Allowed Locations

Please select the locations on which this labware type may be placed.

Filter locations: <all selected>

Allowed locations:

- Carousel_L_Pos
- Carousel_R_Pos
- Cytomat_Tower_10Nest
- Hotel_2Nest_Pos
- Hotel_3Nest_Pos_1_7_13
- Hotel_3Nest_Pos_1_8_15
- Hotel_4Nest
- HotelDWP_Pos
- Incub4Shake_Pos
- LeftTower_P128_4Nest
- LeftTower_P172_3Nest
- LeftTower_P23_21Nest
- LeftTower_P50_10Nest
- LeftTower_P69_7Nest
- LeftTower_P83_6Nest
- LiCONiC_Cassette_P23_S22
- LiCONiC_Cassette_P50_S10
- Liconic_Tower_22Nest
- LiconicSTX88_Pos
- LiconicSTX88_Pos_1
- Mag_block_pos
- MCA_Passive_Nest
- Nest61mm_Pos
- Nest7mm_Diti_Pos
- Nest7mm_HoldingPos
- Nest7mm_Pos
- Nest7mm_Pos_1
- Nest7mm_TempPark_Pos
- NestCooled_Pos
- NestRunner_Pos
- NestRunnerPortrait_Pos
- OffDeckWaste_Pos
- PassiveStack12_Pos
- PassiveStack15_Pos
- PosClip_Nest
- Regrip_Pos
- RightTower_P128_4Nest
- RightTower_P172_3Nest
- RightTower_P23_21Nest
- RightTower_P23_21Nest_1
- RightTower_P50_10Nest
- RightTower_P69_7Nest
- RightTower_P83_6Nest
- Rotanta_LoadPos
- Rotanta_Pos
- ShakerHeating_Pos
- ShakerNest_Pos
- StackerLeft_Pos
- StackerRight_Pos
- StackerTower_16Nest
- StackerTower_40Nest
- StackerTower_Left_16Nest
- StackerTower_Left_40Nest
- TeChrom_MTP_Spacer_Pos
- TeVacs_Back
- TeVacs_Front
- TeVacs_Spacer_Back
- TeVacs_Spacer_Front
- Tevacs_Spacer_Pos
- ThruDeckWaste_Pos
- Ziath_Scan_Pos

FCA Specific

Pierce through septum: ☐

Piercing with	10.0	[mm / s]	Acceleration	1.0	[mm / s²]	Deceleration	1.0 [mm / s²]
Retract with	60.0	[mm / s]	Acceleration	1.0	[mm / s²]	Deceleration	1.0 [mm / s²]
Force	30	[0=low ... 249=high]					

Robot Specific

Grip Narrow: 85.5 [mm] Wide: 127.8 [mm]

	X	Y	Z	
Stacked Plate to Plate offset	0.0	0.0	18.0	[mm]

Labware can have lid: ☑

Select a Lid: Lid_9mm

	X	Y	Z	
Plate to Lid offset	0.0	0.0	13.0	[mm]

Custom Attributes

Attribute Name	Data Type	Attribute Value
Force	Integer	20 [%]
Locking	Integer	1
Grip_delta	Floating Point	4.00 [mm]
StackerType	Integer	13
MapTemplateName	String	24 Well Flat Greiner 662102
WorkspaceDeltaLevel	Integer	-1