## General Settings

Labware name: 96 Well nonskirt adapter LnZ 3.2
Type: Regular
Labware footprint: Microplate
Tips per well: 1
Sort in Labware group: Microplate

## Positions / Teaching

| | X | Y | Z | |
|---|---|---|---|---|
| Dimensions | 127.8 | 85.5 | 23.0 | [mm] |

| Bounding Boxes | X-Offset | Y-Offset | Z-Offset | X-Size | Y-Size | Z-Size | |
|---|---|---|---|---|---|---|---|
| | 0.0 | 0.0 | 0.0 | 127.8 | 85.5 | 23.0 | [mm] |
| | | | | | | | [mm] |

Number of wells: 12 × 8
Well alignment: regular / interleaved
Compartments per well: 1

Please select a labware on the worktable to be used for teaching ...

Current compartment: 1

Pipetting positions: All values are specified in [mm].

| | X | Y | | |
|---|---|---|---|---|
| (position #1) | 14.38 | 74.24 | << Transfer | Move to |
| (position #96) | 113.38 | 11.24 | << Transfer | Move to |




| | X | Y | Z | | |
|---|---|---|---|---|---|
| (adjustment) | | | | Show delta | Save |
| | | | | Show delta | Save |

The adjustment only affects the site hosting the currently selected labware.

Z-Coordinates

| | | | |
|---|---|---|---|
| Z-Travel | 29.11 | << Transfer | Move to |
| Z-Start | 21.01 | << Transfer | Move to |
| Z-Dispense | 19.11 | << Transfer | Move to |
| Z-Max | 7.21 | << Transfer | Move to |
| Z-Bottom | 6.21 | << Transfer | Move to |

Y-distance of wells

## Compartment Definition

Select shape: Add

All dimensions in the table below are specified in [mm].
Square measures are defined in [mm²].

| | Height | Diameter | |
|---|---|---|---|
| Cylinder | 4.00 | 5.50 | |

| | Height | Bottom diameter | Top diameter |
|---|---|---|---|
| (Truncated) Cone | 9.00 | 3.00 | 5.50 |

| | Height | Diameter |
|---|---|---|
| Partial Sphere | 1.50 | 3.00 |




## Miscellaneous

Color:
Mesh or Texture File: plate96 Browse... Import...
Vendor Name:
Part Number: 30122298, 30122304, 30122306
Link to: http://lifesciences.tecan.com/products/consumables/labware/microplates Go to...

## Allowed Labware Types

Please select the labware types which may be placed on this labware.

Filter labware by footprint:

Labware Types 96Microplate_CoverSite_5

**Allowed Locations**

Please select the locations on which this labware type may be placed.

Filter locations: <all selected>

Allowed locations:

- Carousel_L_Pos
- Carousel_R_Pos
- Cooled_Pos
- Cytomat_Tower_10Nest
- Downholder_Pos
- DownholderActive_Pos
- Hotel_2Nest_Pos
- Hotel_3Nest_Pos_1_7_13
- Hotel_3Nest_Pos_1_8_15
- Hotel_4Nest
- HotelDWP_Pos
- HotelMP_Pos
- Hydroflex_Pos
- Hydrospeed_Pos
- Incub4Shake_Pos
- Incub6_Pos
- Incub6Shake_Pos
- InfiniteF50_Pos
- InfiniteM1000_Pos
- InfiniteM200_Pos
- InfiniteM500_Pos
- LeftTower_P128_4Nest
- LeftTower_P172_3Nest
- LeftTower_P19_25Nest
- LeftTower_P23_21Nest
- LeftTower_P50_10Nest
- LeftTower_P69_7Nest
- LeftTower_P83_6Nest
- LiCONiC_Cassette_P23_S22
- LiCONiC_Cassette_P50_S10
- Liconic_Tower_22Nest
- LiconicSTX88_Pos
- LiconicSTX88_Pos_1
- Mag_block_pos
- Mag_pos
- MCA_Passive_Nest
- Metter_OWS1_Pos
- Metter_OWS1_Pos_1
- Metter_OWS2_Pos
- Metter_OWS2_Pos_1
- Nest61mm_Pos
- Nest7mm_Diti_Pos
- Nest7mm_HoldingPos
- Nest7mm_Pos
- Nest7mm_Pos_1
- Nest7mm_TempPark_Pos
- NestCooled_Pos
- NestRunner_Pos
- NestRunnerPortrait_Pos
- ODTC_96_1
- OffDeckWaste_Pos
- PassiveStack12_Pos
- PassiveStack15_Pos
- PosClip_Nest
- Regrip_Pos
- RightTower_P128_4Nest
- RightTower_P172_3Nest
- RightTower_P19_25Nest
- RightTower_P23_21Nest
- RightTower_P23_21Nest_1
- RightTower_P50_10Nest
- RightTower_P69_7Nest
- RightTower_P83_6Nest
- Rotanta_LoadPos
- Rotanta_Pos
- Shaker_Pos
- ShakerHeating_Pos
- ShakerNest_Pos
- Spark_Pos
- StackerLeft_Pos
- StackerRight_Pos
- StackerTower_16Nest
- StackerTower_40Nest
- StackerTower_Left_16Nest
- StackerTower_Left_40Nest
- Sunrise_Pos
- TeChrom_Collection_Pos
- TeChrom_MTP_Spacer_Pos
- TeVacs_Back
- Tevacs_Cap_Pos
- Tevacs_Cap_Pos_1
- Tevacs_Cap_Pos_2
- TeVacs_Front
- TeVacs_Spacer_Back
- TeVacs_Spacer_Front
- Tevacs_Spacer_Pos
- ThruDeckWaste_Pos
- Torrey_Pos
- TorreyPos
- Ziath_Scan_Pos

**FCA Specific**

Pierce through septum: ☐

Piercing with 10.0 [mm / s] Acceleration 1.0 [mm / s²] Deceleration 1.0 [mm / s²]

Retract with 60.0 [mm / s] Acceleration 1.0 [mm / s²] Deceleration 1.0 [mm / s²]

Force 30 [0=low ... 249=high]

**Robot Specific**

Grip Narrow 85.5 [mm] Wide 127.8 [mm]

| | X | Y | Z | |
|---|---|---|---|---|
| Stacked Plate to Plate offset | 0.0 | 0.0 | 13.0 | [mm] |

Labware can have lid: ☑

Select a Lid: Lid_9mm

| | X | Y | Z | |
|---|---|---|---|---|
| Plate to Lid offset | 0.0 | 0.0 | 7.6 | [mm] |

**Custom Attributes**

| Attribute Name | Data Type | Attribute Value |
|---|---|---|
| Force | Integer | 20 [%] |
| Locking | Integer | 1 |
| Grip_delta | Floating Point | 4.00 [mm] |
| StackerType | Integer | 1 |
| ClldSensitivityCalculationLabware | String | WellFlat |
| MapTemplateName | String | 96 Well Flat |
| WorkspaceDeltaLevel | Integer | -1 |