DCA
DRUGS CONTROL ADMINISTRATION
TELANGANA

DRUGS CONTROL ADMINISTRATION
Government of Telangana

FORM MD-5

[See sub-rule (4) of rule 20 and sub-rule (6) of rule 20]

Licence to Manufacture for Sale or for Distribution of Class A or Class B medical device

Licence Number: **MFG/MD/2025/000200.**

1. **M/s DESBRO BIOSCIENCES PRIVATE LIMITED , First And Second Floor , Plot No 43, Survey No. 376/A, Rajiv Gandhi Nagar , Kukatpally, Hyderabad , Medchal-Malkajgiri, Telangana (India) - 500072 Telephone No.: 7013470789 FAX: 7013470789** has been licenced to manufacture for sale or for distribution the below listed medical device(s) at the premises situated at **M/s DESRBRO BIOSCIENCES PRIVATE LIMITED, First and Second Floor Plot No 43, Survey No.376 and A, Rajiv Gandhi Nagar , Kukatpally , Hyderabad, Medchal-Malkajgiri, Telangana (India) - 500072 Telephone No.: 9885444435 FAX: 9885444435.**

2. Details of medical device(s) [Annexed]

3. This licence is subject to the provisions of the Medical Devices Rules, 2017 and conditions prescribed therein.

ANNEXURE

S.No.	Details of Device(s)
1	**Generic Name:** Sterile Skin Stapler (Sterilzation - EO/ Gamma Radiation) **Model No.:** NIL **Intended Use:** Skin Stapler is intended to be applied to close lacerations on the outer layer of the dermis. **Class of medical device:** Class B **Material of construction:** steel, aluminum, or stainless steel **Dimension(if any):**a standard skin staple is made of 316 L stainless steel and has a diameter of 0.5 mm, a span of 6 to 10 mm, and a leg length of 3.5 mm. Several staplers come with wide staples that span from 7 to 13 mm and have a leg length of 4 mm. **Shelflife:**3 Years **Sterile or Non sterile:** Sterilized **Brand Name(if registered under the Trade Marks Act, 1999):**SMART STAPLER (TM)
2	**Generic Name:** Removable skin staple (Sterilzation - EO/ Gamma Radiation) **Model No.:** NIL **Intended Use:** A removable skin staple is a staple-like device intended to connect external tissues temporarily to aid healing. It is not absorbable **Class of medical device:** Class B **Material of construction:** Stainless Steel **Dimension(if any):**Thickness 3 - 5 mm and Length: 4.2 Inch. **Shelflife:**3 Years **Sterile or Non sterile:** Sterilized **Brand Name(if registered under the Trade Marks Act, 1999):**SMART REMOVER (TM)

Joint Director - O/o Director, Drugs Control Adm. Hyderabad Telangana State

27/3/25

3	**Generic Name:** Disposable Skin Stapler (Sterilzation - EO/ Gamma Radiation) **Model No.:** NIL **Intended Use:** Surgical staples are specialized staples used in surgery in place of sutures to close skin wounds, connect or remove parts of body during surgery. **Class of medical device:** Class B **Material of construction:** steel, aluminum, or stainless steel. **Dimension(if any):**a standard skin staple is made of 316 L stainless steel and has a diameter of 0.5 mm, a span of 6 to 10 mm, and a leg length of 3.5 mm. Several staplers come with wide staples that span from 7 to 13 mm and have a leg length of 4 mm. **Shelflife:**3 Year **Sterile or Non sterile:** Sterilized **Brand Name(if registered under the Trade Marks Act, 1999):**SMART STAPLER (TM)

Joint Director - O/o Director, Drugs Control Adm. Hyderabad Telangana State

Place:Hyderabad

Date:27-Mar-25

27/3/2025

State Licensing Authority

JOINT DIRECTOR (Enf)
Licensing & Controlling Authority
DRUGS CONTROL ADMINISTRATION
Government of Telangana
Hyderabad