
Table of Contents

Balance Calibration Summary	1
Facility Information	1
Load Calibration Data	1
Display Calibration Details	1
Plot Calibration Performance	2

Balance Calibration Summary

```
clc; clear; close all;
```

Facility Information

```
d.fac.name = 'SWT';  
d.fac.BalanceType = 'Internal';  
d.fac.BalanceConfig = 'Force'; % or 'Force'
```

Load Calibration Data

```
file = 'kenSN29850_E99_July2023.vol';  
d.cal = read_vol_file(file);  
d.cal = calc_coeffs(d, 'Cubic');
```

Display Calibration Details

```
display_cal_info(d)
```

```
---- Calibration Information ----  
Date: 7/24/2023  
Calibration Performed By: Ken Ostasiewski
```

```
---- Balance Description ----  
Balance Type: 5 Force/1 Moment  
Serial Number: 100 lb Balance  
Outer Diameter: 1.00 in.
```

```
---- Maximal Balance Loads ----  
N1: 100 lb  
N2: 100 lb  
Y1: 50 lb  
Y2: 50 lb  
Ax: 25 lb  
Mx: 50 in-lb
```

---- Bias Error [dimensional] ----

Balance Channel Names: N1, N2, Y1, Y2, Ax, Mx

Balance Bias Errors: 0.051087 0.049722 0.1036 0.09779 0.10606
0.10928


---- Bias as Percentage of Full Scale ----





Balance Channel Names: N1, N2, Y1, Y2, Ax, Mx





Bias as Percentage: 0.051087 0.049722 0.20719 0.19558 0.42425
0.21856 %


Plot Calibration Performance

plot_force_cal_perf(d.cal)









Published with MATLAB® R2023a