



	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT	F-02/06
		Issue Date: 11/11/19 Rev No: 06


ACTIVITY:	☒ Assessment	☐ Surveillance	NCR No: 01 TL
	☐ Re-Assessment	☐ Scope extension	Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category: Very serious ☐ Essential ☒
	ORGANIZATION: Quality control Lab vital Agri. Nutrients (Pvt) Ltd.		Time limit for correction:
Department:	Labs		03 months
Team Leader.	Mr. Ateeq Ur Rehman	Mngt. Rep.	Mr. Himmyat Hussain
Assessor	Dr. Muhammad Aftab		
Expert			
DESCRIPTION: the lab has also developed a format for inspection of purchased products, but has not evaluated as per standard requirements, (NC) lab has also not defined about the re-evaluation and lab not practicing the reevaluation of vendors (NC)			Requirement of: 6.6 ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
10-01-2025 Date	 Signature Assessor/Expert		Ref. Organization's doc.
CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcomings on both the NCs has been taken by revising the format for inspection of purchased products as per standard requirements. Lab has revised the procedure (QCL-LSP-06) and defined the criteria about the re-evaluation and start practicing the reevaluation of vendors.			☐ It is not necessary to attach documentation
Action are documented in the amendment no: <u>CA-11-02-2025</u> <u>07-02-25</u> Date			 Signature (Management Representative)
REASON FOR CLOSING: (To be filled in by the Team Leader.)			
☐ The non-conformity is closed based on satisfactory documentation from the organization ☐ The non-conformity is closed based on recommendation from the technical assessor ☐ Implementation of the corrective actions will be followed up at the next visit			
Remarks (if any) ----- ----- -----			
The non-conformity is closed: _____ Date Signature (Team Leader.)			

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-11-02-2025
Designation: AQCM	CAR Initiated Date: 01-02-2025	NC No. 1 TL

CAR Initiated due to (Please tick the relevant appropriate box below)

- ☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) The lab has also developed a format for inspection of purchased products, but has not evaluated as per standard requirements, (NC). Lab has also not defined about the re-evaluation and lab not practicing the reevaluation of vendors (NC).	 Signatures
--	--

Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 01-02-2025

* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.

B. Root Cause Analysis (where required) 1. The team didn't properly check if the inspection format met the standard requirements. This happened because there wasn't a clear process, proper training, or oversight in place to ensure compliance. 2. The lab management failed to recognize the importance of re-evaluating vendors and did not include it in their processes. This happened because there was no proper training or focus on vendor management requirements during procedure development.
--

C. Proposed Action	Target Date:	07-02-2025
Corrective action for shortcomings on both the NCs has been taken by revising the format for inspection of purchased products as per standard requirements. Lab has revised the procedure (QCL-LSP-06) and defined the criteria about the re-evaluation and start practicing the reevaluation of vendors.		

 Name or Signatures (Person who carried out Root Cause Analysis and Proposed Corrective Action)

D. Verification by Manager QC for implementation & Effectiveness	
Is Proposed Action Implemented? ☒ YES ☐ NO	Is it Effective? ☒ YES ☐ NO


Remarks (If any):

CAR closing date: 07-02-25
 Manager QC Sign: 

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab has also developed a format for inspection of purchased products, but has not evaluated as per standard requirements, (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	01-02-2025	Submission Date	07-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	☒

Cause Effect Diagram



3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why was the product not evaluated as per standard requirements?	The inspection format was developed but not reviewed or cross-checked against the standard requirements.
WHY	Why was the format not reviewed or cross-checked?	There was no formal process or mechanism to ensure alignment with standards during the format's development.
WHY	Why was there no formal process for alignment with standards?	Lack of training or awareness among the team about standard requirements and compliance procedures.
WHY	Why was there a lack of training or awareness?	Training sessions or standard operating procedures (SOPs) were not adequately planned or prioritized by management.
WHY	Why were training sessions or SOPs not prioritized?	Insufficient allocation of resources and oversight in quality assurance planning.
How	How can this NC be resolved?	Review the inspection format against the applicable standard requirements.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

		- Identify gaps and revise the format accordingly. - Establish a formal review process to ensure compliance before using the format.
How	How can this NC be prevented in the future?	- Implement a training program for the team on standard requirements and compliance. - Develop an SOP that includes mandatory review and approval steps for new formats and procedures. - Assign a dedicated quality assurance team or representative to oversee such tasks.
Root Cause: The team didn't properly check if the inspection format met the standard requirements. This happened because there wasn't a clear process, proper training, or oversight in place to ensure compliance.		

RCA Statement: The team didn't properly check if the inspection format met the standard requirements. This happened because there wasn't a clear process, proper training, or oversight in place to ensure compliance.


Reviewed by: 

Evaluated by: 

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
Lab has also not defined about the re-evaluation and lab not practicing the reevaluation of vendors (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	01-02-2025	Submission Date	07-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5W HY analysis	☐
		3W2H / 5W2H	☒

Cause Effect Diagram



3W2H /5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why has the laboratory not re-evaluated vendors?	There is no defined procedure for vendor re-evaluation.
WHY	Why is there no defined procedure for vendor re-evaluation?	The importance of periodic re-evaluation was not identified during the development of vendor management procedures.
WHY	Why was the importance of re-evaluation not identified?	Because there was no risk-based approach or detailed quality control review in the vendor management system.
WHY	Why was there no risk-based approach implemented?	The laboratory staff were not trained adequately on vendor quality management and ISO compliance requirements.
WHY	Why was staff not trained on vendor quality management?	Because management did not prioritize training or allocate resources for vendor management compliance.
How	How to prevent recurrence of this NC?	Develop and implement a documented procedure for vendor re-evaluation, specifying


	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

		frequency, criteria, and methods for assessment. • Provide training to relevant staff on the importance and procedures for vendor re-evaluation.
How	How to ensure the effectiveness of the corrective actions?	• Conduct internal audits to verify that vendor re-evaluation practices are being followed. • Use a risk-based approach to categorize vendors and apply stricter evaluation for high-risk suppliers.

Root Cause: The lab management failed to recognize the importance of re-evaluating vendors and did not include it in their processes. This happened because there was no proper training or focus on vendor management requirements during procedure development.

RCA Statement: The lab management failed to recognize the importance of re-evaluating vendors and did not include it in their processes. This happened because there was no proper training or focus on vendor management requirements during procedure development.

Reviewed by: 

Evaluated by: 

	DOCUMENT CHANGE REQUEST	Doc.#	QCL-FRM-7.02
		Issue Date	01-02-2025
		Rev#	02

CONTROLLED COPY

NO. Section A


DCR No.	05-25	Date	01-02-25
Request Raised By	Masab Rasheed	Section	QC Lab
Document Title		Document Status	NEW EXISTING ☒
Reason(s)	☐	New Process Documentation	New Form Documentation
	☒	Addition in Document	Correction in Document
	☐	Deletion within Document	Modification in Document

Section B

Description of Requirement: Revision required to address the shortcomings observed in the PNAC audit.

Effected Sections (Not Applicable on New Documents): Addition of evaluation and re-evaluation criteria and revise products inspection form

Section C


Review by	Himayat Hussain	Review Status (Choose only one)	☒ Draft Reviewed
Designation			☐ Draft Rejected
			☐ Draft Revision Suggested
Reason for Rejection (Choose only one)	☐	Conflict with Policy	Conflict with Procedure
	☐	Conflict with Client Requirement	Conflict with Regulation
	☐	Incorrect Draft	Draft Does not reflect "Description of Requirement"
Final Remarks	Recommended for Final Approval		
Submitted for Final Approval on (Date)			
Approved By		Designation	

	Externally provided products and services	Doc. No.	QCL-LSP-06
		Issue Date	01-02-2025
		Rev#	02

MASTER **CONTROLLED COPY**
NO. _____

Standard Operating Procedure for Externally provided products and services

ISO/IEC 17025:2017 – Clause 6.6

Prepared By: Masab Rasheed	Reviewed By: Himayat Hussain	Approved By: Ahmad Umair
AQCM	QCM	CEO
		
Date: 01-02-25	Date: 01-02-25	Date: 01-02-2025

Copyright

The copyright of this procedure, together with all confidential information contained herein is the sole property of **QC LAB**.
It may be copied in full or in parts only by the Management / personnel and only for the purpose of Company related activities. Disclosure and copying of any information contained within this procedure to any person (s) outside the employee of company without written permission/authorization of the Chief Executive Officer is strictly prohibited.

Amendment Sheet

Date	Revision #	Amendment Detail	Section	DCR	Amendment Approved by
01-10-24	01	Replace issue# with revision#	Header	10-24	Chief Executive Officer
01-10-24	01	Deletion of page number from header	Header	10-24	Chief Executive Officer
01-10-24	01	QC Lab reporting to CEO for Final Approval	Document Approval	10-24	Chief Executive Officer
01-02-25	02	Addition of evaluation & re-evaluation criteria and Revise products inspection form	Document	05-25	Chief Executive Officer

MASTER Externally provided products and services	Doc. No.	QCL-LSP-06
	Issue Date	01-02-2025
	Rev#	02

3.0 Purpose:

The purpose of this procedure is to explain the purchase procedure of QC LAB. All the laboratory purchase activities will also be followed by this procedure

4.0 Scope:

This procedure is applicable to all activities related to purchase process of all required material.

5. Definition, Abbreviation and Concepts:

QC LAB	Quality Control LAB
QMS	Quality Management System
QCM	Quality Control Manager

4.0 Responsibilities and Authorities

- 4.1 Chief Executive Officer
- 4.2 Quality Control Manager
- 4.3 Manager Supply Chain

5.0 Procedure:

- (a) **Vendors for Equipment and Accessories:**
Fulfil technical Specification requirement and provide required documentation (such as equipment manual). This should also include training services against installation and operation of new equipment.
- (b) **Vendor for External Calibration Services:**
In order to meet accreditation requirement, the vendor must be accredited with PNAC. Non-accredited is only recommended if there is no accredited service provider found in the region or calibration is required for out of scope equipment. Calibration service provider have the capability to provide calibration report and calibration certificate including traceability, method, reference, measurement of uncertainty and range of calibration.
- (c) **Vendor for Proficiency Testing (PT) Services:**
Vendor must be accredited for PT services (ISO/IEC 17043) services of non-accredited vendor are not acquired by QC LAB
- (d) **Vendor for Training Services for ISO/IEC17025:2017**
Experience and Qualification of External Trainers, with preference given to trainers having PNAC training and experience of providing their services in similar lab experience.
- (e) **Vendors for Chemical Supply:**
Chemical/ Solution suppliers to provide supporting documentation and designated lab personal needs to ensure availability of required MSDS. They meet the required specification and quantity.

Doc. No.	QCL-LSP-06
Issue Date	01-02-2025
Rev#	02

MASTER

- (f) **Vendors for Standard Reference Material (Certified Reference Material):**
Suppliers must provide traceable certificate of analysis with it, otherwise the product will not be accepted. In addition, preference will be given to supplier who can provide value of uncertainty associated with the product (where applicable).
- (g) **Suppliers for Lab Accreditation:**
Accreditation services are provided exclusively by PNAC as they are only authorized for Lab Accreditation in Pakistan. QC LAB accepts their terms and conditions in order to meet accreditation requirements.

5.1 Criteria for Evaluation and Selection of external provider

Initially new supplier is evaluated/selected for product and services as per following criteria mentioned in section A & B;

Section A: General Evaluation

Section-A: General Evaluation (For New Supplier)					
		Rating			
		Poor	Fair	Good	Excellent
Market	Image and reputation	0	1	2	3
Feedback	Market share		1	2	3
Operational	Technical Capacity*	0	1	2	3
Performance	Lead Time	0	1	2	3
	Product Quality/Services	0	1	2	3
	Order Processing	0	1	2	3
	After sale services	0	1	2	3
		Total Score Achieved			
Grading		Poor:0-6	Fair:7-14	Good: 15-18	Excellent: 19-21
Minimum score to qualify is 15					
*In case of supplies and service provider					

Section B: Technical Evaluation (For New Supplier)

Technical acceptability Criteria	
Products	
CRM	ISO 17034:2016, ISO/IEC 17025:2017, ISO 9001:2015
Solvents	Analytical Grade, Karl Fischer Grade, COA, Authorized Dealer
Chemicals	Analytical Grade, COA, Authorized Dealer
Glassware	ISO/IEC 17025:2017, COA, class A
Flame photometer	Authorized Dealer
Karl Fischer	Authorized Dealer
Analytical Balance	Authorized Dealer

**MASTER**

Externally provided products and services

CONTROLLED COPY
NO. _____

Doc. No.	QCL-LSP-06
Issue Date	01-02-2025
Rev#	02

pH Meter	Authorized Dealer
EC Meter	Authorized Dealer
Services	
Karl Fischer (Maintenance)	Trained and Certified Engineer
Flame Photometer (Maintenance)	Trained and Certified Engineer
Analytical Balance (Maintenance)	Trained and Certified Engineer
pH Meter (Maintenance)	Trained and Certified Engineer
EC Meter (Maintenance)	Trained and Certified Engineer
PT Provider	ISO/IEC 17043
PNAC (Accreditation)	Accreditation from ILAC & APAC
Calibration	ISO/IEC 17025:2017 accredited for calibration
Training	Institution and Trainer is trained and certified from well reputed organization. Trainer also have experience of training.

Vital Agri Nutrients (Private) Limited testing laboratory has defined acceptability criteria as specification for products include measurement standards and equipment, auxiliary equipment, consumable materials and reference materials and maintained etc.

Acceptance criteria is also defined for services including calibration services, facility and equipment maintenance services, proficiency testing services, assessment and auditing services in product/services Inspection form having Doc # QCL-FRM-6.08.

Pre-existing Supplier (product/services) are re-evaluated/monitored as per following criteria given in section C annually.

Section C: Existing Supplier Performance Re-evaluation/ monitoring

Section-C Existing External Provider Performance Re-evaluation/monitoring			
1	Has the supplier regularly met the commitment with respect to: (Yes/No)		
	a. Quality /services requirement?		
	b. Quantity requirement?		
	c. On time delivery/services?		
	d. After sales support and service?		
2	Is the supplier prompt in responding queries?	☐ Yes	☐ No
3	Does the supplier enjoy market reputation?	☐ Yes	☐ No
4	Product/Services been approved by QC section?	☐ Yes	☐ No



Externally provided products and services

Doc. No.	QCL-LSP-06
Issue Date	01-02-2025
Rev#	02

MASTER

CONTROLLED COPY NO. _____

Section D: Approval

Section-C: Approval			
Decision	☐ Approved	☐ Rejected	☐ Continue
Remarks (If any)			
Evaluated By		Date	
Approved By		Date	

Supplier are evaluated at the time of first engagement and every year afterwards. Supplier evaluation form is divided into four sections A, B, C and D. Section A and B is for the initial assessment of suppliers in case of new suppliers while section C will not be filled in this case. Section C is used for re-evaluation/monitoring of the existing supplier and Section A and B will not be filled in this case.

Suppliers are evaluated for Products/services they have provided based on defined criteria in Supplier Evaluation Form by AQCM.

After evaluation, if suppliers meet the defined criteria, they are approved by Quality Control Manager.

When supplier is approved after successful evaluation, it is included in the list of approved suppliers which is maintained by AQCM.

Actions arising from evaluation, Monitoring and re-evaluation of external provider

If the quality of product or service provided by supplier fails to meet the acceptance criteria, AQCM fills corrective action form and after approval from Quality Control Manager, AQCM will communicate it to the supplier and get written corrective action from the supplier.

A copy of corrective action form and return response from supplier will be maintained by VAN QC Testing Laboratory. A supplier may lose approved supplier status if complaints are not resolved.

5.2 Type of Purchasing and their requirement identification

Following products purchasing is performed in QC LAB.

S/N o.	Product / Service	Purchase Follow-up by	Need Identify and Approval By
1.	Apparatus, glassware's, chemicals, reference materials	AQCM	Quality Control Manager
2.	Services like (consultancy, training & auditing, computer maintenance, Calibration Services, PT service, Cargo & clearing and	Quality Control Manager	CEO

 Externally provided products and services MASTER CONTROLLED COPY NO. _____	Doc. No.	QCL-LSP-06
	Issue Date	01-02-2025
	Rev#	02

	forwarding etc.)		
3.	Maintenance services	AQCM	QCM/AQCM

5.3 Material Issue and Purchase Requisition:

- Lab personal rises demand note before the item reaches its reorder level or in case of any emergent need of supplies or services or purchase of any equipment, material, or service.
- If the item is available in store, then it is issued to lab otherwise demand is raised and forwarded to accounts department /procurement section after getting required approval.
- In case of equipment, pre-approval from CEO is required. After approval supply chain manager contacts different suppliers for quotation and upon receiving different quotations, a comparative study is conducted Quality Control Manager presents his comparative study to CEO for final approval.
- After final approval, Supply Chain Manager places order to relevant supplier.

5.4 Receiving, Inspection and Verification

On receiving, the items are extensively inspected by the lab authorized person. The parameters / criteria against which the inspection has to be made on Purchase inspection form/ Purchase Order by the AQCM for which the purchase has been made or by the required applicable standard.

For those items that have passed the acceptance criteria, Material Receipt Note / Material Inspection Note (MRN/MIN) is prepared, and inventory records are maintained. Items failing to meet required standards are sent back to the supplier.

5.5 Disposal of damaged or expired items:

- Damaged and expired items are disposed in time according to 5.4.3
- Disposal is done with the permission of Quality Manager.
- All chemical waste is collected in containers and sent for recycling.
- Near expiry items are replaced by the supplier (if possible).

6. Rejection Criteria and Acceptance criteria

QC LAB accepts and reject laboratory glassware's, chemicals and apparatus as per below standard

VAN

MASTER

Externally provided products and services

Doc. No.

QCL-LSP-06

Issue Date

01-02-2025

Rev#

02

CONTROLLED COPY

NO. _____

6.1. Glassware's

- Save packaging of glassware's
- Transported under the proper conditions
- A grade material if not available provide calibration certificate
- Glass is resistant to high temperatures and most chemicals
- Readable scale
- On-time delivery
- Market reputed
- Measuring Accuracy

6.2. Chemicals

- COA/ MSDS must available all chemical
- Vendor is responsible to provide production & expiry date
- Chemical strength as per demand
- Moisture free packing
- Seal insured, packing box is not broken
- On-time delivery
- Market reputed

6.3. Apparatus

- Measuring Accuracy
- Clearly readable scale
- On-time delivery
- Market reputed
- pH meter automatic temperature calibration, range 0-14, accuracy ± 0.01 , digital
- Refrigerator/ Heat and dry oven
- ✓ Reliable



MASTER

Externally provided products and services

Doc. No.	QCL-LSP-06
Issue Date	01-02-2025
Rev#	02

CONTROLLED

- ✓ Quality
- ✓ H& DO as specified in purchase order
- Analytical balance
 - ✓ Range as specified in purchase order
 - ✓ Accuracy as specified in purchase order
- Thermometer
 - ✓ Range as specified in purchase order
 - ✓ Accuracy as specified in purchase order

6.4. Services

- PNAC Accredited
- Qualified (QMS)
- Market reputed
- Reliable result
- Low price
- On-time services

7.0. Receiving and Issuance of Packaging Material

7.0.1 Purchased items are received by related Store In charge.

7.0.2 On the basis of Demand, Store Incharge should issues packaging material to Production department on FIFO (First in First Out) basis.

7.1 Storage

Items received are shifted to designated storage locations according to their storage requirements as mentioned in COA of products or in any other standard applicable.



Externally provided products and services

Doc. No.

QCL-LSP-06

Issue Date

01-02-2025

Rev#

02

MASTER

CONTROLLED COPY
NO. _____

8.0. Related Documents

Description	Doc#
Supplier Selection Form	QCL-FRM-6.01
Approved Supplier/ Services Provider List	QCL-FRM-6.02
External Provider Evaluation & Re-evaluation Form	QCL-FRM-6.03
Supplier Performance Monitoring	QCL-FRM-6.04
Comparative statement	QCL-FRM-6.05
Supplier Evaluation Form	QCL-FRM-6.06
Store Purchase Demand (SPD)	QCL-FRM-6.07
Products Services Inspection Form	QCL-FRM-6.08

	EXTERNAL PROVIDER EVALUATION & RE-EVALUATION FORM CONTROLLED COPY	Doc. #	QCL-FRM-6.03
		Issue Date:	1/2/2025
		Rev.#	02

☐ New Supplier	NO.	☐ Existing Supplier
---------------------------------------	------------	--

Supplier's Name			
Product/Service			
Postal Address			
Telephone #		Fax #	
Email		Website	

Section-A: General Evaluation(For New Supplier)

		Rating			
		Poor	Fair	Good	Excellence
Market Feedback	Image and reputation	0	1	2	3
	Market share	0	1	2	3
Operational Performance	Technical Capacity*	0	1	2	3
	Lead Time	0	1	2	3
	Product Quality/Services	0	1	2	3
	Order Processing	0	1	2	3
	After sale Services	0	1	2	3
		Total Score Achieved			
Grading		Poor:0-6	Fair:7-12	Good: 15-18	Excellent: 19-21
Minimum score to qualify is 15					
*In case of supplies and service provider					

Section B: Technical Evaluation (For New Supplier)

Technical Requirement			
Fullfill Technical Requirement	☐ Yes	☐ No	☐ N/A

Section-C Existing External Provider Performance Re-evaluation/monitoring

1	Has the supplier regularly met the comitment with respect to:		
	a. Quality /services requirement?		
	b. Quantity requirement?		
	c. On time delivery/services?		
	d. After sales support and service		
2	Is the supplier prompt in responding queries?	☐ Yes	☐ No
3	Does the supplier enjoy market reputation?	☐ Yes	☐ No
4	Has sample of product/services been approved by QC section?	☐ Yes	☐ No

Section-C: Approval

Decision	☐ Approved	☐ Rejected	☐ Continue
Remrks (If any)			
Evaluated By		Date	
Approved By		Date	

	PRODUCTS SERVICES INSPECTION FORM	Doc. #	QCL-FRM-6.08
		Issue Date	01-02-2023
		Rev #	02

CONTROLLED COPY

Type of Purchase:

☒ ~~Products~~

☐ Services

1. Products / Services Info:

01	Products/Services	
02	External Provider	
03	Provider Status	☐ Newly engaged ☐ Pre-existing
04	Gate Pass #	
05	PR No./PO No.	
06	Invoice #	
07	Date of receipt	

Products and Services Acceptance/Rejection Criteria

Sr.No	Requirements for Products	Remarks
01	Product COA	
02	Specification as required	
03	Desired Quantity	
04	Verification of Lot No.	
05	MFG/EXP	
06	Model/Make	

Sr.No	Requirements for Services	Remarks
01	Traceability of accreditation	
02	Scope of Accreditation	
03	Traceability of Training	
04	Response to queries	
05	Trainer/engineer skills	
06	Validity of certificate	

Accepted

☐

Rejected

☐

Note: If expiry and lot No. do not comply to COA of product then VAN QC Lab reject the external provider products.

If Traceability, scope of accreditation and after sale sport is not adequate as required then VAN QC Lab will reject external provider Services.

Received by

Inspected by