

	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT		F-02/06
			Issue Date: 11/11/19
			Rev No: 06

ACTIVITY:	☒ Assessment ☐ Surveillance		NCR No: 2
	☐ Re-Assessment ☐ Scope extension		Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category: Very serious ☐ Essential ☒
			Time limit for correction:
ORGANIZATION:	Quality control Lab vital Agri. Nutrients (Pvt) ltd.		6 weeks
Department:	Labs		
Team Leader:	Mr. Atique Ur Rehman	Mngt. Rep.	Mr. Hammyat Hussain
Assessor	Dr. Muhammad Aftab		
Expert			
DESCRIPTION: The lab is monitoring the temperature and humidity by using thermo hygrometer (VAN/PFTL/EQP-001) in main lab hall while in sample storage room by using the Thermo hygrometer (VAN/PFTL/EQP-002) all are calibrated but the lab has not authorized the person for this activity (NC). The lab has not developed and maintained the log books for analytical weighing balance (QCL-INS-01) and Block heater (QC-INS-023) (NC).			Requirement of: ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
10-01-2025 Date Signature Assessor/Expert Signature (Management Representative)			Ref. Organization's doc.
CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcomings on Both NCs has been taken by revising the Authorization permit of Lab Assistant (QCL-FRM-2.01). Lab has maintained the log books for analytical weighing balance (QCL-LB-07) and Block heater (QCL-LB-08). Action are documented in the amendment no: <u>CA-02-02-2025</u> <u>08-02-25</u> Signature (Management Representative)			☐ It is not necessary to attach documentation
REASON FOR CLOSING: (To be filled in by the Team Leader.) ☐ The non-conformity is closed based on satisfactory documentation from the organization ☐ The non-conformity is closed based on recommendation from the technical assessor ☐ Implementation of the corrective actions will be followed up at the next visit Remarks (if any) ----- ----- ----- The non-conformity is closed: _____ Date _____ Signature (Team Leader.)			

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-02-02-2025
Designation: AQCM	CAR Initiated Date: 03-02-2025	NC No. 02

CAR Initiated due to (Please tick the relevant appropriate box below)

☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) The lab is monitoring the temperature and humidity by using thermo hygrometer (VAN/PFTL/EQP-001) in main lab hall while in sample storage room by using the Thermo hygrometer (VAN/PFTL/EQP-002) all are calibrated but the lab has not authorized the person for this activity (NC). The lab has not developed and maintained the log books for analytical weighing balance (QCL-INS-01) and Block heater (QC-INS-023) (NC).	
 Signatures	

Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 03-02-2025


* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.

B. Root Cause Analysis (where required)

1. Lack of a formal process or system for documenting and authorizing personnel responsible for monitoring equipment, resulting from insufficient awareness or oversight regarding the quality management system requirements for personnel authorization.
2. Lack of awareness and clarity among responsible personnel regarding the requirement to develop and maintain logbooks for analytical weighing balance (QCL-INS-01) and block heater (QC-INS-023), primarily due to gaps in training, LSP communication, and oversight mechanisms.

C. Proposed Action	Target Date:	08-02-2025
---------------------------	---------------------	-------------------

Corrective action for shortcomings on Both NCs has been taken by revising the Authorization permit of Lab Assistant (QCL-FRM-2.01). Lab has maintained the log books for analytical weighing balance (QCL-LB-07) and Block heater (QCL-LB-08).


Name or Signatures

(Person who carried out Root Cause Analysis and Proposed Corrective Action)

D. Verification by Manager QC for implementation & Effectiveness

Is Proposed Action Implemented?	☒ YES ☐ NO	Is it Effective?	☒ YES ☐ NO
---------------------------------	---	------------------	---

Remarks (If any):


CAR closing date: 08-02-25

Manager QC Sign: 

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab is monitoring the temperature and humidity by using thermo hygrometer (VAN/PFTL/EQP-001) in main lab hall while in sample storage room by using the Thermo hygrometer (VAN/PFTL/EQP-002) all are calibrated but the lab has not authorized the person for this activity (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	03-02-2025	Submission Date	08-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	✓

Cause Effect Diagram



3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why was the NC raised?	The lab has not authorized the person responsible for monitoring the thermo-hygrometers.
WHY	Why was the person not authorized?	There was no formal process or system in place to identify, authorize, and document personnel responsibilities for monitoring equipment.
WHY	Why was there no formal process or system for authorization?	The lab management assumed verbal instructions were sufficient and did not prioritize implementing a formal authorization process.
WHY	Why did the management assume verbal instructions were sufficient?	Lack of awareness or oversight regarding the importance of documented authorization as part of the lab's quality management system requirements.
WHY	Why was there a lack of awareness or oversight?	Insufficient training or orientation of lab management and staff on the requirements for personnel authorization under the lab's quality management system.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

How	How did this happen?	• No documented Lab system procedure (LSP) was in place for assigning and authorizing personnel responsible for equipment monitoring. • Management relied on informal practices, leading to an oversight in formalizing the authorization process.
How	How can this be prevented in the future?	• Develop and implement an LSP for assigning, authorizing, and documenting personnel responsibilities for equipment monitoring, ensuring it aligns with quality management system requirements. • Train all relevant staff on the importance of documented authorization and the associated LSP to prevent recurrence. • Periodically audit and review personnel authorization records to ensure compliance.
Root Cause: Lack of a formal process or system for documenting and authorizing personnel responsible for monitoring equipment, resulting from insufficient awareness or oversight regarding the quality management system requirements for personnel authorization.		

RCA Statement: Lack of a formal process or system for documenting and authorizing personnel responsible for monitoring equipment, resulting from insufficient awareness or oversight regarding the quality management system requirements for personnel authorization.

Reviewed by: 

Evaluated by: 




**PERSONNEL AUTHORIZATION
PERMIT**

Doc. No.	QCL/FRM-2.01
Issue Date	01-02-2025
Rev#	02

**CONTROLLED COPY
NO.**

1. General

Employee Name	Muhammad Mursaleen
Employee ID	N-069
Designation	Lab Assistant
Section	QC LAB
Laboratory	QC LAB
Academic Qualification	BA
Years of Experience	8 Years
Contact Number	03016806205
Email ID	m.mursaleen15322@gmail.com
Verification/Validation of Testing Methods	No
Approval & Signing of Analysis Report	No

Prepared By: Masab Rasheed AQCM	Reviewed By: Himayat Hussain QCM	Approved By: Ahmad Umair CEO
		
Date: 01-02-25	Date: 01-02-25	Date: 01/02/2025



**PERSONNEL AUTHORIZATION
PERMIT**

CONTROLLED COPY

Doc. No. QCL/FRM-2.01

Issue Date 01-02-2025

Rev# 02

3. Authorization: Lab Assistant NO. _____ Employee Name: Muhammad Mursaleen

Sr. #	Authorize to following activities	Work Type	Supervision	Relevant Data Access
1	House Keeping	LAB	AQCM/QCM	Lab Staff
2	Stock Management	All Lab Chemicals and Glassware's	AQCM/QCM	Lab Staff
3	Assist Density Test	All Liquid Products	AQCM/QCM	Lab Staff
4	Assist pH Test	For Liquids and Solids Products	AQCM/QCM	Lab Staff
5	Assist Conductivity	For Liquids and Solids Products	AQCM/QCM	Lab Staff
6	Assist to Inspects Raw Materials for Lab Use	All Materials	AQCM/QCM	Lab Staff
7	Sample Receiving	All Products	AQCM/QCM	Lab Staff
8	Identification of Non-Conformity	All Type	AQCM/QCM	Lab Staff
9	Identification of Need for improvement or Deviation	All Type	AQCM/QCM	Lab Staff
10	Assist in Operation of Analytical Balance	All ranges	AQCM/QCM	Lab Staff
11	Assist in Operation of Flame photometer	Potash	AQCM/QCM	Lab Staff
12	Assist in Operation of pH Meter	pH	AQCM/QCM	Lab Staff
13	Assist in Operation of TDS Meter	Conductivity/TDS	AQCM/QCM	Lab Staff
14	Assist in Operation of Multiparameter Photometer	All tests	AQCM/QCM	Lab Staff
15	Assist in Preparing the Analysis Reports	All Analysis Reports	AQCM/QCM	Lab Staff
16	Assist in Prepare the Control Charts	All Control Charts	AQCM/QCM	Lab Staff
17	Assist in Prepare the temperature/humidity charts	All types of Charts	AQCM/QCM	Lab Staff
18	Monitoring of Environmental conditions	Daily Environmental Check Sheet	AQCM/QCM	Lab Staff
19	Monitoring the temperature and humidity	Daily Environmental Check Sheet	AQCM/QCM	Lab Staff

**PERSONNEL AUTHORIZATION****PERMIT
CONTROLLED COPY
NO. _____**


Doc. No.	QCL/FRM-2 01
Issue Date	01-02-2025
Rev#	02

19	Perform the Cleaning Inspections	House Keeping	AQCM/QCM	Lab Staff
20	Perform the Intermediates Checks	All Intermediates Checks of Instruments	AQCM/QCM	Lab Staff
21	Sample Dispose	All Fertilizer Samples	AQCM/QCM	Lab Staff
22	Sample Retain	All Fertilizer Samples	AQCM/QCM	Lab Staff
23	Housekeeping	All	AQCM/QCM	Lab Staff

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab has not developed and maintained the log books for analytical weighing balance (QCL-INS-01) and Block heater (QC-INS-023) (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	03-02-2025	Submission Date	08-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	☒

Cause Effect Diagram



3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why is the logbook not developed and maintained?	Lack of clarity in responsibilities for logbook creation and maintenance.
WHY	Why is there a lack of clarity in responsibilities?	There is no proper training or communication regarding LSP requirements for equipment logbooks.
WHY	Why is there no proper training or communication?	The training schedule does not cover all aspects of equipment management, including logbook requirements.
WHY	Why does the training schedule not cover these aspects?	Training modules are outdated and have not been revised to include the requirement for specific equipment logbooks.
WHY	Why are the training modules outdated?	No formal mechanism is in place for periodic review and updating of training content.
How	How to prevent this issue from recurring?	- Revise SOPs to explicitly include the requirement for logbook creation and maintenance for all equipment. - Provide a template or standardized format for logbooks.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

		• Assign responsibility to a specific individual (e.g., equipment owner or quality officer) for managing logbooks.
How	How to ensure compliance moving forward?	• Establish a routine monitoring and audit process to check the development and maintenance of logbooks. • Take immediate corrective action for any identified non-compliance and communicate the findings to the team.
Root Cause: Lack of awareness and clarity among responsible personnel regarding the requirement to develop and maintain logbooks for analytical weighing balance (QCL-INS-01) and block heater (QC-INS-023), primarily due to gaps in training, LSP communication, and oversight mechanisms.		

RCA Statement: Lack of awareness and clarity among responsible personnel regarding the requirement to develop and maintain logbooks for analytical weighing balance (QCL-INS-01) and block heater (QC-INS-023), primarily due to gaps in training, LSP communication, and oversight mechanisms.

Reviewed by: 

Evaluated by: 



Vital Agri Nutrients Pvt Ltd

CONTROLLED COPY
NO. _____

LOG BOOK

	
Log Book Code	QCL-LB-07
Log Book Weighing Balance	
Volume #	01
Date Started	February 01, 2025
Date Closed	
Keeper	QCM
File Location	Station No. 01



Custodian QCM

CONTROLLED COPY

Equipment Name: Weighing Balance NO. QC/INS/001

Sr. No.	Date	Test	Sample ID	Start Time	End Time	Performed By	Verified by	Remark
001	01-02-25	K ₂ O	RM0546	9:15 AM	9:18 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
002	01-02-25	Moisture	RT0204	12:10pm	12:25pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
003	01-02-25	Humic	"	"	"	Mubeen	<i>[Signature]</i>	OK
004	01-02-25	K ₂ O	"	"	"	<i>[Signature]</i>	<i>[Signature]</i>	OK
005	01-02-25	Moisture, K ₂ O, Humic	RT0205	12:25pm	12:40pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
006	01-02-25	Moisture, K ₂ O, Humic	RT0206	12:40pm	12:50pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
007	01-02-25	Moisture, K ₂ O, Humic	RT0207	12:50pm	1:05pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
008	01-02-25	Moisture, K ₂ O, Humic	TP0005	9:50 AM	10:10 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
009	01-02-25	Sulfur	PQ0301	3:10 pm	3:15 pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
010	01-02-25	Mg	RM0548	3:12 pm	3:17 pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
011	01-02-25	Humic Acid	RM0550	3:20 pm	3:30 pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
012	01-02-25	Moisture	"	"	"	<i>[Signature]</i>	<i>[Signature]</i>	OK
013	02-02-25	Extraction Solution	STBL3357	9:10 AM	9:20 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
014	02-02-25	Moisture, K ₂ O, Humic	SFO097	2:30 AM	2:45 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
015	02-02-25	K ₂ O, Nitrogen	TP0490	10:00 AM	10:20 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
016	02-02-25	P.H, Moisture	"	"	"	Mubeen	<i>[Signature]</i>	OK
017	02-02-25	Moisture, Humic, K ₂ O, PH	RT0208	10:20 AM	10:30 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
018	02-02-25	Moisture, Humic, K ₂ O, PH	RT0209	10:40 AM	10:50 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK



Vital Agri Nutrients Pvt Ltd

CONTROLLED COPY
NO.

	
Log Book Code	QCL-LB-08
Log Book Block Heater/Nitrogen Digester	
Volume #	01
Date Started	February 01, 2025
Date Closed	
Keeper	QCM
File Location	Station No. 07



Custodian

QCM

CONTROLLED COPY

Equipment Name: Block Heaver

和

Equipment ID: QC/INS/023

[illegible]