




 PNAC Pakistan National Accreditation Council	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT	F-02/06
		Issue Date: 11/11/19
		Rev No: 06


ACTIVITY:	☒ Assessment	☐ Surveillance	NCR No: 3
	☐ Re-Assessment	☐ Scope extension	Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category: Very serious ☐ Essential ☒
			Time limit for correction:
ORGANIZATION:	Quality control Lab vital Agri. Nutrients (Pvt) ltd.		6 weeks
Department:	Labs		Mr. Hammyat Hussain
Team Leader:	Mr. Atique Ur Rehman	Mngt. Rep.	
Assessor	Dr. Muhammad Aftab		
Expert			
DESCRIPTION: All the equipment/apparatus have unique identification number and calibration sticker, wherever required and the equipment logbook analytical balance (QC-INS-001) not maintained (NC). The Calibration plan (QCL-FRM-4.04) for the year 2024 is available in file no (QCL-FIL-22). the lab has develop the Equipment maintenance records sheet (QCL-FRM-4.03) and records related to Equipment maintenance is not maintained (NC). 10-01-2025 Date  Signature Assessor/Expert  Signature (Management Representative)			Requirement of: ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcomings on Both NCs has been taken by maintaining the log book for analytical weighing balance (QCL-LB-07). Lab has started to maintain the equipment maintenance records. Action are documented in the amendment no: <u>CA-03-02-2025</u> <u>07-02-25</u> Date  Signature (Management Representative)			☐ It is not necessary to attach documentation
REASON FOR CLOSING: (To be filled in by the Team Leader.) ☐ The non-conformity is closed based on satisfactory documentation from the organization ☐ The non-conformity is closed based on recommendation from the technical assessor ☐ Implementation of the corrective actions will be followed up at the next visit Remarks (if any) ----- 			
The non-conformity is closed: _____ Date Signature (Team Leader.)			

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-03-02-2025
Designation: AQCM	CAR Initiated Date: 03-02-2025	NC No. 03

CAR Initiated due to (Please tick the relevant appropriate box below)


- ☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) All the equipment/apparatus have unique identification number and calibration sticker, wherever required and the equipment logbook analytical balance (QC-INS-001) not maintained (NC). The Calibration plan (QCL-FRM-4.04) for the year 2024 is available in file no (QCL-FIL-22). the lab has developed the Equipment maintenance records sheet (QCL-FRM-4.03) and records related to Equipment maintenance is not maintained (NC).		
 Signatures		
Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance	
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 03-02-2025	
* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.		
B. Root Cause Analysis (where required) 1. The lack of a systematic process for ensuring logbook entries is due to insufficient awareness and training on LSP adherence, coupled with inadequate oversight and accountability mechanisms within the Laboratory Management System (LMS). 2. The absence of a systematic review and training program on the maintenance procedure and record-keeping requirements.		
C. Proposed Action	Target Date:	07-02-2025
Corrective action for shortcomings on Both NCs has been taken by maintaining the log book for analytical weighing balance (QCL-LB-07). Lab has started to maintain the equipment maintenance records.		
 Name or Signatures (Person who carried out Root Cause Analysis and Proposed Corrective Action)		
D. Verification by Manager QC for implementation & Effectiveness		
Is Proposed Action Implemented?	☒ YES ☐ NO	Is it Effective? ☒ YES ☐ NO
Remarks (If any):		
CAR closing date: <u>07-02-25</u> Manager QC Sign: 		

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description					
All the equipment/apparatus have unique identification number and calibration sticker, wherever required and the equipment logbook analytical balance (QC-INS-001) not maintained (NC).					
RCA Team Members		Himayat Hussain (QCM), Masab Rasheed (AQCM)			
RCA Team Lead / Head		Himayat Hussain (QCM)			
RCA Start Date		03-02-2025	Submission Date		07-02-2025
Methodology / Tools Used					
Cause & Effect Diagram	☐	5WHY analysis	☐	3W2H / 5W2H	✓

Cause Effect Diagram



3W2H /5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why was the logbook not maintained?	The personnel responsible did not record the usage details or updates.
WHY	Why did the personnel not record the details?	Lack of awareness about the importance of maintaining the logbook and the procedural requirement.
WHY	Why was there a lack of awareness?	Training sessions did not emphasize the logbook requirement clearly, or refresher training was not conducted.
WHY	Why was training sessions insufficient or not updated?	There is no established mechanism for periodic review and updating of LSP, including training content.
WHY	Why is there no periodic review mechanism?	Oversight in the lab management system (LMS), and no accountability was assigned for ensuring compliance with procedural updates.
How	How can we correct this issue?	Immediately conduct training or refresher sessions for all QC personnel on the importance of logbook maintenance and adherence to LSP.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

		• Designate a responsible person to monitor logbook entries daily or weekly. • Conduct an audit of other equipment logbooks to ensure compliance across the lab.
How	How can we prevent this issue from recurring?	• Update the training schedule to include regular refresher training focused on equipment logbook maintenance. • Revise the LMS to include a mechanism for periodic review of LSP and their implementation status. • Assign accountability within the team for reviewing logbook entries and maintaining proper documentation. • Establish a checklist or digital tracking system to ensure all equipment has corresponding updated logbooks.

Root Cause: The lack of a systematic process for ensuring logbook entries is due to insufficient awareness and training on LSP adherence, coupled with inadequate oversight and accountability mechanisms within the Laboratory Management System (LMS).

RCA Statement: The lack of a systematic process for ensuring logbook entries is due to insufficient awareness and training on LSP adherence, coupled with inadequate oversight and accountability mechanisms within the Laboratory Management System (LMS).

Reviewed by: 

Evaluated by: 

VAN

Vital Agri Nutrients Pvt Ltd

CONTROLLED COPY
NO. _____

LOG BOOK

	
Log Book Code	QCL-LB-07
Log Book Weighing Balance	
Volume #	01
Date Started	February 01, 2025
Date Closed	
Keeper	QCM
File Location	Station No. 01



CONTROLLED COPY


Custodian QCM

Equipment Name: <u>Weighing Balance</u>				Equipment ID: <u>QC/INS/001</u>				
No. _____								
Sr. No.	Date	Test	Sample ID	Start Time	End Time	Performed By	Verified by	Remark
001	01-02-25	K ₂ O	RM0546	9:15 AM	9:18 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
002	01-02-25	Moisture	RT0204	12:10pm	12:25pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
003	01-02-25	Humic	"	"	"	Mubeen	<i>[Signature]</i>	OK
004	01-02-25	K ₂ O	"	"	"	<i>[Signature]</i>	<i>[Signature]</i>	OK
005	01-02-25	Moisture, K ₂ O, Humic	RT0205	12:25pm	12:40pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
006	01-02-25	Moisture, K ₂ O, Humic	RT0206	12:40pm	12:50pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
007	01-02-25	Moisture, K ₂ O, Humic	RT0207	12:50pm	1:05pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
008	01-02-25	Moisture, K ₂ O, Humic	TP0005	9:50 AM	10:10 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
009	01-02-25	Sulfur	PQ0301	3:10 pm	3:15pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
010	01-02-25	Mg	RM0548	3:12 pm	3:17pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
011	01-02-25	Humic Acid	RM0550	3:20pm	3:30pm	<i>[Signature]</i>	<i>[Signature]</i>	OK
012	01-02-25	Moisture	"	"	"	<i>[Signature]</i>	<i>[Signature]</i>	OK
013	02-02-25	Extraction Solution	STBL3357	9:10 AM	9:20 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
014	02-02-25	Moisture, K ₂ O, Humic	SF0097	2:30 AM	2:45 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
015	02-02-25	K ₂ O, Nitrogen	TP0490	10:00 AM	10:20 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
016	02-02-25	P.H, Moisture	"	"	"	Mubeen	<i>[Signature]</i>	OK
017	02-02-25	Moisture, Humic, K ₂ O, PH	RT0208	10:20 AM	10:30 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK
018	02-02-25	Moisture, Humic, K ₂ O, PH	RT0209	10:40 AM	10:50 AM	<i>[Signature]</i>	<i>[Signature]</i>	OK

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The Calibration plan (QCL-FRM-4.04) for the year 2024 is available in file no (QCL-FIL-22). the lab has developed the Equipment maintenance records sheet (QCL-FRM-4.03) and records related to Equipment maintenance is not maintained (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	03-02-2025	Submission Date	07-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	✓

Cause Effect Diagram



3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why are the maintenance records not maintained?	Maintenance activities are performed, but the team fails to document them.
WHY	Why does the team fail to document maintenance activities?	The team is unaware or not trained on the requirement to maintain records.
WHY	Why are they unaware or untrained about the requirement?	There is a lack of communication and training regarding the importance of record-keeping practices.
WHY	Why is there a lack of communication and training?	The management did not prioritize training or include record-keeping in the Laboratory system procedures (LSPs).
WHY	Why was record-keeping not prioritized in LSP or training?	No formal review mechanism exists to identify gaps in the implementation of maintenance procedures

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

How	How can the issue of unmaintained records be resolved?	Implement immediate training for the team on the importance of maintaining equipment maintenance records. Ensure all team members understand how to use the Equipment Maintenance Records Sheet (QCL-FRM-4.03)
How	How can the issue be prevented in the future?	Develop a formal LSP and checklist that mandates the completion of maintenance records after every activity. Establish a regular review and audit system to ensure compliance with record-keeping requirements.
Root Cause: The absence of a systematic review and training program on the maintenance procedure and record-keeping requirements.		

RCA Statement: The absence of a systematic review and training program on the maintenance procedure and record-keeping requirements.

Reviewed by:  _____

Evaluated by:  _____

CONTROLLED COPY

EQUIPMENT MAINTENANCE RECORD

For the Year: 2024-2025

Sr. #	Name of Equipment	Equipment Identification #	Make & Model / Serial Number	Location of Equipment & Manual	Maintenance Frequency	Maintenance Date	Parts Repaired/ Replaced (if any)	Maintenance By
1	Analytical Weigh Balance	QC-INS-001	S/N2303025	Station# 01	Cleaning (Quarterly)	15/02/25	N/A	Rays Technology Maintenance team
					Calibration (Yearly)	-		
					Compressor check (Bimonthly)	15/02/25		
					Cylinder (Quarterly)	15/02/25		
					Filter Check (Quarterly)	15/02/25		
					Temperature prob (Bimonthly)	15/02/25		
2	Flame Photometer	QC-INS-002	31372	Station# 06	Maintenance (yearly)	-	N/A	Rays Technology Maintenance team
3	Oven	QC-INS-004	China & WT3020	Station# 01	Maintenance of Temperature sensors (yearly)	-	N/A	Rays Technology Maintenance team
4	Vacume Filtration Assembly	QC-INS-005	China & AS20	Station# 06	Calibration (Yearly)	-	-	
5	Heating Mental (Kjeldhal's Apparatus)	QC-INS-006	221234	Station# 05		-	-	

EQUIPMENT MAINTENANCE RECORD

CONTROLLED COPY

NO. _____

Doc. No.	QCL-FRM-4.03
Issue Date:	20-09-2024
Rev. #	01


6	Magnetic Stirrer with hot Plate	QC-INS-007	MS-H280	Station# 04	Maintenance of Temperature sensors (yearly)	-	-	
					Maintenance of Stirring motor	-	-	
7	UV-Vis Spectrophotometer	QC-INS-008	752	Station# 03	Cuvette (As Needed)	-	-	
					Calibration (Yearly)	-	-	
8	Thermo Hygrometer	VAN/PFTL/EQP-001	HTC-1	Station# 04	Calibration (As Needed)	-	-	
9	Set of Masses	VAN-H-1/18	China	Station# 01	Calibration (Yearly)	-	-	
10	Desicator	QC-INS-011	NA	Station# 01	NA	-	-	
11	Muffle Furnace	QC-INS-012	NA	Station# 06	Calibration (Yearly)	-	-	
12	Ultra-Sonic Bath	QC-INS-013	NA	Station# 01	NA	-	-	
13	Distillation Assembly	QC-INS-014	NA	Station# 05	NA	-	-	
14	Refrigerator	QC-INS-015, 016	Haier	Station# 01	Maintenance of Temperature sensors (yearly)	-	-	
15	Printer	QC-INS-017	HP Laser MFP137	QC Office	NA	-	-	
16	pH Meter	QC-INS-018	HI2210	Station# 04	Calibration (Yearly)	-	-	
17	EC Meter	QC-INS-019	HI2300	Station# 04	Calibration (Yearly)	-	-	
18	Karl Fischer	QC-INS-020	HI903	Station# 02	Calibration (Yearly)	-	-	

EQUIPMENT MAINTENANCE RECORD

CONTROLLED COPY
NO. _____

Doc. No.	QCL-FRM-4.03
Issue Date:	20-09-2024
Rev. #	01

19	Multiparameter Photometer	QC-INS-021	HI83300	Station# 03	Calibration (Yearly)	-	-	
20	Mechanical Shaker	QC-INS-022	HY-4A	Station# 03	Calibration (Yearly)	-	-	
21	Block Heater	QC-INS-023	00000D1024	Station# 07	Maintenance of Temperature sensors (yearly)	-	-	
					Calibration (Yearly)	-	-	

Prepared By: Masab Rasheed AQC	Reviewed By: Himayat Hussain QCM	Approved By: Ahmad Umair CEO
		
Date: 01-02-25	Date: 01-02-25	Date: 01/02/2025