

 Pakistan National Accreditation Council	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT		F-02/06	
			Issue Date: 11/11/19	
			Rev No: 06	

ACTIVITY:	☒ Assessment	☐ Surveillance	NCR No: 5
	☐ Re-Assessment	☐ Scope extension	Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category:
			Very serious ☐
			Essential ☒


ORGANIZATION:	Quality control Lab vital Agri. Nutrients (Pvt) ltd.	Time limit for correction:
Department:	Labs	6 weeks

Team Leader:	Mr. Atique Ur Rehman	Mngt. Rep.
Assessor	Dr. Muhammad Aftab	
Expert		

DESCRIPTION: The lab has developed the simple receiving form (QCL-FRM-12.01) and mentioned the criteria for sample acceptance/rejection like improper sample receiving, costumer/sample information is not appropriate, capability and resources do not meet the requirement, out of the detection limit of the equipment etc but the acceptance criteria is not maintained in the procedure, lab needs to revise the procedure (QCL-LSP-22) (NC).	Requirement of: ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
--	---

10-01-2025 Date	 Signature Assessor/Expert	 Signature (Management Representative)	Ref. Organization's doc.
--------------------	---	---	--------------------------

CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcoming on NC has been taken by revising the procedure (QCL-LSP-22) and maintained the acceptance criteria in the procedure.	☐ It is not necessary to attach documentation
--	--

Action are documented in the amendment no: <u>CA-05-02-2025</u> <u>10-02-25</u> Date	 Signature (Management Representative)
--	---

REASON FOR CLOSING: (To be filled in by the Team Leader.)	
☐	The non-conformity is closed based on satisfactory documentation from the organization
☐	The non-conformity is closed based on recommendation from the technical assessor
☐	Implementation of the corrective actions will be followed up at the next visit

Remarks (if any) -----




The non-conformity is closed:	Date	Signature (Team Leader.)
-------------------------------	------	--------------------------

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-05-02-2025
Designation: AQCM	CAR Initiated Date: 04-02-2025	NC No. 05

CAR Initiated due to (Please tick the relevant appropriate box below)


- ☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) The lab has developed the sample receiving form (QCL-FRM-12.01) and mentioned the criteria for sample acceptance/rejection like improper sample receiving, costumer/sample information is not appropriate, capability and resources do not meet the requirement, out of the detection limit of the equipment etc. but the acceptance criteria is not maintained in the procedure, lab needs to revise the procedure (QCL-LSP-22) (NC).		
		 Signatures
Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance	
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 04-02-2025	
* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.		
B. Root Cause Analysis (where required) The lab does not have a structured review and training process for ensuring that procedural updates align with new forms and criteria.		
C. Proposed Action	Target Date:	10-02-2025
Corrective action for shortcomings on NC has been taken by revising the procedure (QCL-LSP-22) and maintained the acceptance criteria in the procedure.		
		 Name or Signatures (Person who carried out Root Cause Analysis and Proposed Corrective Action)
D. Verification by Manager QC for implementation & Effectiveness		
Is Proposed Action Implemented?	☒ YES ☐ NO	Is it Effective? ☒ YES ☐ NO
Remarks (If any):		
CAR closing date: <u>10-02-25</u>		Manager QC Sign: 

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab has developed the simple receiving form (QCL-FRM-12.01) and mentioned the criteria for sample acceptance/rejection like improper sample receiving, costumer/sample information is not appropriate, capability and resources do not meet the requirement, out of the detection limit of the equipment etc. but the acceptance criteria is not maintained in the procedure, lab needs to revise the procedure (QCL-LSP-22) (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	04-02-2025	Submission Date	10-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	☒

Cause Effect Diagram




3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why are the acceptance criteria not mentioned in the procedure (QCL-LSP-22)?	The lab focused on developing the form (QCL-FRM-12.01) but overlooked updating the procedure
WHY	Why was the procedure (QCL-LSP-22) overlooked during the form development?	There was no defined process for ensuring alignment between forms and procedures.
WHY	Why is there no defined process for alignment?	The lab lacks a review mechanism to ensure procedural updates accompany new forms or documents.
WHY	Why is the review mechanism absent?	There is inadequate training or awareness among personnel about maintaining document consistency.
WHY	Why is there inadequate training?	The lab does not have a formal training program or checklist for document control and procedural updates.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

How	How can the lab prevent a similar issue in the future?	Implement a document control procedure that mandates reviewing and updating related procedures whenever new forms or criteria are developed.
How	How can the lab ensure personnel are aware of the need for consistency?	Establish a training program to educate personnel on document control practices, emphasizing the importance of aligning forms, procedures, and criteria.
Root Cause: The lab does not have a structured review and training process for ensuring that procedural updates align with new forms and criteria.		

RCA Statement: The lab does not have a structured review and training process for ensuring that procedural updates align with new forms and criteria.

Reviewed by: 

Evaluated by: 

	DOCUMENT CHANGE REQUEST	Doc.#	QCL-FRM-7.02
		Issue Date	01-02-2025
		Rev#	02

CONTROLLED COPY


NO. Section A

DCR No.	06-25	Date	05-02-25
Request Raised By	Masab Rasheed	Section	QC Lab
Document Title	Handling of Test and Calibration items	Document Status	NEW EXISTING ☒
Reason(s)	☐	New Process Documentation	New Form Documentation
	☒	Addition in Document	Correction in Document
	☐	Deletion within Document	Modification in Document

Section B

Description of Requirement	Revision required to address shortcomings observed in PNAC Audit.
Effected Sections (Not Applicable on New Documents)	Addition of Acceptance criteria of sample and appropriate information in the procedure.

Section C



Review by	Himayat Hussain	Review Status (Choose only one)	☒	Draft Reviewed
Designation			☐	Draft Rejected
			☐	Draft Revision Suggested
Reason for Rejection (Choose only one)	☐	Conflict with Policy	☐	Conflict with Procedure
	☐	Conflict with Client Requirement	☐	Conflict with Regulation
	☐	Incorrect Draft	☐	Draft Does not reflect "Description of Requirement"
Final Remarks	Recommended for final approval.			
Submitted for Final Approval on (Date)				
Approved By		Designation	CEO	

 MASTER	HANDLING OF TEST AND CALIBRATION ITEMS	Doc. No.	QCL-LSP-22
		Issue Date	05-02-2025
		Rev#	01

Procedure for Handling of Test and Calibration Items

ISO/IEC 17025:2017 Clause 7.4

CONTROLLED COPY
NO. _____

Prepared By: Masab Rasheed	Reviewed By: Himayat Hussain	Approved By: Ahmad Umair
AQCM	QCM	CEO
		
Date: 05-02-25	Date: 05-02-25	Date: 05/02/2025

Copyright

The copyright of this procedure, together with all confidential information contained herein is the sole property of QC LAB.

It may be copied in full or in parts only by the Management / personnel and only for the purpose of Company related activities. Disclosure and copying of any information contained within this procedure to any person (s) outside the employee of company without written permission/authorization of the Chief Executive Officer is strictly prohibited.

 MASTER CONTROLLED COPY	HANDLING OF TEST AND CALIBRATION ITEMS		Doc. No.	QCL-LSP-22
			Issue Date	05-02-2025
			Rev#	01

PURPOSE:

NO. _____

The purpose of this document is to define the process for handling of test and calibration items and conclusions as prescribed by the organization.

SCOPE:

This procedure is applicable for VAN QC Lab's transport, receipt, handling, protection, storage, retention, and disposal or return of test items.

PROCEDURE:


Acceptance Criteria for Sample Receiving:

The laboratory will only accept samples that meet the following requirements:

- ✓ **Proper Condition**
 - Samples must be well-packaged, undamaged, and uncontaminated.
- ✓ **Complete Information**
 - The submission form must include all necessary details, such as:
 - Customer name and contact information.
 - Sample identification.
 - Collection date and requested tests.
 - Forms with missing or incorrect details will not be accepted.
- ✓ **Lab Capability**
 - The lab must have the equipment, expertise, and resources to handle the requested tests.
 - If the lab cannot perform the test, the sample will not be accepted.
- ✓ **Equipment Detection Limits**
 - Tests must fall within the detection range of the lab's equipment.
 - If not, the sample will be rejected.
- ✓ **Compliance with Rules**
 - Samples must follow lab safety and regulatory standards.
 - Non-compliant samples will not be accepted.
- ✓ **Timely Submission**
 - Samples must arrive within the required timeframe to ensure accurate testing.
 - Late or expired samples will be rejected.
- ✓ **Special Requirements**
 - Samples needing special handling (e.g., refrigeration or hazardous materials) must meet the required conditions.
 - If not, they will be rejected.

1. Receipt

- QC Lab receives the sample from Customer and signs on the covering document (Customer agreement and analysis request) available with the Lab.
- Before receiving, QC Lab must inspect the samples packaging and physical state.
- Receiving is recorded in the sample receipt register available with QC Lab.

	Doc. No.	QCL-LSP-22
	Issue Date	05-02-2025
	Rev#	01

2. Transport

- Lab Representative/Incharge sends the sample to the respective analyst immediately after receiving.

3. Handling

- Lab staff receives the sample and records it in the sample log register.
- The samples are placed at their appropriate specified places until initiation of analytical process.
- Perishable samples are placed in the refrigerator while non-perishable samples are placed in dry, isolated storage.
- The samples are homogenized before analysis.
- One part of the homogenized sample is used for analysis.

4. Protection

- The samples have remained within the lab premises throughout its testing/storage cycle.
- The lab access is limited and specified while the lab remains locked during off hours.

5. Retained

- The samples are retained after analysis at specified, isolated and dry place while perishable samples are stored in refrigerator.

6. Retention

- Sample retention period is 02 months after issuance of report.

7. Disposal/Discard or return

- After specified retention period, the samples are disposed of appropriately by lab Attendant and disposal record is maintained. Returning of sample is generally not allowed; however, high value samples could be returned based on prior agreement and receiving must be signed by the client on sample return form (QCL-FRM-22.04)

8. Precautions:

The following precautions must be taken to avoid any damage to the sample: -

- The collected sample must be representative of the whole. When the sample is received, the customer should ensure that the sample is homogenized and the aliquot that is going to the lab must be a representative of that whole sample.
- Samples must be stored in a sealed container to avoid any cross-contamination.
- Proper sample records of the test samples and those that need to be returned must be maintained to avoid any loss.
- Storage conditions of the samples must be maintained to avoid any deterioration.

9. Related documents

- Customer agreement and Analysis Request
- Assignment, Summary and Review Form
- Sample Return Form
- Sample Discard Register
- Retaining Sample Register

(QCL-FRM-12.01)
(QCL-FRM-19.01)
(QCL-FRM-22.01)
(QCL-REG-08)
(QCL-REG-09)