


	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT	F-02/06
		Issue Date: 11/11/19
		Rev No: 06

ACTIVITY:	☒ Assessment	☐ Surveillance	NCR No: 8
	☐ Re-Assessment	☐ Scope extension	Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category: Very serious ☐ Essential ☒
	ORGANIZATION: Quality control Lab vital Agri. Nutrients (Pvt) ltd.		Time limit for correction: 6 weeks
Department: Labs	Team Leader: Mr. Atique Ur Rehman		Mngt. Rep. Mr. Hammyat Hussain
	Assessor: Dr. Muhammad Aftab		
Expert:			

DESCRIPTION: The lab analyst involved in reporting of Uncertainty of measurements have not sufficient knowledge over MU calculation lab staff involved in the calculation of MU needs training (NC). The raw data related to uncertainty measurements Nitrogen Phosphorus is not properly available (NC)	Requirement of: ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
10-01-2025 Date	 Signature (Management Representative)
CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcomings on both the NCs has been taken by conducting internal training of Lab Staff on Measurement of Uncertainty Certificate of training attached and raw data related to uncertainty measurements of Nitrogen, phosphorus is maintained.	
Action are documented in the amendment no: <u>CA-08-02-2025</u> <u>15-02-25</u> Date	
Signature (Management Representative)	


REASON FOR CLOSING: (To be filled in by the Team Leader.) ☐ The non-conformity is closed based on satisfactory documentation from the organization ☐ The non-conformity is closed based on recommendation from the technical assessor ☐ Implementation of the corrective actions will be followed up at the next visit	☐ It is not necessary to attach documentation
Remarks (if any) -----	
The non-conformity is closed: _____ Date	
Signature (Team Leader.)	

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-08-02-2025
Designation: AQCM	CAR Initiated Date: 10-02-2025	NC No. 08

CAR Initiated due to (Please tick the relevant appropriate box below)

- ☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) The lab analyst involved in reporting of Uncertainty of measurements have not sufficient knowledge over MU calculation lab staff involved in the calculation of MU needs training (NC). The raw data related to uncertainty measurements Nitrogen Phosphorus is not properly available (NC)	
 Signatures	


Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 10-02-2025

* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.

B. Root Cause Analysis (where required) 1. The lack of a structured competency assessment and training program for MU calculations. This was due to assumptions about staff knowledge and a failure to prioritize skill development for this critical task. 2. Lack of a structured system to ensure procedure compliance, training, and monitoring for quality control.

C. Proposed Action	Target Date:	15-02-2025
---------------------------	---------------------	------------

Corrective action for shortcomings on both the NCs has been taken by conducting internal training of Lab Staff on Measurement of Uncertainty Certificate of training attached and raw data related to uncertainty measurements of Nitrogen, phosphorus is maintained.


Name or Signatures

(Person who carried out Root Cause Analysis and Proposed Corrective Action)

D. Verification by Manager QC for implementation & Effectiveness			
Is Proposed Action Implemented?	☒ YES ☐ NO	Is it Effective?	☒ YES ☐ NO

Remarks (If any):


CAR closing date: 15-02-25

Manager QC Sign: 


	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab analyst involved in reporting of Uncertainty of measurements have not sufficient knowledge over MU calculation lab staff involved in the calculation of MU needs training (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	10-02-2025	Submission Date	15-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	✓

Cause Effect Diagram





3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why is there insufficient knowledge about MU calculation?	The lab staff did not receive proper training on MU calculation methods.
WHY	Why did the staff not receive proper training?	There was no structured training program or plan for MU calculations in place.
WHY	Why was there no training program?	The management did not prioritize MU training as a critical competency for staff.
WHY	Why did the management not prioritize this training?	Management assumed the existing knowledge of staff or basic qualifications were sufficient.
WHY	Why was this assumption made?	There was no formal assessment of staff competency for MU calculations before assigning the task.
How	How can we prevent this issue?	- Implement a structured training program specifically targeting MU calculations. - Conduct periodic competency assessments to identify gaps in knowledge.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

How	How can we monitor and ensure improvement?	• Develop a checklist or audit process to verify that staff are proficient in MU calculations. • Assign a trained and experienced mentor to guide staff until they demonstrate full competency.
Root Cause: The lack of a structured competency assessment and training program for MU calculations. This was due to assumptions about staff knowledge and a failure to prioritize skill development for this critical task.		

RCA Statement: The lack of a structured competency assessment and training program for MU calculations. This was due to assumptions about staff knowledge and a failure to prioritize skill development for this critical task.

Reviewed by: 

Evaluated by: 

Certificate

This is to certify that

CONTROLLED COPY
NO. _____

Mr. Masab Rasheed

has participated in one day training on


**Measurement Uncertainty Training
course based on ISO/IEC 17025:2017**

*(General requirements for measurement
uncertainties of testing)*


Conducted by

Vital Agri Nutrients (Private) Limited

on February 12, 2025



HIMAYAT HUSSAIN
Trainer/QCM
Vital Agri Nutrients




AHMAD UMAIR
Chief Executive Officer
Vital Agri Nutrients

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The raw data related to uncertainty measurements Nitrogen Phosphorus is not properly available (NC)			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	10-02-2025	Submission Date	15-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	✓

Cause Effect Diagram





3W2H / 5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why is the raw data for uncertainty measurements of Nitrogen and Phosphorus not properly available?	The data collection process was not consistently followed.
WHY	Why was the data collection process not consistently followed?	Because there was insufficient training or understanding of the procedure by the personnel involved.
WHY	Why was there insufficient training or understanding of the procedure?	Because no formal training or refreshers were conducted after the procedure was updated or developed.
WHY	Why were no formal training or refreshers conducted?	There was no dedicated process to ensure that updated procedures are communicated and implemented effectively.
WHY	Why is there no dedicated process to ensure procedure implementation?	The organization lacks a structured system to track and monitor procedural compliance for quality control.
How	How can the immediate issue (non-compliance with procedure) be corrected?	- Conduct a thorough review of the procedure for measuring Nitrogen and Phosphorus uncertainty. - Organize immediate training sessions for relevant personnel to

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

		ensure they understand the procedure. • Perform a follow-up audit to verify compliance and accuracy of data collection.
How	How can this issue be prevented in the future?	• Implement a system for tracking procedural updates and their communication to all relevant staff. • Develop a regular training program, including refreshers and assessments, for procedures involving critical measurements. • Establish a quality assurance process to periodically review data collection and procedure adherence.
Root Cause: Lack of a structured system to ensure procedure compliance, training, and monitoring for quality control.		

RCA Statement: Lack of a structured system to ensure procedure compliance, training, and monitoring for quality control.

Reviewed by:  by ASAP

Evaluated by: 

RIZVI

CONTROLLED COPY
NO. _____

REGISTER



REG Code

OCL-REG-05

Uncertainty Calculation Register

Volume #

01

Date Started

12 December 2023

Date Closed

Keeper

Analyst

File Location

Office=R

RIZVI PAPER PRODUCTS, LHR

10.06.24

VL - NPK

$K_2O = ?$

Sample vol. 1ml/100ml
100L/10000L

Reading on Flamephotometer.

conc	Abs
------	-----

0	0
---	---

10	54
----	----

15	79
----	----

20	100
----	-----

Sample	29
--------	----

$$K = 0.1976 (29) - 0.2623$$

$$= 5.4681 \times DF$$

$$= 5.4681$$

$$K_2O = 5.4681 \times CF$$

$$= 5.4681 \times 1.204$$

$$K_2O = 6.58 \%$$

Performed By

M. J. J.

CONTROLLED COPY

NO. _____

10.16.20

VL - NPK

K₂O?

Sample vol 10ml
1ml/10ml = 1

Reading on Flamephotometer.

Conc	Abs
0	0
10	53
15	77
20	100
Sample	28

$$K = 0.1991 (28) = 0.1968$$

$$= 5.378 \times DF$$

$$= 5.378$$

$$K_2O = 5.378 \times CF$$

$$= 5.378 \times 1.204$$

$$K_2O = 6.47 \%$$

Performed By
m. J. J.

CONTROLLED COPY
NO. _____

11-06-24

VL - NPK


N = ?

Sample vol - 0.5 mL

0.1 N H_2SO_4 used 3.03 mL

$$\begin{aligned} \text{Calculation} &= 3.03 \text{ mL} \times CF \\ &= 3.03 \times 2.82 \end{aligned}$$

$$N = 8.54 \%$$

performed by:-


CONTROLLED COPY

NO. _____

11-06-24

VL-NPK

N_a?

Sample vol- 0.5 mL

0.1N H₂SO₄ used = 3.06 mL

$$\begin{aligned}\text{Calculation} &= 3.06 \times CF \\ &= 3.06 \times 2.82\end{aligned}$$

$$N = 8.62 \%$$

performed by:-
[Signature]

CONTROLLED COPY
NO. _____

12-06-24

VL-NPK

$P_2O_5 = ?$

Sample Taken - 0.5 mL

0.1N NaOH used = 20 mL

0.1N H_2SO_4 used = 6 mL

Calculation = $20 - 6$

= 14 mL $\times$ CF


= 14 $\times$ 0.618

P_2O_5 = 8.65 %

CONTROLLED COPY

NO. _____

Performed by:-



12-06-24

VL - NPK

$P_2O_5 = ?$

Sample taken - 0.5ml

0.1N NaOH used = 20 ml

0.1N H_2SO_4 used = 6.2 ml

calculation =

$$= 20 - 6.2$$

$$= 13.8 \times CF$$

$$= 13.8 \times 0.618$$

$$P_2O_5 = 8.52 \%$$

CONTROLLED COPY

NO. _____

Performed by:
 