

	NON-CONFORMITY REPORT FOR ASSESSMENT VISIT	F-02/06
		Issue Date: 11/11/19
		Rev No: 06



ACTIVITY:	☒ Assessment	☐ Surveillance	NCR No: 9
	☐ Re-Assessment	☐ Scope extension	Accr./Appl.No:
	☐ Other Please specify:		Non-conformity category:
			Very serious ☐ Essential ☒
ORGANIZATION:	Quality control Lab vital Agri. Nutrients (Pvt) ltd.		Time limit for correction:
Department:	Labs		6 weeks
Team Leader.	Mr. Atique Ur Rehman	Mngt. Rep.	Mr. Hammyat Hussain
Assessor	Dr. Muhammad Aftab		
Expert			
DESCRIPTION: The lab has developed the procedure for reporting of results (QCL-LSP-19) however lab needs to revise the procedure as per Clause 7.8.1 & 7.8.2 of the standard ISO/IEC 17025:2017 (NC). 10-01-2025 Date Signature Assessor/Expert Signature (Management Representative)			Requirement of: ISO/IEC 17025 ☒ ISO 15189 ☐ ISO/IEC 17021 ☐ ISO/IEC 17020 ☐ Others:
			Ref. Organization's doc.
CORRECTIVE ACTIONS: (Note: It is mandatory to attach Root Cause Analysis) Corrective action for shortcomings on the NC has been taken. Lab has revise the procedure for reporting of results (QCL-LSP-19) as per clause 7.8.1 and 7.8.2 of the standard ISO/IEC 17025:2017. Action are documented in the amendment no: CA-09-02-2025 15-02-25 Date Signature (Management Representative)			☐ It is not necessary to attach documentation
REASON FOR CLOSING: (To be filled in by the Team Leader.) ☐ The non-conformity is closed based on satisfactory documentation from the organization ☐ The non-conformity is closed based on recommendation from the technical assessor ☐ Implementation of the corrective actions will be followed up at the next visit Remarks (if any) ----- ----- ----- The non-conformity is closed: _____ Date Signature (Team Leader.)			

	CORRECTIVE ACTION REQUEST	Doc. No.	QCL-FRM-9.01
		Issue Date	01-10-2024
		Rev#	01

Initiated by: Masab Rasheed	Department: QC Lab	CAR No. CA-09-02-2025
Designation: AQCM	CAR Initiated Date: 10-02-2025	NC No. 09

CAR Initiated due to (Please tick the relevant appropriate box below)


- ☒ PNAC Assessment
 ☐ Complaint
 ☐ Accident/Incident
 ☐ Internal Audit
☐ Suggestion/Improvement
 ☐ Technical Fault
 ☐ Others

A. Description (to be filled by Initiator) The lab has developed the procedure for reporting of results (QCL-LSP-19) however lab needs to revise the procedure as per Clause 7.8.1 & 7.8.2 of the standard ISO/IEC 17025:2017 (NC).		
		 Signatures
Significance/Priority Level	☒ Urgent ☐ Normal ☐ Non-significance	
☒ Accepted ☐ Rejected	Assigned to: Masab Rasheed Date: 10-02-2025	
* If accepted then Manager QC to mark it to concerned person for Root Cause Analysis.		
B. Root Cause Analysis (where required) The root cause of the non-conformance is the lack of a formal process to update staff on changes in relevant standards and regulations.		
C. Proposed Action	Target Date:	15-02-2025
Corrective action for shortcomings on the NC has been taken. Lab has revised the procedure for reporting of results (QCL-LSP-19) as per clause 7.8.1 and 7.8.2 of the standard ISO/IEC 17025:2017.		
		 Name or Signatures
(Person who carried out Root Cause Analysis and Proposed Corrective Action)		
D. Verification by Manager QC for implementation & Effectiveness		
Is Proposed Action Implemented?	☒ YES ☐ NO	Is it Effective? ☒ YES ☐ NO
Remarks (If any):		
CAR closing date: <u>15-02-25</u>		Manager QC Sign: 

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

Non-Conformance Description			
The lab has developed the procedure for reporting of results (QCL-LSP-19) however lab needs to revise the procedure as per Clause 7.8.1 & 7.8.2 of the standard ISO/IEC 17025:2017 (NC).			
RCA Team Members	Himayat Hussain (QCM), Masab Rasheed (AQCM)		
RCA Team Lead / Head	Himayat Hussain (QCM)		
RCA Start Date	10-02-2025	Submission Date	15-02-2025
Methodology / Tools Used			
Cause & Effect Diagram	☐	5WHY analysis	☐
		3W2H / 5W2H	✓

Cause Effect Diagram



3W2H /5W2H	Phenomenon of Problem / Question	Explanation / Answer
WHY	Why was the procedure for reporting results not compliant with Clause 7.8.1 and 7.8.2 of ISO/IEC 17025:2017?	The procedure did not properly outline the requirements for reporting results, including necessary actions related to results validity and traceability as stated in the standard.
WHY	Why did the procedure not address all requirements of Clause 7.8.1 and 7.8.2?	Because the procedure was developed without a comprehensive review of the latest version of ISO/IEC 17025:2017.
WHY	Why was there no comprehensive review of the latest version of the standard?	The team responsible for the procedure development was unaware of the full scope of the updates in the 2017 version of the standard.
WHY	Why was the team unaware of the updates?	There was insufficient training or communication about the changes in the 2017 standard.
WHY	Why was there insufficient training or communication about the changes?	The organization did not have a formal process to ensure that staff are kept up-to-date with changes in relevant standards and regulations.

	ROOT CAUSE ANALYSIS FORM	Doc. No.	QCL-FRM-9.04
		Issue Date	01-10-2024
		Rev#	01

How	How can the lab prevent this non-conformance in the future?	By implementing a formal procedure for regular training and communication to ensure that all personnel are aware of updates to ISO/IEC standards.
How	How can the lab ensure the procedure is compliant with the ISO/IEC 17025:2017?	By conducting a detailed review of all ISO/IEC 17025:2017 clauses and involving experts or consultants during the procedure revision process to ensure full compliance.
Root Cause: The root cause of the non-conformance is the lack of a formal process to update staff on changes in relevant standards and regulations.		


RCA Statement: The root cause of the non-conformance is the lack of a formal process to update staff on changes in relevant standards and regulations.

Reviewed by: 

Evaluated by: 

	DOCUMENT CHANGE REQUEST	Doc.#	QCL-FRM-7.02
		Issue Date	01-02-2025
		Rev#	02

CONTROLLED COPY

NO. Section A					
DCR No.	07-25		Date	15-02-25	
Request Raised By	Masab Rasheed		Section	QC Lab	
Document Title			Document Status	NEW	EXISTING ☒
Reason(s)	☐	New Process Documentation	New Form Documentation		
	☒	Addition in Document	Correction in Document		
	☐	Deletion within Document	Modification in Document		
Section B					
Description of Requirement	Revision of LSP required to Address the shortcomings observed in PNAE audit.				
Effected Sections (Not Applicable on New Documents)	Added division rules, criteria for amended or duplicate report issuance				
Section C					
Review by	Himayat Hussain		Review Status (Choose only one)	☒	Draft Reviewed
Designation				☐	Draft Rejected
				☐	Draft Revision Suggested
Reason for Rejection (Choose only one)	☐	Conflict with Policy	Conflict with Procedure		
	☐	Conflict with Client Requirement	Conflict with Regulation		
	☐	Incorrect Draft	Draft Does not reflect "Description of Requirement"		
Final Remarks	Recommended for final Approval.				
Submitted for Final Approval on (Date)					
Approved By			Designation	CEO	



	PROCEDURE FOR REPORTING OF RESULTS	Doc. No.	QCL-LSP-19
		Issue Date	15-02-2025
		Rev#	02

MASTER

Standard Operating Procedure
For Reporting of Results

ISO/IEC 17025:2017 Clause 7.8

CONTROLLED COPY
NO. _____

Prepared By: Masab Rasheed AQCM	Reviewed By: Himayat Hussain QCM	Approved By: Ahmad Umair CEO
		
Date: 15-02-25	Date: 15-02-25	Date: 15/02/2025

Copyright

The copyright of this procedure, together with all confidential information contained herein is the sole property of **QC LAB**.

It may be copied in full or in parts only by the Management / personnel and only for the purpose of Company related activities. Disclosure and copying of any information contained within this procedure to any person (s) outside the employee of company without written permission/authorization of the Chief Executive Officer is strictly prohibited.



PROCEDURE FOR REPORTING OF RESULTS

Doc. No.	QCL-LSP-19
Issue Date	15-02-2025
Rev#	02

MASTER

CONTROLLED COPY

NO. Amendment Sheet

Date	Revision #	Amendment Detail	Section	DCR	Amendment Approved by
01-10-24	01	Replace issue# with revision#	Header	23-24	Chief Executive Officer
01-10-24	01	Deletion of page number from header	Header	23-24	Chief Executive Officer
01-10-24	01	QC Lab reporting to CEO for Final Approval	Document Approval	23-24	Chief Executive Officer
15-02-25	02	Lab revised the LSP and Added the Decision rule, OK & Not OK Criteria and mentioned amended or Duplicated report issuance	Document	07-25	Chief Executive Officer

	PROCEDURE FOR REPORTING OF RESULTS	Doc. No.	QCL-LSP-19
		Issue Date	15-02-2025
		Rev#	02

MASTER

CONTROLLED COPY
NO. _____

1.0 PURPOSE

The purpose of this document is to define the process for reporting analytical findings and conclusions as prescribed by the organization.

2.0 SCOPE

This procedure is applicable to VAN QC Lab and the release of information.

3.0 Definitions

- **Decision Rule:** A documented rule that defines how measurement uncertainty is considered when determining conformity with specified requirements.
- **Amended Report:** A report reissued to correct errors or update information previously reported.
- **Duplicate Report:** A reissued report identical to the original report, provided upon client request, without any changes.
- **OK:** A result that meets the specified acceptance criteria or tolerance range.
- **Not OK:** A result that does not meet the specified acceptance criteria or tolerance range.

4.0 RESPONSIBILITY

4.1. Quality Control Manager

QCM is responsible to:

1. Ensure timeliness and completeness of all analytical information reported by the laboratory
2. Reviews, or designates reviews of violative worksheets prior to approval by AQCM manager.

4.2. AQCM

Assistant Quality Control Manager is responsible to:

1. Ensure timeliness, accuracy and completeness of all analytical information reported by the analyst
2. Ensure that laboratory analysts comply with the data and information recording procedures.
3. Review analytical findings and supporting records generated by analysts, enter related conclusions, and sample classification codes.
4. In the absence of QCM Ensure timeliness and completeness of all analytical information reported by the laboratory.

 MASTER	PROCEDURE FOR REPORTING OF RESULTS		Doc. No.	QCL-LSP-19
			Issue Date	15-02-2025
			Rev#	02
CONTROL NO.				

4.3. Analysts

Analysts are responsible for capturing and recording accurate and complete analytical information on Daily Lab register/worksheets, continuation sheets, electronic lab notebooks, or other records in accordance with the instructions of this procedure and are to complete worksheets and enter analytical observations and findings.

5.0 PROCEDURE

- Reports of VAN QC LAB tests are reviewed and approved by authorized personnel. All Reports of Special Test are signed by personnel explicitly authorized to do so and it will also include the phrase, "QC.M" under the signature. The list of authorized VAN QC LAB official signatories for the reports is documented in the VAN QC LAB sub-level quality documents.
- Reports of VAN QC LAB is written, issued and retained as technical records. The report includes the visible output of VAN QC LAB that is prepared to ensure that it accurately conveys all information pertaining to an Analysis so that data may be used with maximum benefit by all concerned.
- Reports of VAN QC LAB tests are composed and issued in a specific format as documented below and are not reported in a simplified way.

5.1. Common requirements for reports (test and sampling)

VAN QC LAB Reports of Test are designed by to provide accurate, clear, unambiguous and objective information necessary for the interpretation of the results and all information required by the method used. In addition to the report requirements of ISO/IEC 17025:2017, VAN QC LAB Reports of Test include the following information:

- i. the VAN QC LAB "QC No."
- ii. the sample quantity
- iii. the name of the person(s) performing the analysis
- iv. an explanation of the proper use and interpretation of the reported results, as necessary (this can be in the form of an addendum).

Reports of Test conform to ISO/IEC 17025:2017, for accredited parameters and at a minimum, contain or address the following, unless there are valid reasons for not doing so, as documented in the VAN QC LAB sub-level quality documents:

- a. a title (e.g. "Report of Test);



PROCEDURE FOR REPORTING OF RESULTS

Doc. No.	QCL-LSP-19
Issue Date	15-02-2025
Rev#	02

MASTER

CONTROLLED COPY

- b. at least the first page shall be on ~~VAN QC LAB~~ letterhead or be equivalently identified as originating from VAN QC LAB;
- c. the name and address of the laboratory;
- d. the VAN QC LAB "QC No."
- e. page numbers, with the last page designated as the end of the report;
- f. the name and address of the customer;
- g. identification of the method used;
- h. a description, unambiguous identification, and, when necessary, the condition of the item(s) or samples tested;
- i. the date of receipt of the test item(s) or samples;
- j. the conditions (e.g. environmental) under which the testing was carried out that have an influence on the measurement results;
- k. the date(s) of performance of the laboratory activity;
- l. the date of issue of the report;
- m. a statement to the effect that the results relate only to the items tested, or sampled;
- n. the results with, where appropriate, the units of measurement; an evaluation of the measurement uncertainty associated with each measured value, if requested;
- o. a description of the test method; additions to, deviations, or exclusions from the method;
- p. the name of the person(s) performing the measurements and analysis, and the name, function, and signature of the person(s) authorizing the report;
- q. an explanation of the proper use and interpretation of the reported results; as necessary (this can be in the form of an addendum).

VAN QC LAB is responsible for all information provided in the report, except when information is provided by the customer. Data provided by a customer is identified. In

MASTER**CONTROLLED COPY**

addition, a disclaimer is put on the report when information is supplied by the customer that can affect the validity of results.

5.2. Specific requirements for test reports

Attributes and contents of test reports issued by VAN QC LAB are addressed in 4.1.

VAN QC LAB does not perform sampling as a separate or distinct. Sampling is done as part of testing in laboratory or onsite measurement.

5.3. Specific requirements for calibration certificates

VAN QC LAB does not issue calibration certificates to its clients.

5.4. Reporting sampling

VAN QC LAB does not offer sampling as a distinct service. When done as part of, it details the method of sampling in its test reports.

5.5. Reporting statements of conformity and Decision Rule Employed

- VAN QC LAB issues statements of conformity for the National or Provincial Environmental Quality Standards or other relevant international standards, as required by client.
- Appropriate reference to applied standard is provided in the reports.
- If the result of a test parameter exceeds the specified limits, the statement of conformity states "Not OK" as the case may be.
- If the result of a test parameter does not exceed specified limits, the statement of conformity states "OK".
- the decision rule employed by VAN QC LAB is based on simple acceptance or rejection.
- the decision rule is based on shared risk by both client and VAN QC LAB in equal terms within the bounds of measurement uncertainty at its specified level (most commonly 95% confidence interval).
- VAN QC LAB provides a disclaimer to this effect in measurement/test report that identifies the decision rule employed. The statement is as under

MASTER

NO.

"Uncertainty of measurement (UoM) data will be provided on request, if applicable. The statement of conformity, if provided in the report, is based on the decision rule of simple acceptance or rejection with equal shared risk due to measurement uncertainty".

5.6. Defined Decision Rule:

Decision Rule					
Active Range (%)	Upto 2.5	2.5--10	10--25	25--50	>50
FAD Tolerances	15%	10%	6%	5%	3%

- Results are considered "OK" if they fall within the specified tolerance or acceptance range, accounting for measurement uncertainty.
- Results are considered "Not OK" if they fall outside the specified tolerance or acceptance range.

5.7. Reporting opinions and interpretations

Criteria for "OK" and "Not OK"

1. Criteria for "OK":

- The result meets the acceptance limits defined by the client, regulatory authority, or standard.
- Example: For a tolerance of $100 \pm 5\%$, any result between 95% and 105% is considered "OK."

2. Criteria for "Not OK":

- The result falls outside the defined acceptance limits.
- Example: Any result below 95% or above 105% is considered "Not OK."

5.8. Issuance of Amended Reports

1) When to Issue an Amended Report:

- An amended report is issued if corrections are required due to errors, omissions, or updated results.

2) Process for Issuance:

- The amended report must include:

PROCEDURE FOR REPORTING OF RESULTS

Doc. No.	QCL-LSP-19
Issue Date	15-02-2025
Rev#	02

MASTER

CONTROLLED COPY

NO.

Original report identifier

- Clear marking as "Amended" (e.g., "Issue Status #A,B,C-Amended").
- Reason for the amendment.
- Ensure a complete record of the original report is retained.

3) Authorization:

- Amended reports must be reviewed and approved by the Assistant Quality Control Manager or Quality Control Manager.

4) Communication:

- Notify clients of the amendment and provide the updated report promptly.

5.9. Issuance of Duplicate Reports

1. When to Issue a Duplicate Report:

- A duplicate report may be issued upon client request, provided there are no changes to the original report.

2. Process for Issuance:

- The duplicate report must:
 - Be marked as "Duplicate Copy."
 - Retain the same identifier and issue date as the original report.
- Ensure the reason for the duplicate issuance is recorded.

3. Authorization:

- Duplicate reports must be reviewed and approved by the AQCM/QCM.

4. Record Keeping:

- Maintain a log of all duplicate reports issued, including the reason and date of issuance.

5.10. Amendments to reports

- When VAN QC LAB issued report needs to be changed, amended or re-issued, the change of information is clearly identified and, where appropriate, the reason for the change is included in the report.

PROCEDURE FOR REPORTING OF
RESULTS

Doc. No.	QCL-LSP-19
Issue Date	15-02-2025
Rev#	02

- Amendments to VAN QC LAB issued report are made in the form of a further document, or data transfer, that meets all the requirements of the original report, and includes the statement "Amendment to Report or an equivalent form of wording, with the report number clearly showing as revised (R) as documented in the VAN QC LAB sub-level quality documents.
- When it is necessary to issue a completely new VAN QC LAB report, the new report is uniquely identified and contains a reference to the original that it replaces.

6.0 ASSOCIATED DOCUMENTS AND FORMS

1. Test Results Assignment, Summary, Review (QCL-FRM-19.01)
2. Authorization Form
3. Common Requirements for Reports
4. Requirements for Calibration Certificates
5. Requirements for Reporting Sampling
6. Conditions for Statements of Conformity
7. Conditions for Reporting Opinions and Interpretations

CONTROLLED COPY
NO. _____