

Grade 22-0-12
Sample # 230311

Analyte Report Card
Lab Code 0584

Magruder Fertilizer Proficiency Testing Program
Regular PT Scheme

Proficiency Testing For 2 Analytes

Issue Date : 04/30/2023

Analyte Group	Analyte Group (Units)	Lab Card Data		Analyte Values					Magruder Z Score	Your Method	Flag
		Value	range	Rob Mean	Rob SD	%RSD	R-bar	# Tests			
001	Ammoniacal N (%)	10.38	0.1500	11.59	0.8768	7.57%	0.1804	25	-1.38	001.99	0
050	Soluble K ₂ O (12%)	12.23	0.1500	12.87	0.5556	4.32%	0.2331	78	-1.16	050.61	0

Interpreting Z Scores and Flags:

Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and Green = Z < 2 and >-2 (OK at 95%). Flags indicate data usage: 0 = Used, 1 = rejected for duplicates too far apart, 2 = rejected as extreme outlier, 3 = removed from stats, 4 = rejected due to 0s submitted, 5 = LOD. The 9 Flag indicates a data problem - scores not calculated.

Notes:

Robust statistics not used if < 6 labs used in calculations, in this case Values and Z Scores are grayed out and should be interpreted cautiously. Your offset values are higher (‡) or lower (†) than the Robust Analyte value ± 1A. Analyte codes in light green indicate a guaranteed analyte. Lab values below detection limits reported solely for this PT program.



Grade 22-0-12
Analyte Report Card
Z-Score Box Plot and Guarantees
Sample # 230311


Magruder Fertilizer Proficiency Testing Program

Regular PT Scheme

Proficiency Testing For 2 Analytes

Issue Date : 04/30/2023

Z-Score Box and Whisker Plots for Lab # 0584 <guaranteed analyte>



Guaranteed in This Sample			
Analyte	Code	Value*	IA
N	010	22.299	0.754
K2O	050	12.868	0.825
CI	190	9.386	0.944

* Value (if guaranteed in Sample!) is the Robust Analyte Value estimated using the primary Analyte Codes for this sample and may not reflect the Label value.

Notes: The Methods you used are indicated above and the # Labs involved are below the Box and Whisker. Your Z-Score is indicated by the Black Dot. If you do not see a Black Dot your score is off the chart. Dots between the Green lines are acceptable Z-Scores. Dots outside the Red lines are actionable. The Bar, Box and Whisker represent Median, 25% to 75% percentile and 5% to 95% percentile respectively.