

DAP 18-46-0
Sample # 230714

Analyte ReportCard
Lab Code 0584



Magruder Fertilizer Proficiency Testing Program

Regular PT Scheme

Proficiency Testing For 2 Analytes

Issue Date : 05/31/2023

Analyte Group	Analyte Group (Units)	Lab 0584 Data		*Robust Analyte Values					Magruder Z Score	Your Method	Flag
		Value	range	Rob Mean	Rob SD	%RSD	Rob R-bar (n)	# Tests			
001	Ammoniacal N (%)	18.26	0.1400	17.11	0.6366	3.72%	0.1581 (22)	26	1.81	001.99	0
041	Direct Available P2O5 (46%)	46.51	0.3100	46.14	0.5394	1.17%	0.2865 (38)	40	0.67	041.99	0

Notes and Flags:

1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and a PASS at Green, where Z < 2 and Z >-2 (OK at 95%). The 9 Flag indicates a data problem - scores not calculated.

2: Flags indicate data usage: 0 = Used, 1 = rejected for duplicates too far apart, 2 = rejected as extreme outlier, 3 = removed from stats, 4 = rejected due to 0s submitted, 5 = LOD. The 9 Flag indicates a data problem - scores not calculated.

3: Analyte codes in light green indicate a guaranteed analyte. Your offset values are higher (‡) or lower (†) than the Robust Analyte value ± 1A.

4: *Robust statistics not used if < 6 labs used in calculations, in this case Values and Z Scores are grayed out and should be interpreted cautiously. R-bar is robustified to provide a fair measure of the mean range. Identical duplicates not included in calculation of Rob R-bar. The (n) in Rob R-bar (n) = number of tests used to calculate Rob R-bar.




DAP 18-46-0
Analyte ReportCard
Z-Score Box Plot and Guarantees
Sample # 230714



Magruder Fertilizer Proficiency Testing Program
Regular PT Scheme Sample Doubles with P Scheme
Proficiency Testing For 2 Analytes

Issue Date : 08/31/2023

Z-Score Box and Whisker Plots for Lab # 0584 <guaranteed analyte>



Guaranteed in This Sample			
Analyte	Code	Value*	IA
N	010	17.395	0.691
P2O5	041	46.141	1.002

* Value (if guaranteed in Sample!) is the Robust Analyte Value estimated using the primary Analyte Codes for this sample and may not reflect the Label value.

Notes: The Methods you used are indicated above and the # Labs involved are below the Box and Whisker. Your Z-Score is indicated by the Black Dot. If you do not see a Black Dot your score is off the chart. Dots between the Green lines are acceptable Z-Scores. Dots outside the Red lines are actionable. The Bar, Box and Whisker represent Median, 25% to 75% percentile and 5% to 95% percentile respectively.