

High Micros
Sample # 231111

Analyte ReportCard
Lab Code 0584



Magruder Fertilizer Proficiency Testing Program

Regular PT Scheme

Proficiency Testing For 1 Analyte

Issue Date : 12/31/2023

Analyte Group	Analyte Group (Units)	Lab 0584 Data		*Robust Analyte Values					Magruder Z Score	Your Method	Flag
		Value	range	Rob Mean	Rob SD	%RSD	Rob R-bar (n)	# Tests			
171	Water Soluble B (%)	2.500	0.0200	2.492	0.1907	7.65%	0.0457 (9)	9	0.04	171.10	0

Notes and Flags:

1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and a PASS at Green, where $Z < 2$ and $Z > -2$ (OK at 95%). The 9 Flag indicates a data problem - scores not calculated.

2: Flags indicate data usage: 0 = Used, 1 = rejected for duplicates too far apart, 2 = rejected as extreme outlier, 3 = removed from stats, 4 = rejected due to 0s submitted, 5 = LOD. The 9 Flag indicates a data problem - scores not calculated.

3: Analyte codes in light green indicate a guaranteed analyte. Your offset values are higher ($\dagger$) or lower ($\ddagger$) than the Robust Analyte value $\pm 1A$.

4: *Robust statistics not used if < 6 labs used in calculations, in this case Values and Z Scores are grayed out and should be interpreted cautiously. R-bar is robustified to provide a fair measure of the mean range. Identical duplicates not included in calculation of Rob R-bar. The (n) in Rob R-bar (n) = number of tests used to calculate Rob R-bar.



High Micros
Analyte ReportCard
Z-Score Box Plot and Guarantees
Sample # 231111




Magruder Fertilizer Proficiency Testing Program

Regular PT Scheme

Proficiency Testing For 1 Analyte

Issue Date : 12/31/2023

Z-Score Box and Whisker Plots for Lab # 0584 <guaranteed analyte>



Guaranteed in This Sample

Analyte	Code	Value*	IA
B	165	2.491	0.377
Cu	221	10.147	1.000
Fe	241	8.484	0.853
Mn	261	8.137	0.819

* Value (if guaranteed in Sample!) is the Robust Analyte Value estimated using the primary Analyte Codes for this sample and may not reflect the Label value.

Notes: The Methods you used are indicated above and the # Labs involved are below the Box and Whisker. Your Z-Score is indicated by the Black Dot. If you do not see a Black Dot your score is off the chart. Dots between the Green lines are acceptable Z-Scores. Dots outside the Red lines are actionable. The Bar, Box and Whisker represent Median, 25% to 75% percentile and 5% to 95% percentile respectively.