



K-Mg-SO₄
Sample # 240851

Analyte ReportCard
Lab Code 0584



Magruder Fertilizer Proficiency Testing Program

Potassium Select PT Scheme

Proficiency Testing For 3 Analytes

Issue Date : 09/30/2024

Analyte Group	Analyte Group (Units)	Lab 0584 Data		*Robust Analyte Values					Magruder Z Score	Your Method	Flag
		Value	range	Rob Mean	Rob SD	%RSD	Rob R-bar (n)	# Tests			
050	Soluble K ₂ O (21.5%)	21.56	0.2200	21.91	0.5892	2.69%	0.1512 (15)	17	-0.60	050.99	0
121	Acid Soluble Mg (10.8%)	10.86	0.0900	11.14	0.4537	4.07%	0.2022 (12)	12	-0.62	121.70	0
143	Elemental S (%)	22.04	0.1200					2		143.00	0

Notes and Flags:

1: Interpreting Z Scores: Red indicates a normally distributed Z value >3 or <-3 (requires action), Orange = Z between 2 and 3 or -2 and -3 (warning) and a PASS at Green, where Z < 2 and Z >-2 (OK at 95%). The 9 Flag indicates a data problem - scores not calculated.

2: Flags indicate data usage: 0 = Used, 1 = rejected for duplicates too far apart, 2 = rejected as extreme outlier, 3 = removed from stats, 4 = rejected due to 0s submitted, 5 = LOD. The 9 Flag indicates a data problem - scores not calculated.

3: Analyte codes in light green indicate a guaranteed analyte. Your offset values are higher (±) or lower (†) than the Robust Analyte value ± 1A.

4: *Robust statistics not used if < 6 labs used in calculations, in this case Values and Z Scores are grayed out and should be interpreted cautiously. R-bar is robustified to provide a fair measure of the mean range. Identical duplicates not included in calculation of Rob R-bar. The (n) in Rob R-bar (n) = number of tests used to calculate Rob R-bar.



K-Mg-SO4
 Analyte ReportCard
 Z-Score Box Plot and Guarantees
 Sample # 240851




Magruder Fertilizer Proficiency Testing Program

Potassium Select PT Scheme

Proficiency Testing For 3 Analytes

Issue Date : 09/30/2024

Z-Score Box and Whisker Plots for Lab # 0584 <guaranteed analyte>



Guaranteed in This Sample

Analyte	Code	Value*	IA
K2O	050	21.913	1.147
Mg	121	11.135	0.757
S	148	22.846	1.000
Cl	190	0.763	0.081

* Value (if guaranteed in Sample!) is the Robust Analyte Value estimated using the primary Analyte Codes for this sample and may not reflect the Label value.

Notes: The Methods you used are indicated above and the # Labs involved are below the Box and Whisker. Your Z-Score is indicated by the Black Dot. If you do not see a Black Dot your score is off the chart. Dots between the Green lines are acceptable Z-Scores. Dots outside the Red lines are actionable. The Bar, Box and Whisker represent Median, 25% to 75% percentile and 5% to 95% percentile respectively.