

**CAHAYA HORNBILL SDN BHD**

Lot 10591, Block 10 Kuala Baram Land District Pujut 7 Commercial Centre,  
Jalan Dato Sri Joseph Balan Seling, 98000, Miri Sarawak  
Phone: +6 085 656 655 | Fax : +6 085 640 055 | Email: business@cahayahornbill.com.my

## **CERTIFICATE OF CALIBRATION**

1/002/2025.SRV CHSB-WO-25-1

Model ISC, Radius BZ1  
Serial Number 2210017-002  
Calibration Date January 14, 2025  
Due Calibration June 12, 2025  
Cylinder Lot# 602662

| Set Alarm Points | METHANE<br>0-100% LEL | O2<br>0-30% | CO<br>0-1500PPM | H2S<br>0-500PPM |
|------------------|-----------------------|-------------|-----------------|-----------------|
| Low              | 10% LEL               | 19.50%      | 35 PPM          | 5 PPM           |
| High             | 20%LEL                | 23.50%      | 70 PPM          | 10 PPM          |
| STEL             |                       |             | 200 PPM         | 15 PPM          |
| TWA              |                       |             | 35 PPM          | 10 PPM          |
| Calibrated Value | CH4 50% LEL           | O2 18% VOL  | CO 100 PPM      | H2S 25 PPM      |
| Specification    | 48-52% LEL            | 16-20% PPM  | 98-102% PPM     | 23-27% PPM      |

All applicable inspections, testing, and calibrations were performed using traceable equipment, where available, in accordance with CHSB's ISO 9001 Certified Quality System. Each material, component, and/or instrument must be installed, operated, and maintained in strict accordance with its labels, cautions, warning instructions, and within the limitations stated in the supplied instruction manual. Routine calibration checks, equipment inspections, and applicable preventative maintenance measures must be performed to verify that the material, components, and/or instrument are operating properly. Failure to perform these tasks on a routine basis, or at suggested intervals, with specified equipment or methods, may result in inaccurate readings.

*Remark: Please send back the portable unit to CHSB service center/Authorized Distributor for calibration certification before the due date.*

Calibrated By

.....  
Service Technician