

CAHAYA HORNBILL SDN BHD

Lot 10591, Block 10 Kuala Baram Land District Pujut 7 Commercial Centre,
Jalan Dato Sri Joseph Balan Seling, 98000, Miri Sarawak
Phone: +6 085 656 655 | Fax : +6 085 640 055 | Email: business@cahayahornbill.com.my

CERTIFICATE OF CALIBRATION

1/002/2025.SRV CHSB-WO-25-1

Model ISC, Radius BZ1
Serial Number 2210017-002
Calibration Date January 14, 2025
Due Calibration June 12, 2025
Cylinder Lot# 602662

Set Alarm Points	METHANE 0-100% LEL	O2 0-30%	CO 0-1500PPM	H2S 0-500PPM
Low	10% LEL	19.50%	35 PPM	5 PPM
High	20%LEL	23.50%	70 PPM	10 PPM
STEL			200 PPM	15 PPM
TWA			35 PPM	10 PPM
Calibrated Value	CH4 50% LEL	O2 18% VOL	CO 100 PPM	H2S 25 PPM
Specification	48-52% LEL	16-20% PPM	98-102% PPM	23-27% PPM

All applicable inspections, testing, and calibrations were performed using traceable equipment, where available, in accordance with CHSB's ISO 9001 Certified Quality System. Each material, component, and/or instrument must be installed, operated, and maintained in strict accordance with its labels, cautions, warning instructions, and within the limitations stated in the supplied instruction manual. Routine calibration checks, equipment inspections, and applicable preventative maintenance measures must be performed to verify that the material, components, and/or instrument are operating properly. Failure to perform these tasks on a routine basis, or at suggested intervals, with specified equipment or methods, may result in inaccurate readings.

Remark: Please send back the portable unit to CHSB service center/Authorized Distributor for calibration certification before the due date.

Calibrated By

.....
Service Technician