REV	DESCRIPTION	DFTR	CHKR	ENGR APPV	DATE
C1	REVISED FOR CORRECTNESS	WSG	JO	RC	08/03
C2	REVISED FOR CORRECTNESS	WSG	JO	RC	09/03
C3	REVISED FOR CORRECTNESS	WSG	JO	RC	10/03

HOFFMAN 12X10

DC Voltage 1P7

DC Voltage 2P8

XFMR ARC P6

LEV-2

OIL-LEVEL

P5

MS3108E18-1S

CROWBAR

BELDEN 88761 SHIELDED TWISTED PAIR

BELDEN 88761 SHIELDED TWISTED PAIR

COAX RG-58

T20-I-509 THERMOCOUPLE

RED

ORG/BLK

GRN/BLK

GRN/WHT

RED/BLK

RED/WHT

BLK/WHT

BLK

BLU/BLK

BLU/WHT

BLU

WHT/BLK

WHT

SHIELD

TS-3

PWR SUP MONITORS

XFMR ARC

BNC-2

SLOT3

TC-4

AIR TEMP

COM

IN 3-

IN 3+

COM

IN 2-

IN 2+

COM

IN 1-

IN 1+

COM

IN Ø-

IN Ø+

AB-1794-THERMC

TS-5

CONTACTOR CONTROLS

PPS-S

J1

P1

GØB128BPNE

BELDEN 83715 15C *16 TEFLON

PPS

CONTACTOR READY

ENABLE

CONTACTOR CLOSED

AUX TRIP

RESET

PPS

PPS COM

ON

COMMON

PERSONEL PROTECTION SYSTEM

CONTACTOR DISCONNECT

TB2

S6

TEMP

S5

CONTACTS PPS

S4

CLOSE READY

S3

CONTACTOR

S2

OVERCURRENT

S1

BLOCKING RELAY

GRN

RED

BLU

ORG/BLK

WHT/BLK

ORG

BLU/WHT

GRN/BLK

BLK/WHT

BLU/BLK

RED/WHT

BLK

RED/BLK

GRN/WHT

WHT

MX

ON/OFF

RR

K-4

RESET

PPS

Label swapped

TS

120V

COM

GND

XXX

N

XX

Y

BB

CC

EE

5

6

7

8

9

10

EE

W

DD

14

15

16

EMPTY

EMPTY

EMPTY

HH

W1

CC1

II

20

21

22

23

24

TERMINAL STRIP LAYOUT

120V

COM

PWR SPLY CONTROLS

TS8

DISPLAY

WHT

BLK

BLU

RED

ORG

GRN

RED/BLK

WHT/BLK

TS-NWL

NWL TRANSFORMER

R44 500Ω 25W

R47 50Ω 25W

R42 500Ω 25W

R46 50Ω 25W

R33 500Ω 25W

R36 50Ω 25W

R32 500Ω 25W

R35 50Ω 25W

R31 500Ω 25W

R34 50Ω 25W

R23 500Ω 25W

R26 50Ω 25W

R22 500Ω 25W

R25 50Ω 25W

R21 500Ω 25W

R24 50Ω 25W

R13 500Ω 25W

R16 50Ω 25W

R12 500Ω 25W

R15 50Ω 25W

R11 500Ω 25W

R14 50Ω 25W

R19 500Ω 25W

R18 50Ω 25W

R17 500Ω 25W

D3

D2

D1

120V 20A

R2 50Ω 50W

SLAC-1

GRN

GRN/WHT

BLK/WHT

BLU/WHT

BLU/BLK

BLU

RED/WHT

RED/BLK

RED

ORG/BLK

GRN/BLK

WHT/BLK

ORG

WHT

BLK

WHT

SHIELD

BELDEN 83715 15C *16 TEFLON

GRN/WHT

RED/BLK

RED/WHT

BLK/WHT

BLK

BLU/BLK

BLU/WHT

WHT

WHT/BLK

GRN/BLK

ORG/BLK

BLU

SHIELD

CROWBAR SCR DRIVERS

88761 TWISTED SHIELDED

RED

BLK

SHIELD

BLK

WHT

ORG

WHT/BLK

GRN/BLK

ORG/BLK

RED

RED/BLK

RED/WHT

BLU

BLU/BLK

BLU/WHT

BLK/WHT

GRN/WHT

GRN

TS-7

TRANSFORMER MONITORS

BELDEN 83715 15C *16 TEFLON

TS-NWL

20 RED

19 GRN

18 BLU

17

16

15 ORG/BLK

14

13 ORG

12 WHT/BLK

11

10 BLK/WHT

9 GRN/BLK

8

7 GRN/WHT

6 BLU/BLK

5

4 BLU/WHT

3 RED/BLK

2

1 RED/WHT

SHIELD

NWL- TRANSFORMER

NWL *39308

TRANSFORMER INTERLOCKS

SHIELD

BLK

WHT

RED/WHT

RED/BLK

BLU/WHT

BLU/BLK

GRN/WHT

GRN/BLK

BLK/WHT

WHT/BLK

ORG

ORG/BLK

BLU

GRN

RED

TS-4

AC CURRENT

TEMPERTURE

SUDDEN PRESSURE

OIL LEVEL LOW

TRANSFORMER INTERLOCKS

BELDEN 83715 15C *16 TEFLON

PWR SUPP REGULATOR

HOFFMAN BOX 34X42

T20-I-509 THERMOCOUPLE

T20-I-509 THERMOCOUPLE

TC-1

TOP OIL A

TOP OIL B

TC-2

PHASE SCR'S

PHASE SCR'S

BELDEN 83715 15C *16 TEF

BELDEN 83715 15C *16 TEF

SHIELD

GRN

BLU

BLU/WHT

BLU/BLK

ORG WHT BLK

ORG-B WHT-B BLK/WHT

RED

RED/WHT

RED/BLK

GRN/WHT

GRN/BLK

P1

P2

MS3108E24-20S

TS8

PERMITS

8 WHT/BLK

7 RED/BLK

6 GRN

5 ORG

4 RED

3 BLU

2 BLK

1 WHT

88761 TWISTED

RED

BLK

LEV-1

OIL-LEVEL

SCR TIGGERS

HOFFMAN 12X10

TS-6

18

17

16

15 BLK

14 RED

13

12 RED-B

11 ORG

10 WHT/BLK

9 BLK

8 BLK

7 RED

6 GRN-BLK

5 WHT

4 RED

3 BLU

2 GRN

1 BLK

GND TANK

RED-BLK ORG

BELDEN 83709 9C *16 TEF

BELDEN 83709 9C *16 TEF

TERMINATION TANK

SHIELD

RED

WHT

BLK

WHT/BLK

GRN

BLU

RED-BLK

GRN-BLK

ORG

P5

88761 TWISTED SHIELDED

LEV-3

OIL-LEVEL

RED-BLK

ORG

GRN-BLK

GRN

BLU

BLK

WHT/BLK

WHT

RED

J2

CON9

BNC-1

KLYSTRON ARC

COAX RG-58

BNC1

KLYSTRON ARC

WHT/BLK

STANFORD LINEAR ACCELERATOR CENTER U.S. DEPARTMENT OF ENERGY STANFORD UNIVERSITY STANFORD, CALIFORNIA	CAD FILE NAME: 73079001.WD3	SHEET 1 OF 1
PROPRIETARY DATA OF STANFORD UNIVERSITY AND/OR U.S. DEPARTMENT OF ENERGY, RECIPIENT SHALL NOT PUBLISH THE INFORMATION WITHIN UNLESS GRANTED SPECIFIC PERMISSION OF STANFORD UNIVERSITY.	PEPII RF SYSTEMS 2MW KLYSTRON PWR SPLY INTERCONNECTION WIRING	
	WD-730-790-01-C3	D

		DATE	APPROVAL
ENGR	R. CASSEL		R. CASSEL
DFTR	W. GORECKI	03/02/2000	
CHKR	S. LOWE		

PEP2 2MW KLY PS INTCON

Rev.C3