PCTEST ENGINEERING LABORATORY, INC.



7185 Oakland Mills Road, Columbia, MD 21046 USA Tel. 410.290.6652 / Fax 410.290.6654 http://www.pctestlab.com



MEASUREMENT REPORT FCC Part 22 & 24

Applicant Name: MedicPen AB Box 803 SE-30118 Halmstad Sweden

Date of Testing: 12/22/2014, 4/15/2016 Test Site/Location:

PCTEST Lab., Columbia, MD, USA **Test Report Serial No.:**

0Y1603150550.2ABTH

FCC ID: 2ABTHC200W

APPLICANT: **MEDICPEN AB**

Application Type: Certification Model(s): C200W

EUT Type: Portable Pill Dispenser

PCS Licensed Transmitter (PCB) **FCC Classification:**

FCC Rule Part(s): §2 §22(H) §24(E)

ANSI/TIA-603-D:2010, KDB 971168 v02r02 Test Procedure(s):

Test Device Serial No.: identical prototype [S/N: 163F2]

			ERP/EIRP		
Mode	Tx Frequency (MHz)	Emission Designator	Max. Power (W)	Max. Power (dBm)	
CDMA850	824.70 - 848.31	1M27F9W	0.110	20.42	
CDMA1900	1851.25 - 1908.75	1M28F9W	1.307	31.16	

This equipment has been shown to be capable of compliance with the applicable technical standards as indicated in the measurement report and was tested in accordance with the measurement procedures specified in §2.947. Test results reported herein relate only to the item(s) tested.

I attest to the accuracy of data. All measurements reported herein were performed by me or were made under my supervision and are correct to the best of my knowledge and belief. I assume full responsibility for the completeness of these measurements and vouch for the qualifications of all persons taking them.







FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 1 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 1 of 41




TABLE OF CONTENTS

FCC F	PART 2	22 & 24 MEASUREMENT REPORT	3
1.0		RODUCTION	
	1.1	Scope	4
	1.2	Testing Facility	4
2.0	PRC	DUCT INFORMATION	5
	2.1	Equipment Description	5
	2.2	Device Capabilities	
	2.3	Test Configuration	5
	2.4	EMI Suppression Device(s)/Modifications	5
3.0	DES	CRIPTION OF TESTS	6
	3.1	Evaluation Procedure	6
	3.2	Cellular - Base Frequency Blocks	6
	3.3	Cellular - Mobile Frequency Blocks	6
	3.4	PCS - Base Frequency Blocks	6
	3.5	PCS - Mobile Frequency Blocks	7
	3.6	Radiated Measurements	7
4.0	MEA	SUREMENT UNCERTAINTY	8
5.0	TES	T EQUIPMENT CALIBRATION DATA	9
6.0	SAM	IPLE CALCULATIONS	10
7.0	TES	T RESULTS	11
	7.1	Summary	
	7.2	Occupied Bandwidth	
	7.3	Spurious and Harmonic Emissions at Antenna Terminal	
	7.4	Band Edge Emissions at Antenna Terminal	
	7.5	Peak-Average Ratio	26
	7.6	Radiated Power (ERP/EIRP)	28
	7.7	Radiated Spurious Emissions Measurements	31
	7.8	Frequency Stability / Temperature Variation	36
8.0	CON	ICLUSION	41

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 2 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 2 of 41





MEASUREMENT REPORT FCC Part 22 & 24



§2.1033 General Information

APPLICANT: MedicPen AB **APPLICANT ADDRESS:** Box 803

SE-30118 Halmstad, Sweden

PCTEST ENGINEERING LABORATORY, INC. **TEST SITE: TEST SITE ADDRESS:** 7185 Oakland Mills Road, Columbia, MD 21046 USA

FCC RULE PART(S): §2 §22(H) §24(E)

BASE MODEL: C200W

FCC ID: 2ABTHC200W

FCC CLASSIFICATION: PCS Licensed Transmitter (PCB)

MODE: **CDMA**

FREQUENCY TOLERANCE: ±0.00025 % (2.5 ppm)

☐ Production ☐ Pre-Production ☐ Engineering **Test Device Serial No.:** 163F2

DATE(S) OF TEST: 12/22/2014. 4/15/2016 **TEST REPORT S/N:** 0Y1603150550.2ABTH

Test Facility / Accreditations

Measurements were performed at PCTEST Engineering Lab located in Columbia, MD 21046, U.S.A.

- PCTEST facility is an FCC registered (PCTEST Reg. No. 159966) test facility with the site description report on file and has met all the requirements specified in Section 2.948 of the FCC Rules and Industry Canada (2451B-1).
- PCTEST Lab is accredited to ISO 17025 by U.S. National Institute of Standards and Technology (NIST) under the National Voluntary Laboratory Accreditation Program (NVLAP Lab code: 100431-0) in EMC, FCC and Telecommunications.
- PCTEST Lab is accredited to ISO 17025-2005 by the American Association for Laboratory Accreditation (A2LA) in Specific Absorption Rate (SAR) testing, Hearing Aid Compatibility (HAC) testing, CTIA Test Plans, and wireless testing for FCC and Industry Canada Rules.
- PCTEST Lab is a recognized U.S. Conformity Assessment Body (CAB) in EMC and R&TTE (n.b. 0982) under the U.S.-EU Mutual Recognition Agreement (MRA).
- PCTEST TCB is a Telecommunication Certification Body (TCB) accredited to ISO/IEC Guide 65 by the American National Standards Institute (ANSI) in all scopes of FCC Rules and Industry Canada Standards (RSS).
- PCTEST facility is an IC registered (2451B-1) test laboratory with the site description on file at Industry Canada.
- PCTEST is a CTIA Authorized Test Laboratory (CATL) for AMPS, CDMA, and EvDO wireless devices and for Over-the-Air (OTA) Antenna Performance testing for AMPS, CDMA, GPRS, GPRS, EGPRS, UMTS (W-CDMA), CDMA 1xEVDO, and CDMA 1xRTT.





FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 2 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 3 of 41
© 2016 DCTEST Engineering Loberston, Inc.				



INTRODUCTION 1.0

1.1 Scope

Measurement and determination of electromagnetic emissions (EME) of radio frequency devices including intentional and/or unintentional radiators for compliance with the technical rules and regulations of the Federal Communications Commission and the Industry Canada Certification and Engineering Bureau.

1.2 **Testing Facility**

The map below shows the location of the PCTEST LABORATORY, its proximity to the FCC Laboratory, the Columbia vicinity, the Baltimore-Washington Internt'I (BWI) airport, the city of Baltimore and the Washington, DC area. (See Figure 1-1).

These measurement tests were conducted at the PCTEST Engineering Laboratory, Inc. facility located at 7185 Oakland Mills Road, Columbia, MD 21046. The site coordinates are 39° 10'23" N latitude and 76° 49'50" W longitude. The facility is 0.4 miles North of the FCC laboratory, and the ambient signal and ambient signal strength are approximately equal to those of the FCC laboratory. The detailed description of the measurement facility was found to be in compliance with the requirements of § 2.948 according to ANSI C63.4-2014 on January 22, 2015.




Figure 1-1. Map of the Greater Baltimore and Metropolitan Washington, D.C. area

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogg 4 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 4 of 41

© 2016 PCTEST Engineering Laboratory, Inc.



PRODUCT INFORMATION

2.1 **Equipment Description**

The Equipment Under Test (EUT) is the MedicPen Portable Pill Dispenser FCC ID: 2ABTHC200W. This device contains a previously certified CDMA module (Telit Communications, FCC ID: RI7CE910-DUAL). The test data contained in this report pertains only to the emissions due to the EUT's 2G/3G licensed transmitters.

2.2 **Device Capabilities**

This device contains the following capabilities:

850/1900 CDMA (BC0, BC1)

2.3 **Test Configuration**

The MedicPen Portable Pill Dispenser FCC ID: 2ABTHC200W was tested per the guidance of ANSI/TIA-603-D:2010 and KDB 971168 v02r02. See Section 7.0 of this test report for a description of the radiated and antenna port conducted emissions tests.

2.4 **EMI Suppression Device(s)/Modifications**

No EMI suppression device(s) were added and no modifications were made during testing.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo F of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 5 of 41




3.0 DESCRIPTION OF TESTS

3.1 Evaluation Procedure

The measurement procedures described in the "Land Mobile FM or PM – Communications Equipment – Measurements and Performance Standards" (ANSI/TIA-603-D:2010) and "Measurement Guidance for Certification of Licensed Digital Transmitters" (KDB 971168 v02r02) were used in the measurement of the **MedicPen Portable Pill Dispenser FCC ID: 2ABTHC200W.**

Deviation from Measurement Procedure......None

3.2 Cellular - Base Frequency Blocks §22.905




BLOCK 1: 869 - 880 MHz (A* Low + A)

BLOCK 3: 890 - 891.5 MHz (A* High)

BLOCK 2: 880 - 890 MHz (B)

BLOCK 4: 891.5 - 894 MHz (B*)

3.3 Cellular - Mobile Frequency Blocks




BLOCK 1: 824 - 835 MHz (A* Low + A)

BLOCK 3: 845 – 846.5 MHz (A* High)

BLOCK 2: 835 – 845 MHz (B)

BLOCK 4: 846.5 - 849 MHz (B*)

3.4 PCS - Base Frequency Blocks



BLOCK 1: 1930 - 1945 MHz (A)

BLOCK 4: 1965 - 1970 MHz (E)

BLOCK 2: 1945 - 1950 MHz (D)

BLOCK 5: 1970 - 1975 MHz (F)


BLOCK 3: 1950 - 1965 MHz (B)

BLOCK 6: 1975 - 1990 MHz (C)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dage 6 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 6 of 41



3.5 **PCS - Mobile Frequency Blocks**



BLOCK 2: 1865 - 1870 MHz (D) BLOCK 5: 1890 - 1895 MHz (F)

BLOCK 3: 1870 - 1885 MHz (B) BLOCK 6: 1895 - 1910 MHz (C)

3.6 Radiated Measurements §2.1053 §22.913(a.2) §22.917(a) §24.232(c) §24.238(a)

The radiated test facilities consisted of an indoor 3 meter semi-anechoic chamber used for final measurements and exploratory measurements, when necessary. The measurement area is contained within the semi-anechoic chamber which is shielded from any ambient interference. The test site inside the chamber is a 6m x 5.2m elliptical, obstruction-free area in accordance with Clause 5, Figure 5.7 of ANSI C63.4-2014. For measurements above 1GHz absorbers are arranged on the floor between the turn table and the antenna mast in such a way so as to maximize the reduction of reflections. For measurements below 1GHz, the absorbers are removed. An ETS Lindgren Model 2188 raised turntable is used for radiated measurement. It is a continuously rotatable, remote-controlled, metallic turntable and 2 meters (6.56 ft.) in diameter. The turn table is flush with the raised floor of the chamber in order to maintain its function as a ground plane. A 78cm high PVC support structure is placed on top of the turntable. A 3/4" (~1.9cm) sheet of high density polyethylene is used as the table top and is placed on top of the PVC supports to bring the total height of the table to 80cm.

The equipment under test was transmitting while connected to its integral antenna and is placed on a turntable 3 meters from the receive antenna. The receive antenna height is adjusted between 1 and 4 meter height, the turntable is rotated through 360 degrees, and the EUT is manipulated through all orthogonal planes representative of its typical use to achieve the highest reading on the receive spectrum analyzer.

Per the guidance of ANSI/TIA-603-D:2010, a half-wave dipole is then substituted in place of the EUT. For emissions above 1GHz, a horn antenna is substituted in place of the EUT. The substitute antenna is driven by a signal generator with the level of the signal generator being adjusted to obtain the same receive spectrum analyzer level previously recorded from the spurious emission from the EUT. The power of the emission is calculated using the following formula:

$$P_{d [dBm]} = P_{g [dBm]} - cable loss [dB] + antenna gain [dBd/dBi]$$

Where, P_d is the dipole equivalent power, P_q is the generator output into the substitution antenna, and the antenna gain is the gain of the substitute antenna used relative to either a half-wave dipole (dBd) or an isotropic source (dBi). The substitute level is equal to Pg [dBm] - cable loss [dB].

Radiated power and radiated spurious emissions levels are investigated with the receive antenna horizontally and vertically polarized per ANSI/TIA-603-D:2010.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 7 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 7 of 41



MEASUREMENT UNCERTAINTY

The measurement uncertainties shown below were calculated in accordance with the requirements of ANSI C63.4-2014. All measurement uncertainty values are shown with a coverage factor of k = 2 to indicate a 95% level of confidence. The measurement data shown herein meets or exceeds the U_{CISPR} measurement uncertainty values specified in CISPR 16-4-2 and, thus, can be compared directly to specified limits to determine compliance.

Contribution	Expanded Uncertainty (±dB)
Conducted Bench Top Measurements	1.13
Radiated Disturbance (<1GHz)	4.98
Radiated Disturbance (>1GHz)	5.07
Radiated Disturbance (>18GHz)	5.09

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 9 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 8 of 41



TEST EQUIPMENT CALIBRATION DATA

Test Equipment Calibration is traceable to the National Institute of Standards and Technology (NIST).

Manufacturer	Model	Description	Cal Date	Cal Interval	Cal Due	Serial Number
-	LTx1	Licensed Transmitter Cable Set	1/29/2014	Annual	1/29/2015	N/A
-	RE1	Radiated Emissions Cable Set (UHF/EHF)	5/29/2014	Annual	5/29/2015	N/A
Agilent	8447D	Broadband Amplifier	5/30/2014	Annual	5/30/2015	2443A01900
Agilent	N9020A	MXA Signal Analyzer	10/27/2014	Annual	10/27/2015	US46470561
Anritsu	MA2411B	Pulse Sensor	4/8/2014	Biennial	4/8/2016	846215
Com-Power	AL-130	9kHz - 30MHz Loop Antenna	6/26/2013	Biennial	6/26/2015	121034
Espec	ESX-2CA	Environmental Chamber	4/16/2014	Annual	4/16/2015	17620
ETS Lindgren	3117	1-18 GHz DRG Horn (Medium)	4/8/2014	Biennial	4/8/2016	125518
ETS Lindgren	3160-09	18-26.5 GHz Standard Gain Horn	6/17/2014	Biennial	6/17/2016	135427
ETS Lindgren	3164-08	Quad Ridge Horn Antenna	3/12/2014	Biennial	3/12/2016	128337
K & L	11SH10-3075/U18000	High Pass Filter	12/1/2014	Annual	12/1/2015	2
K & L	13SH10-1000/U1000	N Type High Pass Filter	12/1/2014	Annual	12/1/2015	1
Rohde & Schwarz	CMU200	Base Station Simulator	6/6/2014	Annual	6/6/2015	109892
Rohde & Schwarz	ESU26	EMI Test Receiver (26.5GHz)	1/27/2014	Annual	1/27/2015	100342
Rohde & Schwarz	TS-PR18	1-18 GHz Pre-Amplifier	3/5/2014	Annual	3/5/2015	100071
Rohde & Schwarz	TS-PR26	18-26.5 GHz Pre-Amplifier	3/12/2014	Annual	3/12/2015	100040
Schwarzbeck	UHA 9105	Dipole Antenna (400 - 1GHz) Rx	11/21/2013	Biennial	11/21/2015	9105-2404
Schwarzbeck	UHA 9105	Dipole Antenna (400 - 1GHz) Tx	11/21/2013	Biennial	11/21/2015	9105-2403
Sunol	JB5	Bi-Log Antenna (30M - 5GHz)	1/28/2014	Biennial	1/28/2016	A051107

Table 5-1. 12/22/2016 Test Equipment

Manufacturer	Model	Description	Cal Date	Cal Interval	Cal Due	Serial Number
-	RE1	Radiated Emissions Cable Set (UHF/EHF)	4/28/2015	Annual	4/28/2016	RE1
Agilent	8447D	Broadband Amplifier	6/12/2015	Annual	6/12/2016	1937A03348
Agilent	E5515C	Wireless Communications Test Set		N/A		GB46310798
Com-Power	AL-130	9kHz - 30MHz Loop Antenna	7/30/2015	Biennial	7/30/2017	121034
Emco	3115	Horn Antenna (1-18GHz)	3/10/2016	Biennial	3/10/2018	9704-5182
Espec	ESX-2CA	Environmental Chamber	3/4/2016	Annual	3/4/2017	17620
ETS Lindgren	3160-09	18-26.5 GHz Standard Gain Horn	6/17/2014	Biennial	6/17/2016	135427
ETS Lindgren	3164-08	Quad Ridge Horn Antenna	10/22/2014	Biennial	10/22/2016	128338
K & L	11SH10-3075/U18000	High Pass Filter	7/18/2015	Annual	7/18/2016	11SH10-3075/U18000-2
K & L	13SH10-1000/U1000	N Type High Pass Filter	7/18/2015	Annual	7/18/2016	13SH10-1000/U1000-2
Mini Circuits	PWR-SEN-4GHS	USB Power Sensor	3/4/2016	Annual	3/4/2017	11401010036
Mini-Circuits	SSG-4000HP	Synthesized Signal Generator		N/A		11208010032
Mini-Circuits	TVA-11-422	RF Power Amp		N/A		QA1303002
Rhode & Schwarz	TS-PR18	Pre-Amplifier	3/7/2016	Annual	3/7/2017	101622
Rohde & Schwarz	FSW67	Signal / Spectrum Analyzer	6/2/2015	Annual	6/2/2016	103200
Rohde & Schwarz	TS-PR18	1-18 GHz Pre-Amplifier	3/7/2016	Annual	3/7/2017	100071
Rohde & Schwarz	TS-PR26	18-26.5 GHz Pre-Amplifier	3/7/2016	Annual	3/7/2017	100040
Schwarzbeck	UHA 9105	Dipole Antenna (400 - 1GHz) Rx	11/18/2015	Biennial	11/18/2017	91052523RX
Seekonk	NC-100	Torque Wrench 5/16", 8" lbs	3/2/2016	Biennial	3/2/2018	N/A
Sunol	JB5	Bi-Log Antenna (30M - 5GHz)	3/14/2016	Biennial	3/14/2018	A051107

Table 5-2. 4/15/2016 Test Equipment

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Page 9 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Fage 9 01 41



SAMPLE CALCULATIONS

CDMA Emission Designator

Emission Designator = 1M25F9W

CDMA BW = 1.25 MHz F = Frequency Modulation 9 = Composite Digital Info W = Combination (Audio/Data)

Spurious Radiated Emission

Example: Spurious emission at 3700.40 MHz

The receive spectrum analyzer reading at 3 meters with the EUT on the turntable was -81.0 dBm. The gain of the substituted antenna is 8.1 dBi. The signal generator connected to the substituted antenna terminals is adjusted to produce a reading of -81.0 dBm on the spectrum analyzer. The loss of the cable between the signal generator and the terminals of the substituted antenna is 2.0 dB at 3700.40 MHz. So 6.1 dB is added to the signal generator reading of -30.9 dBm yielding -24.80 dBm. The fundamental EIRP was 25.50 dBm so this harmonic was 25.50 dBm - (-24.80) = 50.3 dBc.

FCC ID: 2ABTHC200W	PCTEST*	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		, ,
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 10 of 41



TEST RESULTS

7.1 Summary

Company Name: MedicPen AB

FCC ID: 2ABTHC200W

FCC Classification: PCS Licensed Transmitter (PCB)

Mode(s): **CDMA**

FCC Part Section(s)	Test Description	Test Limit	Test Condition	Test Result	Reference
TRANSMITTER	MODE (TX)				
2.1049	Occupied Bandwidth	N/A		PASS	Section 7.2
2.1051 22.917(a) 24.238(a)	Conducted Band Edge / Spurious Emissions	> 43 + log ₁₀ (P[Watts]) at Band Edge and for all out-of-band emissions		PASS	Sections 7.3, 7.4
24.232(d)	Peak-Average Ratio	< 13 dB	CONDUCTED	PASS	Section 7.5
2.1046	Transmitter Conducted Output Power	N/A		PASS	RF Exposure Report
2.1055 22.355 24.235	Frequency Stability	< 2.5 ppm (Part 22) Emission must remain in band (Part 24)		PASS	Section 7.8
22.913(a.2)	Effective Radiated Power	< 7 Watts max. ERP		PASS	Section 7.6
24.232(c)	Equivalent Isotropic Radiated Power	< 2 Watts max. EIRP	RADIATED	PASS	Section 7.6
2.1053 22.917(a) 24.238(a)	Radiated Spurious Emissions	> 43 + log ₁₀ (P[Watts]) for all out-of-band emissions		PASS	Section 7.7

Table 7-1. Summary of Test Results

Notes:

- 1) All modes of operation and data rates were investigated. The test results shown in the following sections represent the worst case emissions.
- 2) The analyzer plots were all taken with a correction table loaded into the analyzer. The correction table was used to account for the losses of the cables, directional couplers, and attenuators used as part of the system to maintain a link between the call box and the EUT at all frequencies of interest.
- 3) All antenna port conducted emissions testing was performed on a test bench with the antenna port of the EUT connected to the spectrum analyzer through calibrated cables, attenuators, and couplers.
- 4) For conducted spurious emissions, automated test software was used to measure emissions and capture the corresponding plots necessary to show compliance. The measurement software utilized is PCTEST "2G/3G Automation," Version 3.2.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 11 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 11 of 41



Occupied Bandwidth §2.1049

Test Overview

The occupied bandwidth, that is the frequency bandwidth such that, below its lower and above its upper frequency limits, the mean powers radiated are each equal to 0.5 percent of the total mean power radiated by a given emission shall be measured. All modes of operation were investigated and the worst case configuration results are reported in this section.

Test Procedure Used

KDB 971168 v02r02 - Section 4.2

Test Settings

- 1. The signal analyzer's automatic bandwidth measurement capability was used to perform the 99% occupied bandwidth and the 26dB bandwidth. The bandwidth measurement was not influenced by any intermediate power nulls in the fundamental emission.
- 2. RBW = 1 5% of the expected OBW
- 3. VBW \geq 3 x RBW
- 4. Detector = Peak
- 5. Trace mode = max hold
- 6. Sweep = auto couple
- 7. The trace was allowed to stabilize
- 8. If necessary, steps 2 7 were repeated after changing the RBW such that it would be within 1 – 5% of the 99% occupied bandwidth observed in Step 7

Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.




Figure 7-1. Test Instrument & Measurement Setup

Test Notes

None.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 12 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 12 of 41





Plot 7-1. Occupied Bandwidth Plot (Cellular CDMA Mode - Ch. 384)



Plot 7-2. Occupied Bandwidth Plot (PCS CDMA Mode - Ch. 600)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 12 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 13 of 41



Spurious and Harmonic Emissions at Antenna Terminal §2.1051 §22.917(a) §24.238(a)

Test Overview

The level of the carrier and the various conducted spurious and harmonic frequencies is measured by means of a calibrated spectrum analyzer. The spectrum is scanned from the lowest frequency generated in the equipment up to a frequency including its 10th harmonic. All out of band emissions are measured with a spectrum analyzer connected to the antenna terminal of the EUT while the EUT is operating at maximum power, and at the appropriate frequencies. All data rates were investigated to determine the worst case configuration. All modes of operation were investigated and the worst case configuration results are reported in this section.

The minimum permissible attenuation level of any spurious emission is $43 + \log_{10}(P_{[Watts]})$, where P is the transmitter power in Watts.

Test Procedure Used

KDB 971168 v02r02 - Section 6.0

Test Settings

- 1. Start frequency was set to 30MHz and stop frequency was set to 10GHz for Cell, 20GHz for PCS (separated into at least two plots per channel)
- 2. Detector = RMS
- 3. Trace mode = max hold
- 4. Sweep time = auto couple
- 5. The trace was allowed to stabilize
- 6. Please see test notes below for RBW and VBW settings

Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.




Figure 7-2. Test Instrument & Measurement Setup

Test Notes


Compliance with the applicable limits is based on the use of measurement instrumentation employing a resolution bandwidth of 100 kHz or greater for Part 22 and 1 MHz or greater for Part 24. However, in the 1 MHz bands immediately outside and adjacent to the frequency block a resolution bandwidth of at least one percent of the emission bandwidth of the fundamental emission of the transmitter may be employed. The emission bandwidth is defined as the width of the signal between two points, one below the carrier center frequency and one above the carrier center frequency, outside of which all emission are attenuated at least 26 dB below the transmitter power.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 14 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 14 of 41






Plot 7-3. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 1013)




Plot 7-4. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 1013)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 15 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 15 of 41





Plot 7-5. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 384)




Plot 7-6. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 384)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Daga 16 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 16 of 41






Plot 7-7. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 777)




Plot 7-8. Conducted Spurious Plot (Cellular CDMA Mode - Ch. 777)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogg 17 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 17 of 41






Plot 7-9. Conducted Spurious Plot (PCS CDMA Mode - Ch. 25)




Plot 7-10. Conducted Spurious Plot (PCS CDMA Mode - Ch. 25)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 10 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 18 of 41






Plot 7-11. Conducted Spurious Plot (PCS CDMA Mode - Ch. 600)




Plot 7-12. Conducted Spurious Plot (PCS CDMA Mode - Ch. 600)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 10 of 11
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 19 of 41





Plot 7-13. Conducted Spurious Plot (PCS CDMA Mode - Ch. 1175)



Plot 7-14. Conducted Spurious Plot (PCS CDMA Mode - Ch. 1175)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 20 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 20 of 41



Band Edge Emissions at Antenna Terminal §2.1051 §22.917(a) §24.238(a)

Test Overview

All out of band emissions are measured with a spectrum analyzer connected to the antenna terminal of the EUT while the EUT is operating at maximum power, and at the appropriate frequencies. All data rates were investigated to determine the worst case configuration. All modes of operation were investigated and the worst case configuration results are reported in this section.

The minimum permissible attenuation level of any spurious emission is 43 + log₁₀(P_[Watts]), where P is the transmitter power in Watts.

Test Procedure Used

KDB 971168 v02r02 - Section 6.0

Test Settings

- 1. Start and stop frequency were set such that the band edge would be placed in the center of the plot
- 2. Span was set large enough so as to capture all out of band emissions near the band edge
- 3. RBW > 1% of the emission bandwidth
- 4. VBW > 3 x RBW
- 5. Detector = RMS
- 6. Number of sweep points ≥ 2 x Span/RBW
- 7. Trace mode = trace average
- 8. Sweep time = auto couple
- 9. The trace was allowed to stabilize

Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.




Figure 7-3. Test Instrument & Measurement Setup

Test Notes

Per 22.917(b), 24.238(b), in the 1 MHz bands immediately outside and adjacent to the frequency block a resolution bandwidth of at least one percent of the emission bandwidth of the fundamental emission of the transmitter may be employed to demonstrate compliance with the out-of-band emissions limit. The emission bandwidth is defined as the width of the signal between two points, one below the carrier center frequency and one above the carrier center frequency, outside of which all emission are attenuated at least 26 dB below the transmitter power.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	Reviewed by: Quality Manager	
Test Report S/N:	Test Dates:	EUT Type:		Dags 21 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 21 of 41





Plot 7-15. Band Edge Plot (Cellular CDMA Mode - Ch. 1013)



Plot 7-16. 4MHz Span Plot (Cellular CDMA Mode - Ch. 1013)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 22 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 22 of 41





Plot 7-17. Band Edge Plot (Cellular CDMA Mode – Ch. 777)



Plot 7-18. 4MHz Span Plot (Cellular CDMA Mode - Ch. 777)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicaea		
Test Report S/N:	Test Dates:	EUT Type:		Dags 22 of 41	
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 23 of 41	






Plot 7-19. Band Edge Plot (PCS CDMA Mode - Ch. 25)



Plot 7-20. 4MHz Span Plot (PCS CDMA Mode - Ch. 25)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager	
Test Report S/N:	Test Dates:	EUT Type:		Dogo 24 of 41	
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 24 of 41	





Plot 7-21. Band Edge Plot (PCS CDMA Mode - Ch. 1175)



Plot 7-22. 4MHz Span Plot (PCS CDMA Mode - Ch. 1175)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 25 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser	Page 25 of 41	



Peak-Average Ratio §24.232(d)

Test Overview

A peak to average ratio measurement is performed at the conducted port of the EUT. The spectrum analyzers Complementary Cumulative Distribution Function (CCDF) measurement profile is used to determine the largest deviation between the average and the peak power of the EUT in a given bandwidth. The CCDF curve shows how much time the peak waveform spends at or above a given average power level. The percent of time the signal spends at or above the level defines the probability for that particular power level.

Test Procedure Used

KDB 971168 v02r02 - Section 5.7.1

Test Settings

- 1. The signal analyzer's CCDF measurement profile is enabled
- 2. Frequency = carrier center frequency
- 3. Measurement BW > Emission bandwidth of signal
- 4. The signal analyzer was set to collect one million samples to generate the CCDF curve
- 5. The measurement interval was set depending on the type of signal analyzed. For continuous signals (>98% duty cycle), the measurement interval was set to 1ms.

Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.





Figure 7-4. Test Instrument & Measurement Setup

Test Notes

None

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	Reviewed by: Quality Manager	
Test Report S/N:	Test Dates:	EUT Type:		Dags 26 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 26 of 41





Plot 7-23. Peak-Average Ratio Plot (PCS CDMA Mode - Ch. 600)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 27 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser	Page 27 of 41	



Radiated Power (ERP/EIRP) 7.6 §22.913(a)(2) 24.232(c)

Test Overview

Effective Radiated Power (ERP) and Equivalent Isotropic Radiated Power (EIRP) measurements are performed using the substitution method described in ANSI/TIA-603-D:2010 with the EUT transmitting into an integral antenna. Measurements on signals operating below 1GHz are performed using vertically polarized tuned dipole antennas. Measurements on signals operating above 1GHz are performed using vertically polarized broadband horn antennas. All measurements are performed as RMS average measurements while the EUT is operating at maximum power, and at the appropriate frequencies.

Test Procedures Used

KDB 971168 v02r02 - Section 5.2.1

ANSI/TIA-603-D:2010 - Section 2.2.17

Test Settings

- 1. Radiated power measurements are performed using the signal analyzer's "channel power" measurement capability for signals with continuous operation.
- 2. RBW = 1 5% of the expected OBW, not to exceed 1MHz
- 3. VBW ≥ 3 x RBW
- 4. Span = 1.5 times the OBW
- 5. No. of sweep points $\geq 2 \times \text{span} / \text{RBW}$
- 6. Detector = RMS
- 7. Trigger is set to "free run" for signals with continuous operation with the sweep times set to "auto".
- 8. The integration bandwidth was roughly set equal to the measured OBW of the signal for signals with continuous operation.
- 9. Trace mode = trace averaging (RMS) over 100 sweeps
- 10. The trace was allowed to stabilize

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 20 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 28 of 41



Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.




Figure 7-5. Radiated Test Setup <1GHz




Figure 7-6. Radiated Test Setup >1GHz

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 20 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 29 of 41



Test Notes

- 1) This device was tested under all RC and SO combinations and the worst case is reported with RC3/SO55 with "All Up" power control bits.
- 2) This unit was tested with its standard battery.
- 3) The EUT was tested in three orthogonal planes and in all possible test configurations and positioning. The worst case setup is reported in the tables below.

Frequency [MHz]	Mode	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Substitute Level [dBm]	Ant. Gain [dBd]	ERP [dBm]	ERP [Watts]	ERP Limit [dBm]	Margin [dB]
824.70	CDMA850	Н	204	330	15.41	5.01	20.42	0.110	38.45	-18.03
836.52	CDMA850	Н	353	322	14.41	5.16	19.57	0.091	38.45	-18.88
848.31	CDMA850	Н	187	341	13.48	5.30	18.78	0.076	38.45	-19.67

Table 7-2. ERP (Cellular CDMA)

Frequency [MHz]	Mode	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Substitute Level [dBm]	Ant. Gain [dBi]	EIRP [dBm]	EIRP [Watts]	EIRP Limit [dBm]	Margin [dB]
1851.25	CDMA1900	Н	100	266	20.85	9.35	30.20	1.047	33.01	-2.81
1880.00	CDMA1900	Н	250	269	21.89	9.27	31.16	1.307	33.01	-1.85
1908.75	CDMA1900	Н	235	260	21.50	9.25	30.75	1.188	33.01	-2.26

Table 7-3. EIRP (PCS CDMA)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Daga 20 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 30 of 41



Radiated Spurious Emissions Measurements §2.1053 §22.917(a) 24.238(a)

Test Overview

Radiated spurious emissions measurements are performed using the substitution method described in ANSI/TIA-603-D:2010 with the EUT transmitting into an integral antenna. Measurements on signals operating below 1GHz are performed using horizontally and vertically polarized tuned dipole antennas. Measurements on signals operating above 1GHz are performed using vertically and horizontally polarized broadband horn antennas. All measurements are performed as peak measurements while the EUT is operating at maximum power, and at the appropriate frequencies.

Test Procedures Used

KDB 971168 v02r02 - Section 5.8

ANSI/TIA-603-D:2010 - Section 2.2.12

Test Settings

- 1. RBW = 100kHz for emissions below 1GHz and 1MHz for emissions above 1GHz
- 2. VBW ≥ 3 x RBW
- 3. Span = 1.5 times the OBW
- 4. No. of sweep points ≥ 2 x span / RBW
- 5. Detector = Peak
- 6. Trace mode = max hold
- 7. The trace was allowed to stabilize

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogg 21 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 31 of 41



Test Setup

The EUT and measurement equipment were set up as shown in the diagram below.




Figure 7-7. Test Instrument & Measurement Setup

Test Notes

- 1) This device was tested under all RC and SO combinations and the worst case is reported with RC3/SO55 with "All Up" power control bits.
- 2) This unit was tested with its standard battery.
- 3) The EUT was tested in three orthogonal planes and in all possible test configurations and positioning. The worst case setup is reported in the tables below.
- 4) The spectrum is measured from 9kHz to the 10th harmonic of the fundamental frequency of the transmitter. The worst-case emissions are reported.
- 5) Emissions below 18GHz were measured at a 3 meter test distance while emissions above 18GHz were measured at a 1 meter test distance with the application of a distance correction factor.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dago 22 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 32 of 41



OPERATING FREQUENCY: 824.70 MHz

CHANNEL: 1013

MEASURED OUTPUT POWER: 20.42 dBm = 0.110 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 33.42$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBd]	Spurious Emission Level [dBm]	[dBc]
1649.40	Н	205	138	-61.26	6.56	-54.70	75.1
2474.10	Н	-	-	-63.59	7.30	-56.29	76.7
3298.80	Н	-	-	-60.92	7.37	-53.55	74.0
4123.50	Н	-	-	-58.01	8.02	-49.99	70.4

Table 7-4. Radiated Spurious Data (Cellular CDMA Mode - Ch. 1013)

OPERATING FREQUENCY: 836.52 MHz

CHANNEL: 384

MEASURED OUTPUT POWER: 19.57 dBm = 0.091 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 32.57$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBd]	Spurious Emission Level [dBm]	[dBc]
1673.04	Н	111	132	-56.57	6.55	-50.02	69.6
2509.56	Н	-	-	-63.82	7.34	-56.48	76.0
3346.08	Н	-	-	-60.63	7.44	-53.19	72.8
4182.60	Н	-	-	-58.21	8.20	-50.01	69.6

Table 7-5. Radiated Spurious Data (Cellular CDMA Mode – Ch. 384)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 22 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 33 of 41



OPERATING FREQUENCY: 848.31 MHz

CHANNEL: 777

MEASURED OUTPUT POWER: 18.78 dBm = 0.076 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 31.78$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBd]	Spurious Emission Level [dBm]	[dBc]
1696.62	Н	-	-	-63.84	6.55	-57.30	76.1
2544.93	Н	-	-	-63.40	7.36	-56.04	74.8
3393.24	Н	-	-	-60.63	7.51	-53.12	71.9
4241.55	Н	-	-	-58.45	8.39	-50.06	68.8

Table 7-6. Radiated Spurious Data (Cellular CDMA Mode – Ch. 777)

OPERATING FREQUENCY: 1851.25 MHz

CHANNEL: 25

MEASURED OUTPUT POWER: 30.20 dBm = 1.047 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 43.20$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Height	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBi]	Spurious Emission Level [dBm]	[dBc]
3702.50	Н	-	-	-57.65	9.43	-48.22	78.4
5553.75	Н	-	-	-55.98	10.79	-45.19	75.4
7405.00	Н	-	-	-52.33	10.70	-41.63	71.8

Table 7-7. Radiated Spurious Data (PCS CDMA Mode - Ch. 25)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 24 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 34 of 41



OPERATING FREQUENCY: 1880.00 MHz

CHANNEL: 600

MEASURED OUTPUT POWER: 31.16 dBm = 1.307 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 44.16$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Antenna Height [cm]	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBi]	Spurious Emission Level [dBm]	[dBc]
3760.00	Н	-	-	-57.39	9.28	-48.11	78.3
5640.00	Н	-	-	-56.62	11.03	-45.59	75.8
7520.00	Н	-	-	-53.50	10.97	-42.53	72.7

Table 7-8. Radiated Spurious Data (PCS CDMA Mode - Ch. 600)

OPERATING FREQUENCY: 1908.75 MHz

CHANNEL: 1175

MEASURED OUTPUT POWER: 30.75 dBm = 1.188 W

MODULATION SIGNAL: CDMA

DISTANCE: 3 meters

LIMIT: $43 + 10 \log_{10} (W) = 43.75$ dBc

Frequency [MHz]	Ant. Pol. [H/V]	Height	Turntable Azimuth [degree]	Level at Antenna Terminals [dBm]	Substitute Antenna Gain [dBi]	Spurious Emission Level [dBm]	[dBc]
3817.50	Н	-	-	-57.06	9.19	-47.87	78.1
5726.25	Н	-	-	-57.42	11.28	-46.14	76.3
7635.00	Н	-	-	-53.41	11.17	-42.23	72.4

Table 7-9. Radiated Spurious Data (PCS CDMA Mode - Ch. 1175)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 25 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 35 of 41



Frequency Stability / Temperature Variation §2.1055 §22.355 §24.235

Test Overview and Limit

Frequency stability testing is performed in accordance with the guidelines of ANSI/TIA-603-D:2010. The frequency stability of the transmitter is measured by:

- Temperature: The temperature is varied from -30°C to +50°C in 10°C increments using an a.) environmental chamber.
- b.) Primary Supply Voltage: The primary supply voltage is varied from 85% to 115% of the nominal value for non hand-carried battery and AC powered equipment. For hand-carried, battery-powered equipment, primary supply voltage is reduced to the battery operating end point which shall be specified by the manufacturer.

For Part 22, the frequency stability of the transmitter shall be maintained within ±0.00025% (±2.5 ppm) of the center frequency. For Part 24 the frequency stability shall be sufficient to ensure that the fundamental emission stays within the authorized frequency block.

Test Procedure Used

ANSI/TIA-603-D:2010

Test Settings

- 1. The carrier frequency of the transmitter is measured at room temperature (20°C to provide a reference).
- 2. The equipment is turned on in a "standby" condition for fifteen minutes before applying power to the transmitter. Measurement of the carrier frequency of the transmitter is made within one minute after applying power to the transmitter.
- 3. Frequency measurements are made at 10°C intervals ranging from -30°C to +50°C. A period of at least one half-hour is provided to allow stabilization of the equipment at each temperature level.

Test Setup

The EUT was connected via an RF cable to a spectrum analyzer with the EUT placed inside an environmental chamber.

Test Notes

None

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 26 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 36 of 41



Frequency Stability / Temperature Variation §2.1055 §22.355

OPERATING FREQUENCY: 836,520,000 Hz

> CHANNEL: 384

3.00 VDC REFERENCE VOLTAGE:

DEVIATION LIMIT: ± 0.00025 % or 2.5 ppm

VOLTAGE (%)	POWER (VDC)	TEMP (°C)	FREQUENCY (Hz)	Freq. Dev. (Hz)	Deviation (%)
100 %	3.00	+ 20 (Ref)	836,520,079	79	0.0000094
100 %		- 30	836,519,891	-109	-0.0000130
100 %		- 20	836,520,018	18	0.0000022
100 %		- 10	836,519,711	-289	-0.0000345
100 %		0	836,519,948	-52	-0.0000062
100 %		+ 10	836,520,163	163	0.0000195
100 %		+ 20	836,520,002	2	0.0000002
100 %		+ 30	836,519,993	-7	-0.0000008
100 %		+ 40	836,519,742	-258	-0.0000308
100 %		+ 50	836,519,996	-4	-0.000005
85 %	2.55	+ 20	836,520,026	26	0.0000031

Table 7-10. Frequency Stability Data (Cellular CDMA Mode - Ch. 384)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 27 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 37 of 41



Frequency Stability / Temperature Variation §2.1055 §22.355




Figure 7-8. Frequency Stability Graph (Cellular CDMA Mode – Ch. 384)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 20 of 44
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 38 of 41



Frequency Stability / Temperature Variation §2.1055 §24.235

OPERATING FREQUENCY: 1,880,000,000 Hz

> CHANNEL: 600

REFERENCE VOLTAGE: 3.00 **VDC**

VOLTAGE (%)	POWER (VDC)	TEMP (°C)	FREQUENCY (Hz)	Freq. Dev. (Hz)	Deviation (%)
100 %	3.00	+ 20 (Ref)	1,880,000,042	42	0.0000022
100 %		- 30	1,880,000,312	312	0.0000166
100 %		- 20	1,879,999,979	-21	-0.0000011
100 %		- 10	1,880,000,089	89	0.0000047
100 %		0	1,879,999,938	-62	-0.0000033
100 %		+ 10	1,879,999,885	-115	-0.0000061
100 %		+ 20	1,879,999,687	-313	-0.0000166
100 %		+ 30	1,880,000,000	0	0.0000000
100 %		+ 40	1,879,999,775	-225	-0.0000120
100 %		+ 50	1,880,000,206	206	0.0000110
85 %	2.55	+ 20	1,880,000,280	280	0.0000149

Table 7-11. Frequency Stability Data (PCS CDMA Mode - Ch. 600)

Note:

Based on the results of the frequency stability test at the center channel the frequency deviation results measured are very small. As such it is determined that the channels at the band edge would remain inband when the maximum measured frequency deviation noted during the frequency stability tests is applied. Therefore the device is determined to remain operating in band over the temperature and voltage range as tested.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)		Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Page 39 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		raye 39 01 4 1



Frequency Stability / Temperature Variation §2.1055 §24.235




Figure 7-9. Frequency Stability Graph (PCS CDMA Mode – Ch. 600)

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)	medicpen	Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dags 40 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 40 of 41



CONCLUSION

The data collected relate only to the item(s) tested and show that the MedicPen Portable Pill Dispenser FCC ID: 2ABTHC200W complies with all the requirements of Parts 22 & 24 of the FCC rules.

FCC ID: 2ABTHC200W	PCTEST	FCC Pt. 22 & 24 CDMA MEASUREMENT REPORT (CERTIFICATION)		Reviewed by: Quality Manager
Test Report S/N:	Test Dates:	EUT Type:		Dogo 41 of 41
0Y1603150550.2ABTH	12/22/2014, 4/15/2016	Portable Pill Dispenser		Page 41 of 41