

FCC & IC Labelling submission

ResMed Confidential

Proprietary Information, not to be reproduced or made available to third parties without prior consent from ResMed and not to be used in any unauthorised way. © ResMed I td

SPEC NO.	eLB0925-070
REV	1.0
PAGE	1 of 1

Mark Bertinetti 16 July 2014

Common labels/moulded in artwork



BRAND

ResMed branding chrome or colour foiled on front of product




"BF APPLIED PART" Body Fibrillation Icon & Text

Moulded into plastic case using an insert near rear Outlet



"REFER TO MANUAL" Label Man reading book Icon & Text

White pre-printed label stuck next to power connector on <u>rear</u> of chassis



REGULATORY Label White pre-printed label stuck on bottom of chassis


Product Specific Labels















SERIAL NUMBER LABELS Thermally printed labels attached to rear of product

PRODUCT NAME PLAQUES
 Foiled product variant badges
 "permanently" attached to front of product

FCC & IC Compliance Labels
Thermally printed labels
attached to bottom of product





Figure 1 Front of CPAP





Figure 3 Underneath of CPAP