

Report No.: EED32K00143402 Page 1 of 7

# RF Exposure Evaluation Report

**Product** : Pulse Oximeter

Trade mark : JUMPER

Model/Type reference : JPD-500G, JPD-500H

Serial Number : N/A

Report Number : EED32K00143402 FCC ID : 2ADYL-JPD500G

Date of Issue : Aug. 01, 2018

47 CFR Part 1.1307

**Test Standards** 47 CFR Part 2.1093

KDB 447498 D01v06

Test result : PASS

#### Prepared for:

Shenzhen Jumper Medical Equipment Co., Ltd D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzhen, Guangdong, China

Prepared by:

Centre Testing International Group Co., Ltd. Hongwei Industrial Zone, Bao'an 70 District, Shenzhen, Guangdong, China

TEL: +86-755-3368 3668 FAX: +86-755-3368 3385

Report Sea

Tested By:

10M-chen

Tom chen (Test Project)

Reviewed by:

Date:

Keun (my

Kevin yang (Reviewer)

Aug. 01, 2018

Kevin lan (Project Engineer)

Sheek Luo (Lab supervisor)

Check No.: 3177403017

Hotline: 400-6788-333 www.cti-cert.com E-mail: info@cti-cert.com Complaint call: 0755-33681700 Complaint E-mail: complaint@cti-cert.com









Page 2 of 7

Report No. : EED32K00143402

# 2 Version

Version No.	Date		Description	
00	Aug. 01, 2018		Original	
/				(3)
1		(6)	6	0

















































































Page 3 of 7

Report No.: EED32K00143402

3 Contents

		Page
1 COVER PAGE		 1
2 VERSION		 2
3 CONTENTS	()	 3
4 GENERAL INFORMATION		4
4.1 CLIENT INFORMATION	EUT	4
	UBJECTIVE TO THIS STANDARD	
	DS	
4.6 ABNORMALITIES FROM STAP	NDARD CONDITIONS	 5
	JESTED BY THE CUSTOMER	
5 SAR EVALUATION		 6
5.1 RF EXPOSURE COMPLIANC	E REQUIREMENT	6
	nt	
PHOTOGRAPHS OF EUT CON	STRUCTIONAL DETAILS	 7

















































Report No.: EED32K00143402 Page 4 of 7

# 4 General Information

#### 4.1 Client Information

Applicant:	Shenzhen Jumper Medical Equipment Co., Ltd	
Address of Applicant:	D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzhen, Guangdong, China	
Manufacturer:	Shenzhen Jumper Medical Equipment Co., Ltd	
Address of Manufacturer:  D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzh Guangdong, China		
Factory:	Shenzhen Jumper Medical Equipment Co., Ltd	
Address of Factory:	D Building, No. 71, Xintian Road, Fuyong Street, Baoan, Shenzhen, Guangdong, China	

## 4.2 General Description of EUT

Product Name:	Pulse Oximeter
Model No.(EUT):	JPD-500G, JPD-500H
Test Model No.:	JPD-500G
Trade mark:	JUMPER
EUT Supports Radios application:	BT 4.2 Single mode, 2402-2480MHz
Hardware Version of the sample:	V2.1(manufacturer declare)
Firmware version of the sample:	JPD_500G_BT(manufacturer declare)

# 4.3 Product Specification subjective to this standard

Operation Frequency:	2402MHz~2480MHz		
Bluetooth Version:	4.2		
Modulation Technique:	DSSS		
Number of Channel:	40		
Test Power Grade:	N/A(manufacturer declare)		
Test Software of EUT:	nRFgo Studio(manufacturer declare)		
Antenna Type:	PCB Antenna		
Antenna Gain:	0dBi		
Power Supply:	Battery: 2x1.5V(AAA)=3.0V		
Conducted Book Outroit	-3.695dBm		
Conducted Peak Output Power:	The Conducted Peak Output Power data refer to the report EED32K00143401		
Sample Received Date:	Jun. 8, 2018		
Sample tested Date:	Jun. 8, 2018 to Aug. 1, 2018		

The tested sample(s) and the sample information are provided by the client.

Model No.: JPD-500G, JPD-500H

Only the model JPD-500G was tested, since the electrical circuit design, layout, components used and internal wiring were identical for the above models, with difference being appearance and model name.

Hotline: 400-6788-333 www.cti-cert.com E-mail: info@cti-cert.com Complaint call: 0755-33681700 Complaint E-mail: complaint@cti-cert.com









Report No.: EED32K00143402 Page 5 of 7

#### 4.4 Test Location

All tests were performed at:

Centre Testing International Group Co., Ltd.

Building C, Hongwei Industrial Park Block 70, Bao'an District, Shenzhen, Guangdong, China 518101

Telephone: +86 (0) 755 33683668 Fax:+86 (0) 755 33683385

No tests were sub-contracted. FCC Designation No.: CN1164

### 4.5 Deviation from Standards

None

### 4.6 Abnormalities from Standard Conditions

None.

# 4.7 Other Information Requested by the Customer

None.















































































Report No.: EED32K00143402 Page 6 of 7

#### SAR Evaluation

## 5.1 RF Exposure Compliance Requirement

#### 5.1.1 Standard Requirement

According to KDB447498D01 General RF Exposure Guidance v06 Standalone SAR test exclusion considerations

Unless specifically required by the published RF exposure KDB procedures, standalone 1-g head or body and 10-g extremity SAR evaluation for general population exposure conditions, by measurement or numerical simulation, is not required when the corresponding SAR Exclusion Threshold condition, listed below, is satisfied.

#### **5.1.2 Limits**

The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

[(max. power of channel, including tune-up tolerance, mW)/(min. test separation distance, mm)]  $[\sqrt{f(GHz)}] \le 3.0$  for 1-g SAR and  $\le 7.5$  for 10-g extremity SAR, where

f(GHz) is the RF channel transmit frequency in GHz

Power and distance are rounded to the nearest mW and mm before calculation<sup>17</sup>

The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minimum test separation distance is ≤ 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion

#### 5.1.3 EUT RF Exposure

The Max Conducted Peak Output Power is -3.695dBm in highest channel(2.480GHz);

The best case gain of the antenna is 0dBi.

EIRP = -3.695dBm + 0dBi = -3.695dBm

-3.695dBm logarithmic terms convert to numeric result is nearly 0.43mW

According to the formula. calculate the EIRP test result:

[(max. power of channel, including tune-up tolerance, mW)/(min. test separation distance, mm)] · [√f(GHz)]

General RF Exposure = (0.43mW / 5 mm ) x  $\sqrt{2.480}$ GHz = 0.14 ①

SAR requirement:

S = 3.0

(2):

(1) < (2).

So the SAR report is not required.







































Report No.: EED32K00143402 Page 7 of 7

# **PHOTOGRAPHS OF EUT Constructional Details**

Refer to Report No. EED32K00143401 for EUT external and internal photos.



The test report is effective only with both signature and specialized stamp, The result(s) shown in this report refer only to the sample(s) tested. Without written approval of CTI, this report can't be reproduced except in full.

