intelcl:nic smart engineering

Intelclinic Sp. z o.o. Niegolewskiego 17/1 01-570 Warszawa tel: 22 633 87 05 info@intelclinic.com www.intelclinic.com NIP: 7010369609 REGON: 146531603 KRS: 0000450090 Sąd rejonowy dla m. st. Warszawy Kapital Zakładowy: 5 250 zł opłacony w całości

RF Exposure Analysis

FCC ID: 2AF73-NEUR1

Analysis for FCC portable use

Standalone SAR test exclusion considerations are defined in KDB 447498 D01, Chapter 4.3.1 where the 1-g head or body and 10-g extremity SAR exclusion threshold is defined by the following formula:

[(max. power of channel, including tune-up tolerance, mW) / (min. test separation distance, mm)] * $[\sqrt{f(GHz)}] \le 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR

- f(GHz) is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison

For the Neuroon the maximum conducted output power is +1.75dBm (1.5mW) at 2.4GHz

Applying the above data using the given KDB 447498 D01 formula, and minimum separation distance of 5mm, the following results:

 $(1.5 \text{mW} / 5 \text{ mm}) \text{ x } \sqrt{2.4 \text{ GHz}} = 0.46$

(i.e.: \leq 3.0 for 1-g SAR and \leq 7.5 for 10-g extremity SAR)

Conclusion

This demonstrates the Neuroon meets the criteria for 1-g head / body and 10-g extremity SAR test exemption.