

# Maximum Permissive Exposure

FCC ID: 2AFQC22008

Product Name: MOCAheart Heart Health Monitor


Model No: MOCAheart

Pursuant to KDB 447498 D01 v05r02 that SAR testing is not required when EUT has output power related to distance satisfy with following table; or when distance is less than 5 mm and output power threshold calculated from  $[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})] \cdot [\sqrt{f_{\text{(GHz)}}}] \leq 3.0$ .

MHz	5	10	15	20	25	mm
150	39	77	116	155	194	SAR Test Exclusion Threshold (mW)
300	27	55	82	110	137	
450	22	45	67	89	112	
835	16	33	49	66	82	
900	16	32	47	63	79	
1500	12	24	37	49	61	
1900	11	22	33	44	54	
2450	10	19	29	38	48	
3600	8	16	24	32	40	
5200	7	13	20	26	33	
5400	6	13	19	26	32	
5800	6	12	19	25	31	
MHz	30	35	40	45	50	mm
150	232	271	310	349	387	SAR Test Exclusion Threshold (mW)
300	164	192	219	246	274	
450	134	157	179	201	224	
835	98	115	131	148	164	
900	95	111	126	142	158	
1500	73	86	98	110	122	
1900	65	76	87	98	109	
2450	57	67	77	86	96	
3600	47	55	63	71	79	
5200	39	46	53	59	66	
5400	39	45	52	58	65	
5800	37	44	50	56	62	

This application FCC ID: 2AFQC22008 is equipped PCB antenna. Thus the most closely distance from antenna to user is **1 mm**(indicates in next page) and the max output power (e.i.r.p.) is 1.184 mW which is satisfy with the exclusion level **1.91 mW**.

Max Output Power (dBm)	Antenna Gain (dBi)	e.i.r.p. (dBm)	e.i.r.p. (mW)	Power Exclusion Threshold (mW)
-1.128	1.86	0.732	1.184	1.91



*Ben Cheng*

---

Mr. Ben Cheng  
Manager  
AUDIX Technology Corporation