



REPORT No. : SZ15090001S01

RF EXPOSURE EVALUATION REPORT

APPLICANT : Shenzhen CardioCycle Technology Co.,Ltd .

PRODUCT NAME : SenseON

MODEL NAME : SN-1512

TRADE NAME : CardioCycle

BRAND NAME : CardioCycle

FCC ID : 2AG6S-SN1512

STANDARD(S) : 47CFR 2.1093
KDB 447498 D01 General RF Exposure
Guidance v05r02

ISSUE DATE : 2016-01-12



SHENZHEN MORLAB COMMUNICATIONS TECHNOLOGY Co., Ltd.

NOTE: This document is issued by MORLAB, the test report shall not be reproduced except in full without prior written permission of the company. The test results apply only to the particular sample(s) tested and to the specific tests carried out which is available on request for validation and information confirmed at our website.

MORLAB GROUP

FL1-3, Building A, FeiYang Science Park, No.8 LongChang Road,
Block67, BaoAn District, ShenZhen , GuangDong Province, P. R. China

Tel: 86-755-36698555
Http://www.morlab.com

Fax: 86-755-36698525
E-mail: service@morlab.cn



DIRECTORY

TEST REPORT DECLARATION	3
1. TECHNICAL INFORMATION	4
1.1. IDENTIFICATION OF APPLICANT	4
1.2. IDENTIFICATION OF MANUFACTURER	4
1.3. EQUIPMENT UNDER TEST (EUT)	4
1.3.1. PHOTOGRAPHS OF THE EUT	5
1.3.2. IDENTIFICATION OF ALL USED EUT	6
1.4. APPLIED REFERENCE DOCUMENTS	6
2.DEVICE CATEGORY AND RF EXPOSURE LIMIT	7
3.MEASUREMENT OF CONDUCTED PEAK OUTPUT POWER	8
4. RF EXPOSURE EVALUATION	8
ANNEX A GENERAL INFORMATION	9

Change History		
Issue	Date	Reason for change
1.0	2016-01-12	First edition



REPORT No. : SZ15090001S01

TEST REPORT DECLARATION

Applicant	Shenzhen CardioCycle Technology Co.,Ltd .
Applicant Address	422 Bldg 4,Shenzhen International Software park,Nanshan District,Shenzhen,China
Manufacturer	Shenzhen CardioCycle Technology Co.,Ltd .
Manufacturer Address	422 Bldg 4,Shenzhen International Software park,Nanshan District,Shenzhen,China
Product Name	SenseON
Model Name	SN-1512
Brand Name	CardioCycle
HW Version	2.16
SW Version	Beta1
Test Standards	47CFR 2.1093; KDB 447498 D01 General RF Exposure Guidance v05r02
Issue Date	2016-01-12
SAR Evaluation	Not Required

Tested by : Liu Jun
Liu Jun

Reviewed by : Zhu Zhan
Zhu Zhan

Approved by : Zeng Dexin
Zeng Dexin



1. TECHNICAL INFORMATION

Note: the following data is based on the information by the applicant.

1.1. Identification of Applicant

Company Name:	Shenzhen CardioCycle Technology Co.,Ltd .
Address:	422 Bldg 4,Shenzhen International Software park,Nanshan District,Shenzhen,China

1.2. Identification of Manufacturer

Company Name:	Shenzhen CardioCycle Technology Co.,Ltd .
Address:	422 Bldg 4,Shenzhen International Software park,Nanshan District,Shenzhen,China

1.3. Equipment Under Test (EUT)

Model Name:	SN-1512
Trade Name:	CardioCycle
Brand Name:	N/A
Hardware Version:	2.16
Software Version:	Beta1
Frequency Bands:	Bluetooth4.0:2402-2480MHz;
Modulation Mode:	Bluetooth4.0: GFSK
Antenna type:	Fixed Internal Antenna
Development Stage:	Identical prototype



REPORT No. : SZ15090001S01

1.3.1. Photographs of the EUT

1. EUT front view





1.3.2. Identification of all used EUT

The EUT identity consists of numerical and letter characters, the letter character indicates the test sample, and the following two numerical characters indicate the software version of the test sample.

EUT Identity	Hardware Version	Software Version
1#	2.16	Beta1

1.4. Applied Reference Documents

Leading reference documents for testing:

No.	Identity	Document Title
1	47 CFR§2.1093	Radiofrequency Radiation Exposure Evaluation: portable devices
2	KDB 447498 D01v05r02	General RF Exposure Guidance



2. DEVICE CATEGORY AND RF EXPOSURE LIMIT

Per user manual, this device is a Bluetooth heartbeat meter. Based on 47CFR 2.1093, this device belongs to portable device category with General Population/Uncontrolled exposure.

Portable Devices:

47CFR 2.1093(b)

For purposes of this section, a portable device is defined as a transmitting device designed to be used so that the radiating structure(s) of the device is/are within 20 centimeters of the body of the user.

GENERAL POPULATION / UNCONTROLLED EXPOSURE

47CFR 2.1093(d) (2)

Limits for General Population/Uncontrolled exposure: 0.08 W/kg as averaged over the whole-body and spatial peak SAR not exceeding 1.6 W/kg as averaged over any 1 gram of tissue (defined as a tissue volume in the shape of a cube). Exceptions are the hands, wrists, feet and ankles where the spatial peak SAR shall not exceed 4 W/kg, as averaged over any 10 grams of tissue (defined as a tissue volume in the shape of a cube). General Population/Uncontrolled limits apply when the general public may be exposed, or when persons that are exposed as a consequence of their employment may not be fully aware of the potential for exposure or do not exercise control over their exposure. Warning labels placed on consumer devices such as cellular telephones will not be sufficient reason to allow these devices to be evaluated subject to limits for occupational/controlled exposure in paragraph (d)(1) of this section.



3. MEASUREMENT OF CONDUCTED PEAK OUTPUT POWER

1. BT 4.0 peak output power

Band	Channel	Frequency (MHz)	Output Power(dBm)
			GFSK
BT	0	2402	-3.62
	19	2440	-2.79
	39	2480	-2.37

4. RF EXPOSURE EVALUATION

The device only incorporates a Bluetooth transmitter, so standalone SAR evaluation is required for Bluetooth and simultaneous SAR is not required.

Standalone transmission SAR evaluation

According to KDB 447498 section 4.3.1, the 1-g SAR test exclusion thresholds at test separation Distances ≤ 50 mm are determined by:

$[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})] \cdot [\sqrt{f(\text{GHz})}] \leq 3.0$

The maximum tune-up limit power is **0.63mW @ 2.48GHz**

BT antenna spacing 0mm from body, so use **5mm** as the most conservative minimum test separation distance,

$[(\text{max. power of channel, including tune-up tolerance, mW})/(\text{min. test separation distance, mm})] \cdot [\sqrt{f(\text{GHz})}] = \mathbf{0.195} \leq 3.0$

So SAR evaluation is not required for this device.



REPORT No. : SZ15090001S01

ANNEX A GENERAL INFORMATION

1. Identification of the Responsible Testing Laboratory

Company Name:	Shenzhen Morlab Communications Technology Co., Ltd.
Department:	Morlab Laboratory
Address:	FL.3, Building A, FeiYang Science Park, No.8 LongChang Road, Block 67, BaoAn District, ShenZhen, Guangdong Province, P. R. China
Responsible Test Lab Manager:	Mr. Su Feng
Telephone:	+86 755 36698555
Facsimile:	+86 755 36698525

2. Identification of the Responsible Testing Location

Name:	Shenzhen Morlab Communications Technology Co., Ltd. Morlab Laboratory
Address:	FL.3, Building A, FeiYang Science Park, No.8 LongChang Road, Block 67, BaoAn District, ShenZhen, Guangdong Province, P. R. China

***** END OF REPORT *****