
Annex 3 to Test Report # EMCC-150530A, 2016-04-25

PHOTOGRAPHS OF MODIFICATIONS OF EQUIPMENT UNDER TEST (EUT)

EQUIPMENT UNDER TEST:

Device:	AG501
Serial Number:	03-12-10-24
Equipment Class:	Articulograph
Manufacturer:	Carstens Medizinelektronik GmbH
Address:	Nelkenweg 8 37120 Bovenden Germany
Phone:	+49 5593 1697
Fax:	+49 5593 8791

RELEVANT STANDARD(S): 47 CFR §§ 15.207, 15.209

MEASUREMENT PROCEDURE:

☒ ANSI C63.10-2013 ☐ RSS-Gen Issue 4 ☐ Other



Photo A3-1: Modification 1, ferrites at Ethernet cable