10591- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc duty cycle)	X	4.76	66.87	16.69	0.46	130.0	± 9.6 %
		Y	4.63	66.33	16.27		130.0	
		Z	4.59	67.14	16.79		130.0	
10592- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc duty cycle)	Х	4.91	67.21	16.82	0.46	130.0	± 9.6 %
		Y	4.76	66.65	16.40		130.0	
		Z	4.70	67.43	16.91		130.0	
10593- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS2, 90pc duty cycle)	Х	4.83	67.11	16.70	0.46	130.0	± 9.6 %
		Υ	4.68	66.53	16.26		130.0	
		Z	4.62	67.31	16.77		130.0	
10594- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc duty cycle)	Х	4.88	67.28	16.86	0.46	130.0	± 9.6 %
		Y	4.73	66.70	16.42		130.0	
		Z	4.68	67.50	16.94		130.0	
10595- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc duty cycle)	Х	4.85	67.24	16.76	0.46	130.0	± 9.6 %
		Y	4.70	66.66	16.32		130.0	
1055	1555 000 11 1055	Z	4.64	67.49	16.86		130.0	
10596- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS5, 90pc duty cycle)	X	4.78	67.24	16.77	0.46	130.0	± 9.6 %
		Y	4.63	66.64	16.31		130.0	
1055	1555 000 11 (155	Z	4.57	67.44	16.85		130.0	
10597- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc duty cycle)	X	4.73	67.14	16.64	0.46	130.0	± 9.6 %
		Y	4.58	66.51	16.17		130.0	
		Z	4.52	67.30	16.69		130.0	
10598- AAB	IEEE 802.11n (HT Mixed, 20MHz, MCS7, 90pc duty cycle)	X	4.72	67.36	16.90	0.46	130.0	± 9.6 %
		Y	4.56	66.73	16.43		130.0	
		Z	4.53	67.58	16.99		130.0	
10599- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc duty cycle)	X	5.42	67.32	16.85	0.46	130.0	± 9.6 %
		Y	5.31	66.90	16.54		130.0	
		Z	5.27	67.52	16.99		130.0	
10600- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc duty cycle)	X	5.54	67.72	17.03	0.46	130.0	± 9.6 %
		Y	5.45	67.35	16.75		130.0	
		Z	5.36	67.85	17.13		130.0	
10601- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc duty cycle)	X	5.44	67.50	16.94	0.46	130.0	± 9.6 %
		Y	5.33	67.07	16.62		130.0	
		Z	5.28	67.68	17.06		130.0	
10602- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc duty cycle)	X	5.56	67.62	16.91	0.46	130.0	± 9.6 %
		Y	5.46	67.23	16.62		130.0	
		Z	5.35	67.63	16.95		130.0	
10603- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc duty cycle)	X	5.62	67.85	17.16	0.46	130.0	± 9.6 %
		Y	5.52	67.49	16.88		130.0	
1000:		Z	5.41	67.93	17.24		130.0	
10604- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS5, 90pc duty cycle)	X	5.47	67.46	16.95	0.46	130.0	± 9.6 %
		Y	5.41	67.17	16.71		130.0	
10005	1555 000 44 445544 4 45544	Z	5.28	67.47	16.98		130.0	
10605- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc duty cycle)	X	5.54	67.66	17.05	0.46	130.0	± 9.6 %
		Y	5.44	67.26	16.75		130.0	
		Z	5.34	67.69	17.09		130.0	Y
10606- AAB	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc duty cycle)	X	5.28	66.97	16.57	0.46	130.0	± 9.6 %
		Y	5.18	66.52	16.23		130.0	
		Z	5.14	67.17	16.69			

Certificate No: EX3-7350_Dec18

10607- AAB	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	X	4.61	66.25	16.34	0.46	130.0	± 9.6 %
		Y	4.46	65.64	15.88		130.0	
		Z	4.46	66.58	16.49		130.0	
10608- AAB	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	X	4.79	66.64	16.51	0.46	130.0	± 9.6 %
	0000 000, 070.07	Y	4.62	66.00	16.04		130.0	
		Z	4.59	66.90	16.63		130.0	
10609- AAB	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	X	4.68	66.49	16.35	0.46	130.0	± 9.6 %
, , , ,	cope day of old	Y	4.51	65.83	15.86		130.0	
		Z	4.49	66.74	16.45		130.0	
10610- AAB	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	Х	4.73	66.65	16.51	0.46	130.0	± 9.6 %
		Y	4.56	65.99	16.03		130.0	
		Z	4.54	66.93	16.63		130.0	
10611- AAB	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	Х	4.64	66.45	16.36	0.46	130.0	± 9.6 %
		Y	4.48	65.79	15.87		130.0	
		Z	4.45	66.71	16.47		130.0	
10612- AAB	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	Х	4.65	66.62	16.42	0.46	130.0	± 9.6 %
		Y	4.48	65.93	15.91		130.0	
		Z	4.44	66.83	16.51		130.0	
10613- AAB	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	X	4.65	66.48	16.28	0.46	130.0	± 9.6 %
		Y	4.48	65.78	15.77		130.0	
		Z	4.43	66.62	16.33		130.0	
10614- AAB	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	Х	4.60	66.67	16.51	0.46	130.0	± 9.6 %
	1	Y	4.43	65.97	16.01		130.0	
		Z	4.41	66.91	16.62		130.0	
10615- AAB	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	X	4.64	66.30	16.15	0.46	130.0	± 9.6 %
		Y	4.48	65.63	15.65		130.0	
		Z	4.44	66.52	16.22		130.0	
10616- AAB	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	X	5.25	66.64	16.50	0.46	130.0	± 9.6 %
		Y	5.12	66.09	16.12		130.0	
		Z	5.08	66.74	16.59		130.0	
10617- AAB	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	Х	5.32	66.83	16.57	0.46	130.0	± 9.6 %
		Y	5.19	66.31	16.20		130.0	
		Z	5.11	66.84	16.62		130.0	
10618- AAB	IEEE 802.11ac WiFi (40MHz, MCS2, 90pc duty cycle)	X	5.21	66.85	16.59	0.46	130.0	± 9.6 %
		Y	5.08	66.31	16.22		130.0	
		Z	5.03	66.94	16.68		130.0	
10619- AAB	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	Х	5.22	66.63	16.42	0.46	130.0	± 9.6 %
		Y	5.09	66.09	16.05		130.0	
		Z	5.06	66.79	16.54		130.0	
10620- AAB	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	X	5.31	66.66	16.48	0.46	130.0	± 9.6 %
		Y	5.18	66.13	16.11		130.0	
		Z	5.12	66.73	16.55		130.0	
10621- AAB	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	Х	5.31	66.79	16.66	0.46	130.0	± 9.6 %
		Y	5.19	66.27	16.30		130.0	
		Z	5.13	66.86	16.74		130.0	
10622- AAB	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.33	66.97	16.75	0.46	130.0	± 9.6 %
		Y	5.20	66.43	16.38		130.0	
		Z	5.12	66.96	16.79	_	130.0	

Certificate No: EX3-7350_Dec18 Page 36 of 39

10623- AAB	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc duty cycle)	X	5.20	66.49	16.39	0.46	130.0	± 9.6 %
		Y	5.07	65.92	15.99		130.0	
		Z	5.01	66.51	16.42		130.0	
10624- AAB	IEEE 802.11ac WiFi (40MHz, MCS8, 90pc duty cycle)	Х	5.39	66.68	16.54	0.46	130.0	± 9.6 %
		Y	5.26	66.16	16.18		130.0	
		Z	5.20	66.75	16.60		130.0	
10625- AAB	IEEE 802.11ac WiFi (40MHz, MCS9, 90pc duty cycle)	X	5.68	67.45	16.97	0.46	130.0	± 9.6 %
		Y	5.49	66.74	16.52		130.0	
		Z	5.29	66.88	16.73		130.0	
10626- AAB	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	X	5.56	66.68	16.44	0.46	130.0	± 9.6 %
		Y	5.45	66.17	16.10		130.0	
		Z	5.41	66.70	16.50		130.0	
10627- AAB	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	X	5.79	67.23	16.68	0.46	130.0	± 9.6 %
		Y	5.70	66.81	16.39		130.0	
		Z	5.64	67.32	16.78		130.0	
10628- AAB	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	Х	5.58	66.74	16.38	0.46	130.0	± 9.6 %
		Y	5.45	66.19	16.01		130.0	
		Z	5.40	66.67	16.39		130.0	
10629- AAB	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc duty cycle)	X	5.65	66.80	16.40	0.46	130.0	± 9.6 %
		Y	5.54	66.31	16.07		130.0	
		Z	5.52	66.90	16.50		130.0	
10630- AAB	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc duty cycle)	X	6.01	68.09	17.04	0.46	130.0	± 9.6 %
		Y	5.91	67.63	16.73		130.0	
		Z	5.72	67.73	16.92		130.0	
10631- AAB	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	5.95	67.99	17.17	0.46	130.0	± 9.6 %
		Y	5.81	67.43	16.82		130.0	
		Z	5.71	67.83	17.16		130.0	
10632- AAB	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.76	67.30	16.85	0.46	130.0	± 9.6 %
		Y	5.67	66.90	16.57		130.0	
		Z	5.66	67.57	17.05		130.0	
10633- AAB	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	X	5.65	66.93	16.50	0.46	130.0	± 9.6 %
		Y	5.52	66.40	16.15		130.0	
		Z	5.42	66.75	16.46		130.0	
10634- AAB	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	X	5.63	66.95	16.56	0.46	130.0	± 9.6 %
		Y	5.50	66.40	16.20		130.0	
		Z	5.46	66.98	16.63		130.0	
10635- AAB	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	X	5.51	66.29	15.98	0.46	130.0	± 9.6 %
		Y	5.37	65.72	15.59		130.0	
		Z	5.31	66.16	15.94		130.0	
10636- AAC	IEEE 802.11ac WiFi (160MHz, MCS0, 90pc duty cycle)	X	5.98	67.03	16.52	0.46	130.0	± 9.6 %
		Y	5.88	66.56	16.21		130.0	
1000-	1	Z	5.85	67.02	16.56		130.0	
10637- AAC	IEEE 802.11ac WiFi (160MHz, MCS1, 90pc duty cycle)	X	6.12	67.40	16.69	0.46	130.0	± 9.6 %
		Y	6.03	66.94	16.39		130.0	
1000-		Z	5.96	67.31	16.70		130.0	
10638- AAC	IEEE 802.11ac WiFi (160MHz, MCS2, 90pc duty cycle)	X	6.13	67.39	16.66	0.46	130.0	± 9.6 %
				0001	1000			
		Y	6.03	66.91 67.42	16.35 16.73		130.0 130.0	

Certificate No: EX3-7350_Dec18

10639- AAC	IEEE 802.11ac WiFi (160MHz, MCS3, 90pc duty cycle)	Х	6.10	67.32	16.66	0.46	130.0	± 9.6 %
		Y	5.99	66.82	16.34		130.0	
		Z	5.94	67.26	16.69		130.0	
10640- AAC	IEEE 802.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	Х	6.10	67.33	16.62	0.46	130.0	± 9.6 %
0.0	5555 4547 575.57	Y	5.99	66.82	16.29		130.0	
		Z	5.89	67.11	16.55		130.0	
10641- AAC	IEEE 802.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	Х	6.15	67.26	16.60	0.46	130.0	± 9.6 %
		Y	6.06	66.83	16.31		130.0	
		Z	6.00	67.22	16.63		130.0	
10642- AAC	IEEE 802.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	Х	6.19	67.48	16.87	0.46	130.0	± 9.6 %
		Y	6.08	67.01	16.57		130.0	
		Z	6.02	67.42	16.90		130.0	
10643- AAC	IEEE 802.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	Х	6.03	67.19	16.63	0.46	130.0	± 9.6 %
		Y	5.93	66.72	16.32		130.0	
		Z	5.86	67.10	16.63		130.0	
10644- AAC	IEEE 802.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	Х	6.16	67.59	16.85	0.46	130.0	± 9.6 %
		Y	6.03	67.02	16.49		130.0	
		Z	5.92	67.27	16.74		130.0	
10645- AAC	IEEE 802.11ac WiFi (160MHz, MCS9, 90pc duty cycle)	Х	6.32	67.70	16.87	0.46	130.0	± 9.6 %
		Y	6.18	67.13	16.51		130.0	
		Z	6.04	67.32	16.72		130.0	
10646- AAF	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,7)	Х	27.21	126.02	43.98	9.30	60.0	± 9.6 %
7011	GI GI G G G G G G G G G G G G G G G G G	Y	9.45	98.48	34.90		60.0	
		Z	9.20	102.60	37.36		60.0	
10647- AAF	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	X	20.56	119.72	42.33	9.30	60.0	± 9.6 %
700	at on, or outside any	Y	8.20	95.60	34.03		60.0	
		Z	7.43	97.70	35.80		60.0	
10648- AAA	CDMA2000 (1x Advanced)	Х	0.74	65.23	11.62	0.00	150.0	± 9.6 %
, , , ,		Y	0.48	60.68	7.73		150.0	
		Z	0.53	63.13	8.96		150.0	
10652- AAD	LTE-TDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	Х	3.59	67.56	17.09	2.23	80.0	± 9.6 %
7010	Outputs 11707	Y	3.21	65.71	15.87		80.0	
		Z	3.43	68.17	17.07		80.0	
10653- AAD	LTE-TDD (OFDMA, 10 MHz, E-TM 3.1, Clipping 44%)	X	4.05	66.51	17.03	2.23	80.0	± 9.6 %
		Y	3.77	65.22	16.18		80.0	
		Z	3.85	66.64	16.97		80.0	
10654- AAD	LTE-TDD (OFDMA, 15 MHz, E-TM 3.1, Clipping 44%)	X	4.03	66.08	16.99	2.23	80.0	± 9.6 %
		Y	3.78	64.88	16.21		80.0	
		Z	3.85	66.10	16.93		80.0	
10655- AAE	LTE-TDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	X	4.09	66.03	17.02	2.23	80.0	± 9.6 %
		Y	3.84	64.84	16.25		80.0	
		Z	3.91	65.94	16.93		80.0	
10658- AAA	Pulse Waveform (200Hz, 10%)	Х	100.00	109.53	24.88	10.00	50.0	± 9.6 %
		Y	100.00	107.33	23.96		50.0	
		Z	40.45	94.77	19.99		50.0	
10659- AAA	Pulse Waveform (200Hz, 20%)	X	100.00	111.18	24.63	6.99	60.0	± 9.6 %
		Y	100.00	106.37	22.34		60.0	

Certificate No: EX3-7350_Dec18 Page 38 of 39

10660- AAA	Pulse Waveform (200Hz, 40%)	X	100.00	118.30	26.42	3.98	80.0	± 9.6 %
		Y	100.00	104.34	20.10		80.0	
		Z	100.00	107.03	21.02		80.0	
10661- Pulse Waveform (200Hz, 60%	Pulse Waveform (200Hz, 60%)	Х	100.00	133.09	31.12	2.22	100.0	± 9.6 %
		Y	100.00	95.96	15.60		100.0	
		Z	100.00	120.96	25.28		100.0	
10662- AAA	Pulse Waveform (200Hz, 80%)	Х	100.00	188.58	49.89	0.97	120.0	± 9.6 %
		Y	19.43	61.07	1.66		120.0	
		Z	99.98	60.00	307.71		120.0	
10670- AAA	Bluetooth Low Energy	X	100.00	136.51	33.03	2.19	100.0	± 9.6 %
		Y	100.00	107.99	20.68		100.0	
		Z	100.00	149.50	36.92		100.0	

^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7350_Dec18

2 2450 MHz Dipole

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client TuV Rheinland USA

Cortificate No. diSAPA24E0 2402102 Oct19

Accreditation No.: SCS 0108

Calibration procedure(s) Calibration procedure for dipole validation kits above 700 MHz Calibration procedure for dipole validation kits above 700 MHz Calibration procedure for dipole validation kits above 700 MHz Calibration date: October 16, 2018 This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration Power meter NRP Power meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Power sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02672) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02683) Apr-19 Type-N mismatch combination SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Type-N mismatch combination SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 8481A SN: GB37480704 07-Oct-15 (in house check Oct-18) In house check: Oct- Power sensor HP 8481A SN: MY41092317 07-Oct-15 (in house check Oct-18) In house check: Oct- Power sensor HP 8481A SN: MY41092317 07-Oct-15 (in house check Oct-18) In house check: Oct- RF generator R&S SMT-06	Object	diSARA2450 - S	N:2402103	
Calibration procedure for dipole validation kits above 700 MHz Calibration date: October 16, 2018 This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID #	•	4.5, 11 12 100	WE 102 100	
Calibration date: October 16, 2018 This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration) Power meter NRP Power sensor NRP-Z91 SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Power sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02682) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Power sensor HP 8481A SN: 100972 15-Jun-15 (in house check Oct-18) In house check: Oct Name Function Heads in house check: Oct In house check: Oct Patients of the certificate No. In house check: Oct Signature Patients of the certificate No. In house check: Oct-18 (In house check: Oct Patients of the certificate No. In house check: Oct Signature	Calibration procedure(s)	QA CAL-05.v10		
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration) Prower meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Power sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 5047.2 / 06327 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec-17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct-18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Orower sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Name Function Signature		Calibration proce	edure for dipole validation kits ab	ove 700 MHz
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration) Prower meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Power sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Vower sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Residence R				
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibrate Prower meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Prower sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Prower sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 Reference Probe EX3DV4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Recondary Standards ID # Check Date (in house) Scheduled Check Prower sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Residence Resid				
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibrat Power meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Sower sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter EPM-442A SN: GB37480704 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: W341092317 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US	Calibration date:	October 16, 201	8	
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibrat Power meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 Sower sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter EPM-442A SN: GB37480704 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: W341092317 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP Reference Ressor HP 8481A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Reference Ressor HP 8481A SN: US				
All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Primary Standards ID #	This calibration certificate docume	ents the traceability to nat	tional standards, which realize the physical ur	nits of measurements (SI).
D # Cal Date (Certificate No.) Scheduled Calibration	The measurements and the uncer	tainties with confidence p	probability are given on the following pages ar	nd are part of the certificate.
D # Cal Date (Certificate No.) Scheduled Calibration				
Cal Date (Certificate No.) Scheduled Calibrate	All calibrations have been conduct	ted in the closed laborato	ry facility: environment temperature (22 ± 3)°	C and humidity < 70%.
D# Cal Date (Certificate No.) Scheduled Calibrate				
Power meter NRP SN: 104778 04-Apr-18 (No. 217-02672/02673) Apr-19 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Apr-19 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Apr-19 SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Apr-19 SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Apr-19 SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Apr-19 SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Apr-19 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Dec-18 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards SN: 601 04-Oct-18 (No. DAE4-601_Oct18) In house check: Oct-19 Secondary Standards SN: GB37480704 07-Oct-15 (in house check Oct-18) In house check: Oct-20 Oct-20 SN: MY41092317 07-Oct-15 (in house check Oct-18) In house check: Oct-20 SN: 100972 15-Jun-15 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct-20 SN: US41080477 31-Mar-14 (in house check Oct-18) In house check:	Calibration Equipment used (M&T	E critical for calibration)		
SN: 104778	Primary Standards	ID#	Cal Date (Certificate No.)	Scheduled Calibration
Power sensor NRP-Z91 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-Z91 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Attenuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Type-N mismatch combination SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Reference Probe EX3DV4 SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18 DAE4 SN: 601 04-Oct-18 (No. DAE4-601_Oct18) Oct-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter EPM-442A SN: GB37480704 07-Oct-15 (in house check Oct-18) In house check: Oct Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oct RF generator R&S SMT-06 SN: 100972 15-Jun-15 (in house check Oct-18) In house check: Oct Network Analyzer Agilent E8358A Name Function Signature Tabletate Name Function Signature	Power meter NRP	SN: 104778		
SN: 103245 O4-Apr-18 (No. 217-02673) Apr-19				1.401
SN: 5058 (20k) O4-Apr-18 (No. 217-02682) Apr-19		SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
SN: 5047.2 / 06327	Power sensor NRP-Z91	1 (1000)		
SN: 7349 30-Dec-17 (No. EX3-7349_Dec17) Dec-18	Power sensor NRP-Z91 Power sensor NRP-Z91	SN: 103245	04-Apr-18 (No. 217-02673)	Apr-19
D# Check Date (in house) Scheduled Check	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	SN: 103245 SN: 5058 (20k)	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682)	Apr-19 Apr-19
Power meter EPM-442A SN: GB37480704 O7-Oct-15 (in house check Oct-18) In house check: Oc Ower sensor HP 8481A SN: US37292783 O7-Oct-15 (in house check Oct-18) In house check: Oc Ower sensor HP 8481A SN: MY41092317 O7-Oct-15 (in house check Oct-18) In house check: Oc In house check: Oc In house check: Oct Notwork Analyzer Agilent E8358A SN: US41080477 SN: US4108047	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683)	Apr-19 Apr-19 Apr-19
Power sensor HP 8481A SN: US37292783 07-Oct-15 (in house check Oct-18) In house check: Oc SN: MY41092317 07-Oct-15 (in house check Oct-18) In house check: Oc In house check: Oc Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct In house	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17)	Apr-19 Apr-19 Apr-19 Dec-18
Power sensor HP 8481A SN: MY41092317 07-Oct-15 (in house check: Oct-18) In house check: Oct- RF generator R&S SMT-06 SN: 100972 15-Jun-15 (in house check Oct-18) In house check: Oct- Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-18) In house check: Oct- Name Function Signature	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19
RF generator R&S SMT-06 SN: 100972 15-Jun-15 (in house check Oct-18) In house check: Oct Network Analyzer Agilent E8358A SN: US41080477 S1-Mar-14 (in house check Oct-18) Name Function Signature	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check
Name Function Signature Name Function Signature	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20
Name Function Signature	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
Calibrated by: Michael Woher Loberston Tachmining	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
Calibrated by: Michael Weber Laboratory Technician	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Rype-N mismatch combination Reference Probe EX3DV4 DAE4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A RF generator R&S SMT-06	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19
17.11076	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A RF generator R&S SMT-06	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 17-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20
	Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Power sensor NRP-Z91 Reference Probe EX3DV4 PAE4 Recondary Standards Reconda	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19 Signature
Approved by: Katja Pokovic Technical Manager	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18)	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19 Signature
EX 43	Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 DAE4 Secondary Standards Power meter EPM-442A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB37480704 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Michael Weber	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 30-Dec-17 (No. EX3-7349_Dec17) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18) Function Laboratory Technician	Apr-19 Apr-19 Apr-19 Dec-18 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-19 Signature

Certificate No: diSARA2450-2402103_Oct18

Page 1 of 8

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
C Service suisse d'étalonnage
Servizio svizzero di taratura
S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA

Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL

tissue simulating liquid

ConvF N/A sensitivity in TSL / NORM x,y,z not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point.
 No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: diSARA2450-2402103_Oct18

Page 2 of 8

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy , $dz = 5 mm$	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.8 ± 6 %	1.85 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.9 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	54.4 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	6.31 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.9 W/kg ± 16.5 % (k=2)

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	51.2 ± 6 %	2.01 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.0 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	50.9 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.93 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	23.4 W/kg ± 16.5 % (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	$50.3 \Omega + 4.4 j\Omega$	
Return Loss	- 27.2 dB	

Antenna Parameters with Body TSL

Impedance, transformed to feed point	47.4 Ω + 1.9 jΩ	
Return Loss	- 29.6 dB	

General Antenna Parameters and Design

Electrical Delay (one direction)	1.044 ns
----------------------------------	----------

Dipole designed and manufactured by IMST. Please see details on http://www.imst.com

Additional EUT Data

Manufactured by	IMST
Manufactured on	Unknown

Certificate No: diSARA2450-2402103_Oct18 Page 4 of 8

DASY5 Validation Report for Head TSL

Date: 16.10.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: IMST Dipole; Serial: diSARA2450 - SN:2402103

Communication System: UID 0 - CW; Frequency: 2450 MHz

Medium parameters used: f = 2450 MHz; $\sigma = 1.85 \text{ S/m}$; $\varepsilon_r = 37.8$; $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

Probe: EX3DV4 - SN7349; ConvF(7.88, 7.88, 7.88) @ 2450 MHz; Calibrated: 30.12.2017

• Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Electronics: DAE4 Sn601; Calibrated: 04.10.2018

Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001

• DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 119.9 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 28.1 W/kg


SAR(1 g) = 13.9 W/kg; SAR(10 g) = 6.31 W/kg

Maximum value of SAR (measured) = 23.1 W/kg



0 dB = 23.1 W/kg = 13.64 dBW/kg

Impedance Measurement Plot for Head TSL



DASY5 Validation Report for Body TSL

Date: 16.10.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: IMST Dipole; Serial: diSARA2450 - SN:2402103

Communication System: UID 0 - CW; Frequency: 2450 MHz

Medium parameters used: f = 2450 MHz; $\sigma = 2.01$ S/m; $\varepsilon_r = 51.2$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.01, 8.01, 8.01) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.10.2018
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 108.2 V/m; Power Drift = -0.03 dB

Peak SAR (extrapolated) = 26.4 W/kg

SAR(1 g) = 13 W/kg; SAR(10 g) = 5.93 W/kg

Maximum value of SAR (measured) = 21.1 W/kg



0 dB = 21.1 W/kg = 13.24 dBW/kg

Impedance Measurement Plot for Body TSL

