



174, Gajeong-ro, Yuseong-gu, Daejeon, KOREA
 305-350
SELVAS Healthcare, Inc.
Tel No : +82-42-864-4460
Fax No : +82-42-864-4462

Date: June 13, 2017

Dear Whom it may concern,

The model number H532B uses the following power setting on the wireless LAN and Bluetooth.

And also current software version in this model operates under the following power settings.

IC Name	Wireless	Mode	Frequency(MHz)	Result(dBm)
EMMY-W163	2.4GHz WLAN	b mode	2412	8.77
			2437	8.63
			2462	8.24
		g mode	2412	8.68
			2437	8.13
			2462	8.54
		n20 mode	2412	8.66
			2437	8.28
			2462	8.45
		n40 mode	2422	8.13
			2437	7.85
			2452	7.96
	5GHz WLAN	ac20 mode Band 1	5180	-0.29
			5200	-0.10
			5240	-0.61
		ac20 mode Band 4	5745	-1.18
			5785	-0.84
			5825	0.41
	Bluetooth	BLE	2402	7.63
			2440	7.76
			2480	7.52
		DH5	2402	0.27
			2441	0.60
			2480	0.77
		2DH5	2402	-1.87
			2441	-1.88
			2480	-2.20
		3DH5	2402	-1.89
			2441	-1.90
			2480	-2.22

HBG2X3N	Bluetooth	DH5	2402	2.91
			2441	2.07
			2480	1.72
		2DH5	2402	-2.27
			2441	-3.37
			2480	-3.87
		3DH5	2402	-2.28
			2441	-3.37
			2480	-3.89

SELVAS Healthcare, Inc.

