

Test Report No.50078182 001

Appendix D: Radiated and Mains Spurious Emission Data

(File: 50078182AppendixD)



Contents

| | |
|---|----|
| Spurious Emissions, Band Edges, 2.35-2.5G | 2 |
| Spurious Emissions, TX Mode, 1-18G..... | 6 |
| Spurious Emissions, TX Mode, 18-26G..... | 12 |
| Spurious Emissions, TX Mode, 30M-1G | 18 |
| Spurious Emissions, TX Mode, 9K-30M | 20 |
| Spurious Emissions, RX Mode, 1-18G | 22 |
| Spurious Emissions, RX Mode, 18-26G | 24 |
| Spurious Emissions, RX Mode, 30M-1G..... | 26 |
| Spurious Emissions, RX Mode, 9K-30M..... | 28 |

Spurious Emissions, Band Edges, 2.35-2.5G



Radiated Bandedge (GFSK)

Low Channel (Hor)



Spurious Emissions, Band Edge, 2.35-2.5G



Low Channel (Ver)



Spurious Emissions, Band Edge, 2.35-2.5G



Radiated Bandedge (GFSK)

High Channel (Hor)



Spurious Emissions, Band Edge, 2.35-2.5G

High Channel (Ver)



Spurious Emissions, Band Edge, 2.35-2.5G


Spurious Emissions, TX Mode, 1-18G



TUV Taiwan


11F., No.758, Sec.4 Bade Road, Songshan Dist, Taipei City 105

Tel:+886-2172-7000 fax:+886-2528-0018




| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:36:02 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2402-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 1969.000 | -7.35 | 51.79 | 44.44 | 74.00 | -29.56 | peak | 100 | 210 | |
| 2 | 4804.000 | -2.66 | 48.86 | 46.20 | 74.00 | -27.80 | peak | | | |
| 3 | 7206.000 | 5.58 | 45.62 | 51.20 | 74.00 | -22.80 | peak | | | |




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:37:04 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2402-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1969.000 | -7.35 | 51.68 | 44.33 | 74.00 | -29.67 | peak | 100 | 70 | |
| 2 | 4804.000 | -2.66 | 48.65 | 45.99 | 74.00 | -28.01 | peak | | | |
| 3 | 7206.000 | 5.58 | 45.88 | 51.46 | 74.00 | -22.54 | peak | | | |




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:31:53 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1986.000 | -7.23 | 52.57 | 45.34 | 74.00 | -28.66 | peak | 100 | 360 | |
| 2 | 4880.000 | -2.48 | 50.36 | 47.88 | 74.00 | -26.12 | peak | | | |
| 3 | 7320.000 | 5.93 | 45.15 | 51.08 | 74.00 | -22.92 | peak | | | |




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:32:55 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | x | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1986.000 | -7.23 | 50.86 | 43.63 | 74.00 | -30.37 | peak | 100 | 358 | |
| 2 | 4880.000 | -2.48 | 49.55 | 47.07 | 74.00 | -26.93 | peak | | | |
| 3 | 7320.000 | 5.93 | 44.71 | 50.64 | 74.00 | -23.36 | peak | | | |



| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:27:59 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2480-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1748.000 | -8.75 | 56.01 | 47.26 | 74.00 | -26.74 | peak | 100 | 207 | |
| 2 | 4960.000 | -2.31 | 48.65 | 46.34 | 74.00 | -27.66 | peak | | 0 | |
| 3 | 7440.000 | 6.28 | 45.38 | 51.66 | 74.00 | -22.34 | peak | | 0 | |



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:29:01 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2480-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1765.000 | -8.64 | 53.22 | 44.58 | 74.00 | -29.42 | peak | 100 | 359 | |
| 2 | 4960.000 | -2.31 | 49.44 | 47.13 | 74.00 | -26.87 | peak | | | |
| 3 | 7440.000 | 6.28 | 45.24 | 51.52 | 74.00 | -22.48 | peak | | | |


Spurious Emissions, TX Mode, 18-26G



TUV Taiwan


11F., No.758, Sec.4 Bade Road, Songshan Dist, Taipei City 105

Tel:+886-2172-7000 fax:+886-2528-0018




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:52:04 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2402-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 18986.000 | 30.47 | 19.08 | 49.55 | 74.00 | -24.45 | peak | | | |




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:52:23 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2402-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 18765.000 | 29.97 | 19.53 | 49.50 | 74.00 | -24.50 | peak | | | |




| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:52:47 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 18323.000 | 29.13 | 20.60 | 49.73 | 74.00 | -24.27 | peak | | | |




| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:53:04 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 19785.000 | 31.36 | 16.40 | 47.76 | 74.00 | -26.24 | peak | | | |



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:53:27 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2480-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 19249.500 | 30.85 | 18.47 | 49.32 | 74.00 | -24.68 | peak | | | |



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:53:44 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2480-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 19955.000 | 31.45 | 17.86 | 49.31 | 74.00 | -24.69 | peak | | | |


Spurious Emissions, TX Mode, 30M-1G



TUV Taiwan

11F., No.758, Sec.4 Bade Road. Songshan Dist, Taipei City 105


Tel:+886-2172-7000 fax:+886-2528-0018



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Class B 3M Radiation | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 21:04:25 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%) | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 31.9400 | -1.18 | 23.87 | 22.69 | 40.00 | -17.31 | QP | 100 | 246 | |
| 2 | 131.8499 | -6.97 | 25.91 | 18.94 | 43.50 | -24.56 | QP | 100 | 136 | |
| 3 | 370.4700 | -3.13 | 26.38 | 23.25 | 46.00 | -22.75 | QP | 100 | 318 | |
| 4 | 872.9299 | 3.90 | 26.77 | 30.67 | 46.00 | -15.33 | QP | 100 | 271 | |
| 5 | 958.2899 | 5.45 | 26.62 | 32.07 | 46.00 | -13.93 | QP | 100 | 91 | |
| 6 | 973.8099 | 5.60 | 25.79 | 31.39 | 54.00 | -22.61 | QP | 100 | 86 | |

Spurious Emissions, TX Mode, 30M-1G



| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Class B 3M Radiation | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 21:05:28 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-TX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 30.0000 | -0.07 | 25.29 | 25.22 | 40.00 | -14.78 | QP | 100 | 0 | |
| 2 | 44.5500 | -8.40 | 31.94 | 23.54 | 40.00 | -16.46 | QP | 100 | 360 | |
| 3 | 105.6600 | -8.75 | 34.11 | 25.36 | 43.50 | -18.14 | QP | 100 | 210 | |
| 4 | 570.2900 | -0.46 | 27.03 | 26.57 | 46.00 | -19.43 | QP | 100 | 1 | |
| 5 | 896.2100 | 4.12 | 32.87 | 36.99 | 46.00 | -9.01 | QP | 100 | 319 | |
| 6 | 988.3600 | 5.73 | 25.83 | 31.56 | 54.00 | -22.44 | QP | 100 | 277 | |


Spurious Emissions, TX Mode, 9K-30M



TUV Taiwan

11F., No.758, Sec.4 Bade Road, Songshan Dist, Taipei City 105


Tel: +886-2172-7000 fax: +886-2528-0018



| | | | |
|----------------|---|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC15.209_9k-1G_3m | Ant. Polarization: | |
| Test item: | Radiation Emission | Test Time: | 2017/4/10 15:14:06 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%) | 21.5(°C)/48% |
| Model No.: | HB-8301 | Test Engineer: | George Yang |
| Test Mode: | 2440-TX | | |
| Remark: | | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 0.0096 | 20.01 | 34.02 | 54.03 | 127.94 | -73.91 | QP | 100 | 208 | |
| 2 | 0.0483 | 19.31 | 26.24 | 45.55 | 113.91 | -68.36 | QP | 100 | 320 | |
| 3 | 0.0952 | 18.88 | 34.40 | 53.28 | 108.02 | -54.74 | QP | 100 | 224 | |
| 4 | 0.0964 | 18.87 | 34.78 | 53.65 | 107.91 | -54.26 | QP | 100 | 85 | |
| 5 | 0.1430 | 18.81 | 23.38 | 42.19 | 104.49 | -62.30 | QP | 100 | 357 | |
| 6 | 0.1446 | 18.81 | 23.51 | 42.32 | 104.39 | -62.07 | QP | 100 | 80 | |

Spurious Emissions, TX Mode, 30M-1G



| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC15.209_9k-1G_3m | Ant. Polarization: | |
| Test item: | Radiation Emission | Test Time: | 2017/4/10 15:10:46 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 21.5(°C)/48% |
| Model No.: | HB-8301 | Test Engineer: | George Yang |
| Test Mode: | 2440-TX | | |
| Remark: | | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 0.1799 | 18.77 | 27.63 | 46.40 | 102.50 | -56.10 | QP | 100 | 70 | |
| 2 | 4.5678 | 19.44 | 20.51 | 39.95 | 69.50 | -29.55 | QP | 100 | 103 | |
| 3 | 10.2990 | 20.42 | 20.51 | 40.93 | 69.50 | -28.57 | QP | 100 | 200 | |
| 4 | 14.5377 | 20.92 | 19.97 | 40.89 | 69.50 | -28.61 | QP | 100 | 186 | |
| 5 | 20.0898 | 21.56 | 20.77 | 42.33 | 69.50 | -27.17 | QP | 100 | 313 | |
| 6 | 27.2837 | 22.08 | 21.33 | 43.41 | 69.50 | -26.09 | QP | 100 | 84 | |


Spurious Emissions, RX Mode, 1-18G



TUV Taiwan

11F., No.758, Sec.4 Bade Road. Songshan Dist, Taipei City 105


Tel:+886-2172-7000 fax:+886-2528-0018



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:43:44 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%) | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 1850.000 | -9.06 | 53.01 | 43.95 | 74.00 | -30.05 | peak | 100 | 1 | |
| 2 | 2785.000 | -7.47 | 51.25 | 43.78 | 74.00 | -30.22 | peak | 100 | 255 | |
| 3 | 3567.000 | -6.45 | 51.62 | 45.17 | 74.00 | -28.83 | peak | 100 | 52 | |

Spurious Emissions, RX Mode, 1-18G



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:44:46 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | x | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 2122.000 | -8.06 | 52.32 | 44.26 | 74.00 | -29.74 | peak | 100 | 29 | |
| 2 | 3295.000 | -6.81 | 51.38 | 44.57 | 74.00 | -29.43 | peak | 100 | 39 | |
| 3 | 4570.000 | -3.72 | 50.41 | 46.69 | 74.00 | -27.31 | peak | 100 | 31 | |


Spurious Emissions, RX Mode, 18-26G



TUV Taiwan


11F., No.758, Sec.4 Bade Road. Songshan Dist, Taipei City 105

Tel:+886-2172-7000 fax:+886-2528-0018



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:51:24 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 18782.000 | 30.01 | 20.10 | 50.11 | 74.00 | -23.89 | peak | | | |



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Above 1G PEAK | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 20:51:38 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 18153.000 | 28.90 | 20.66 | 49.56 | 74.00 | -24.44 | peak | | | |


Spurious Emissions, RX Mode, 30M-1G



TUV Taiwan

11F., No.758, Sec.4 Bade Road, Songshan Dist, Taipei City 105


Tel:+886-2172-7000 fax:+886-2528-0018



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Class B 3M Radiation | Ant. Polarization: | Horizontal |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 21:07:45 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 30.0000 | -0.07 | 23.29 | 23.22 | 40.00 | -16.78 | QP | 100 | 274 | |
| 2 | 133.7899 | -6.97 | 25.45 | 18.48 | 43.50 | -25.02 | QP | 100 | 94 | |
| 3 | 428.6700 | -2.28 | 26.62 | 24.34 | 46.00 | -21.66 | QP | 100 | 85 | |
| 4 | 844.7999 | 3.60 | 26.87 | 30.47 | 46.00 | -15.53 | QP | 100 | 5 | |
| 5 | 963.1399 | 5.50 | 26.73 | 32.23 | 54.00 | -21.77 | QP | 100 | 0 | |
| 6 | 980.6000 | 5.65 | 26.79 | 32.44 | 54.00 | -21.56 | QP | 100 | 246 | |

Spurious Emissions, RX Mode, 30M-1G



| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC Class B 3M Radiation | Ant. Polarization: | Vertical |
| Test item: | Radiation Emission | Test Time: | 2017/2/15 21:08:48 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 19.3(°C)/50% |
| Model No.: | HB-8301 | Test Engineer: | Freeman Wang |
| Test Mode: | 2440-RX | | |
| Remark: | X | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 30.0000 | -0.07 | 25.05 | 24.98 | 40.00 | -15.02 | QP | 100 | 23 | |
| 2 | 44.5500 | -8.40 | 31.49 | 23.09 | 40.00 | -16.91 | QP | 100 | 6 | |
| 3 | 105.6600 | -8.75 | 29.60 | 20.85 | 43.50 | -22.65 | QP | 100 | 244 | |
| 4 | 593.5700 | -0.09 | 27.47 | 27.38 | 46.00 | -18.62 | QP | 100 | 316 | |
| 5 | 875.8400 | 3.93 | 26.49 | 30.42 | 46.00 | -15.58 | QP | 100 | 73 | |
| 6 | 973.8100 | 5.60 | 26.52 | 32.12 | 54.00 | -21.88 | QP | 100 | 31 | |


Spurious Emissions, RX Mode, 9K-30M



TUV Taiwan

11F., No.758, Sec.4 Bade Road. Songshan Dist, Taipei City 105


Tel:+886-2172-7000 fax:+886-2528-0018



| | | | |
|----------------|--|--------------------|--------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC15.209_9k-1G_3m | Ant. Polarization: | |
| Test item: | Radiation Emission | Test Time: | 2017/4/10 15:15:31 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%) | 21.5(°C)/48% |
| Model No.: | HB-8301 | Test Engineer: | George Yang |
| Test Mode: | 2440-RX | | |
| Remark: | | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|-------------|--------|
| 1 | 0.0097 | 20.09 | 34.19 | 54.28 | 127.85 | -73.57 | QP | 100 | 259 | |
| 2 | 0.0451 | 19.47 | 26.31 | 45.78 | 114.51 | -68.73 | QP | 100 | 359 | |
| 3 | 0.0485 | 19.30 | 26.21 | 45.51 | 113.88 | -68.37 | QP | 100 | 104 | |
| 4 | 0.0954 | 18.88 | 33.70 | 52.58 | 108.01 | -55.43 | QP | 100 | 83 | |
| 5 | 0.0968 | 18.87 | 34.52 | 53.39 | 107.88 | -54.49 | QP | 100 | 2 | |
| 6 | 0.1453 | 18.80 | 24.57 | 43.37 | 104.35 | -60.98 | QP | 100 | 38 | |

Spurious Emissions, RX Mode, 30M-1G



| | | | |
|-----------------------|---|---------------------------|---------------------------|
| Service No.: | 114058234 | Test Distance: | 3m |
| Test Standard: | FCC15.209_9k-1G_3m | Ant. Polarization: | |
| Test item: | Radiation Emission | Test Time: | 2017/4/10 15:12:09 |
| Applicant: | H2 | Test Rating: | DC 3V |
| Product: | 藍牙血糖數據傳輸配件/ Health2Sync Bluetooth Dongle | Temp.(°C)/Hum.(%): | 21.5(°C)/48% |
| Model No.: | HB-8301 | Test Engineer: | George Yang |
| Test Mode: | 2440-RX | | |
| Remark: | | | |

| No. | Frequency (MHz) | Factor (dB/m) | Reading (dBuV) | Level (dBuV/m) | Limit (dBuV/m) | Margin (dB) | Det. | Height (cm) | Azimuth (°) | Remark |
|-----|-----------------|---------------|----------------|----------------|----------------|-------------|------|-------------|---------------|--------|
| 1 | 0.1799 | 18.77 | 27.51 | 46.28 | 102.50 | -56.22 | QP | 100 | 72 | |
| 2 | 4.5380 | 19.43 | 21.23 | 40.66 | 69.50 | -28.84 | QP | 100 | 3 | |
| 3 | 8.6573 | 20.18 | 20.06 | 40.24 | 69.50 | -29.26 | QP | 100 | 238 | |
| 4 | 12.5079 | 20.69 | 21.11 | 41.80 | 69.50 | -27.70 | QP | 100 | 224 | |
| 5 | 21.5226 | 21.66 | 21.74 | 43.40 | 69.50 | -26.10 | QP | 100 | 263 | |
| 6 | 28.8060 | 22.19 | 22.08 | 44.27 | 69.50 | -25.23 | QP | 100 | 302 | |