

Additional Conversion Factors

for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1377

Place of Assessment:

Zurich

Date of Assessment:


October 29, 2004

Probe Calibration Date:

September 29, 2004

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1377Conversion factor (\pm standard deviation)**900 MHz** ConvF **$5.9 \pm 8\%$**

$\alpha = 55.0 \pm 5\%$
 $\sigma = 1.05 \pm 5\%$ mho/m
(body tissue)

1800 MHz ConvF **$4.7 \pm 8\%$**

$\alpha = 53.3 \pm 5\%$
 $\sigma = 1.52 \pm 5\%$ mho/m
(body tissue)

Important Note:

For numerically assessed probe conversion factors, parameters Alpha and Delta in the DASY software must have the following entries: Alpha = 0 and Delta = 1.

Please see also Section 4.7 of the DASY4 Manual.