

TRaC Global
100 Frobisher Business Park
Leigh Sinton Road
Malvern
Worcestershire
WR14 1BX
UK

Widex A/S
Nymoellevej 6
DK-3540 Lyngø
Denmark

Tel.: (+45) 44 35 56 00
Fax: (+45) 44 35 56 01
widex@widex.com
<http://www.widex.com>
CVR. Nr. 1577 1100

Date: 7 February, 2012

FCC Labeling of Widex model D-FS

FCC ID: TTY-DFS

To Whom It May Concern:

The dimensions of the hearing aid are too small to allow the FCC ID to be printed on the device in a font size that is readable. Consequently, the FCC ID is printed in the user instructions together with the FCC regulatory statements.

Justification is provided by 3 photographs below that show the D-FS in 3 different angles together with a printout of the FCC ID in font size 4 point that is the minimum readable font size as defined in FCC OET Labelling Guidelines.

Bottom view



Side view



Top view



Thank you for your attention to this matter.

Sincerely,
Widex A/S

Hans-Otto Bindeballe