APPENDIX A: SAR TEST DATA

DUT: CASIO-HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902 -0751 Conducted Power: 24.5 dBm

Communication System: Cellular CDMA; Frequency: 848.31 MHz;Duty Cycle: 1:1 Medium: 835 Brain (σ = 0.87 mho/m, $\varepsilon_{\rm r}$ = 40.37, ρ = 1000 kg/m³) Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN3589; ConvF(8.36, 8.36, 8.36); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006 Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: CDMA PTT, Face SAR, Flip open, High.ch, Standard Battery


Area Scan (7x15x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.06 V/m

Peak SAR (extrapolated) = 0.086 W/kg

SAR(1 g) = 0.063 mW/g; SAR(10 g) = 0.045 mW/g



DUT: CASIO-HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902 -0751 Conducted Power: 24.5 dBm

Communication System: Cellular CDMA; Frequency: 848.31 MHz;Duty Cycle: 1:1 Medium: 835 Brain (σ = 0.87 mho/m, $\varepsilon_{\rm r}$ = 40.37, ρ = 1000 kg/m³) Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN3589; ConvF(8.36, 8.36, 8.36); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006 Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: CDMA PTT, Face SAR, Flip close, High ch, Standard Battery


Area Scan (7x11x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.49 V/m

Peak SAR (extrapolated) = 0.060 W/kg

SAR(1 g) = 0.044 mW/g; SAR(10 g) = 0.031 mW/g



DUT: CASIO-HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902 -0751 Conducted Power: 24.5 dBm

Communication System: PCS CDMA; Frequency: 1880 MHz; Duty Cycle: 1:1 Medium: 1900 Brain (σ = 1.46 mho/m, ϵ_r = 38.20, ρ = 1000 kg/m³) Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.2°C

Probe: EX3DV4 - SN3589; ConvF(7.11, 7.11, 7.11); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006 Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS PTT, Face SAR, Flip open, Mid.ch, Standard Battery


Area Scan (7x15x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.06 V/m

Peak SAR (extrapolated) = 0.776 W/kg

SAR(1 g) = 0.494 mW/g; SAR(10 g) = 0.306 mW/g



DUT: CASIO-HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902 -0751 Conducted Power: 24.5 dBm

Communication System: PCS CDMA; Frequency: 1880 MHz; Duty Cycle: 1:1 Medium: 1900 Brain (σ = 1.46 mho/m, ε_r = 38.20, ρ = 1000 kg/m³) Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.2°C

Probe: EX3DV4 - SN3589; ConvF(7.11, 7.11, 7.11); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006 Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS PTT, Face SAR, Flip close, Mid ch, Standard Battery


Area Scan (7x11x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 14.7 V/m

Peak SAR (extrapolated) = 0.774 W/kg

SAR(1 g) = 0.48 mW/g; SAR(10 g) = 0.287 mW/g



DUT: CASIO HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902-0751 Conducted Power: 24.5 dBm

Communication System: Cellular CDMA; Frequency: 848.31 MHz; Duty Cycle: 1:1

Medium: 835 Brain ($\sigma = 0.87 \text{ mho/m}, \, \varepsilon_r = 40.37, \, \rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN3589; ConvF(8.36, 8.36, 8.36); Calibrated: 7/14/2006

Sensor-Surface: 3mm (Mechanical Surface Detection)
Electronics: DAE4 Sn704; Calibrated: 6/1/2006

Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: CDMA PTT, Face SAR, Flip open, High.ch, Standard Battery


Area Scan (7x15x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 8.06 V/m

Peak SAR (extrapolated) = 0.086 W/kg

SAR(1 g) = 0.063 mW/g; SAR(10 g) = 0.045 mW/g



DUT: CASIO-HITACHI Gzone; Type: CDMA Dual Band Phone w/ PTT Mode; SN: 0902 -0751 Conducted Power: 24.5 dBm

Communication System: PCS CDMA; Frequency: 1880 MHz;Duty Cycle: 1:1 Medium: 1900 Brain (σ = 1.46 mho/m, ϵ_r = 38.20, ρ = 1000 kg/m³) Phantom section: Flat Section; Space: 2.5 cm from DUT to Flat Phantom

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN3589; ConvF(7.11, 7.11, 7.11); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006

Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

Mode: PCS PTT, Face SAR, Flip open, Mid.ch, Standard Battery


Area Scan (7x15x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Reference Value = 7.06 V/m

Peak SAR (extrapolated) = 0.776 W/kg

SAR(1 g) = 0.494 mW/g; SAR(10 g) = 0.306 mW/g



APPENDIX B: DIPOLE VALIDATION

DUT: Dipole 835 MHz; Type: D835V2; Serial: 4d026

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium: 835 Brain (σ = 0.87 mho/m, ϵ_r = 40.37, ρ = 1000 kg/m³)

Phantom section: Flat Section

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.6°C

Probe: EX3DV4 - SN3589; ConvF(8.36, 8.36, 8.36); Calibrated: 7/14/2006


Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006 Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

835MHz Dipole Validation

Area Scan (7x13x1): Measurement grid: dx=15mm, dy=15mm **Zoom Scan (5x5x7)/Cube 0:** Measurement grid: dx=8mm, dy=8mm, dz=5mm

Input Power = 24.0 dBm (250 mW) **SAR(1 g) = 2.39 mW/g; SAR(10 g) = 1.63 mW/g**Target SAR(1g) = 2.25 mW/g; Deviation = + 6.1 %



DUT: Dipole 1900 MHz; Type: D1900V2; Serial: 5d080

Communication System: CW; Frequency: 1900 MHz; Duty Cycle: 1:1 Medium: 1900 Brain ($\sigma = 1.46 \text{ mho/m}$, $\varepsilon_r = 38.20$, $\rho = 1000 \text{ kg/m}^3$)

Phantom section: Flat Section

Test Date: 03-23-2007; Ambient Temp: 23.8°C; Tissue Temp: 21.2°C

Probe: EX3DV4 - SN3589; ConvF(7.11, 7.11, 7.11); Calibrated: 7/14/2006 Sensor-Surface: 3mm (Mechanical Surface Detection) Electronics: DAE4 Sn704; Calibrated: 6/1/2006

Phantom: SAM Main; Type: SAM 4.0; Serial: TP:1197

Measurement SW: DASY4, V4.7 Build 53; Postprocessing SW: SEMCAD, V1.8 Build 172

1900MHz Dipole Validation


Area Scan (5x7x1): Measurement grid: dx=15mm, dy=15mm

Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=8mm, dy=8mm, dz=5mm

Input Power = 20.0 dBm (100 mW)

SAR(1 g) = 3.87 mW/g; SAR(10 g) = 2.03 mW/g

Target SAR(1g) = 3.77 mW/g; Deviation = +2.65 %



APPENDIX C: PROBE CALIBRATION

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst
Service suisse d'étalonnage
Servizio svizzero di taratura
S swiss Calibration Service

Accredited by the Swiss Federal Office of Metrology and Accreditation
The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 108

Client

PC Test

Certificate No: EX3-3589_Jul06

CALIBRATION CERTIFICATE

Object EX3DV4 - SN:3589

Calibration procedure(s) QA CAL-01.v5 and QA CAL-14.v3

Calibration procedure for dosimetric E-field probes

Calibration date: July 14, 2006

Condition of the calibrated item In Tolerance

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID#	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power meter E4419B	GB41293874	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Power sensor E4412A	MY41495277	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Power sensor E4412A	MY41498087	5-Apr-06 (METAS, No. 251-00557)	Apr-07
Reference 3 dB Attenuator	SN: S5054 (3c)	11-Aug-05 (METAS, No. 251-00499)	Aug-06
Reference 20 dB Attenuator	SN: S5086 (20b)	4-Apr-06 (METAS, No. 251-00558)	Apr-07
Reference 30 dB Attenuator	SN: S5129 (30b)	11-Aug-05 (METAS, No. 251-00500)	Aug-06
Reference Probe ES3DV2	SN: 3013	2-Jan-06 (SPEAG, No. ES3-3013_Jan06)	Jan-07
DAE4	SN: 654	21-Jun-06 (SPEAG, No. DAE4-654_Jun06)	Jun-07
Secondary Standards	ID#	Check Date (in house)	Scheduled Check
RF generator HP 8648C	US3642U01700	4-Aug-99 (SPEAG, in house check Nov-05)	In house check: Nov-07
Network Analyzer HP 8753E	US37390585	18-Oct-01 (SPEAG, in house check Nov-05)	In house check: Nov 06
	Name	Function	Signature
Calibrated by:	Katja Pokovic	Technical Manager	Hour Hade
			111
Approved by:	Niels Kuster	Quality Manager	1/1/
			1 WM

Issued: July 17, 2006

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: EX3-3589_Jul06

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst S

Service suisse d'étalonnage C Servizio svizzero di taratura

Swiss Calibration Service

Accreditation No.: SCS 108 Accredited by the Swiss Federal Office of Metrology and Accreditation The Swiss Accreditation Service is one of the signatories to the EA

Glossary:

TSL tissue simulating liquid sensitivity in free space NORMx,y,z

sensitivity in TSL / NORMx,y,z ConF

DCP diode compression point φ rotation around probe axis

Multilateral Agreement for the recognition of calibration certificates

9 rotation around an axis that is in the plane normal to probe axis (at Polarization 9

measurement center), i.e., $\vartheta = 0$ is normal to probe axis

Calibration is Performed According to the Following Standards:

a) IEEE Std 1528-2003, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", December 2003

b) CENELEC EN 50361, "Basic standard for the measurement of Specific Absorption Rate related to human exposure to electromagnetic fields from mobile phones (300 MHz - 3 GHz), July 2001

Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization $\vartheta = 0$ (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx, y, z does not effect the E^2 -field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx.v.z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.

Probe EX3DV4

SN:3589

Manufactured: March 30, 2006 Calibrated: July 14, 2006

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: EX3-3589_Jul06 Page 3 of 9

DASY - Parameters of Probe: EX3DV4 SN:3589

Sensitivity in Free Space^A

Diode Compression^B

NormX	0.460 ± 10.1%	$\mu V/(V/m)^2$	DCP X	90 mV
NormY	0.400 ± 10.1%	μ V/(V/m) ²	DCP Y	90 mV
NormZ	0.370 ± 10.1%	$\mu V/(V/m)^2$	DCP Z	90 mV

Sensitivity in Tissue Simulating Liquid (Conversion Factors)

Please see Page 8.

Boundary Effect

TSL

835 MHz

Typical SAR gradient: 5 % per mm

Sensor Center to	3.7 mm	4.7 mm	
SAR _{be} [%]	Without Correction Algorithm	4.5	1.6
SAR _{be} [%]	With Correction Algorithm	0.3	0.5

TSL

1900 MHz

Typical SAR gradient: 10 % per mm

Sensor Center to Phantom Surface Distance			4.7 mm
SAR _{be} [%]	Without Correction Algorithm	2.5	1.0
SAR _{be} [%]	With Correction Algorithm	0.2	0.4

Sensor Offset

Probe Tip to Sensor Center

1.0 mm


The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of NormX,Y,Z do not affect the E²-field uncertainty inside TSL (see Page 8).

⁸ Numerical linearization parameter: uncertainty not required.


Frequency Response of E-Field


(TEM-Cell:ifi110 EXX, Waveguide: R22)




Uncertainty of Frequency Response of E-field: \pm 6.3% (k=2)

Receiving Pattern (ϕ), θ = 0°








Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


Dynamic Range f(SAR_{head})


(Waveguide R22, f = 1800 MHz)



Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Conversion Factor Assessment






f [MHz]	Validity [MHz] ^c	TSL	Permittivity	Conductivity	Alpha	Depth	ConvF Uncertainty
835	± 50 / ± 100	Head	41.5 ± 5%	0.90 ± 5%	0.84	0.69	8.36 ± 11.0% (k=2)
1900	± 50 / ± 100	Head	40.0 ± 5%	1.40 ± 5%	0.45	0.74	7.11 ± 11.0% (k=2)
2450	± 50 / ± 100	Head	39.2 ± 5%	1.80 ± 5%	0.60	0.75	6.27 ± 11.8% (k=2)
5200	± 50 / ± 100	Head	36.0 ± 5%	4.66 ± 5%	0.45	1.80	4.45 ± 13.1% (k=2)
5800	± 50 / ± 100	Head	35.3 ± 5%	5.27 ± 5%	0.45	1.80	4.05 ± 13.1% (k=2)
835	± 50 / ± 100	Body	55.2 ± 5%	0.97 ± 5%	0.92	0.62	8.15 ± 11.0% (k=2)
1900	± 50 / ± 100	Body	53.3 ± 5%	1.52 ± 5%	0.20	1.13	6.64 ± 11.0% (k=2)
2450	± 50 / ± 100	Body	52.7 ± 5%	1.95 ± 5%	0.65	0.75	6.37 ± 11.8% (k=2)
5200	± 50 / ± 100	Body	49.0 ± 5%	5.30 ± 5%	0.47	1.70	4.08 ± 13.1% (k=2)
5800	± 50 / ± 100	Body	48.2 ± 5%	6.00 ± 5%	0.45	1.70	3.86 ± 13.1% (k=2)

^c The validity of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2). The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band.

Deviation from Isotropy in HSL

Error (ϕ , ϑ), f = 900 MHz



Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)