



NOTES:

- 1) MATERIAL: .010" THICK MATTE VELVET POLYCARBONATE, WITH 3M 468 PRESSURE SENSITIVE ADHESIVE, FARSIDE.
- 2) LABEL TO BE DIE CUT TO SHAPE WITH PEEL OFF LINER.
- 3) COLORS: BACKGROUND TO BE SILVER LINING (PANTONE 428), TEXT AND SYMBOLS TO BE ECLIPSE EXCEPT WHERE INDICATED.
- 4) SHALL BE AN APPROVED MARKING AND LABELING SYSTEM MEETING THE REQUIREMENTS OF UL 969 AND CSA 22.2 NO. 0.15 AND SHALL BEAR THE MANUFACTURERS IDENTIFYING APPROVAL MARK AND MATERIAL TYPE (EX. IV).
- 5) SHALL SATISFY THE DURABILITY REQUIREMENTS OF UL60601-1.
- 6) LABEL TO BE SERIALIZED STARTING WITH 0101 AND RUNNING SEQUENTIALLY. VENDOR TO TRACK NUMBERS.
- 7) PLACE WRITE-ON OVERLAMINATE OVER ENTIRE SURFACE.
- 8) SEE PACKAGING GROUP FOR ARTWORK.

NO. 8G1598 SHEET 1 NEXT SHEET 0		
CHG. NO. DATE REL	REVISIONS	DR BY APPS.
REV A 10/17/08	INITIAL RELEASE	DR
REV B 2684-013 11/17/08	STANDARD DATAPLATE UPDATES	DR JEM
UNLESS OTHERWISE SPECIFIED LINEAR TOLERANCE ± 1.6 ANGULAR TOLERANCE \pm — DIMENSIONS ARE IN MILLIMETERS		MD
CARESTREAM HEALTH, INC. ROCHESTER, NY 14608		
DR. WESTON	DATE 11-17-08	CK
DFTG. JEM	DES. ENG. M. CALVO	MFG. ENG. M. CALVO
NAME DATAPLATE - CARESTREAM DRX-1 SYSTEM DETECTOR		
SCALE 1X SIZE		DWG SIZE A
NO. 8G1598 SHEET 1 NEXT SHEET 0		