

Postfach 529 • D-79005 Freiburg  
Bötzingen Straße 41 • D-79111 Freiburg  
t: (+49) 761 4512 0 f: (+49) 761 4512 120

**stryker®**

Navigation

## **Request for Class II Permissive Change**

PHOENIX TESTLAB GmbH  
Product Certification  
Königswinkel 10

Date: 2009/08/11

D 32825 Blomberg

Ref: FCC Class II Permissive change for FCC ID: U92-1

Applicant: Stryker Leibinger GmbH & Co. KG

Dear Examiner,

This is to request a Class II Permissive change for the RFID device with  
FCC ID: U92-1 originally granted on 07/03/2007.

Following modifications have been implemented:

- USB Interface modified  
(USB to parallel FIFO bidirectional data transfer interface)  
Parts and component placements changed
- Several components removed (resister networks, LEDs) and re-arranged
- Housing changed (mixture of metal & plastic; hence no EMC ferrite on the USB line is needed)

We hereby confirm that no changes are made to the basic frequency determining and stabilizing  
circuitry, frequency multiplication stages, basic modulator  
circuit or maximum power or field strength ratings.

Sincerely

Applicant's Company : Stryker Leibinger GmbH & Co. KG  
address : Bötzingen Straße 41  
D-79111 Freiburg

Signature : *C. A. ASHILL*

Job Title and Dept. : *Software Engineer*

*i. V. A. SARVESTANI*

*i. V. MANAGER QA/RA*  
*27.8.09*

Stryker Leibinger GmbH & Co. KG  
Bank: Bank of America N.A. Frankfurt 17 362 010 (BLZ 500 109 00)  
IBAN: DE04 5001 0900 0017 3620 10  
Eintrag: Amtsgericht Freiburg HRA 4375  
Komplementärin: Stryker Beteiligungs GmbH  
Geschäftsführer: Dr. Heinrich W. Dreyer, Klaus Welte  
Eintrag: Amtsgericht Freiburg HRB 5441, EAR Nr. 92587730