

Test Equipment List with Calibration Data / Test date: June 2006

No.	Type	Inv.-No.	Model	Serial Number	Manufacturer	Calibration Organization	Date of last calibration	Date of next calibration ¹
1.	Test receiver	1028	ESHS 10	860043/016	Rohde & Schwarz	Rohde & Schwarz	01/2006	07/2007
2.	Test receiver	1025	ESVP	881120/024	Rohde & Schwarz	Rohde & Schwarz	06/2005	12/2006
3.	Test receiver	1711	ESPI7	836914/0002	Rohde & Schwarz	Rohde & Schwarz	08/2005	02/2007
4.	EMI test receiver	1569	ESMI	839379/013 839587/006	Rohde & Schwarz	Rohde & Schwarz	12/2005	06/2007
5.	Spectrum analyzer	1078	R3271A	05050023	Advantest	Rohde & Schwarz	03/2006	09/2008
6.	Spectrum analyzer	1666	FSP 30	100063	Rohde & Schwarz	Rohde & Schwarz	02/2006	08/2007
7.	Preamplifier	1142	R14601	13120026	Advantest	Rohde & Schwarz	05/2005	11/2006
	Matrix RELAX consists:	1941						
8.	Preamplifier	1651	CPA9231A1	3393	Schaffner	Schaffner	05/2005	11/2006
9.	Preamplifier	1684	AFS3-00100800-32-LN	847743	Miteq	Senton GmbH	04/2005	10/2006
10.	V-network	1059	ESH3-Z5	894785/005	Rohde & Schwarz	Senton GmbH	12/2005	12/2008
11.	V-network	1060	ESH3-Z5	862770/021	Rohde & Schwarz	Senton GmbH	01/2006	01/2009
12.	V-network	1218	ESH3-Z5	830952/025	Rohde & Schwarz	Senton GmbH	05/2005	05/2008
13.	Loop antenna	1016	HFH2-Z2	882964/0001	Rohde & Schwarz	Rohde & Schwarz	01/2006	07/2007
14.	Trilog wide band antenna Chamber 2	1722	VULB 9163	9163-188	Schwarzbeck	Schwarzbeck	05/2005	11/2006
15.	Trilog wide band antenna Chamber 3	1802	VULB 9163	9163-214	Schwarzbeck	Schwarzbeck	01/2006	07/2007
16.	Biconical antenna EG 1	1518	HK 116	842204/001	Rohde & Schwarz	Rohde & Schwarz	12/2005	06/2007
17.	Log. per. antenna EG 1	1519	HL 223	841516/023	Rohde & Schwarz	Rohde & Schwarz	12/2005	06/2007
18.	Horn Antenna	1516	3115	9508-4553	Emco	ARC-Seibersdorf	05/2005	05/2008

¹ Contains maximum tolerance of six months for devices with calibration cycles less than three years.