

Dräger Medical AG & Co. KG, 23542 Lübeck

To whom it may concern

Unser Zeichen

LOC_FCCIC2009

+49 451 882-4684

+49 451 882-74684

Rene.Weyergraf@draeger.com

18.06.2009

Letter of Change

This application is for the amendment of the modular approval of the module M21/M21.2. The module can be equipped with further antennas (Antenna WAL MP02091, Antenna CO2 MP02131 and Antenna Set Primus 8607311) which can be additionally used to the already licensed antennas (Insp. Exp. Antenna MP02021 and Spirolog Antenna MP02011). The module itself was not modified only the antennas for further applications are introduced. Hence, we consider that the module M21/M21.12 still qualifies the certification.

The RFID functionality is used with typical installations on anesthesia machine, like CO2 absorber, water trap and breathing circuits for e.g. exchange control.

The application contains a FCC/IC report which is attached. The report is valid for all antennas

Test Report for Antenna CO2 Watertrap, Antenna Set Test Report No. 50420-080059-4 editon 1

(DRA080059-4 FCC15-225 (M21+all Antennas) ed1.pdf)

Sincerely

Dräger Medical AG & Co. KG

René Weyergraf

Senior Manager Regulatory Affairs