

**Model: ARCHIMEDE**

**SN: SSA 1020**

**Classe I - Instrument with applied part type B.**



**IP 41**

**Voltage: 90 VAC 84W 1A  
250 VAC 81W 0.4A  
47-63 Hz 150 VA Fuses 2 AT**



**R.A.E.E. IT08020000001950**

**MAC ADDRESS**

**00-20-4A-9E-03-23**



**MOELCA s.r.l.**

**Via del Lavoro, 19**

**22070 Limido Comasco (CO) ITALY**

**Tel.: 031-3520153 3520279 3524739**

**Fax: 031- 3524732 Email: [info@moelca.it](mailto:info@moelca.it)**



**Activity**



**FCC ID: V991504**

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

## **CAUTION**

Disconnect power before changing fuses. To reduce the risk of electric shock do not remove cover. Refer servicing to qualified service staff. Disconnect the system from supply before servicing.

Revision 1.02