

tobii **dynavox**



FCC ID: W5M-TDI110
IC: 5534A-TDI110



E351386
13YG

MEDICAL-GENERAL MEDICAL EQUIPMENT AS
TO ELECTRICAL SHOCK, FIRE AND
MECHANICAL HAZARDS ONLY IN
ACCORDANCE WITH ANSI/AAMI ES 60601-1
(2005)+AMD(2012) AND CAN/CSA-C22.2 NO.
60601-1 (2008)+(2014)



Model:I-110



For use power supply : ADAPTER TECH. /ATM036T-A150

INPUT:100-240V ~, 50-60Hz 1A-0.45A

OUTPUT:15V == 2.4A

For use Battery Pack : 9789OTG101

Rating:11.4Vdc == , 4090mAh/46.62Wh www.tobiidynavox.com

Tobii Dynavox

Karlsrovägen 2D, 182 53 Danderyd, Sweden



This device complies with Part 15 of the FCC Rules.

Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference and
- (2) this device must accept any interference received,
including interference that may cause undesired operation.