Customer Specification

HL1916

19 Inch LCD Display

| Name | Department/Title | Date | Signature |
|---------------|------------------|------|-----------|
| Cheng HuanJun | SSME CT PL | | |
| | | | |
| | | | |
| | Bigtide R&D VP | | |
| Xu Bin | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 1 of 32 |
|--|--------|--------------|
| rights reserved | | |

CONTENTS

| 1. | SCOPE | . 4 |
|-----------|---|-----|
| 2. | ELECTRICAL PERFORMANCE | . 5 |
| | 2.1 Power Supply | 5 |
| | 2.2 Power Management | 5 |
| | 2.3 Signal Interface | 5 |
| | 2.3.1 Signal Specifications | .6 |
| | 2.3.2 D-SUB Connector and Pin Assignment (Figure 1) | 6 |
| | 2.3.3 Digital Visual Interface and Pin Assignment (DVI) | 7 |
| | 2.3.4 Control Interface | .7 |
| | 2.4 Product Features | 9 |
| | 2.5 Screen Performance | 10 |
| | 2.5.1 Standard Testing Conditions | 10 |
| | 2.5.2 Brightness | 11 |
| | 2.5.3 View angle | 11 |
| | 2.5.4 Brightness Uniformity | 11 |
| | 2.5.5 Contrast radio | 11 |
| | 2.5.6 White Color Coordinates | 11 |
| | 2.5.7 Response Time | 11 |
| | 2.5.8 Color Gray | 11 |
| | 2.5.9 Gamma Curve | 12 |
| 3. | OPERATING GUIDE | 14 |
| | 3.1 Keys assignment | 14 |
| | 3.2 Key Functions without active OSD Menu | 15 |
| | 3.3 Key Functions in the OSD Menu | 15 |
| | 3.4 Submenu Calls | 15 |
| | 3.5 Locking of OSD Menu | 15 |
| | 3.6 Description of OSD Menu | 15 |
| 4. | MECHANICAL SPECIFICATIONS | 18 |
| | 4.1 Outline dimensions & weight | 18 |
| | 4.2 Screen Quality | 19 |
| | 4.2.1 H/V outline position | 19 |
| | 4.2.2 Outline edge position | 20 |
| | 4.2.3 Structure width position | 20 |
| | 4.3 Packaging | 21 |
| | 4.3.1 Package dimension and weight | |
| 5. | ENVIRONMENT CONDITONS | 21 |
| | 5.1 Operation Temperature | 21 |

Last change: 27th, Oct. 2005 Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved HL1916 Page 2 of 32

| | 5.2 Transport and storage (Packed) | 22 |
|----|------------------------------------|----|
| | 5.3 Mechanical requirements | 22 |
| | 5.4 Drop Test (packed) | 23 |
| | 5.5 Safety specifications | |
| | 5.6 Electromagnetic compatibility | 24 |
| | 5.7 MTBF | 24 |
| 6. | DEFECT, SCRATCH and DUST | 25 |
| | 6.1 Condition | 25 |
| | 6.2 Dot Defect | 25 |
| | 6.3 Polarize Defects | 26 |
| | 6.4 Foreign Material | 26 |
| | 6.5 Line Defect | 27 |
| | 6.6 Others | 27 |
| 7. | NOTICE FOR HANDING | 27 |
| | 7.1 Handing | 27 |
| | <u></u> | |
| | 11 4 D 4 W 1 | |
| Ap | ppendix 1 Preset Timings | 29 |

Federal Communications Commission (FCC) Statement

You are cautioned that changes or modifications not expressly approved by the part responsible for compliance could void the user's authority to operate the equipment

FCC- Class B

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a communications. However, there is no guarantee that interference will not occur in particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

1. SCOPE

This document defines the performance requirements for a 19.0-inch TFT LCD color monitor for medical use. This product is controlled by model name; any change will be recorded in the list and confirmed by SIEMENS MED.

This high-resolution color display is specifically designed to meet the rigorous performance standards needed for diagnostic, interventional radiology, and other medical applications. To guarantee image integrity, features include accurate signal conversion and a wide range of interfacing options.

This monitor is factory calibrated to achieve DICOM part 3.14 compliance and Gamma CIE at the factory set point. The luminance stabilization circuit employs a built in photo sensor to keep the back-light lamps at a constant luminance for consistent calibration over the life of the display and can control the back light system automatically to extend the life of the monitor and achieve very short warming up time.

The surface of the monitor has an anti-glare coating to minimize reflection and a hard coating to reduce scratch.

With the deliberate designed bracket the monitor can stick on the desktop firmly.

| rights reserved | | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 4 of 32 |
|-----------------|--|--|--------|--------------|
|-----------------|--|--|--------|--------------|

2. ELECTRICAL PERFORMANCE

2.1 Power Supply

-Input Voltage : $AC100-240 \pm 10\%$

- Current (max) : 0.25A(AC220-240V); 0.55A(AC100-120V)

- Frequency : 50/60Hz \pm 3Hz

- Power Consumption : <54W

2.2 Power Management


Power Management condition and status for ANALOG Input mode

| State | | SIGNALS | | Power | LED |
|------------|------------|----------|---------|-------------|--------|
| | Horizontal | Vertical | Video | Consumption | Status |
| ON | ON | ON | Active | <54W | Green |
| | OFF | ON | Blanked | | Orange |
| Active off | ON | OFF | Blanked | <2W | Orange |
| | OFF | OFF | Blanked | | Orange |

Power Management condition and status for DIGITAL Input mode

| State | | SIGNA | ALS | | Power | LED |
|------------|--------------|------------|----------|---------|-------------|--------|
| | DE | Horizontal | Vertical | Video | Consumption | Status |
| ON | Pulses | ON | ON | Active | <54W | Green |
| Active off | NO Pulses | N/A | N/A | Blanked | <2W | Orange |

2.3 Signal Interface



| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 5 of 32 |
|--|--------|--------------|
| rights reserved | | |

2.3.1 Signal Specifications

| | Item | SPEC |
|--------------|--------------|-----------------------------|
| | | Analog: |
| | Frequency | H 30 ~ 82kHz |
| Signal Input | | V 50 ~ 85Hz |
| (Analog) | Pixel clock | 25165MHz |
| D-SUB | Video Innut | Analog 0.7Vpp |
| Video Input | | Input Impedance 75 Ohm |
| | Signal Input | Separate Sync, TTL (N or P) |
| CVS Signal | | Video Level: 0.51.0V |
| C v S Signai | | Sync level: 0.20.3V |
| DVI-I Input | | Via DVI-I to VGA connector |
| DVI D Innut | | DVI-Digital DVI-Digital |
| DVI-D Input | | DDC via DVI |

2.3.2 D-SUB Connector and Pin Assignment (Figure 1)





Figure 1

15-pin Min D-type Female Connector

Table 1

| Pin - Assignment of 15-pin D-sub: | | | | | |
|-----------------------------------|-------------|----|--------------|----|------------------|
| 1 | Red Video | 6 | Red Ground | 11 | Monitor Ground |
| 2 | Green Video | 7 | Green Ground | 12 | DDC-Serial Data |
| 3 | Blue Video | 8 | Blue Ground | 13 | H-Sync. |
| 4 | N/C | 9 | NC | 14 | V-Sync. |
| 5 | GND | 10 | Logic Ground | 15 | DDC-Serial Clock |

| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 6 of 32 |
|--|--------|--------------|
| rights reserved | | |



DVI-D Receptacle Connector

Figure 2

29-pin DVI-D Female Connector

Table 2

| Pin - Assignment of 29-pin DVI-D/DVI-I Female Connector | | | | | |
|---|-------------------------------|----------------------------|--|--|--|
| Pin 1 - TMDS Data 2- | Pin 12 - TMDS Data 3- | Pin 22 - TMDS Clock Shield | | | |
| Pin 2 - TMDS Data 2+ | Pin 13 - TMDS Data 3+ | Pin 23 - TMDS Clock+ | | | |
| Pin 3 - TMDS Data 2/4 Shield | Pin 14 - +5 V Power | Pin 24 - TMDS Clock- | | | |
| Pin 4 - TMDS Data 4- | Pin 15 – Ground | | | | |
| Pin 5 - TMDS Data 4+ | Pin 16 - Hot Plug Detect | | | | |
| Pin 6 - DDC Clock | Pin 17 - TMDS Data 0- | | | | |
| Pin 7 - DDC Data | Pin 18 - TMDS Data 0+ | | | | |
| Pin 8 - Analog Vertical Sync | Pin 19 - TMDS Data 0/5 Shield | | | | |
| Pin 9 - TMDS Data 1- | Pin 20 - TMDS Data 5- | | | | |
| Pin 10 - TMDS Data 1+ | Pin 21 - TMDS Data 5+ | | | | |
| Pin 11 - TMDS Data 1/3 Shield | | | | | |

2.3.4 Control Interface

Beside the OSD, we also support RS-232 interface to update and control the monitor by software. When the monitor is connected with a CT or other iatrical diagnosis instruments, we provide additional interface for the service technician to get the status of the monitor and update software. When use RS-232 for software update, first put 5V switch (near the RS-232 input) to on position. And then load the soft ware from computer. After complete software update, slide the switch to off position. And then power off the monitor. Power on again the software update is completed.

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 7 of 32 |
|--|---|--------|--------------|
|--|---|--------|--------------|




Figure 3 Female 9-PIN D-SUB connector

Table3

| Pin | Signal | Pin | Signal |
|-----|---------------------|-----|-----------------|
| 1 | Data Carrier Detect | 6 | Data Set Ready |
| 2 | Received Data | 7 | Request to Send |
| 3 | Transmitted Data | 8 | Clear to Send |
| 4 | Data Terminal Ready | 9 | Ring Indicator |
| 5 | Signal Ground | | |

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 8 of 32 |
|---|---|--------|--------------|
| | rights reserved | | |

2.4 Product Features

| Item | | | Specification | | |
|------------------|----------------------|-------|---|--------------------------------------|--|
| Item | item | | Analog Input | Digital Input | |
| LCD | Panel Module Size | | LG.PHILIPS LM190E05-SL02 | | |
| | | | 19.0" (48 cm diagonal) | | |
| | Active Display A | rea | 376.32 (H) x 301.056 (V) r | nm | |
| | Resolution | | 1280 x 1024 dots (SXGA) | | |
| | Pixel Pitch | | 0.294(H) x 0.294(V) mm | | |
| | Color Depth | | 16.7M true 8 bit | | |
| | Luminance | | 280 cd/m2 (typical); 230cd | /m2 (min.) | |
| | Viewing Angle | | Н | L 89 degree | |
| | (Type.) | CR>10 | П | R 89 degree | |
| | | CR>10 | V | H 89 degree | |
| | | | V | L 89 degree | |
| | Contrast Ratio | | 600:1 (typical); 400:1 (min.) | | |
| | Back Light | | CCFL x 6pcs. | | |
| | Horizontal frequency | | 31kHz -82kHz | 31kHz -82kHz | |
| | Vertical frequency | | 50.0Hz - 85.0 Hz (Non-Interlaced) | 50.0Hz - 85.0 Hz (Non-Interlaced) | |
| | Video Signal | | Analog RGB | Digital RGB | |
| Input Signals | Sync. Signal | | Separate Sync. (TTL) Composite Sync. Sync on green | TMDS | |
| | Pixel Clock | | 25.0MHz -165.0MHz | 25.0MHz -165.0MHz | |
| | Input connector | | Mini D-sub 15Pin | DVI-I (D) & DVI-D | |
| Preset Timing | SS | | Factory preset: 39 / User preset: 10 * | | |
| | Control key | | Select, up, down, exit | | |
| Functions | OSD | | Backlight Brightness, Contrast, Color control, Position, Size, Phase, Gamma, etc. | | |
| Regulations | Safety | | IEC60601-1, CCC | | |
| | EMC | | IEC60601-1-2 | | |
| | Power Management | | VESA DPMS, EPA, | | |

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 9 of 32 |
|-----------------------------|--|--------|--------------|
| Last change. 27 , Oct. 2005 | rights reserved | | |

| | Plug and Play | | VESA DDC2B |
|--|--------------------------------|------------------|---------------------------------------|
| Environment | 1 | | 5-40 degree C |
| Condition | Humidity | | 30-80% (without condensation) |
| | Input Voltage | | AC100-240V, 50 / 60Hz; <0.5A |
| Power Power | | Normal operation | <54W |
| Supply Consumption | Power saving | <2W | |
| Input Connector | | or | 3P IEC Type |
| VESA compatible arm mounting interface | | ing interface | 100mm x 100mm |
| Height up/dov | vn | | 60mm—80mm (around) |
| Tilt adjust | | | Up & Down –515 degree |
| | AC Power cord (Extension Type) | | 3.0 m |
| Accessories | Signal Cable | | 3.0 m: D-sub15pin VGA; 3.0 m: DVI-D |
| | Others | | User's manual, DVI-I to VGA connector |

• Remark: The monitor shall recognize preset modes within a range of ±1KHz for horizontal and ±1Hz for vertical. (See appendix 1)

2.5 Screen Performance

2.5.1 Standard Testing Conditions

- Warm up time ≥ 20 minutes.

- AC supply voltage 100-240VAC, 50/60Hz

- Ambient temperature 20°C -25°C
 - Relative Humidity 30% --80%

- Video signal 1280 x 1024 @ 60Hz; DVI-D

- Ambient Environment Dark

- Setting Set to Gamma CIE factory preset

- Luminance meter Minolta CA-210

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 10 of 32 |
|--|---|--------|---------------|
|--|---|--------|---------------|

2.5.2 Brightness

◆ Factory Preset (Gamma CIE): Lmin = 0.36+0.15/-0.1 cd/m²

 $Lmax = 137+5/-5 \text{ cd/m}^2$

• Factory Preset (DICOM): Lmin = 0.45+0.15/-0.1 cd/m²

 $Lmax = 137+5/-5 \text{ cd/m}^2$

Max. Brightness: ≥230cd/m2 (adjust CCFL, Contrast and Brightness to Maximum)

◆ Test Condition: White Luminance (L max) is defined as a luminance of L255 Gray

Level at the center point on LCD surface. Also Black Luminance (L min) is defined as a Luminance of L0 Gray level at the center point on LCD surface. (See Note 1 Note 4)

2.5.3 View angle

Left/Right and Up/Down typical 178 degree (CR≥10) (Note 2)

2.5.4 Brightness Uniformity

Deviation less than 20% (Note 5)

Comply with DIN6868-57 Class B

2.5.5 Contrast radio

Over 400:1 (note 3)

2.5.6 White Color Coordinates

 $X=0.313 \pm 0.03$

 $Y=0.329 \pm 0.03$

(Note 7)

2.5.7 Response Time

Typical (Tr+Td): 18ms (Note 6)

2.5.8 Color Gray

64 gray level should seen clearly

256 gray levels should be seen smoothly

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 11 of 32 |
|----------------------------|---|--------|---------------|
| Last change. 27, Oct. 2003 | rights reserved | | |

2.5.9 Gamma Curve

Within \pm 10% tolerance of calculated value


Factory Preset gamma value: Gamma CIE; DICOM; Gamma 2.0; 2.2; 2.4; NATIVE

(Gamma CIE is default setting)

Note1: Test Equipment Setup

The measurement should be executed in a stable, windless and dark room between 20 minutes after the backlight at the given temperature for stabilization of the backlight. This should be measured in the center of screen. Test equipment should be equivalent with the following equipment.


Environment condition: Ta=25±2°C



Optical Measuring Equipment Setup

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 12 of 32 | |
|----------------------------|---|--------|---------------|--|
| Last change. 27, Oct. 2005 | rights reserved | | | |

Note2: Viewing angle is measured as follow:



Note 3: Definition of contrast Ratio (CR):

Ratio of gray max (Gmax) & gray min (Gmin) at the center point of the panel.

$$CR = \frac{G \max}{G \min}$$

Gmax: Luminance with all pixels white

Gmin: Luminance with all pixels black

Note 4: Definition of Luminance of White: Luminance of white at center point.

Note 5: Definition of brightness uniformity

$$Buni = 100*\frac{(B\max - B\min)}{B\max}$$

Bmax: Maximum brightness


Bmin: minimum brightness

| Last change: 27 th , Oct. 2005 | Medical Equipment Ltd.) All | HL1916 | Page 13 of 32 |
|---|-----------------------------|--------|---------------|
| | rights reserved | | |

Note 6: Definition of response time is as follows:

Sum of TrD, TfR

When the display data is changed from white to black, response time is measured



Note 7: Definition of Color Chromaticity (CIE 1931)

Color coordinate of Red , Green, Blue & White at center point .

3. OPERATING GUIDE

3.1 Keys assignment




Figure 4

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 14 of 32 |
|--|---|--------|---------------|
|--|---|--------|---------------|

3.2 Key Functions without active OSD Menu

| Key | Action |
|------|---------------------------------|
| Menu | Activate OSD |
| Up | Select VGA (DVI-A) input source |
| Down | Select DVI-D input source |

Note: This choice is in case all the signal sources are available. If not, the signal on the any one input will be displayed.

3.3 Key Functions in the OSD Menu

| Keys | Situation | Action |
|------|------------------------|-------------------------|
| Menu | Always | Jump to next line |
| Up | Slide controller | Increase Value |
| | Command | "Enter Key" |
| Down | Slide controller | Decrease value |
| Set | Except "Exit OSD" Menu | One menu level upwards |
| | | (Settings are retained) |
| | In "Exit OSD" Menu | Return to main menu |
| | | (Settings are retained) |

3.4 Submenu Calls

Press the "Menu" key while the OSD is active, the function icon will jump to next line. Pressing the "Up" key, the coordinate submenu will be selected.

3.5 Locking of OSD Menu

| Keys | Action |
|------------------|--|
| 1 time Set key | Lock or unlock OSD |
| 3 times Up key | If the OSD is locked, it is only possible to |
| within 3 seconds | select input source (see 3.1 section). |

3.6 Description of OSD Menu

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 15 of 32 |
|---|---|--------|---------------|
|---|---|--------|---------------|

| Main Menu | Function | Adjustment range | Description |
|-----------------------------|----------------|------------------------------|---|
| | | | It is used to adjust the black level of the monitor. |
| | Brightness | 0100% | This allows the darker area to be seen more distinctly. |
| | | | Note: for DVI-D signals the brightness setting is optimized. Manual changes are not recommended. |
| | | | Adjustment of contrast |
| | Contrast | 0100% | This allows the brighter area to be seen more distinctly. |
| Brightness | Contrast | 010070 | Note: for DVI-D signals the Contrast setting is optimized. Manual changes are not recommended. |
| /Contrast | | 0100% | A diversion and heighten and of LCD height had |
| | Backlight | Recommended setting Max. 80% | Adjustment brightness of LCD backlight to adapt total brightness for room illumination |
| | | 1,2,3 User | |
| | Color | 1, 9300K | Set desired color or Hue |
| | | 2, 7300K | Three fixed color temperatures and one adjustable color temperature are available |
| | | 3, 6500K | The color locations 1 to 3 can not be saved |
| | | User define | The color locations 1 to 3 can not be saved |
| | Set user color | | |
| | Red Gain | -32+32 | Define user color temperature |
| | Green Gain | 32132 | Select Red, Green, Blue gain separately |
| | Blue Gain | | |
| | H Position | 0100% | Shift picture in horizontal direction |
| Position | V Position | 0100% | Shift picture in vertical direction |
| | | | Select source for main display |
| Picture Source | | VGA (DVI-A) | When you enter this OSD menu the current source is highlighted. |
| 1 icture Source | | DVI-D | Following switch off and on of the display, the picture sources are searched one after another. |
| Auto function (Analog input | | | The auto functions are used to assist the automatic setting of parameters. The quality of settings depends on the picture contents and the type of synchronization. |
| only) | | | Corresponding items in the OSD can adjust all settings finely. |

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 16 of 32 |
|---|---|--------|---------------|
|---|---|--------|---------------|

| | Auto brightness | On/Off | Automatically get input signal match with the monitor |
|-----------------|--|-----------------------------|---|
| | /contrast Auto position /phase /frequency | On/Off | Automatically adjusts the image position, H-Size (or V-Size) and Fine settings. Improve focus clarity and image stability. |
| | Execute selected auto function | | The selected auto functions are executed. Note: The quality of the function depends on the applied picture contents. To get better effect it is recommended to apply full screen and including whiteness contents picture. |
| Language | English | | Select the language of OSD menu. |
| Language | Chinese | | Note: English menu is default state. |
| | | | You can make further settings for the picture source |
| | Frequency /Phase | | Adjust the frequency and phase of the input signal. |
| | Sharpness | Interpolation filter 1 to 7 | One of the 7 filters can be selected for the focus setting to reduce scaling artifacts. Interpolation filters depend on the input resolution. A filter is not usually used with 1280X1024 since each pixel is controlled by its own pulse. The user should individually adjust the filter depending on the application. |
| | | | Adjustment of OSD horizontal position |
| Others | OSD setting | H position | Adjustment of OSD vertical position |
| Officis | | V position | (Default state is on the bottom right corner of the screen) |
| | DPMS setting | DPMS on | The backlight is switched off while it is no input signal. (The default state is DPMS on) |
| | | DPMS off | The backlight will not switch off if it is no input signal. Only switches off along with the power switches off. |
| | Status | | Current display settings can be called here in the respective picture mode. |
| | | | Working hours of display, firmware type and version, OSD version, configuration version, current source, current timing of input signal, current LUT |
| Service level 2 | Settings in the | is menu must only be | carried out by service person* |
| | Quit OSD menu | Accept changes | Quick OSD menu and either save or reject the changes. |
| | | Reject changes | If you have reached this menu unintentionally, you can return to the main |

| Medical Equipment Ltd.) All | HL1916 | Page 17 of 32 |
|-----------------------------|--------|---------------|
| rights reserved | | |

| | menu using the Set key. Note: if the OSD menu is quit by changing the timing or by switching off the monitor, the modification you have made are saved. |
|---------------|--|
| OSD is locked | If this message is displayed, you do not have the authority to carry out the changes in the OSD menu. Please contact your servicing partner in this case. |

| Service level 2 | For LUT select and factory settings recall: Brightness; Contrast; Backlight; Color 3, 6500K; OSD position. (The default state is Gamma CIE) | | | |
|-----------------|---|--|----------------------|--|
| | Gamma CIE Select Gamma CIE LUT | | | |
| | | DICOM Select DICOM LUT | | |
| | | Gamma 2.0 | Select Gamma 2.0 LUT | |
| | | Gamma 2.2 Select Gamma 2.2 LUT | | |
| | | Gamma 2.4 | Select Gamma 2.4 LUT | |
| | | Native Select panel native characteristics | | |

4. MECHANICAL SPECIFICATIONS

4.1 Outline dimensions & weight





Figure 5

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 18 of 32 |
|-----------------------------|---|--------|---------------|
| East change. 27 , Oct. 2005 | rights reserved | | |

| Item | Set | | |
|------------------------|--------------------------------------|--------------|--|
| | Width | 425mm | |
| Size of set | Depth | 210mm | |
| | Height | 493 (-60) mm | |
| Tilt | 0 degree – 20degree | | |
| Housing components | Aluminum | | |
| Kensington lock | Yes | | |
| Visible screen surface | Approx. 376mm×301mm | | |
| Ventilation slots | In rear panel | | |
| Degree of protection | IP20 to DIN40050 | | |
| Connection panel | At rear, covered | | |
| Not weight | Approximately 5.5 Kg (without stand) | | |
| Net weight | Approximately 8.8 Kg (with stand) | | |

4.2 Screen Quality

4.2.1 H/V outline position




H: $|L-R| \leq 1.0$ mm

 $V\!\!: |U\text{-}D|\!\leqslant\!\!1.0mm$


| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 19 of 32 |
|--|---|--------|---------------|
|--|---|--------|---------------|

4.2.2 Outline edge position



$$| A-B | \le 1.0 \text{mm}$$
 $| C-D | \le 1.0 \text{mm}$ $| E-F | \le 1.0 \text{mm}$ $| G-H | \le 1.0 \text{mm}$

4.2.3 Structure width position



| A-B | ≤2.0mm

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 20 of 32 |
|----------------------------|---|--------|---------------|
| Last change. 27, Oct. 2003 | rights reserved | | |

4.3 Packaging




Figure 6

4.3.1 Package dimension and weight

| | Box | |
|-------------------|----------------------|-----|
| Outer size of box | Width (mm) | 570 |
| | Depth (mm) | 370 |
| | Height (mm) | 575 |
| Gross weight | Approximately 12.5Kg | |

5. ENVIRONMENT CONDITONS

5.1 Operation Temperature

| Ambient temperature range | +5 +40°C |
|---------------------------|-----------------------------|
| Temperature gradient | Max. 5°C/h, no condensation |
| Atmospheric pressure | 1040 – 674 hPa |

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 21 of 32 |
|---|---|--------|---------------|
|---|---|--------|---------------|

5.2 Transport and storage (Packed)

| Ambient temperature range | -20 +60°C |
|---------------------------|-----------------------------|
| Temperature gradient | Max. 5°C/h, no condensation |
| Atmospheric pressure | 1040 – 674 hPa (0 3048m) |

5.3 Mechanical requirements

Operation

| | According to EN60068-2-6 |
|-----------|--|
| Vibration | 1058 Hz within $\pm~0.075$ mm deflection |
| | 58—500 Hz at 10 m/s2 |
| | According to EN 60068-2-27 (single shock) |
| Shock | 150m/s2, 6ms |
| | No permanent shock allowed in operating conditions |

Packed unit

| | According to EN60068-2-6 |
|-----------|--|
| Vibration | 59 Hz within \pm 3.5 mm deflection |
| | 9—500 Hz at 10 m/s2 |
| | According to EN 60068-2-27 (single shock) |
| Shock | 250m/s2, 6ms(in storage packing) |
| | According to EN 60068-2-29 (permanent shock) |

| Last change: 27 th , Oct. 2005 Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 22 of 32 |
|---|--------|---------------|
|---|--------|---------------|

5.4 Drop Test (packed)

| | Position | Height |
|---|----------|--------|
| Corner | 1 | 76.0cm |
| Edge | 3,2,4 | 76.0cm |
| | A,B,C,D | 76.0cm |
| Surfaces | Е | 76.0cm |
| | F | 76.0cm |
| (Cushion should be changed to new one.) | | |
| Corner | 5 | 76.0cm |
| Edge | 6,7 | 76.0cm |




Figure 7

5.5 Safety specifications

| Safety standards | EN60601, IEC601 |
|-------------------------|----------------------------------|
| Approvals | CAN/CSA – C 22.2 No 601.1 – M 90 |
| Approvais | CSA/us mark, UL 2601-1 |
| Protection class | Protection class 1 |
| Degree of protection to | IP 20 |
| DIN 40050 | IF 20 |
| Conformity | CE |

| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 23 of 32 |
|--|---|--------|---------------|
|--|---|--------|---------------|

5.6 Electromagnetic compatibility

| EMI voltage/ radiated interference | EN 60601-1-2 |
|---------------------------------------|---------------------------------------|
| Voltage variations | EN 610004-11 |
| Burst on power supply lines | EN 61000-4-4 |
| | 1 kV |
| Surge on power supply lines | EN 61000-4-5 |
| | 1 kV symmetric, 2 kV unsymmetric |
| Static discharge on casing | EN 61000-4-2 |
| parts (ESD) | 8 kV air, 4 kV contact |
| RF irradiation | EN 61000-4-3 |
| | 80 MHz 2,5 GHz, 3 V/m 80% AM 1 kHz |
| Noise immunity | EN 61000-4-6 |
| | 150 kHz 80 MHz |
| | 3 V _{eff} 80% AM 1 kHz |
| Magnetic constant fields | EN 61000-4-8 |
| | Max. 4000 A/m |
| Magnetic alternating fields | EN 61000-4-8 |
| | Min. 10 A/m |
| Noise emission | FCC Class A |

5.7 MTBF

Flat screen without Backlight: 50,000 operation hours.

Flat screen with Backlight: 10,000 operation hours.

| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 24 of 32 |
|---|--------|---------------|
| rights reserved | | |

6. DEFECT, SCRATCH and DUST

6.1 Condition

These defects are inspected under the following conditions:

Temperature: $20 \sim 25 \text{ degrees C}$

Humidity: $65 \pm 5\%$ RH

Illumination: Single 20W fluorescent lamp non-directive

(Appearance: 300 to 700 LUX)

Viewing distance: The distance between the LCM and the inspector's eyes shall be at


least 30cm --50 cm.

Viewing angle: The inspection shall be conducted within normal viewing angle range.

6.2 Dot Defect

The number of defect dot is defined as follows:

In the range A, the distance of defect dots from one to each other is at least 5cm. The number is at most 5. In the range B, the distance is at least 1cm. The number is at most 10.



| | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 25 of 32 |
|-----------------------------|---|--------|---------------|
| 2450 01441900 27 , 000 2000 | rights reserved | | |

Note 1: a. every dot herein means Sub- Pixel (Each Red, Green, or Blue Color)

- b. Damaged less than half size of sub-pixels is not counted as defect
- c. Dots darker than half brightness of sub-pixel are not defined as bright dot defect and dots brighter than half brightness of sub-pixels is not defined as dark dot defect.
- d. The definition of range A is included in range B.

Note 2: Panel Pixel Defect comply with ISO 13406-2 class 2

6.3 Polarize Defects

(Unit: mm)

| | Items | Criteria | | |
|-----------|----------|---|--|--|
| Scratches | Linear | 0.02 <= W <=0.05, 1.0<= L <=10.0, N <=5 | | |
| Dent | Circular | 0.15<= D <=0.5, N<=5 | | |

Where W: Width


L: Length

D: Average diameter = (a + b) / 2

Note:

a. Average Diameter

b. Linear: a>2b, Circular: a<=2b



- c. Extraneous substances which can be wiped out, like Finger Print, Particles, are not considered as a defect.
- d. Defect which is on the black matrix (outside of Active Area) are not considered as a defect.

6.4 Foreign Material

(Unit: mm)


| | Items | Criteria | | | |
|-------------------|--------|--------------------------------------|--|--|--|
| Foreign | Linear | 0.02<= W <=0.1, 0.3<= L <=3.0, N <=5 | | | |
| Material Circular | | 0.2<= D <=0.5, N<=5 | | | |

| Last change: 27 th Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 26 of 32 |
|---|---|--------|---------------|
| Last change. 27, Oct. 2003 | rights reserved | | |

Where W: Width

L: Length

D: Average diameter = (a+b)/2



Note)

a. Average Diameter

b. Linear: a>2b, Circular: a<=2b

6.5 Line Defect

All kinds of line defects such as vertical, horizontal or cross are not allowed.

6.6 Others

Issues, which are not defined in these criteria, shall be discussed with both parties, Customer and Supplier, for better solution.

7. NOTICE FOR HANDING

7.1 Handing


- (1) When the module is assembled, it should be attached to the system firmly using all mounting holes. Be careful not to twist or bend the modules.
- (2) Because the inverters use high voltage, power should be disconnected before it is assembled or disassembled.
- (3) Refrain from string mechanical shock and /or any force to the module. In addition to damage, this may cause improper operation or the module and CCFT backlight.
- (4) Note that polarizers are very fragile and could be easily damaged. Do not press or scratch the surface using the harder than a HB pencil lead.
- (5) Wipe off water droplets or oil immediately. If you leave the droplets for a long time, staining and discoloration may occur.
- (6) If the surface of the polarizer is dirty, clean it using some absorbent cotton or soft cloth.
- (7) The desirable cleaners are water, IPA (Isopropyl Alcohol) or Hexane. Do not use Ketone type materials (ex. Acetone), Ethyl alcohol, Toluene, Ethyl acid or Methyl chloride. It might damage to the polarizer due to chemical reaction.


| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 27 of 32 |
|--|--------|---------------|
| rights reserved | | |

- (8) If the liquid crystal material leaks from the panel, it should be kept away from the eyes or mouth. In case of contact with hands, legs or cloths, it must be washed away thoroughly with soap.
- (9) Protect the module from static which may cause damage to the CMOS Gate Array IC.
- (10) Use fingerstalls with soft gloves in order to keep display clean during the incoming inspection and assembly process.
- (11)Do not disassemble the module.
- (12)Do not pull or fold the lamp wire.
- (13)Do not adjust the variable resistor located on the module.
- (14)Protection film for polarizer on the module should be slowly peeled off just before use so that the electrostatic charge can be minimized.
- (15)Pins of I/F connector should not be touched directly with bare hands.

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 28 of 32 |
|---|--|--------|---------------|
| _ | rights reserved | | |

Appendix 1 Preset Timings





| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 29 of 32 |
|---|--------|---------------|
| rights reserved | | |

• Analog Signal Timings

| NO | Timing | Dot freq (Mhz) | H.freq (Khz) | H.dis (us) | H.bac k porch (us) | Hs width (us) | V.freq (Khz) | V.dis (us) | V.back porch (us) | Vs Wid (us) | Hs. Vs pol | Inte rlac e |
|----|-------------------|----------------------|-----------------|------------|-----------------------------|---------------------|-----------------|---------------|-------------------------|-------------------|------------------|-------------------|
| 1 | 1024*768@43 HZ | 44.900 | 35.522 | 22.806 | 1.247 | 3.920 | 86.958 | 10.810 | 0.296 | 0.056 | P P | YES |
| 2 | 1024*768@60 HZ | 65.000 | 48.363 | 15.754 | 2.462 | 2.092 | 60.004 | 15.880 | 0.600 | 0.124 | N N | NO |
| 3 | 1024*768@70 HZ | 75.000 | 56.476 | 13.653 | 1.920 | 1.813 | 70.069 | 13.599 | 0.513 | 0.106 | N N | NO |
| 4 | 1024*768@75 HZ | 78.750 | 60.023 | 13.003 | 2.235 | 1.219 | 75.029 | 12.795 | 0.466 | 0.050 | P P | NO |
| 5 | 1024*768@85 HZ | 94.500 | 68.677 | 10.836 | 2.201 | 1.106 | 84.997 | 11.183 | 0.524 | 0.044 | P P | NO |
| 6 | 1024*768@60 HZ | MAC | | | | | | | | | | |
| 7 | 1024*768@72 HZ | IBM | | | | | | | | | | |
| 8 | 800*600@56 HZ | 36.000 | 35.156 | 22.222 | 3.556 | 2.000 | 56.250 | 17.067 | 0.626 | 0.057 | P P | NO |
| 9 | 800*600@60 HZ | 40.000 | 37.879 | 20.000 | 2.200 | 3.200 | 60.317 | 15.840 | 0.607 | 0.106 | P P | NO |
| 10 | 800*600@72 HZ | 50.000 | 48.077 | 16.000 | 1.280 | 2.400 | 72.188 | 12.480 | 0.478 | 0.125 | P P | NO |
| 11 | 800*600@75 HZ | 49.500 | 46.875 | 16016 2 | 3.232 | 1.616 | 75.000 | 12.800 | 0.448 | 0.064 | P P | NO |
| 12 | 800*600@85 HZ | 56.250 | 53.674 | 14.222 | 2.702 | 1.138 | 85.061 | 11.179 | 0.503 | 0.056 | P P | NO |
| 13 | 832*624@75 HZ | MAC | | | | | | | | | | |
| 14 | 640*480@60 HZ | 25.175 | 31.469 | 25.422 | 1.907 | 3.813 | 59.940 | 15.253 | 1.048 | 0.064 | N P | NO |
| 15 | 640*480@72 HZ | 31.500 | 37.861 | 20.317 | 4.064 | 1.270 | 72.809 | 12.678 | 0.739 | 0.079 | N N | NO |
| 16 | 640*480@75 HZ | 31.500 | 37.500 | 20.317 | 4.064 | 1.270 | 72.809 | 12.678 | 1.739 | 0.079 | N N | NO |
| 17 | 640*480@85 HZ | 36.000 | 43.269 | 17077 8 | 2.222 | 1.556 | 85.008 | 11.093 | 0.578 | 0.069 | N N | NO |
| 18 | 640*480@66 HZ | MAC | | | | | | | | | | |

| ĺ | | | | |
|---|---|-----------------------------|-----------|---------------|
| | | Copyright @ SSME (SIEMENS | HL1916 | Page 30 of 32 |
| | Last change: 27 th , Oct. 2005 | Medical Equipment Ltd.) All | TILL 1710 | |
| | • | rights reserved | | |

| | | | | • | | | • | | | | | |
|----|--------------------|-----------|--------|--------|-------|-------|--------|--------|-------|-------|-----|----|
| 19 | 640*350@70 HZ | 25.175 | 31.469 | 25.422 | 1.907 | 3.813 | 70.087 | 11.122 | 1.907 | 0.064 | P N | NO |
| 20 | 720*350@70 HZ | | | | | | | | | | | |
| 21 | 640*400@70 HZ | 25.175 | 31.469 | 25.422 | 1.907 | 3.813 | 70.087 | 12.711 | 1.112 | 0.064 | N P | NO |
| 22 | 720*400@70 HZ | | | | | | | | | | | |
| 23 | 640*350@85 HZ | 31.500 | 37.86 | 20.317 | 3.048 | 2.032 | 85.080 | 9.244 | 1.585 | 0.079 | P N | NO |
| 24 | 720*350@85 HZ | | | | | | | | | | | |
| 25 | 640*400@85 HZ | 31.500 | 37.86 | 20.317 | 3.048 | 2.032 | 85.080 | 10.565 | 1.083 | 0.079 | N P | NO |
| 26 | 720*400@85 HZ | 35.500 | 37.927 | 20.282 | 3.042 | 2.028 | 85.039 | 10.546 | 1.107 | 0.079 | N P | NO |
| 27 | 1152*864@60 HZ | 80.000 | 54.346 | 14.400 | 2.400 | 1.200 | 60.053 | 15.898 | 0.681 | 0.055 | P P | NO |
| 28 | 1152*864@70 HZ | 94.200 | 63.955 | 12.229 | 2.038 | 1.019 | 70.016 | 13.501 | 0.719 | 0.047 | P P | NO |
| 29 | 1152*864@75 HZ | 108.00 | 67.500 | 10.667 | 2.370 | 1.185 | 75.000 | 12.800 | 0.474 | 0.044 | P P | NO |
| 30 | 1152*870@75 HZ | MAC 21 | | | | | | | | | | |
| 31 | 1152*900@66 HZ | SUN | | | | | | | | | | |
| 32 | 1152*900@76 HZ | | | | | | | | | | | |
| 33 | 1280*960@60 HZ | 108.00 | 60.000 | 11.852 | 2.889 | 1.037 | 60.000 | 16.000 | 0.600 | 0.050 | P P | NO |
| 34 | 1280*960@85 HZ | 148.50 | 85.938 | 8.620 | 1.508 | 1.077 | 85.002 | 11.171 | 0.547 | 0.035 | P P | NO |
| 35 | 1280*1024@ 60HZ | 108.00 | 63.981 | 11.852 | 2.296 | 1.037 | 60.020 | 16.005 | 0.594 | 0.047 | P P | NO |
| 36 | 1280*1024@ 75HZ | 135.00 | 79.976 | 9.481 | 1.837 | 1.067 | 75.025 | 12.804 | 0.475 | 0.038 | P P | NO |
| 37 | 1280*1024@ 85HZ | 157.50 | 91.146 | 8.127 | 1.422 | 1.016 | 85.024 | 11.235 | 0.483 | 0.033 | P P | NO |
| 38 | 1280*768@ 60HZ | | | | | | | | | | | |
| 39 | 1600*1200@ 60HZ | 162.00 | 75.000 | 9.877 | 1.877 | 1.185 | 60.000 | 16.000 | 0.613 | 0.040 | P P | |

| Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All | HL1916 | Page 31 of 32 |
|--|--------|---------------|
| rights reserved | | |

• Digital Signal Timings

| Resolution | Horizontal Frequency (kHz) | Refresh Rate (Hz) | Pixel Frequency (MHz) | Standard Type | Comments |
|--|----------------------------------|-------------------------|-----------------------------|--|----------|
| 640 x 350 | 31.5 | 70 | 25.175 | VGA | Visible* |
| | 37.9 | 85 8 | 31.500 | VESA | Visible* |
| 640 x 400 | 31.5 | 70 | 25.175 | VGA | Visible* |
| | 37.9 | 85 | 31.500 | VESA | Visible* |
| 640 x 480 | 31.5 | 59.9 | 25.175 | VGA | Yes |
| and the second s | | | | and the second s | 110000 |
| | 43.3 | 85.0 | 36.000 | VESA | Yes |
| 720 x 350 | 31.5 | 70.1 | 28.322 | VGA | Yes |
| | 37.9 | 85.0 | 35.500 | VGA | Yes |
| 720 x 400 | 31.5 | 70.1 | 28.322 | VGA | Yes |
| | 37.9 | 85.0 | 35.500 | VESA | Yes |
| 800 x 600 | 35.2 | 56.3 | 36.000 | VESA | Yes |
| | 37.9 | 60.3 | 40.000 | VESA | Yes |
| | 48.1 | 72 | 50.000 | VESA | Yes |
| | 46.9 | 75.0 | 49.500 | VESA | Yes |
| | 53.7 | 85.1 | 56.250 | VESA | Yes |
| 832 x 624 | 49.7 | 74.5 | 57.280 | Mac | Yes |
| 1024 x 768 | 48.4 | 60.0 | 65.000 | VESA | Yes |
| | 56.5 | 70.1 | 75.000 | VESA | Yes |
| | 60.0 | 75.0 | 78.750 | VESA | Yes |
| | 68.7 | 85.0 | 94.500 | VESA | Yes |
| 1152 x 864 | 67.5 | 75.0 | 108.000 | VESA | Yes |
| 1152 x 870 | 68.7 | 75.1 | 100.000 | Mac | Visible* |
| 1152 x 900 | 62.0 | 66.1 | 94.200 | SUN WS | Yes |
| | 71.9 | 76 | 107.500 | SUN WS | Yes |
| 1280 x 960 | 60.0 | 60.0 | 108.000 | VESA | Yes |
| | 75.0 | 75.0 | 126.200 | Mac | Yes |
| 1280 x 1024 | 64.0 | 60.0 | 108.000 | VESA | Yes |
| | 71.7 | 67.2 | 117.000 | SUN WS | Visible* |
| | 80.0 | 75.0 | 135.000 | VESA | Visible* |
| | 81.1 | 76.1 | 135.000 | SUN WS | Visible* |

| Last change: 27 th , Oct. 2005 | Copyright @ SSME (SIEMENS Medical Equipment Ltd.) All rights reserved | HL1916 | Page 32 of 32 |
|---|---|--------|---------------|
|---|---|--------|---------------|